

UNCLASSIFIED

AD 410158

DEFENSE DOCUMENTATION CENTER

FOR

SCIENTIFIC AND TECHNICAL INFORMATION

CAMERON STATION, ALEXANDRIA, VIRGINIA



UNCLASSIFIED

NOTICE: When government or other drawings, specifications or other data are used for any purpose other than in connection with a definitely related government procurement operation, the U. S. Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

N-63-4-3


USNRDL-TR-649
24 April 1963

CATALOGED BY DDC
410158

CONVERSION OF THE AN/PDR-TLB RADIAC TO
A REMOTE AREA MONITOR

by
P. A. Covey
AS AD No.

410158



U.S. NAVAL RADIOLOGICAL
DEFENSE LABORATORY
SAN FRANCISCO 24, CALIFORNIA

TECHNICAL DEVELOPMENTS BRANCH
P. D. LaRiviere, Head

CHEMICAL TECHNOLOGY DIVISION
L. H. Gevantman, Head

ADMINISTRATIVE INFORMATION

The work reported was part of a project sponsored by the Office of Civil Defense and by Defense Atomic Support Agency. The project is described in USNRDL Technical Program Summary for Fiscal Years 1963, 1964, and 1965, 1 November 1962, where it is designated Program A-1, Problem 11.

AVAILABILITY OF COPIES

Requests for additional copies by agencies or activities of the Department of Defense, their contractors certified to DDC (formerly ASTIA), and other government agencies or activities should be directed to the Defense Documentation Center for Scientific and Technical Information, Arlington Hall Station, Attn: TIPCR, Arlington 12, Virginia.

All other persons and organizations should direct requests for this report to the U. S. Department of Commerce, Office of Technical Services, Washington 25, D. C.

Eugene P. Cooper
Eugene P. Cooper
Scientific Director

E. B. Roth
E. B. Roth, CAPT USN
Commanding Officer and Director

ABSTRACT

A reliable remote-area monitor was devised from two AN/PDR-T1B gamma survey meters by coupling the electronics of one unit, located at the readout position, with the ion chamber from a second instrument located at the remote monitoring position. Cable lengths up to 50 ft were used with no appreciable signal degradation. The instrument is useful in fields up to 500 r/hr, and it has produced measurements agreeing with a standard NRDL Model 103 gamma intensity vs time recorder (GITR) to within ± 20 %.

SUMMARY

Problem

In a recent weapon effects test it was necessary for personnel in underground manned stations to know the exterior gamma dose rate at all times. No suitable instrument was available for this purpose.

Findings

A reliable instrument satisfying the requirement was devised from two interconnected AN/PDR-T1B survey meters. One served as the control and readout unit, the other as the remote detector. Separations up to 50 ft were employed without appreciable signal degradation.

INTRODUCTION

The development of radiological countermeasures and the evaluation of the effectiveness of radiological countermeasure systems requires information and data on radioactive fallout produced by nuclear explosions. During a recent weapon effects test, personnel in underground manned stations measured, during fallout, the radiation and deposition dynamics of the event. To assess the progress of the fallout event, personnel in the manned stations required an instrument to record continuously the exterior gamma dose rate.

The two major requirements for the instrument were: an immediate and continuous readout capability, and a radiation intensity range extending to an upper limit of 50 r/hr, or 500 r/hr as specified. The standard NRDL GITR (Gamma Intensity vs Time Recorder)** easily covered the intensity range desired, but this device requires laboratory processing of the magnetic tape record for readout. A low-level self-recording GITR developed at NRDL* offered immediate readout, but was limited to a maximum intensity of .10 r/hr. It was, therefore, necessary to develop an instrument satisfying the above requirements. The required six units were designed, tested and fabricated within three months.

DESCRIPTION OF INSTRUMENT

The radiac training set AN/PDR-T1B was modified to meet the requirements for remote gamma radiation monitoring. Two instruments were required for each installation. One instrument was located in the control or readout position within the manned station and the ion chamber from a second instrument was placed in the remote monitoring locations. A light-tight, moisture-proof and dust-proof container was provided for the remote detector by using the original case of the second instrument.

METHOD OF MODIFICATION

The remote feature was achieved by substituting the remote ion chamber for the chamber of the control radiac. This was accomplished by disconnecting the cathode of the internal chamber of the control radiac at point "A", shown in the circuit diagram reproduced in Fig. 1, and connecting the cathode of the remote detector to the electrometer circuit in the control unit. RG/59U Co-ax cable was used for this connection,

*P. A. Covey, A Low-Range Gamma Intensity Time Recorder, Technical Report in preparation.

**M. M. Bigger, H. R. Rinnert, H. A. Zagorites. Operation Hardtack, Proj. 2.1. WT-1619, January 1960 (CONFIDENTIAL).

in lengths up to 50 ft, with no degradation of signal or appreciable change in time constant. The high voltage lead was disconnected from the internal chamber at point "C" (Fig. 1) and connected through a length of wire (alpha 1553, MIL-W-76A) to the high voltage terminal of the remote chamber.

The remote wiring was installed in conduit to provide electrical and mechanical protection and to prevent movement of the signal cable. Any movement of this cable would have generated some noise voltage which would have indicated on the readout meter, when the more sensitive ranges were being used.

EXTENDED RANGE MODIFICATIONS

Where specified the range of this instrument was extended to 500 r/hr in the following way: Referring to Fig. 1, the range resistor R1 was removed and resistors R2, R3, R4 and R5 were moved one position to the left on the range switch, so that R2 would now occupy position R1, etc. This left a blank terminal at the 50K mr/hr position, R5. A 2.0 megohm, 5 % carbon resistor was installed between this terminal and the range resistor common terminal.

In addition to these changes the voltage supplied to the ion chamber by batteries BT1 and BT6 was increased to 300 volts to insure adequate saturation current at high radiation rates (500 r/hr). A Burgess type U-200, 300 volt battery was convenient for this purpose. The original batteries BT1 and BT6 were disconnected at point Y (Fig. 1), and the U-200 negative terminal was connected to this point with the positive side connected to the remote chamber. This battery did not fit in the space occupied by BT1 and BT6. It was convenient in the manned station application to put this battery adjacent to the control unit.

OPERATION AND CALIBRATION

These modifications necessitated instrument recalibration. The procedure described in the instruction manual TM11-5514 was followed for all ranges except the 500 r/hr. After calibration of the lower ranges was accomplished it was necessary to check the instrument output on the high range, in order to determine whether the new range resistor R5 was the correct value. Tests of several of these systems showed that the value of this resistor varied between 1.8 and 2.2 megohms. This difference cannot be compensated for by readjusting the calibrate pot,

R17, since this would cause maladjustment on all lower ranges. Therefore where necessary, adjustment of the value of R5 was done in the following way: A gamma radiation field equal to $\frac{4}{5}$ of full scale (400 r/hr) was established at the remote detector. The meter in the control unit was then read. If the reading was higher than $\pm 20\%$ of 400 r/hr, R5 was changed to 1.8 meg; if lower, R5 was changed to 2.2 meg. A miniature 2.5 megohm potentiometer might have been used for R5 to facilitate calibration, but this was not done since one of the stated three fixed resistor values adequately adjusted the system to within $\pm 20\%$ of the true gamma dose rate on this range. The zero-set, battery check and range controls were operated in the normal manner, except that with the 500 r/hr modification an X10 Scale Multiplier was applied to the meter readings on all ranges. Every precaution was taken when the electrometer compartment or any part of the signal circuitry was being worked on to avoid any contamination (e.g., fingerprints, dust, or moisture).

RESULTS AND CONCLUSIONS

Six remote reading radiacs were modified as described and installed in manned stations for use in a recent weapon effects test. The performance of all six instruments was excellent. A typical result is shown in Fig. 2 for the remote radiac detector mounted 3 ft above the level of the soil backfill over the manned station, with one from a nearby standard GTR. The points representing individual readings obtained manually with the remote radiac are compared with the GTR record. It can be seen that except in regions where the intensity is changing rapidly the readings agree within $\pm 20\%$. It is concluded that the novel modifications described led to an accurate and reliable remote gamma monitoring device.




Fig. 1 Schematic Diagram




Fig. 2 Readout Comparison

INITIAL DISTRIBUTION

CopiesNAVY

3 Chief, Bureau of Ships (Code 210L)
 1 Chief, Bureau of Ships (Code 320)
 6 Chief, Bureau of Ships (Code 685C)
 1 Chief, Bureau of Medicine and Surgery
 1 Chief, Bureau of Naval Weapons (RRMA-11)
 2 Chief, Bureau of Yards and Docks (Code 74)
 1 Chief, Bureau of Yards and Docks (Code C-400)
 1 Chief, of Naval Operations (Op-07T)
 1 Chief of Naval Research (Code 104)
 1 Commander, New York Naval Shipyard (Material Lab.)
 3 Director, Naval Research Laboratory (Code 2021)
 1 Office of Naval Research (Code 422)
 10 Office of Naval Research, FPO, New York
 1 CO, U.S. Naval Civil Engineering Laboratory
 1 Commander, Naval Air Material Center, Philadelphia
 1 Naval Medical Research Institute
 1 U.S. Naval Postgraduate School, Monterey
 1 Commander, Naval Ordnance Laboratory, Silver Spring
 1 CI, Naval Nuclear Ordnance Evaluation Unit (Code 4011)
 1 Office of Patent Counsel, San Diego
 1 CO, Naval Medical Field Research Laboratory, Camp Lejeune

ARMY

1 Chief of Research and Development (Atomics Division)
 1 Chief of Research and Development (Life Science Division)
 1 Deputy Chief of Staff for Military Operations (DGM)
 1 Deputy Chief of Staff for Military Operations (CBR)
 1 Office of Assistant Chief of Staff, G-2
 1 Chief of Engineers (ENGMC-EB)
 1 Chief of Engineers (ENGMC-DE)
 1 Chief of Engineers (ENG CW)
 1 CG, Army Materiel Command (AMCRD-DE-NE)
 1 CG, Ballistic Research Laboratories
 1 CG, USA CBR Agency
 1 CO, Fort McClellan, Alabama
 1 Commandant, Chemical Corps Schools (Library)

1 CG, CBR Combat Developments Agency
 1 CO, Chemical Research and Development Laboratories
 1 Commander, Chemical Corps Nuclear Defense Laboratory
 1 Hq., Army Environmental Hygiene Agency
 1 CG, Aberdeen Proving Ground
 1 Director, Walter Reed Army Medical Center
 1 Hq., Army Nuclear Medicine Research Detachment, Europe
 1 CG, Combat Developments Command (CDCMR-V)
 1 CG, Quartermaster Res. and Eng. Command
 1 Hq., Dugway Proving Ground
 3 The Surgeon General (MEDNE)
 1 Army Electronic Res. and Dev. Agency
 1 CG, Army Electronic Proving Ground
 1 CG, Engineer Res. and Dev. Laboratory
 1 Director, Office of Special Weapons Development
 1 CO, Watertown Arsenal
 1 CG, Mobility Command
 1 CG, Munitions Command
 1 CO, Frankford Arsenal
 1 CG, Army Missile Command

AIR FORCE

1 Assistant Chief of Staff, Intelligence (AFCIN-3B)
 6 CG, Aeronautical Systems Division (ASAPRD-NS)
 1 Directorate of Civil Engineering (AFOCE-ES)
 1 Director, USAF Project RAND
 1 Commandant, School of Aerospace Medicine, Brooks AFB
 1 Office of the Surgeon (SUP3.1), Strategic Air Command
 1 Office of the Surgeon General
 1 CG, Special Weapons Center, Kirtland AFB
 1 CG, Special Weapons Center (SWRB)
 1 Director, Air University Library, Maxwell AFB
 2 Commander, Technical Training Wing, 3415th TTG
 1 Hq., Second Air Force, Barksdale AFB
 1 Commander, Electronic Systems Division (CRZT)

OTHER DOD ACTIVITIES

3 Chief, Defense Atomic Support Agency (Library)
 1 Commander, FC/DASA, Sandia Base (FCDV)
 1 Commander, FC/DASA, Sandia Base (FCTG5, Library)
 1 Commander, FC/DASA, Sandia Base (FCWT)
 2 Office of Civil Defense, Washington
 2 Civil Defense Unit, Army Library
 10 Armed Services Technical Information Agency

1 Director, Armed Forces Radiobiology Research Institute
 1 Research Analysis Corporation
 1 Texas Instruments, Inc. (Mouser)
 1 Aerojet General, Azusa
 1 Aerojet General, San Ramon
 1 Allis-Chalmers Manufacturing Co., Milwaukee
 1 Allis-Chalmers Manufacturing Co., Washington
 1 Allison Division -GMC
 2 Argonne Cancer Research Hospital
 10 Argonne National Laboratory
 1 Atomic Bomb Casualty Commission
 1 AEC Scientific Representative, France
 3 Atomic Energy Commission, Washington
 4 Atomic Energy of Canada, Limited
 4 Atomics International
 1 Babcock and Wilcox Company
 2 Battelle Memorial Institute
 2 Beers, Roland F., Inc.
 1 Beryllium Corporation
 4 Brookhaven National Laboratory
 1 Bureau of Mines, Albany
 1 Bureau of Mines, Salt Lake City
 1 Carnegie Institute of Technology
 1 Chicago Patent Group
 1 Columbia University (Havens)
 1 Columbia University (Rossi)
 1 Combustion Engineering, Inc.
 1 Combustion Engineering, Inc. (NRD)
 1 Committee on the Effects of Atomic Radiation
 5 Defence Research Member
 1 Denver Research Institute
 1 Dow Chemical Company, Rocky Flats
 4 duPont Company, Aiken
 1 duPont Company, Wilmington
 1 Edgerton, Germeshausen and Grier, Inc., Goleta
 1 Edgerton, Germeshausen and Grier, Inc., Las Vegas
 1 Franklin Institute of Pennsylvania
 1 Fundamental Methods Association
 1 General Atomic Division
 1 General Dynamics/Astronautics (NASA)
 2 General Dynamics, Fort Worth
 2 General Electric Company, Cincinnati
 4 General Electric Company, Richland
 1 General Electric Company, San Jose
 1 General Electric Company, St. Petersburg
 1 General Nuclear Engineering Corporation
 1 General Scientific Corporation

1 Gibbs and Cox, Inc.
 1 Goodyear Atomic Corporation
 1 Hughes Aircraft Company, Culver City
 1 Iowa State University
 1 Jet Propulsion Laboratory
 2 Knolls Atomic Power Laboratory
 1 Lockheed-Georgia Company
 1 Lockheed Missiles and Space Company (NASA)
 2 Los Alamos Scientific Laboratory (Library)
 1 Lovelace Foundation
 1 Maritime Administration
 1 Martin-Marietta Corporation
 2 Midwestern Universities Research Association
 1 Mound Laboratory
 1 NASA, Lewis Research Center
 2 NASA, Scientific and Technical Information Facility
 1 National Bureau of Standards (Library)
 1 National Bureau of Standards (Taylor)
 1 National Lead Company of Ohio
 2 Nevada Operations Office
 1 New Brunswick Area Office
 1 New York Operations Office
 1 New York University (Eisenbud)
 1 Nuclear Materials and Equipment Corporation
 1 Nuclear Metals, Inc.
 1 Office of Assistant General Counsel for Patents
 4 Phillips Petroleum Company
 1 Power Reactor Development Company
 2 Pratt and Whitney Aircraft Division
 1 Princeton University (White)
 2 Public Health Service, Washington
 1 Public Health Service, Las Vegas
 1 Public Health Service, Montgomery
 1 Rensselaer Polytechnic Institute
 1 Sandia Corporation, Albuquerque
 1 Sandia Corporation, Livermore
 1 Space Technology Laboratories, Inc. (NASA)
 1 Stanford University (SLAC)
 1 States Marine Lines, Inc.
 1 Sylvania Electric Products, Inc.
 1 Tennessee Valley Authority
 2 Union Carbide Nuclear Company (ORGDP)
 6 Union Carbide Nuclear Company (ORNL)
 1 Union Carbide Nuclear Company (Paducah Plant)
 1 United Nuclear Corporation (NDA)
 1 U.S. Geological Survey, Denver

1 U.S. Geological Survey, Menlo Park
 1 U.S. Geological Survey, Washington
 2 University of California Lawrence Radiation Lab., Berkeley
 2 University of California Lawrence Radiation Lab., Livermore
 1 University of California, Los Angeles
 1 University of California, San Francisco
 1 University of Chicago Radiation Laboratory
 1 University of Hawaii
 1 University of Puerto Rico
 1 University of Rochester (Atomic Energy Project)
 1 University of Rochester (Marshak)
 1 University of Utah
 1 University of Washington (Geballe)
 1 University of Washington (Rohde)
 4 Westinghouse Bettis Atomic Power Laboratory
 1 Westinghouse Electric Corporation (Rahilly)
 1 Westinghouse Electric Corporation (NASA)
 1 Western Reserve University (Friedell)
 1 Western Reserve University (Major)
 1 Yankee Atomic Electric Company
 25 Technical Information Extension, Oak Ridge

USNRDL

35 USNRDL, Technical Information Division

DISTRIBUTION DATE: 14 June 1963

Naval Radiological Defense Laboratory USNRDL-TR-649 CONVERSION OF THE AN/PDR-T1B RADIAC TO A REMOTE AREA MONITOR by P. A. Covey 24 April 1963 14 p. illus. UNCLASSIFIED A reliable remote-area monitor was devised from two AN/PDR-T1B gamma survey meters by coupling the electronics of one unit, located at the readout position, with the ion chamber from a second instrument located at the remote monitoring position. Cable lengths (over)	1. Radiation measurements systems components. 2. Ionization chambers. 3. Gamma radiation. I. Covey, P. A. II. Title. <u>UNCLASSIFIED</u>
Naval Radiological Defense Laboratory USNRDL-TR-649 CONVERSION OF THE AN/PDR-T1B RADIAC TO A REMOTE AREA MONITOR by P. A. Covey 24 April 1963 14 p. illus. UNCLASSIFIED A reliable remote-area monitor was devised from two AN/PDR-T1B gamma survey meters by coupling the electronics of one unit, located at the readout position, with the ion chamber from a second instrument located at the remote monitoring position. Cable lengths (over)	1. Radiation measurements systems components. 2. Ionization chambers. 3. Gamma radiation. I. Covey, P. A. II. Title. <u>UNCLASSIFIED</u> up to 50 ft were used with no appreciable signal degradation. The instrument is useful in fields up to 500 r/hr, and it has produced measurements agreeing with a standard NRDL Model 103 gamma intensity vs time recorder (GTR) to within + 20%. <u>UNCLASSIFIED</u>