

UNCLASSIFIED

AD 4 5 6 6 8 4

DEFENSE DOCUMENTATION CENTER

FOR

SCIENTIFIC AND TECHNICAL INFORMATION

CAMERON STATION ALEXANDRIA, VIRGINIA



UNCLASSIFIED

NOTICE: When government or other drawings, specifications or other data are used for any purpose other than in connection with a definitely related government procurement operation, the U. S. Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

4 5 6 6 8 4

EXCLUDED BY DDC

AS AD NO. 456684

TECHNICAL MANUSCRIPT 195

A NEW RACK FOR USE IN CELL CULTURE

FEBRUARY 1965

UNITED STATES ARMY
BIOLOGICAL LABORATORIES
FORT DETRICK

U.S. ARMY BIOLOGICAL LABORATORIES
Fort Detrick, Frederick, Maryland

TECHNICAL MANUSCRIPT 195

A NEW RACK FOR USE IN CELL CULTURE

Alan Richter

Sidney Halle

Virus and Rickettsia Division
DIRECTOR OF BIOLOGICAL RESEARCH

Project 1A01450A9101

February 1965

This publication or any portion thereof may not be reproduced without specific authorization from the Commanding Officer, U. S. Army Biological Laboratories, ATTN: Technical Releases Branch, Technical Information Division, Fort Detrick, Frederick, Maryland. 21701. However, DDC is authorized to reproduce the publication for U. S. Government purposes.

The information in this publication has not been cleared for release to the public.

DDC AVAILABILITY NOTICE

Qualified requestors may obtain copies of this publication directly from DDC.

Foreign announcement and dissemination of this publication by DDC is limited.

ABSTRACT

This describes a new rack with a variety of uses and applications in tissue culture and virology. Its invention was prompted by a need for a simple method of cloning cells. The principal feature of the rack is that it is designed to fit onto the stage of the inverted microscope.

A NEW RACK FOR USE IN CELL CULTURE

This describes a new rack with a variety of uses and applications in tissue culture and virology. Its invention was prompted by a need for a simple method of cloning cells. We started by using the tube method of cloning in a rack of the type offered by Microbiological Associates. In that method the tubes are incubated at an angle of 5 degrees to allow the cells to grow on one wall of the tube. The tube can then be removed from the rack, placed on a special observation track, and viewed in the microscope.

It occurred to us that the whole procedure might be simplified if the cells were grown on the floor of the tube rather than on the wall so that the cells could be examined directly in the inverted microscope without removing the tubes from the rack. This necessitated a flat-bottomed tube, i.e., a vial. Accordingly, we designed a rack (Figure 1) to fit onto the stage of the inverted microscope. The rack holds 40 shell vials (Kimble 60930, outside diameter 15 mm, height 45 mm). It is constructed of type 304 stainless steel, 22-gauge, and has holes $19/32$ inch in diameter to position the vials. The holes in the floor of the rack are $9/16$ inch in diameter, just small enough to support the vials without their falling through the rack and yet large enough to permit viewing the entire inside floor of the vial with the inverted microscope. The rows of vials are protected from airborne contamination by 3M Micropore surgical tape which is translucent. The rack is $11\ 9/16$ inches long and 4 inches wide. The total height, including vials, is $1\ 13/16$ inches. There is thus no space problem in the carbon dioxide incubator.




Figure 1. Rack, showing holes, vials, and tape.
(FD Neg C-7642)

Two experiments illustrate applications of the new rack. In the first experiment, a suspension of Syrian hamster 113 cells* was diluted to 4 cells per 10 ml. Two ml of this suspension was distributed to each of 40 vials and the rack was incubated at 37 C in a carbon dioxide incubator. After 7 days the whole rack was placed on the stage of the inverted microscope and each vial was examined for the presence of colonies (simply by moving the rack). The results are shown in Table 1. One vial was contaminated. The 14 vials containing one colony each were separated from the rest by cutting the tape and removing the vials from the rack. The colonies were refed and grown into cultures directly in the vials, thus avoiding any special techniques of isolation from other clones.

TABLE 1. CLONING OF HAMSTER 113 CELLS IN THE VIAL RACK

	Number of colonies ^a / per vial			
	0	1	2	3
Number of vials found with 0, 1, 2 colonies, etc.	23	14	2	0
Number of vials expected ^b /	25	11	3	0

a. Cell input per vial 0.8.

b. Calculated according to the Poisson distribution.

This rack is also useful for other purposes, for example, the titration of viruses by production of cytopathic effect in monolayer cultures. In the second experiment 200,000 cells of strain 113 were transferred to each of 50 vials and the vials were incubated for two hours. A virus sample of known plaque-forming titer was serially diluted and the dilutions were inoculated into the vials in rows of 10 vials each. After 40 to 60 hours the rack was placed on the stage of the inverted microscope and the cultures were scored as showing cytopathic effect or not. The results of the experiment are shown in Table 2.

* Supplied by Dr. Vittorio Defendi, Wistar Institute, Philadelphia, Pa.

TABLE 2. ASSAY OF EASTERN EQUINE ENCEPHALITIS VIRUS BY
PRODUCTION OF CYTOPATHIC EFFECT IN HAMSTER 113 CELLS

Virus dilution	Number of cultures showing cytopathic effect/normal
No virus	0/10
10^{-7}	10/0
10^{-8}	10/0
10^{-9}	3/7
10^{-10}	0/10

The vial rack has been in use in our laboratory for several months and has appreciably reduced the time required to perform a variety of tasks. It should be especially useful in genetic isolation of mutants, both cell and viral.