

UNCLASSIFIED

AD NUMBER
AD464211
NEW LIMITATION CHANGE
TO Approved for public release, distribution unlimited
FROM Distribution authorized to U.S. Gov't. agencies and their contractors; Administrative/Operational Use; MAY 1965. Other requests shall be referred to Commanding Officer, U.S. Army Biological Labs., Attn: Technical Releases Branch/TID, Fort Detrick, Frederick, MD
AUTHORITY
BDRL D/A ltr, 28 Sep 1971

THIS PAGE IS UNCLASSIFIED

UNCLASSIFIED

AD 464211

DEFENSE DOCUMENTATION CENTER

FOR

SCIENTIFIC AND TECHNICAL INFORMATION

CAMERON STATION ALEXANDRIA, VIRGINIA



UNCLASSIFIED

NOTICE: When government or other drawings, specifications or other data are used for any purpose other than in connection with a definitely related government procurement operation, the U. S. Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

AD

TECHNICAL MANUSCRIPT 218

SAMPLING SUBMICRON  
T1 BACTERIOPHAGE AEROSOLS

AVAILABLE COPY WILL NOT PERMIT  
FULLY LEGIBLE REPRODUCTION.  
REPRODUCTION WILL BE MADE   
REQUESTED BY USERS OF DDG.

MAY 1965

UNITED STATES ARMY  
BIOLOGICAL LABORATORIES  
FORT DETRICK

CATALOGED BY: DDG

AS AD No. 464211

464211

This publication or any portion thereof may not be reproduced without specific authorization from the Commanding Officer, U. S. Army Biological Laboratories, ATTN: Technical Releases Branch, Technical Information Division, Fort Detrick, Frederick, Maryland. However, DDC is authorized to reproduce the publication for U.S. Government purposes.

The information in this publication has not been cleared for release to the public.

DDC AVAILABILITY NOTICE

Qualified requestors may obtain copies of this publication directly from DDC.

Foreign announcement and dissemination of this publication by DDC is limited.

U.S. ARMY BIOLOGICAL LABORATORIES  
Fort Detrick, Frederick, Maryland

TECHNICAL MANUSCRIPT 218

SAMPLING SUBMICRON T1 BACTERIOPHAGE AEROSOLS

J. Bruce Harstad

Physical Defense Division  
DIRECTORATE OF MEDICAL RESEARCH

Project 10622401A072

May 1965

#### ACKNOWLEDGMENTS

The author thanks Robert Sine and Billy Passeur for electron microscopy; Roger Scherff for advice and the use of radiological facilities; Joseph Milo, John Happ, and Carl Michael for skilled technical assistance; and M.B. Albert for derivation of the theoretical radiological aerosol concentrations.

#### ABSTRACT

Liquid impingers, filter papers, and fritted bubblers were partial viable collectors of radioactive submicron T1 bacteriophage aerosols at 30, 55, and 85% relative humidity. Sampler differences for viable collection were due to incomplete physical collection (slippage) and killing of phage by the samplers. Dynamic aerosols of a mass median diameter of 0.2 micron were produced with a Dautrebande generator from concentrated aqueous purified phage suspensions containing extracellular soluble radioactive phosphate as a physical tracer. There was considerable destruction of phage by the Dautrebande generator; phage titers of the Dautrebande suspension decreased exponentially but there was a progressive (linear) increase in tracer titers. Liquid impingers recovered the most viable phage but allowed considerable (30 to 48%) slippage, which varies inversely with the aerosol relative humidity. Filter papers were virtually complete physical collectors of submicron particles but were the most destructive. Fritted bubbler slippage was more than 80%. With all samplers, phage kill was highest at 85% relative humidity and lowest at 55% relative humidity. An electrostatic precipitator was used to collect aerosol samples for particle sizing with an electron microscope. The particle size was slightly larger at 85% relative humidity than at 30 or 55% relative humidity.

## CONTENTS

Acknowledgments . . . . .	3
Abstract . . . . .	3
I. INTRODUCTION . . . . .	7
II. MATERIALS AND METHODS . . . . .	8
A. Phage Concentration and Purification . . . . .	8
B. Radioactive Tracer . . . . .	10
C. Aerosol Generator . . . . .	10
D. Aerosol Chamber . . . . .	11
E. Aerosol Samplers . . . . .	11
1. All Glass Impinger (AGI-4) . . . . .	12
2. Capillary Impinger . . . . .	12
3. Fritted Bubbler . . . . .	13
4. Chemical Corps Type 6 Filter Paper . . . . .	13
5. MSA 1106BH Glass Filter Paper . . . . .	13
6. Backup Filter . . . . .	13
F. Aerosol Particle Sizing . . . . .	14
G. Assay . . . . .	15
H. Design of Experiment . . . . .	16
III. EXPERIMENTAL RESULTS . . . . .	17
A. Killing of Phage by the Dautrebande Generator . . . . .	17
B. Radiological Aerosol Concentrations . . . . .	17
C. Phage Aerosol Concentrations . . . . .	20
D. Killing of Phage by the Samplers . . . . .	20
E. Relative Viable Sampler Collection Efficiency . . . . .	21
F. Sampler Physical Slippage . . . . .	21
G. Sampler Viable Slippage . . . . .	22
H. Filter Paper Physical Slippage . . . . .	22
I. Aerosol Particle Size . . . . .	23
IV. DISCUSSION . . . . .	26
Literature Cited . . . . .	27
Distribution List . . . . .	29

FIGURES

1. Electron Micrograph of T1 Phage Particle Prepared by Air-Drying . . . 9
2. Aerosol Apparatus . . . . . 12
3. Size Distribution of Radioactive T1 Phage Aerosol at 55% Relative Humidity Dispersed with Dautrebande D<sub>30</sub><sup>1</sup> Aerosol Generator . . . . . 25
4. Electron Micrograph of Radioactive T1 Phage Aerosol at 55% Relative Humidity Dispersed with Dautrebande D<sub>30</sub><sup>1</sup> Aerosol Generator . . . . . 25

TABLES

1. Effect of Aerosolization and Sampling on the Viability of T1 Phage . . . . . 18
2. Comparison of Air Sampling Devices for Viable and Physical Collection of Radioactive Submicron T1 Phage Aerosols at 30, 55, and 85% Relative Humidity . . . . . 19
3. Slippage of Radioactive Submicron T1 Phage Aerosols Through Type 6 and 1106BH Filter Papers at 30, 55, and 85% Relative Humidity . . . . . 20
4. Particle Size Parameters of Radioactive T1 Phage Aerosols at 30, 55, and 85% Relative Humidity . . . . . 24

## I. INTRODUCTION

The literature on microbial aerosols has been concerned mainly with bacterial aerosols.<sup>1</sup> Less is known about viral aerosols,<sup>2</sup> particularly with regard to the production, particle sizing, and sampling of submicron aerosols of viruses. Submicron aerosols have been studied extensively, but knowledge is limited almost entirely to aerosols of inert materials. Extremely fine aerosols in the millimicron size range (<0.1 micron) have been produced from dilute solutions of dyes and salts and from colloidal suspensions.<sup>3</sup> Stern et al.<sup>4</sup> generated homogeneous millimicron aerosols from dilute colloidal suspension with a modified Vaponefrin nebulizer. Their technique was unique in that the colloids were highly purified viral suspensions. The concentrations of the suspensions were adjusted by dilution with water so that the chance of two virus particles being present simultaneously in a single evaporating water droplet was about 1 to 100. Morris et al.<sup>5</sup> evaluated an electrostatic precipitator for sampling monodispersed T3 coliphage aerosols produced with a modified Collison atomizer from phage broth lysates diluted extensively with water so that the aerosols consisted principally (95%) of single phage particles.

Physical tracers such as radioactive isotopes, dyes, and bacterial spores have been used to study the behavior and sampling of microbial aerosols. Harper, Hood, and Morton<sup>6</sup> and Miller et al.<sup>7</sup> used radioactive isotopes and bacterial spores to study the survival of bacterial aerosols. Harper<sup>8</sup> used radioactive phosphate to study the survival of airborne viruses. Tyler, Ships, and Painter<sup>9</sup> used physical tracers with bacterial aerosols to show that slippage of the aerosol particles through the samplers, particle size discrimination, and destruction of bacteria during the sampling process are sources of sampler differences.

This paper reports on the generation, sampling, and particle sizing of submicron aerosols of T1 phage. The aerosols were produced with a Dautrebande aerosol generator from concentrated aqueous suspensions of purified phage containing soluble radioactive phosphate as a physical tracer. Five air sampling devices were compared for the physical and viable collection of submicron aerosols at relative humidities of 30, 55, and 85% and a temperature of 24 C. The samplers were all-glass impingers (AGI-4), capillary impingers, Chemical Corps Type 6 filter papers, MSA 1106BH filter papers, and fritted bubblers.

## II. MATERIALS AND METHODS

### A. PHAGE CONCENTRATION AND PURIFICATION

T1 and T3 bacteriophage, two of the well known T-series of phages of Escherichia coli, were studied. These phages are relatively stable to chemical and physical manipulations and are relatively easy to obtain in pure form. Aerosolization of concentrated aqueous purified suspensions of both of these phages resulted in aerosols of small particles well within the submicron size range ( $<0.3\mu$ ). However, the aerosols are subject to biological decay, and the purification required to obtain submicron aerosols is detrimental to the survival of the airborne phage. Studies with submicron T3 phage aerosols were discontinued when tests revealed that T1 phage was much more stable in air, especially at the lower relative humidities.

An electron micrograph of a T1 phage particle prepared by the air drying technique is shown in Figure 1. It is tadpole-shaped, with a slender tail about 150 millimicrons long attached to a hexagonal head about 60 millimicrons in diameter. Aqueous purified phage suspensions were used to produce submicron aerosols. The suspension must be free of soluble and particulate contaminants to produce an aerosol of minimum particle size. Clean phage suspensions were prepared by concentrating and purifying large volumes of broth lysates by differential centrifugation and washing with distilled water.

T1 phage lysates were produced in double strength Bacto nutrient broth. Batches of five liters in 18-liter carboys proved convenient. The seed inoculum of E. coli strain B (ATCC 11303) for a carboy was prepared in 100 ml of the same medium in a 500-ml Erlenmeyer flask and incubated at 37 C on a shaker for 16 hours. The carboy, previously warmed to 37 C, was then inoculated with 100 ml of the seed E. coli culture and incubated at 37 C with aeration on a shaking machine. After 3½ hours' incubation the E. coli culture had reached the end of the logarithmic phase of growth ( $1 \times 10^7$  cells per ml). Crude phage lysate, clarified by low-speed centrifugation, was added in the ratio of 2 to 5 phage particles per bacterium and the carboy was shaken 4 to 5 hours. Aqueous purified phage suspensions were not used for seeding coli cultures because higher phage yields were obtained when broth lysates were used for inoculation.<sup>10</sup>

The crude lysates titering  $1 \times 10^{11}$  phage particles per ml were clarified at low speed (10,000 x g) for 10 minutes in a Servall SS-1 centrifuge to remove bacterial and large-particle debris. The supernatant fluid was decanted and centrifuged at high speed (60,000 x g) for 25 minutes in a No. 30 rotor in a Model L Spinco preparative ultracentrifuge to sediment the phage component. The supernatant fluid was drawn off with a flat-tipped needle until the level of the liquid remaining in the bottom of



Figure 1. Electron Micrograph of T1 Phage Particles Prepared by Air-Drying.

the centrifuge tube was just above the virus pellet. The pellets were resuspended in the bottom supernatant fluid by shaking and treated with 5 micrograms per ml of each of the enzymes deoxyribonuclease and ribonuclease for 2 to 4 hours at 37°C in the presence of an  $MgSO_4$  concentration of 0.003 M and at a final pH of 6.6. The enzymatic digestion of extraneous DNA and RNA decreased the viscosity of the suspension and enhanced the separation of the phage component by centrifugation.<sup>11</sup> The suspension was clarified by low-speed centrifugation and then subjected to six complete cycles of differential centrifugation, i.e., alternate high (60,000 x g) and low (10,000 x g) speed centrifugation. The pellets from the high-speed centrifugations were resuspended in the bottom supernatant fluid and washed with triple-distilled water to dilute out the salts. From the original five liters of crude lysate about 70 ml of aqueous purified phage titering up to  $10^{12}$  phage particles per ml was produced. Loss of viable phage particles from the purification process was about 80% in most cases.

For the sampling trials, several purified lots that had been stored at 5 C for periods up to one year were pooled, clarified at low-speed centrifugation, and put through a final cycle of differential centrifugation. The pooled suspension stored very well at 5 C; there was no discernible drop in viable titer during the period of this study, nearly two months.

The physical and biological properties of the radioactive phage suspension used in all trials were:

a. Viable count	$1.55 \times 10^{11}$ phage per ml
b. Radioactivity	10 microcurie ( $\mu\text{c}$ ) per ml
c. Viscosity at 30 C	1.0349 centipoise
d. Specific gravity	1.001
e. Surface tension	74.4 dynes per cm
f. Dry weight	0.01%
g. pH	6.6

#### B. RADIOACTIVE TRACER

A radioactive isotope was selected as the physical tracer because the minute amounts of carrier-free isotope required would not alter the characteristics of the phage suspensions and therefore would not increase the aerosol particle size. Soluble radioactive phosphate was used as an extracellular tracer, i.e., it was added directly to the phage suspension. The isotope, specially prepared for this study by the Oak Ridge National Laboratory, Oak Ridge, Tennessee, was a solution of carrier-free  $\text{H}_2\text{P}^{32}\text{O}_4$  in distilled water of an activity of 1 millicurie ( $\text{mc}$ ) per ml. It had a pH of 6.2, which was compatible with that of the phage suspension, 6.6. The isotope solution was added to the phage suspension at the start of each test. For all tests the amounts were constant (0.1 ml of isotope solution per 9.9 ml of phage). As a consequence the activity of the test phage suspension decreased, because of radioactive decay, from 10  $\mu\text{c}$  per ml for the first test to less than 1  $\mu\text{c}$  per ml for the last test, nearly two months later.

Preliminary experiments revealed that phage inactivation by the radioactive tracer was not detectable over a 24-hour holding period. In the sampling trials, the phage-isotope contact period, i.e., the time between sample collection and assay, was less than four hours.

#### C. AEROSOL GENERATOR

The particle size of aerosols produced from liquid suspensions is primarily a function of the atomizer droplet size and the constituency or amount of solid material in the spray suspension. Phage aerosols of the minimum particle size would therefore be produced from clean aqueous phage suspensions with an atomizer producing the smallest droplets.

The Dautrebande D30<sup>1</sup> aerosol generator\* has been used for the production of submicron aerosols from solutions of dyes and salts. It is a type of atomizer-impactor in which the larger droplets are returned to the solution while the smaller droplets are permitted to escape. This selection is accomplished by means of successive liquid barriers, which is termed obligatory liquid filtration. The obligatory liquid filtration takes place in a hollow cylinder in which the unstable particles coalesce and form a continuous liquid mass. The droplets that are able to pass the barrier are passed through a capillary pipe system where particles that have not attained a sufficient degree of stability coalesce and form another liquid barrier through which only the most stable droplets can traverse. The dispersion process involves some evaporation and consequently there is a progressive (linear) concentration of the generator suspension.<sup>3</sup>

#### D. AEROSOL CHAMBER

The aerosol chamber and apparatus used to test the air-sampling devices are illustrated in Figure 2. Dynamic aerosols were produced by atomizing radioactive aqueous purified T1 phage suspensions with a Dautrebande D30<sup>1</sup> aerosol generator operated at 17.5 psig. The air flow through the generator was 18 liters per minute and the fluid atomization rate was 0.15 ml per minute. The aerosol was diluted and mixed with 30 liters of air of the desired relative humidity in a 4-liter cylindrical glass chamber. From the mixing chamber the aerosol passed into a 12-gallon carboy and finally into a circular (10-cm-diameter) manifold with ten sampling ports. The excess aerosol was bled off through a filter located upstream from the sampling manifold. Aerosol relative humidity was controlled by dryers and humidifiers for the mixing air. Aerosol relative humidity and temperature measurements were obtained from wet and dry bulb thermometers. Air supplies for generation and mixing were filtered, which assured clean particle-free air. Standard operation involved generating the aerosol continuously for a total of 25 minutes: 10 minutes to establish the desired humidity followed by a 5-minute sampling period and a final 10 minutes to recheck the humidity. The samplers removed three-fourths of the total inflow, consisting of 48 liters per minute. In order to minimize the radiological hazard, the system was housed in a ventilated chemical hood.

#### E. AEROSOL SAMPLERS

Five aerosol samplers were evaluated in comparative tests for the physical (radioactive) and viable collection of submicron T1 phage aerosols. They included two types of high-velocity liquid impingers, two ultra-high-efficiency filter papers, and a low-flow bubbler. Slip-page of the aerosol through the samplers was determined from backup filters placed in the sampler exhaust line. The following samplers were evaluated simultaneously for 5-minute sampling periods.

---

\* S.H. Emerson Co., 22 Cottage Park Avenue, Cambridge, Massachusetts.




Figure 2. Aerosol Apparatus.

### 1. All Glass Impinger (AGI-4)

This sampler is a high-velocity liquid impinger evaluated at the maximum (~ sonic) flow rate of 12½ liters per minute with 20 ml of Bacto nutrient broth containing 0.06% Dow Corning Antifoam A. The flow rate is controlled by a capillary orifice at the lower end of the inlet tube. The tip of the capillary is 4 mm from the flask bottom, hence the designation AGI-4.<sup>1</sup> The raised impinger (AGI-30) was not included in this study because preliminary tests with nonradioactive submicron T1 phage aerosols revealed that the AGI-4 was superior for viable collection at all three aerosol relative humidities.

### 2. Capillary Impinger

This sampler is similar in principle to the AGI-4 except that maximum (~ sonic) flow rate is 2.5 liters per minute. It was evaluated at an orifice-to-flask-bottom clearance of 8 mm with 25 ml of broth with antifoam.<sup>1</sup>

### 3. Fritted Bubbler

This sampler is an air washer-bubbler made from a fritted glass disc and a 250-ml flask. The fritted glass disc at the end of the inlet tube provides for uniform distribution of the bubbles and intimate gas-liquid washing. The sampler was operated at a low flow rate of 1.0 liter per minute to enhance collection by diffusion. The flow was controlled by a capillary orifice in the vacuum line. The fritted glass disc, 20 mm in diameter and of coarse porosity with a maximum pore size of 40 to 60 microns, was immersed in 150 ml of broth with antifoam.

### 4. Chemical Corps Type 6 Filter Paper

This material is an ultra-high-efficiency filter paper composed of cellulose, rope, and asbestos fibers. The filter discs (2.5 cm in diameter) were sealed in in-line filter holders with an effective area of 2.0 cm. Sampler flow rate was 1.0 liter per minute and therefore the filter face velocity was 8 cm per second. The flow rate was controlled by a capillary orifice placed in the vacuum line. For assay, the filter discs were placed, immediately after the test, in 100 ml of broth containing 0.1% Tween 20 and shaken vigorously for 10 minutes on a mechanical shaker. The Type 6 paper is easily disintegrated to pulp by shaking.

### 5. MSA 1106BH Glass Filter Paper

This material\* is an ultra-high-efficiency filter paper composed of very finely spun glass fibers and an organic binder to increase tensile strength. Unlike Type 6 paper, it does not break up with shaking.

### 6. Backup Filter

The sampler slippage was determined from in-line filter holders containing Type 6 filter paper backed up in the same holder by Chemical Corps Type 5 filter paper. The Type 5 is a low-efficiency paper of cellulose and asbestos fibers with a backing of cotton scrim. It served as a support to prevent rupture of the Type 6 paper and was assayed with the Type 6 paper.

The toxicity of the filter papers to phage and E. coli was investigated. None of the three filter papers is toxic to phage but Type 6 paper inhibits the growth of E. coli, which causes a reduction in phage plaque counts. Inhibition, presumably a bacteriostatic effect, is exhibited only when the filters are suspended in small volumes of broth and plated without dilution. Inhibition was demonstrated by suspending Type 6 filter discs in 20, 50, and 100 ml of Tween broth of a known phage concentration. Mean phage counts from 1-ml samples of the suspensions were 68% of controls for 20 ml, 87% for 50 ml, and 100% for 100 ml. In the sampling trials 100-ml blanks were used.

---

\* Mine Safety Appliances, Pittsburgh, Pennsylvania.

Sampler air flow rates were determined with a wet test meter. They were calibrated as they were to be used in the tests, i.e., with backup filters. The increase in resistance due to the backup filter reduced the AGI-4 flow rate to 95% of the maximum (~ sonic) flow. For the capillary impinger this value was 99%. The effect of this deviation from sonic velocity on impinger performance for submicron particles was not determined.

In the tests the samplers were attached to a vacuum manifold. The absolute pressure at the manifold was less than 0.4 atmosphere, sufficient to maintain pressure drops exceeding 0.5 atmosphere across the sampler orifices and assuring constant maximum flows.

#### F. AEROSOL PARTICLE SIZING

Preliminary tests of nonradioactive T1 phage aerosols with the Casella cascade impactor<sup>15</sup> and the Andersen sampler,<sup>16</sup> both of which are used for sizing viable microbial aerosols in the micron size range, revealed that T1 phage aerosols were somewhere in the submicron size range but no finite parameters could be calculated because more than 90% of the viable particles passed through these samplers. Solid gelatin, after the work of Dahlgren, Decker, and Harstad,<sup>18</sup> was the collecting medium used in the Andersen sampler. Finite particle size parameters of the radioactive T1 phage aerosols were determined with an electron microscope from aerosol samples collected with an electrostatic precipitator. Electrostatic precipitation sampling is the method of choice for sizing aerosols below the limits of resolution of the optical microscope.<sup>16</sup> The precipitator used in these tests is similar to those described by Billings and Silverman<sup>17</sup> and Morrow and Mercer<sup>18</sup> and consists of a glass tube (1.75 cm in diameter) holding two electrodes. One electrode is a needle. The opposing electrode is a brass holder for an electron microscope specimen grid. A potential of 10,000 volts a.c. is maintained across the electrodes and the distance between the electrodes is adjusted so that a corona is barely visible. This distance was about 1.8 cm for my instrument. An air flow rate of 1.0 liter per minute was used, giving a linear velocity through the precipitator of 7 cm per second. The specimen grids were collodion-coated and shadowed with carbon. A sampling time of 5 minutes was usually required to collect a sufficient number of particles on the grid. The grids were then shadowed with uranium and examined in the electron microscope at a magnification of 6890. Random areas from grids with a particle population of 25 to 100 particles per field were photographed. For sizing, the electron micrographs were projected on a screen of translucent paper to a total magnification of 50,000 (1 mm = 20  $\mu$ ). All particles were measured. It was not possible to distinguish phage-bearing particles from empty particles; therefore, only the physical or total aerosol particle size was determined. The individual particles were measured according to Martin's statistical diameter to the nearest  $\frac{1}{2}$   $\mu$ m. Martin's statistical diameter has the effect of avoiding bias in the measurement of irregular-shaped particles. It is

defined as the mean length of a line intercept by the profile boundary that approximately bisects the area of the profile. The bisecting line is taken parallel to a fixed direction, irrespective of the orientation of each particle.<sup>18</sup> For each size interval, the number of particles, percentage of the total number of particles, and the cumulative percentages were determined. The cumulative percentages, when plotted against particle diameter, gave a good straight-line fit on log probability paper. Therefore, the particle size distribution was log-normal. Populations of 200 to 500 particles were used to determine the following parameters:

- 1) Number median diameter (NMD)
- 2) Mass median diameter (MMD)
- 3) Geometric standard deviation (GSD)
- 4) Per cent by number less than 60  $\mu$  in diameter
- 5) Per cent by mass less than 60  $\mu$  in diameter

The geometric standard deviation, which indicates the degree of heterogeneity of the aerosol, was determined from the following equation:

$$\text{GSD} = \frac{84.13\% \text{ size}}{50\% \text{ size}} = \frac{50\% \text{ size}}{15.87\% \text{ size}}$$

The MMD was calculated from the NMD graphs by using the following equation derived by Hatch,<sup>20</sup> which is valid only if the distribution is log-normal.

$$\log \text{MMD} = \log \text{NMD} + 6.9078 \log^2 \text{GSD}$$

#### G. ASSAY

Aerosol samplers and phage suspensions were assayed in duplicate both for radioactivity and for T1 phage. Dilution for the phage assays reduced the soluble tracer to negligible amounts, terminating the exposure of phage to radiation. Samples were assayed for phage by making duplicate serial dilutions in nutrient broth and plating 1-ml samples of each dilution in triplicate, using an agar layer method similar to that described by Adams.<sup>21</sup> Fresh nutrient broth cultures of E. coli, aerated on a shaking machine at 37 C for 4 to 6 hours, were the seed cultures. Test tubes containing 3 ml of melted dilute Bacto nutrient agar (11 grams per liter) were inoculated with 0.5 ml of the seed culture, and 1 ml of the phage dilution was added. The mixture was swirled briefly and poured over the surface of a solid nutrient agar plate. The plate was rocked gently and allowed to harden. The melted dilute agar tubes were held at 46 C in a water bath prior to seeding with the E. coli culture. The phage plaques were counted after incubation at 37 C for 5 hours.

For the radioactive  $P^{32}$  assays, duplicate 2-ml samples were pipetted into aluminum planchets, evaporated to dryness overnight, and counted the following day in a windowless gas-flow proportional counter of 50% geometry. A strong attempt was made to keep the method of planchet preparation constant. In order to compare the results of the trials, the activity of the samples at zero time was computed from the observed counts by using the decay curve of  $P^{32}$  (half life = 14.3 days). Zero time was selected as the time when the specific (absolute) activity of the isotope solution, which was added to the phage suspension at the start of each trial, was 1.0 mc per ml. The specific activity of the samples was determined by calibrating the proportional counter with a known standard and making the appropriate correction for back-scatter. Self-absorption was negligible.

#### H. DESIGN OF EXPERIMENT

A series of 15 aerosol trials was conducted to compare five air sampling devices for the physical and viable collection of radioactive submicron T1 phage aerosols at relative humidities of 30, 55, and 85% and a temperature of 24 C. Five trials were conducted at each relative humidity, employing duplicate samplers of each type per trial.

### III. EXPERIMENTAL RESULTS

Data from 15 aerosol trials showing the tracer and phage concentrations in the Dautrebande suspension and in the test and backup samplers are summarized in Table 1. From these data the following parameters were calculated and listed in Tables 1, 2, and 3.

#### A. KILLING OF PHAGE BY THE DAUTREBANDE GENERATOR

There was considerable destruction (death) of phage by the Dautrebande  $D_{30}^1$  aerosol generator (Table 1). Radiological and phage assays of Dautrebande suspension at intervals during the 25-minute aerosolization period revealed an exponential decrease in phage titer and a linear increase in radiological titer. The exponential form of the equation that describes the rate of phage inactivation in the Dautrebande suspension is

$$\frac{a}{x} = \frac{a_0 e^{-\frac{t}{T}}}{x_0}$$

where  $a$  is the phage count at any time  $t$ ,  $a_0$  is the original count at zero time,  $x$  is the radioactivity at time  $t$ , and  $x_0$  is the original radioactivity at zero time. Viability, i.e., the fraction of phage surviving in the Dautrebande suspension at any time, is  $(a/x_0)/(a_0/x)$ . At the midpoint of the aerosol sampling period ( $t = 12\frac{1}{2}$  minutes) 54% of the phage particles that were viable at zero time were still viable.

#### B. RADIOLOGICAL AEROSOL CONCENTRATIONS

The Dautrebande generator concentrated the tracer in a linear manner and therefore the radiological aerosol concentrations increased with time (Table 1). The mean theoretical radiological concentration at the midpoint of the sampling period ( $t = 12\frac{1}{2}$  minutes) was  $16.6 \times 10^{-8}$   $\mu\text{c}$  per liter. It was derived from the radioactive and phage counts of the Dautrebande suspensions at  $12\frac{1}{2}$  minutes, fluid atomization rate, air flow rates for generation and mixing, and aerosol chamber volume. The effective radiological aerosol concentration was a mean of the actual total sampler recoveries. Total sampler (test + backup) recovery averaged  $6.3 \times 10^{-8}$   $\mu\text{c}$  per liter for the five trials at each of the three relative humidities. Therefore, about 60% of the aerosol was physically lost on passage through the apparatus.

TABLE 1. EFFECT OF AEROSOLIZATION AND SAMPLING ON THE VIABILITY OF T1 PHAGE

		Radioactivity, µc/ml	$10^{10}$ Phage/ml	Viability, per cent			
Dautrebande D <sub>30</sub> <sup>1</sup> Aerosol Generator <sup>a</sup> /	0 Time	10	15.5	100			
	12½ min	12	10.0	54			
	25 min	14	6.3	29			
Aerosol chamber		$10^{-3}$ µc/liter	$10^8$ phage/liter				
Theoretical <sup>b</sup> /		16.6	138	No basis for determining			
Effective <sup>c</sup> /		6.3	52.5				
<u>Test Sampler<sup>d</sup>/</u>	<u>% RH</u>	<u>Samplers</u>		<u>Samplers</u>		<u>Samplers</u>	
		<u>Test</u>	<u>Backup</u>	<u>Test</u>	<u>Backup</u>	<u>Test</u>	<u>Backup</u>
All-Glass Impinger (AGI-4)	30	3.3	3.0	2.3	0.73	8.4	2.9
	55	4.0	2.3	2.9	0.55	8.6	2.7
	85	4.4	1.9	0.89	0.30	2.4	1.9
Capillary Impinger	30	3.3	3.0	2.3	0.51	8.4	2.0
	55	3.4	2.9	2.8	0.50	9.8	2.1
	85	3.8	2.5	0.87	0.34	2.7	1.6
Type 6 Filter Paper	30	6.3	(0.001) <sup>e</sup> /	1.2	0.0006	2.3	(5) <sup>e</sup> /
	55	6.3	(0.001)	2.1	0.0013	4.1	(10)
	85	6.3	(0.004)	0.70	0.0006	1.3	(2)
1106BH Filter Paper	30	6.0	(<0.0006)	0.23	0.0006	0.5	(>10)
	55	6.0	(<0.0006)	0.34	0.0003	0.7	(>5)
	85	6.0	(0.002)	0.27	0.0003	0.5	(2)
Fritted Rubbler (coarse porosity)	30	1.2	5.1	0.56	0.44	5.6	1.0
	55	1.2	5.1	0.51	0.42	5.1	1.0
	85	1.1	5.2	0.28	0.40	2.9	0.9

a. Means of 15 aerosol trials.

b. Based on radioactive and phage counts of Dautrebande suspension at 12½ minutes, flow rates of suspension and air, and chamber volume.

c. Based on the actual total (test + backup) radioactive recoveries of all samplers, except 1106BH, from the 15 trials.

d. Means of 5 trials with 2 samplers per trial.

e. Approximate values based on additional tests of Table 3 that were not assayed for phage.

TABLE 2. COMPARISON<sup>a</sup> OF AIR SAMPLING DEVICES FOR VIABLE AND PHYSICAL COLLECTION OF RADIOACTIVE SUBMICRON T1 PHAGE AEROSOLS AT 30, 55, AND 85% RELATIVE HUMIDITY

Test Sampler	Flow Rate, liter/min	Relative Humidity, %	Relative		
			Efficiency, %	Viability, <sup>b</sup> %	Physical Slippage, %
All-Glass Impinger (AGI-4)	~ Sonic	30	100	48	24
	12.5	55	100	37	16
		85	100	30	25
Capillary Impinger	~ Sonic	30	100	48	18
	2.5	55	97	46	15
		85	98	40	28
Type 6 Filter Paper	1.0	30	52	-c/	0.05
		55	75		0.06
		85	79		0.09
1106BH Filter Paper	1.0	30	10	-	0.26
		55	12		0.09
		85	30		0.12
Fritted Bubbler (coarse porosity)	1.0	30	24	81	44
		55	18	81	45
		85	31	82	59

a. Each value represents mean of 10 samplers.

b. AGI-4 as the reference sampler:  $\frac{\text{Test sampler}}{\text{AGI-4}} \times 100$

Mean AGI-4 recoveries: 30% RH -  $2.3 \times 10^8$  phage/liter  
 55% RH -  $2.9 \times 10^8$  phage/liter  
 85% RH -  $8.9 \times 10^8$  phage/liter

c. Backup filter recoveries indistinguishable from background radiation.

TABLE 3. SLIPPAGE<sup>a/</sup> OF RADIOACTIVE SUBMICRON T1 PHAGE AEROSOLS THROUGH TYPE 6 AND 1106BH FILTER PAPERS AT 30, 55, AND 85% RELATIVE HUMIDITY

Test Filter <sup>b/</sup>	Flow Rate, liter/min	Relative Humidity, %	Physical Slippage, %
Type 6	1.0	30	0.02
		55	0.02
		85	0.07
1106BH	1.0	30	<0.01
		55	<0.01
		85	0.03

a. Each value represents mean of 5 samplers.

b. Test filter activity:  $7 \times 10^8$  counts per minute.

#### C. PHAGE AEROSOL CONCENTRATIONS

The phage concentrations (Table 1) in the aerosol chamber were calculated from the radiological aerosol concentrations and the ratio of phage to tracer  $a/x$  in the Dautrebande suspension at the midpoint of the sampling period ( $t = 12\frac{1}{2}$  minutes) according to the formula  $A = X(a/x)$ , where  $X$  is the radiological aerosol concentration (theoretical or effective) and  $a/x$  is the ratio of phage to tracer. There was no basis for determining the viability of the phage aerosol. Phage aerosols are subject to biological decay (death); therefore, the number of living phage particles expected to be available in the aerosol for sampling was unknown, but must be less than  $52.5 \times 10^8$  per liter.

#### D. KILLING OF PHAGE BY THE SAMPLERS

There was considerable destruction of phage by the samplers (Table 1). Viability refers to the fraction of phage collected by the samplers that remains viable. It is a measure of the destructiveness of the samplers for T1 phage and was calculated from the ratio of phage to tracer in the sampler and the ratio of phage to tracer in the Dautrebande suspension at  $t = 12\frac{1}{2}$  minutes:

$$\text{Viability} = \frac{\text{Sampler phage count/Sampler radioactivity}}{\text{Dautrebande phage count/Dautrebande radioactivity}}$$

Obviously, this did not compensate for loss of the phage aerosol due to biological decay (death), but gives a basis for comparing sampler kill. The high-velocity impingers, AGI-4 and capillary impinger, were the least destructive, followed by fritted bubblers, Type 6 filter paper, and 1106BH filter paper. Sampler kill was highest at 85% relative humidity and lowest at 55% relative humidity. The 1106BH filter paper was much more destructive than Type 6 filter paper.

#### E. RELATIVE VIABLE SAMPLER COLLECTION EFFICIENCY

Relative viable collection efficiency (Table 2) was based on the phage assay of test samplers and expressed as a percentage of the reference sampler, AGI-4, as given by:

$$\text{Relative viable efficiency} = \frac{\text{test sampler}}{\text{AGI-4}}$$

The high-velocity impingers, AGI-4 and capillary impinger, gave the highest viable recoveries, followed by Type 6 filter paper, fritted bubblers, and 1106BH filter paper. Both types of filter paper improved with an increase in aerosol relative humidity, but Type 6 was superior to 1106BH paper for viable collection. Viable recoveries were highest at 55% relative humidity and lowest at 85% relative humidity.

#### F. SAMPLER PHYSICAL SLIPPAGE

Physical slippage (Table 2) was based on the radiological assay of test samplers and backup filters as given by:

$$\text{Physical slippage} = \frac{\text{Backup filter}}{\text{Backup filter} + \text{test sampler}}$$

The high-velocity impingers, which are widely used for micron-sized aerosols such as bacteria where slippage is less than 1%, were not efficient physical collectors of submicron particles. Impinger slippage decreased with an increase in aerosol relative humidity, presumably because of the change in aerosol particle size with relative humidity. However, the differences in particle size at the three aerosol relative humidities were small. The fritted bubbler, a sampler designed to collect particles by the diffusion principle, was the poorest submicron sampler, in that physical slippage was greater than 80%.

#### G. SAMPLER VIABLE SLIPPAGE

Viable slippage (Table 2) was based on the phage assay of test samplers and backup filters as given by:

$$\text{Viable slippage} = \frac{\text{Backup filter}}{\text{Backup filter} + \text{test sampler}}$$

As expected, viable slippage was less than physical slippage because backup filters (Type 6 paper) as well as the test samplers themselves are destructive to phage. For both types of impingers viable slippage was lowest at 55% relative humidity. The viable slippage parameter, although it is an obvious way of presenting the data, is not very useful; the other parameters are more meaningful in describing sampler performance.

#### H. FILTER PAPER PHYSICAL SLIPPAGE

The filter papers were the most efficient samplers for the physical collection of submicron aerosols (Table 3), but finite slippage values were not obtained because backup recoveries were indistinguishable from background radiation. Additional tests, shown in Table 3, were conducted using much higher effective radiological concentrations in an attempt to obtain finite values. The filter papers were not assayed for phage in these tests. The effective aerosol concentration was increased 120 times by increasing the sampling period to 150 minutes and suspending the filters in a smaller volume (25 ml) of fluid, but again backup recoveries were low. One aerosol trial was conducted at each relative humidity, using five samplers of each filter paper per trial. Finite slippage values are shown only where backup filter recoveries were significantly above background. Both filter papers were very efficient for submicron particles but 1106BH was superior to Type 6 paper for physical collection. Slippage was highest for both papers at 85% relative humidity.

Not all the radioactivity was removed from the 1106BH filters by shaking. The 1106BH paper does not disintegrate into pulp as does Type 6 paper, and recoveries were slightly lower than the total or actual radioactivity. This was demonstrated by centrifugation techniques involving the analysis of the supernatant fluid and sediment and in the sampling trials (Table 1), where 1106BH filter paper recoveries were 95% of Type 6 recoveries, which, by the same techniques, were shown to be representative of the total activity.

## I. AEROSOL PARTICLE SIZE

Aerosol particle size was adequately described by a log-normal distribution. The particle size parameters of radioactive T1 phage aerosols at 30, 55, and 85% relative humidity are shown in Table 4. Typical results are shown in Figures 3 and 4. Figure 3 is a graph of the 55% relative humidity aerosol of Day 1, showing the cumulative percentages by number and by mass plotted against particle diameter on log probability paper. The median diameters (50% intercepts) are 119  $\mu$  MMD and 172  $\mu$  MMD. Figure 4 is an electron micrograph of this aerosol. Differences between the particle size parameters at the three relative humidities were small but the particles were slightly larger at 85% relative humidity.

A radioactive tracer was used in this study to follow both the physical (radioactive) and viable presence of the aerosol particles. However, there was a deviation from this concept because, while all the aerosol particles were radioactive, a small percentage of the aerosol particles were smaller than the size of a single phage particle. In sizing the aerosols from electron micrographs it was not possible to distinguish phage-bearing particles from empty particles, but all particles were measured; therefore, only the total or physical aerosol particle size was determined. Phage assays are a function of the number of particles and radiological assays are a function of mass. Inspection of the MMD graphs reveals that less than 5% of the aerosol particles were smaller than a single phage particle (60  $\mu$  diameter). Therefore, many of the aerosol particles contained more than one phage particle and disintegration of phage clumps by the samplers could be a possible cause of sampler differences for viable collection. However, not all of the aerosol particles larger than 60  $\mu$  contained phage, because aerosol particle size is related to the amount of solid material in the Dautrebande suspension and it cannot be assumed that all of the extraneous material was removed from the phage suspensions by the purification process. The 60- $\mu$  intercept of the MMD graphs was less than 0.2%, i.e., the fraction of the total aerosol radioactivity associated with the known empty particles was less than 0.2%.

TABLE 4. PARTICLE SIZE PARAMETERS OF RADIOACTIVE T1 PHASE AEROSOLS  
AT 30, 55, AND 85% RELATIVE HUMIDITY

Relative Humidity, %	Day	Number of Particles	Median Diameter, $\mu$		Less than 60 $\mu$		Geometric Standard Deviation
			Number	Mass	Number	Mass	
30	1	219	126	167	1	0.04	1.36
	2	382	114	181	5	0.2	1.48
55	1	485	119	172	2	0.1	1.42
	2	387	117	186	4	0.2	1.48
85	1	225	126	194	2	0.08	1.45
	2	272	132	209	2	0.06	1.48




Figure 3. Size Distribution of Radioactive T1 Phage Aerosol at 55% Relative Humidity Dispersed with Bantrolane D<sub>30</sub> Aerosol Generator.




Figure 4. Electron Micrograph of Radioactive T1 Phage Aerosol at 55% Relative Humidity Dispersed with Bantrolane D<sub>30</sub> Aerosol Generator.

#### IV. DISCUSSION

All of the samplers tested were partial collectors of viable submicron phage aerosols. Viable collection, unlike physical slippage, is specific for T1 phage and is a function of both the physical and biological properties of the aerosol. Sampler differences for viable collection were attributed to incomplete physical collection (slippage) and destruction (death) during the sampling process. Other possible causes believed to be of less importance are disintegration of clumps by the sampler and retention in the sampler, e.g., inability to remove particles completely from 1106BH filter paper. Physical slippage and disintegration are related to the aerosol particle size. Destruction may also be a function of particle size. Physical slippage was quantitated but it was not possible to distinguish between destruction and disintegration because (i) phage aerosols are subject to biological decay, therefore the viable aerosol concentration at the moment of sampling was unknown, and (ii) radiological assays are a function of mass and not the number of particles; therefore, they do not reflect breakup of clumps. However, there was strong but indirect evidence of phage destruction by the samplers and phage inactivation by the Dautrebande generator was proved. The high-velocity impingers, AGI-4 and capillary impinger, gave the highest viable recoveries but allowed considerable slippage. Impinger physical slippage varies inversely with relative humidity, presumably because of changes in particle size with relative humidity. However, the particle size differences at the three relative humidities were small, indicating that impinger slippage is extremely sensitive to changes in the size of submicron particles. Type 6 filter paper is virtually a complete physical collector of submicron particles but cannot, in a strict sense, be classified as a representative sampler because destruction may be a function of aerosol particle size.

LITERATURE CITED

1. Wolf, H.W., P. Skaliy, L.B. Hall, M.M. Harris, H.M. Decker, L.M. Buchanan, and C.M. Dahlgren. 1959. Sampling microbiological aerosols. Public Health Monograph No. 60. U.S. Government Printing Office, Washington, D.C.
2. Webb, S.J., R. Bather, and R.W. Hodges. 1963. The effect of relative humidity and inositol on airborne viruses. *Can. J. Microbiol.* 9:87-92.
3. Dautrebande, L. 1962. Microaerosols, p. 1-22. Academic Press, Inc., New York.
4. Stern, S.C., J.S. Baumstark, A.I. Schekman, and R.K. Olson. 1959. Simple technique for generation of homogeneous millimicron aerosols. *J. Appl. Physics* 30:952-953.
5. Morris, E.J., H.M. Darlow, J.F.H. Peal, and W.C. Wright. 1961. The quantitative assay of mono-dispersed aerosols of bacteria and bacteriophage by electrostatic precipitation. *J. Hyg.* 59:487-496.
6. Harper, G.J., A.M. Hood, and J.D. Morton. 1958. Airborne microorganisms: A technique for studying their survival. *J. Hyg.* 56: 364-370.
7. Miller, W.S., R.A. Scherff, C.R. Piepoli, and L.S. Idoine. 1961. Physical tracers for bacterial aerosols. *Appl. Microbiol.* 9:248-251.
8. Harper, G.J. 1961. Airborne microorganisms: Survival tests with four viruses. *J. Hyg.* 59:479-486.
9. Tyler, M.E., E.L. Shipe, Jr., and R.B. Painter. 1959. Bacterial aerosol samplers: III. Comparison of biological and physical effects in liquid impinger samplers. *Appl. Microbiol.* 7:355-362.
10. Puck, T.T. 1949. A reversible transformation of T1 bacteriophage. *J. Bacteriol.* 57:647-655.
11. Herriott, R.M., and J.L. Barlow. 1952. Preparation, purification and properties of *E. coli* virus T2. *J. Gen. Physiol.* 36:17-28.
12. Decker, H.M., L.M. Buchanan, L.B. Hall, and K.R. Goddard. 1962. Air filtration of microbial particles. Public Health Service Publication No. 953. U.S. Government Printing Office, Washington, D.C.
13. Sonkin, L.S. 1950. Application of the cascade impactor to studies of bacterial aerosols. *Amer. J. Hyg.* 51:319-342.

14. Andersen, A.A. 1958. New sampler for collection, sizing, and enumeration of viable airborne particles. *J. Bact.* 76:471-484.
  15. Dahlgren, C.M., H.M. Decker, and J.B. Harstad. 1961. A slit sampler for collecting T3 bacteriophage and Venezuelan Equine Encephalomyelitis virus: I. Studies with T3 bacteriophage. *Appl. Microbiol.* 9:103-105.
  16. Fraser, D.A. 1956. The collection of submicron particles by electrostatic precipitation. *Amer. Ind. Hyg. Assoc. Quart.* 17:75-79.
  17. Billings, C.E., and L. Silverman. 1962. Aerosol sampling for electron microscopy. *J. Air Pollut. Control Assoc.* 12:586-590.
  18. Morrow, P.E., and T.T. Mercer. 1964. A point-to-plane electrostatic precipitator for particle size sampling. *Amer. Ind. Hyg. Assoc. J.* 25:8-14.
  19. Herdan, G. 1953. Small particle statistics, p. 65-68. Elsevier Publishing Co., Amsterdam.
  20. Hatch, T. 1933. Determination of average particle size from screen analysis of non-uniform particulate substances. *J. Franklin Inst.* 207:27.
  21. Adams, M.H. 1959. Bacteriophages, p. 430-451. Interscience Publishers, Inc., New York.
-

DISTRIBUTION LIST

<u>ADDRESSEE</u>	<u>NUMBER OF COPIES</u>
Assistant Scientific Director Building 812	1
Directorate of Biological Research Building 560	1
Directorate of Industrial Health & Safety Building 550	1
Acting Director of Medical Research Building 538	1
Chief, Program Coordination Office Building 812	1
Chief, Aerobiology Division Building 459	1
Chief, Medical Bacteriology Division Building 560	1
Chief, Medical Investigation Building 604	1
Chief, Physical Sciences Division Building 568	2
Chief, Process Development Division Building 469	1
Chief, Technical Evaluation Division Building 568	1
Chief, Physical Defense Division Building 521	10
Chief, Biomathematics Division Building 1422	1
Documents, Technical Library Building 426	2

<u>ADDRESSEE</u>	<u>NUMBER OF COPIES</u>
Test Chamber Branch Technical Evaluation Division Building 1412	1
Technical Releases Branch Technical Information Division Building 426	10
Editorial Branch Building 816	1
U.S. Army Medical Unit Building 120	1
Liaison Representative Animal Disease Investigations Building 1301	1
U.S. Public Health Service Liaison Office Building 1301	6
Commanding Officer U.S. Naval Unit Building 125	3
Commanding General U.S. Army Edgewood Arsenal ATTN: SMUEA-CS Edgewood Arsenal, Maryland, 21010	1
Commanding Officer U.S. Army Chemical Research & Development Laboratories ATTN: Librarian Edgewood Arsenal, Maryland, 21010	2
Commanding General U.S. Army Munitions Command ATTN: AMSMU-SS-CS Dover, New Jersey, 07801	1
Commanding General U.S. Army Munitions Command ATTN: AMSMU-RE-R Dover, New Jersey, 07801	1
Commandant U.S. Army CBR Weapons Orientation Course Dugway Proving Ground Dugway, Utah, 84022	1

ADDRESSEENUMBER OF COPIES

Commanding General  
Deseret Test Center  
ATTN: Technical Library  
Fort Douglas, Utah, 84113

2

Commanding General  
U.S. Army Materiel Command  
Research Division, AMGRD-RC  
R&D Directorate  
Washington, D.C., 20315

1

Defense Documentation Center  
Cameron Station  
Alexandria, Virginia, 22314

20

AFRSTA, Hq. USAF  
ATTN: Mr. C.R. Nixon, Jr.  
Washington, D.C., 20330

1

Detachment 4, RTD (ATCB)  
Eglin Air Force Base, Florida, 32542

1

APGC (PGBAP-1)  
Eglin Air Force Base, Florida, 32542

1

6570 AMRL  
MRMP14 (D. S. A. London)  
Wright-Patterson Air Force Base, Ohio, 45433

1

Dr. S.H. Madin  
Scientific Director  
Naval Biological Laboratory  
Naval Supply Center  
Oakland, California, 94614

1

Commander (Code 4036)  
U.S. Naval Ordnance Test Station  
China Lake, California, 93557

1

Commanding Officer and Director  
U.S. Naval Applied Science Laboratory  
Naval Base, Code 9440  
Brooklyn, New York, 11251

1

U.S. Army Medical R&D Command  
Office of the Surgeon General  
ATTN: MEDDHC  
Main Navy Building, Room 2526  
Washington, D.C., 20315

1

<u>ADDRESSEE</u>	<u>NUMBER OF COPIES</u>
Commandant USA3miGen & Sch, ATTN: Bio Br Fort McJannet, Alabama, 36205	1
U.S. Army Standardization Group - Canada Office, Senior Standardization Rep. c/o Director of Equipment Policy Canadian Army Headquarters Ottawa 4, Canada	1
Munitions/TW Defence Research Staff British Embassy 3100 Massachusetts Avenue, N.W. Washington 8, D.C.	3
Canadian Liaison Office (CBR) Building 5101 Edgewood Arsenal, Maryland, 21010	3
Australian Embassy ATTN: Lt. Col. P.D. Yonge Australian Army Staff (W) 2001 Connecticut Avenue, N.W. Washington 7, D.C.	2

Best Available Copy

**DOCUMENT CONTROL DATA - R&D**

*(Security classification of title, body of abstract and indexing annotation must be entered when the security report is classified)*

1. ORIGINATING ACTIVITY (Corporate author) U.S. Army Biological Laboratories Fort Detrick, Frederick, Maryland, 21701		2a. REPORT SECURITY CLASSIFICATION Unclassified	
		2b. GROUP	
3. REPORT TITLE SAMPLING SUBMICRON T1 BACTERIOPHAGE AEROSOLS			
4. DESCRIPTIVE NOTES (Type of report and inclusive dates)			
5. AUTHOR(S) (Last name, first name, initial) Harstad, J. Bruce			
6. REPORT DATE May 1965		7a. TOTAL NO. OF PAGES 32	7b. NO. OF REFS 21
8a. CONTRACT OR GRANT NO.		9a. ORIGINATOR'S REPORT NUMBER(S) Technical Manuscript 218	
b. PROJECT NO. IC622401A072		9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report)	
c.			
d.			
10. AVAILABILITY/LIMITATION NOTICES Qualified requestors may obtain copies of this report from DDC. Foreign announcement and dissemination of this report by DDC is not authorized.			
11. SUPPLEMENTARY NOTES		12. SPONSORING MILITARY ACTIVITY U.S. Army Biological Laboratories Fort Detrick, Frederick, Maryland, 21701	
13. ABSTRACT Liquid impingers, filter papers, and fritted bubblers were partial viable collectors of radioactive submicron T1 bacteriophage aerosols at 30, 55, and 85% relative humidity. Sampler differences for viable collection were due to incomplete physical collection (slippage) and killing of phage by the samplers. Dynamic aerosols of a mass median diameter of 0.2 micron were produced with a Dautrebande generator from concentrated aqueous purified phage suspensions containing extracellular soluble radioactive phosphate as a physical tracer. There was considerable destruction of phage by the Dautrebande generator; phage titers of the Dautrebande suspension decreased exponentially but there was a progressive (linear) increase in tracer titers. Liquid impingers recovered the most viable phage but allowed considerable (30 to 48%) slippage, which varies inversely with the aerosol relative humidity. Filter papers were virtually complete physical collectors of submicron particles but were the most destructive. Fritted bubbler slippage was more than 80%. With all samplers, phage kill was highest at 85% relative humidity and lowest at 55% relative humidity. An electrostatic precipitator was used to collect aerosol samples for particle sizing with an electron microscope. The particle size was slightly larger at 85% relative humidity than at 30 or 55% relative humidity.			