

Award Number: W81XWH-11-1-0197

TITLE: Eye TVR: Eye Trauma & Visual Restoration Team

PRINCIPAL INVESTIGATOR: Dr. Shuliang Jiao

CONTRACTING ORGANIZATION: University of Southern California
Los Angeles, CA 90089

REPORT DATE: March 2012

TYPE OF REPORT: Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for public release; distribution unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

REPORT DOCUMENTATION PAGE				Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.					
1. REPORT DATE (DD-MM-YYYY) March 2012		2. REPORT TYPE Annual		3. DATES COVERED (From - To) 18 Feb 2011 - 17 Feb 2012	
4. TITLE AND SUBTITLE Eye TVR: Eye Trauma & Visual Restoration Team				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER W81XWH-11-1-0197	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S) Dr. Shuliang Jiao E-Mail: sjiao@usc.edu				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) University of Southern California Los Angeles, CA 90089				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION / AVAILABILITY STATEMENT Approved for Public Release; Distribution Unlimited					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT The EyeTVR team conducts applied research aimed at saving and restoring sight to the war fighter who suffers traumatic eye injury. The larger goal is also to help veterans and their family members who suffer from eye injury, eye infection, and inherited or age-related eye disease. The research achieved the following aims: (1) developed a dual-band OCT system for investigating the spectral contrast for the diagnosis of early damage to the retinal nerve fiber layer; (2) investigated millimeter wave stimulation for functional retinal diagnostics; (3) conducted experiments on active sensor for mobility aid to blind; (4) developed an advanced neural interface system for a cortical visual prosthesis.					
15. SUBJECT TERMS Optical coherence tomography, image segmentation, image registration					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT	18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON
a. REPORT U	b. ABSTRACT U	c. THIS PAGE U			USAMRMC
			UU	18	19b. TELEPHONE NUMBER (include area code)

TABLE OF CONTENTS

	Page
Introduction.....	4
Body	4
Reportable outcomes.....	11
Conclusion	11
APPENDICES	12

INTRODUCTION

A March 9, 2008 article in the *San Francisco Chronicle* (“*New treatments for traumatic eye injuries*”) highlighted the incidence of eye injuries in the Iraq war (more than double those in World War II). Eye wounds have become among the most common - and devastating - form of battlefield injury, with an estimated 10 to 13% of wounded Iraq veterans sustaining direct, penetrating eye damage. Serious eye wounds have accumulated at almost twice the rate as wounds requiring amputations. In a recent article in the *Army Times*, it was estimated that 1,400 troops have vision problems caused by combat injuries. According to the Blinded Veterans Association, 75% of service members with traumatic brain injuries also have vision problems including double vision, lack of peripheral vision, color blindness and total blindness.

The EyeTVR team conducts applied research aimed at saving and restoring sight to the war fighter who suffers traumatic eye injury. The larger goal is also to help veterans and their family members who suffer from eye injury, eye infection, and inherited or age-related eye disease. The research projects are in the following two groups:

1. **Functional Imaging** — to develop innovative methods to visualize retinal structure to determine whether cells of the retina and optic nerve are functioning properly.
Tasks: a. Spectral contrast for the diagnosis of early damage to the retinal nerve fiber layer.
b. Millimeter wave stimulation for functional retinal diagnostics.
2. **Visual Prosthesis and Biomaterials** — to develop materials to promote eye wound healing and attachment of devices to the retina and the visual centers of the brain, enabling the functional placement of visual prosthesis or drug delivery systems.
Tasks: a. Active sensor for mobility aid to blind.
b. An advanced neural interface system for a cortical visual prosthesis.

BODY

Task 1a: Spectral contrast for the diagnosis of early damage to the retinal nerve fiber layer




Fig. 1 Calibrated OCT parameters in the VIS and NIR spectra. (a): the measured VIS spectrum; (b): the measured NIR spectrum; (c) the measured point spread function (PSF) of the VIS OCT; (d) the measured PSF of the NIR OCT.

In this year we worked on building the OCT systems in the visible (VIS) and near infrared (NIR) spectra. The light source for the two OCT systems is a broadband Ti:Sapphire ultra-fast laser centered at 800 nm (Micra-5, Coherence, USA). We use a BBO crystal (CASTECH Inc, China) to double the light frequency. To increase the bandwidth of the frequency doubled light, the fundamental light is focused onto the crystal by an objective lens with a focal length of 14 mm. We used a long-pass filter (FEL0600, Thorlabs, USA) to separate the frequency-doubled light and the residue fundamental light. After the long-pass filter the separated VIS and NIR light beams are coupled into the source arms of their corresponding fiber-based Michelson interferometers. The spectrum of the VIS light is shown in Fig. 1a, which has a center wavelength of 415 nm and a FWHM bandwidth of about 10 nm. The measured point spread function (PSF) of the VIS OCT is shown in Fig. 1c. The calibrated depth resolution is 12.2 μm in air. The spectrum of the NIR light is shown in Fig. 1b, which has a center wavelength of 800 nm and a FWHM bandwidth of about 100 nm. The measured PSF of the NIR OCT is shown in Fig. 1d. The calibrated depth resolution is 4.7 μm in air.

We then integrated the two OCT systems on a slit lamp. After the animal protocol was approved by ACURO we conducted animal experiments. The system was applied to imaging the normal rat retina *in vivo*. The animals (Sprague Dawley rats, body weight: 600~750 g, Charles Rivers) were anesthetized by intraperitoneal injection of a cocktail containing ketamine (54 mg/kg body weight) and xylazine (6 mg/kg body weight). In the meantime, the pupils were dilated with 10% phenylephrine solution. Drops of artificial tears were applied to the eyes every two minutes to prevent dehydration of the cornea and cataract formation. After anesthetization, the animals were restrained in a mounting tube, which was fixed on a five-axis platform. Raster scans with the fast axis along the horizontal direction were performed. All experiments were performed in compliance with the guidelines of the University of Southern California's Institutional Animal Care and Use Committee.

In the experiment, we collected the NIR and VIS-OCT images sequentially. The initial alignment including finding the area of interest was done with the NIR OCT. After the acquisition of the NIR OCT image we mechanically switched the VIS OCT into position and acquired the VIS OCT images. In our current study the laser energy after the objective lens was measured to be 450 μW for the visible and 750 μW for the NIR light, which are both below the ANSI safety limits for eye imaging. Since the OCT images of the two different bands were not acquired simultaneously they are not completely registered due to movement of the animal during the switching process. The advantage of using the NIR-OCT for alignment is to reduce the retinal exposure to the visible light. The higher axial resolution and greater imaging depth of the NIR-OCT also aided the alignment.




Fig. 2 Sample images of a rat retina acquired with the NIR OCT and VIS OCT. Images consist of 2048 A-lines. Bar: 100 μm .

Figure 2 shows one pair of the acquired NIR and VIS OCT images. The images consist of 2048 A-scans and cover a length of about 1.3 mm across the retina. The differences between the NIR and VIS OCT images are significant. As expected the blood vessels blocked the VIS probing light completely with only reflections from their front boundaries, while the NIR-OCT still revealed some features beneath the vessels. This difference comes from the significant difference of the absorption coefficients of hemoglobin at the two bands: at 415 nm $\mu_a = 2801.52 \text{ (cm}^{-1}\text{)}$ for oxy-hemoglobin (HbO_2) and $\mu_a = 1891.79 \text{ (cm}^{-1}\text{)}$ for deoxy-hemoglobin (Hb); at 808 nm $\mu_a = 4.58 \text{ (cm}^{-1}\text{)}$ for HbO_2 and $\mu_a = 3.88 \text{ (cm}^{-1}\text{)}$ for Hb .

More detailed description on the animal experiments can be found in the attached peer reviewed publication in Optics Express.

Task 1b: Millimeter wave stimulation for functional retinal diagnostics

The overall goal of this project is to develop a technology for non-invasive neuromodulation of retinal activity. Our approach is to measure the neuronal activity in vivo while the eye is exposed to the focused beam of millimeter waves (MMW). Considerable progress was made toward all the milestones set in the sub-project.

1. Development of real-time user-friendly interface for measuring evoked retinal responses

The software for optical/MMW stimulation and multichannel neural data acquisition was developed using the Visual Basic 6.0 (Microsoft Corp.) and Measurement Studio ActiveX components (National Instruments Inc.). Fig. 3 shows a sample screenshot of the graphical user interface:



Fig. 3. Graphical user interface for neural data acquisition.

The interface accommodates multiple electrode array types, ranging from 4 to 32 channels. The data acquisition is hardware-synchronized with the generation of the TTL pulse that can be used to trigger an optical or MMW stimulation. The second program was created for off-line viewing of the recorded data and its export to Matlab for further analysis.

2. Measurement of retinal responses to calibrated optical flash stimuli with 10% or less within-animal within-session error

Our original goal was to use the electroretinography (ERG), however we quickly realized that we would like to get the information about the spatial distribution of the stimulus in the retina, which is not available in the single-channel ERG. We have examined three techniques potentially suitable for recording spatially-resolved retinal responses:

- a. epiretinal recordings,
- b. subretinal recordings,
- c. visual cortex recordings.

To enable such recordings, we have evaluated three types of flexible multi-electrode arrays (MEAs): 4-channel with 200-um diameter sites, 15-channel with 400-um diameter sites, and 15-channel with 50-um

diameter sites. These MEAs feature different geometries and flexibility. Fig. 4 shows the sample insertions of the 15-channel MEA into the epiretinal and subretinal space in the rat.




Fig. 4. Photographs of the epiretinal (A) and subretinal (B) placement of the flexible MEA.

Disappointingly, neural recordings could not be obtained from epiretinal and subretinal spaces. We speculate that our inability to record is caused by insufficient proximity of the MEA to the retina on the epiretinal side and by a lack of firing neuronal cells on the subretinal side (photoreceptors and bipolar cells do not fire many action potentials). We then explored whether the flexible MEAs can be used for subdural recordings from the V1 cortical region. Considering the 1-mm distance of the MEA to the layer 3 of the cortex, it is practically impossible to record the multi-unit neuronal activity. Instead, we decided to record the evoked local field potentials (eLFPs). Initial calibration of our software and hardware setup was successfully done by electrically stimulating the hindlimb muscle while recording in the contralateral M1 motor cortex with the MEA. In the second year, we plan to apply the optical and focused MMW stimulation of the retina to evaluate the feasibility of recording eLFPs in the V1 cortex.

3. Fabrication of the MMW delivery system




Fig. 5. Photograph of the MMW delivery system.

The fabricated MMW exposure system is based on a commercial 60-GHz source with a built-in power amplifier, providing the power levels between 1 μ W and 30 mW (Fig. 5). The narrow-angle far-field horn has been purchased to provide an efficient coupling between the source and the exposed eye with an aberration-free Gaussian distribution of the energy from the horn opening. The selected power supply (E3630A, Agilent) provides a stable switching-transient-free operation of the MMW source. The MMW delivery system is mounted on the Z-axis manipulator to allow precise positioning of the horn opening relative to the eye.

Task 2a: Active sensor for mobility aid to blind

RGB-D Camera based SLAM: Accurate disparity maps are a key element in carrying out successful navigation and obstacle avoidance tasks. Most stereo matching algorithms that generate disparity maps are formulated to minimize a pre-defined energy function. Although many advanced stereo matching have been proposed, none of them provide reliable results in textureless environments. Recently, a cost-effective RGB-D sensor developed by Primesense was introduced. It is the reference design for the Microsoft Kinect game interface. The RGB-D camera uses Infra-Red to obtain depth information registered to RGB images, and streams both a depth map and RGB images at 30 FPS. The sensor is shown in Fig. 6. The sensor was used to provide more accurate disparity information of indoor textureless regions in our application. The two main goals of augmentation of the sensor in our application are as follows. First, we aim to have baseline for performance comparison of stereo sensors in textureless Indoor environments. Second, we wish to test the feasibility of a stereo and RGB-D camera hybrid sensor for the navigation tasks.



Fig. 6. Primesensor, a RGB-D sensor.




Fig. 7. Disparity map experiments in low textured areas. Left: RGB camera images. Center: Disparity maps generated by a stereo camera. Right: Disparity maps generated by a RGB-D camera.

As depicted in Fig. 7, a stereo camera based system does not provide correct disparity information in low texture scenes as shown in the 2nd column compared to those generated by Primesensor indicated in the 3rd

column. As a consequence, a blind subject may receive incorrect feedback information, heading to possible collision. In order to verify the performance of a RGB-D sensor in a guidance system, experiments were conducted for a simple navigation task.

A subject starts from one end of a corridor and aims to reach the other end of the corridor. As shown in Fig. 8, the experiment environment has solid walls and reflective floors, which is quite common in indoor environments. Fig. 8 shows that the 3D depth map from the stereo camera is relatively inaccurate and results in inconsistency of the traversability map in the experiment area. From the 2nd column of the 2nd row of the left side of Fig. 8 and on, errors accumulate when building the traversability map of an environment consisting of red and green dots. This spurious information in the map causes the navigation algorithm to fail to find an appropriate way point and a safe path. As described in the 4th column of the left side of Fig. 8, the stereo camera based navigation system hallucinates obstacles that do not exist, and steers a subject in the wrong direction. In the right side of Fig. 8, we replace the stereo system by a RGB-D camera. The dense traversability map is significantly more accurate and consistent with the real corridor environment. It is worth noting that the low textured experiment area still affects the visual odometry performance of the system since we perform visual odometry on RGB images. However, traversability map and navigation worked much more robustly than the navigation system with a stereo camera pair. The performance and reliability of the system improved. In conclusion, RGB-D camera sensor can be a successful alternative of a stereo camera based system when navigating indoors, while a stereo camera based system can still work effectively for outdoor navigation tasks, the sensors effectively complementing each other.




Fig. 8. Four snapshots of traversability map of the system with a pair of stereo camera (left) and with Primesensor (right). More accurate 3D estimation of the indoor environment was achieved by using RGB-D sensor which leads more robust guidance.

Task 2b: An advanced neural interface system for a cortical visual prosthesis

A visual prosthesis that employs multiple arrays of microelectrodes implanted into the primary visual cortex affords the opportunity for restoring useful vision to persons who have become blind due to injuries that render them unsuitable for a retinal level visual prosthesis. However, there is little data on the optimal depth(s) or design for intracortical stimulation electrodes in humans. Our task is to design a 4-shank, 16-site microstimulating array with integrated cable suitable for animal testing, and validate an expected lifetime of at least 2 years *in vivo* by accelerated soak-testing at 70°C. Begin long-term soak tests at 37°C (YR 01). The ultimate goal is a device with a project lifetime of 20 years *in vivo*.

Subtask 2b.1.1 Design a 4-shank, 16-site microstimulating array

A preliminary design of the microstimulating probe suitable for implantation into cat cerebral cortex has been completed (Fig. 9). The array will contain 4 silicon shanks, each with 4 electrode sites. Each shank will be 2 mm in length, to span the full depth of the feline sensory-motor cortex. Unlike the thin “Michigan probes” we will use mechanically-sturdy probe shanks fabricated by DRIE and with a special tip-shaping technique which can penetrate through the pia mater covering the human cerebral cortex. The surface areas of the stimulating sites will be $2,000 \mu\text{m}^2$. Each of the electrode sites will be coated with sputtered iridium oxide film (SIROF) to increase their charge injection capacity.

Subtask 2b.1.2 Develop integrated cable suitable for chronic implantation

We are developing an integrated flexible Parylene cable that will be interfaced to the microelectrode array. Unlike the conventional way, the integrated cable is fabricated by batch processing on silicon wafers, and reduces the bulk and stiffness of cables with many conductors. Our approach is to coil the batch-processed integrated cables around a segment of .025-inch diameter silicone tubing, and then overcoat the assembly with a layer of silicone elastomer. This approach will enhance the durability and longevity of the electrical connections, and allow the cable to flex in all directions, unlike a flat cable which cannot flex laterally in the same plane. We evaluated several design parameters (Fig. 10).




Fig. 9: Schematic of cortical visual prosthetics array based on thick silicon substrate.




Fig. 10: Preliminary designs and simulation of integrated cables: (a) sample placements of slightly curved cables, placed in 4-in diameter space, (b) Trial parameters for the planar cable to expedite winding into a helix (c) a view of flexible cable after it is wound around a section of silicone tubing .

We conducted a complete wafer fabrication cycle for the parylene cable, in order to validate all of the process steps. This run was to establish the processing parameters for subsequent runs which will yield usable cables (Figure 11). The run encompassed the following operations: 1. Fiducial etching onto the silicon wafer, 2. Coating the wafer with the base Parylene-C layer, 3. deposition of gold electrode sites and conductor traces (Metallization) 4. Deposition of the top Parylene-C layer, 5. Opening of the top layer to expose electrodes sites, and 6. Shaping of the device by etching through the two Parylene C layers. This run will serve as a model for the subsequent rung that will produce cables that can be integrated with the penetrating silicon electrode carrier and attached to a percutaneous connector.

Subtask 2b.1.3 Develop a method of bonding electrode array and the integrated cable

Because of the soft nature of these cable substrates, conventional bonding methods such as wedge or ball bonding alone will cannot be used to bond the flexible conductors to the silicon electrode arrays. We tested anisotropic adhesives as an alternative to mechanical pressure bonding as well as ball-bonding as part of rivet bonding (Fig. 12). The anisotropic




Figure 11: Electrodes and metal traces fabricated between two flexible Parylene C layers. The rectangular electrodes are free of the top Parylene C layer.

adhesive must be approximately 25 μm or thinner when cured to be electrically conductive, which requires application of considerable compressive force without disturbing the alignment between the points of electrical contact. This was somewhat challenging given the small dimension of the bond pads on the Parylene device (75 μm x 100 μm). In the case of board-to-board condition (Fig. 11(a)), the connection was successful with less than 1 Ω contact resistance and no lateral current flow, illustrating the potential of anisotropic bonding. Another bonding method of rivet bonding is under testing (Fig. 12 (b-c)), and its testing scheme will be included in the next fabrication cycle.




Fig. 12: Evaluation of bonding methods: (a) Anisotropic conductive bonding, (b) ball-bonding on silicon probe, and (c) ball-bonding on Parylene substrate. (a) is a parallel process with a self-space-filling feature, while (b-c) are serial and require a separate space-filling following bonding.

REPORTABLE OUTCOMES

Journal article

1). X. Y. Zhang, J. Hu, R. W. Knighton, X-R Huang, C. A. Puliafito, and S. Jiao, “Dual-band spectral-domain optical coherence tomography for *in vivo* imaging the spectral contrasts of the retinal nerve fiber layer”, *Optics Express* **19**, 19653-19659 (2011).

CONCLUSION

We can draw the following conclusion for the projects: (1) the dual-band OCT provides a tool for studying the spectral contrast for imaging the RNFL. Next step is to apply the imaging tool in imaging animal models. (2) The enabling technologies, including the millimeter wave source and the flexible multielectrode array, have been developed for non-invasive neuromodulation of retinal activity. Further work is required to validate the feasibility of the proposed neuromodulation approach. (3) The strategy of joining a multisite intracortical silicon-substrate stimulating electrode to a flexible multi-conductor parylene-C cable appears to be feasible. The helically-wound paralyze cable can be fabricated and then coiled so as to remain very flexible in all planes and should provide a methods of routing numerous multi-conductor cables to numerous intracortical array, as will be required for a cortical-level visual prosthesis. The best method of attaching the parylene cables to the silicon probes (rivet bonds vs. anisotropic epoxy) remains to be determined.

Dual-band spectral-domain optical coherence tomography for *in vivo* imaging the spectral contrasts of the retinal nerve fiber layer

Xiangyang Zhang,¹ Jianming Hu,^{1,3} Robert W. Knighton,² Xiang-Run Huang,² Carmen A. Puliafito,³ and Shuliang Jiao^{1,*}

¹Department of Ophthalmology, Keck School of Medicine, University of Southern California, Los Angeles, CA 90033, USA

²Bascom Palmer Eye Institute, University of Miami Miller School of Medicine, Miami, FL 33136, USA

³Chongqing Key Laboratory of Optics and Engineering, Chongqing Normal University, Chongqing 40047, China
*jiao@usc.edu

Abstract: The ultimate goal of the study is to provide an imaging tool to detect the earliest signs of glaucoma before clinically visible damage occurs to the retinal nerve fiber layer (RNFL). Studies have shown that the optical reflectance of the damaged RNFL at short wavelength (<560nm) is reduced much more than that at long wavelength, which provides spectral contrast for imaging the earliest damage to the RNFL. To image the spectral contrast we built a dual-band spectral-domain optical coherence tomography (SD-OCT) centered at 808nm (NIR) and 415nm (VIS). The light at the two bands was provided by the fundamental and frequency-doubled outputs of a broadband Ti:Sapphire laser. The depth resolution of the NIR and VIS OCT systems are 4.7μm and 12.2μm in the air, respectively. The system was applied to imaging the rat retina *in vivo*. Significantly different appearances between the OCT cross sectional images at the two bands were observed. The ratio of the light reflected from the RNFL over that reflected from the entire retina at the two bands were quantitatively compared. The experimental results showed that the dual-band OCT system is feasible for imaging the spectral contrasts of the RNFL.

©2011 Optical Society of America

OCIS codes: (170.0170) Medical optics and biotechnology; (170.1610) Clinical applications; (170.4460) Ophthalmic optics and devices; (170.4500) Optical coherence tomography; (170.4580) Optical diagnostics for medicine.

References and links

1. H. A. Quigley, "Glaucoma," *Lancet* 377(9774), 1367–1377 (2011).
2. L. A. Kerrigan-Baumrind, H. A. Quigley, M. E. Pease, D. F. Kerrigan, and R. S. Mitchell, "Number of ganglion cells in glaucoma eyes compared with threshold visual field tests in the same persons," *Invest. Ophthalmol. Vis. Sci.* 41(3), 741–748 (2000).
3. M. L. Gabriele, G. Wollstein, H. Ishikawa, L. Kagemann, J. Xu, L. S. Folio, and J. S. Schuman, "Optical coherence tomography: history, current status, and laboratory work," *Invest. Ophthalmol. Vis. Sci.* 52(5), 2425–2436 (2011).
4. H. G. Lamm and N. J. Reus, "New developments in scanning laser polarimetry for glaucoma," *Curr. Opin. Ophthalmol.* 19(2), 136–140 (2008).
5. R. W. Knighton, X. Huang, and Q. Zhou, "Microtubule contribution to the reflectance of the retinal nerve fiber layer," *Invest. Ophthalmol. Vis. Sci.* 39(1), 189–193 (1998).
6. X.-R. Huang and R. W. Knighton, "Microtubules contribute to the birefringence of the retinal nerve fiber layer," *Invest. Ophthalmol. Vis. Sci.* 46(12), 4588–4593 (2005).
7. B. Fortune, L. Wang, G. Cull, and G. A. Cioffi, "Intravitreal colchicine causes decreased RNFL birefringence without altering RNFL thickness," *Invest. Ophthalmol. Vis. Sci.* 49(1), 255–261 (2008).
8. G. M. Pocock, R. G. Arambur, N. J. Kamp, C. S. Spacht, M. K. Markey, and H. G. Rylander 3rd, "The relationship between retinal ganglion cell axon constituents and retinal nerve fiber layer birefringence in the primate," *Invest. Ophthalmol. Vis. Sci.* 50(11), 5238–5246 (2009).
9. X.-R. Huang and R. W. Knighton, "Altered F-actin distribution in retinal nerve fiber layer of a rat model of glaucoma," *Exp. Eye Res.* 88(6), 1107–1114 (2009).

#152410 - \$15.00 USD Received 4 Aug 2011; revised 12 Sep 2011; accepted 15 Sep 2011; published 22 Sep 2011
(C) 2011 OSA 26 September 2011 / Vol. 19, No. 20 / OPTICS EXPRESS 19653

10. X.-R. Huang, W. Kong, Y. Zhou, and G. Gregori, "Distortion of axonal cytoskeleton: an early sign of glaucomatous damage," *Invest. Ophthalmol. Vis. Sci.* **52**(6), 2879–2888 (2011).
11. X.-R. Huang, Y. Zhou, W. Kong, and R. W. Knighton, "Reflectance decrease prior to thickness change of the retinal nerve fiber layer in glaucomatous retinas", *Invest. Ophthalmol. Vis. Sci.*, published ahead of print July 5, 2011.
12. R. W. Knighton and X.-R. Huang, "Directional and spectral reflectance of the rat retinal nerve fiber layer," *Invest. Ophthalmol. Vis. Sci.* **40**(3), 639–647 (1999).
13. X.-R. Huang, Y. Zhou, W. Kong, and R. W. Knighton, "Change of retinal nerve fiber layer reflectance correlated with cytostructural change in glaucoma," *ARVO Meeting Abstracts* (Fort Lauderdale, Florida 2011), 52:2442.
14. M. Ruggeri, G. Tsechpenakis, S. Jiao, M. E. Jockovich, C. Cebulla, E. Hernandez, T. G. Murray, and C. A. Puliafito, "Retinal tumor imaging and volume quantification in mouse model using spectral-domain optical coherence tomography," *Opt. Express* **17**(5), 4074–4083 (2009).
15. X. Zhang, H. F. Zhang, C. A. Puliafito, and S. Jiao, "Simultaneous in vivo imaging of melanin and lipofuscin in the retina with photoacoustic ophthalmoscopy and autofluorescence imaging," *J. Biomed. Opt.* **16**(8), 080504 (2011).
16. R. W. Knighton and X.-R. Huang, "Visible and near-infrared imaging of the nerve fiber layer of the isolated rat retina," *J. Glaucoma* **8**(1), 31–37 (1999).
17. Scott Prahl, "Optical Absorption of Hemoglobin," <http://omlc.ogi.edu/spectra/hemoglobin/summary.html>.

1. Introduction

Glaucoma, one of the leading causes of irreversible blindness worldwide, occurs when the retinal ganglion cells and their axons in the retinal nerve fiber layer (RNFL) and optic nerve degenerate in a characteristic and identifiable pattern called glaucomatous optic neuropathy (recently reviewed in Ref. 1). Untreated glaucoma may lead to permanent loss of vision, which motivates attempts to detect the earliest signs of the disease. Early diagnosis will permit early medical intervention to prevent vision loss. Current clinical methods for glaucoma diagnosis detect degenerative damage after it has already occurred, either structural damage (tissue loss in the optic disc, RNFL or macula) or functional damage (loss of visual sensitivity). Some data suggest that structural damage can occur many years before detectable loss of visual sensitivity [2] and a variety of optical approaches to structural assessment have been pursued. In particular, recent advances in optical coherence tomography (OCT) for glaucoma diagnosis have mostly focused on structural measurements [3].

In the case of the RNFL, some clinical and laboratory methods have progressed beyond the measurement of structure to the assessment of the optical properties of the tissue. For example, the commercially available technology of scanning laser polarimetry uses retardation by RNFL birefringence as a measure of RNFL integrity [4]. The optical properties of the RNFL depend on axonal structure at the submicroscopic level; its birefringence, for example, arises from parallel intracellular fibrils called microtubules [5–8]. Importantly, studies of the RNFL in a rat model of glaucoma show that the axonal cytoskeleton can change before the decrease in RNFL thickness [9,10] and in pilot studies we used high resolution small animal OCT operating at a center wavelength of 830 nm to confirm that there was no decrease in the RNFL thickness measured *in vivo* for the same rats (data unpublished). Even more promising, RNFL reflectance can decline before histologically detectable change in the axonal cytoskeleton [11]. Thus, the measurement of RNFL optical properties may allow a means to detect changes in cellular substructure that occur before axonal degeneration. This in turn could lead to diagnostic methods that reveal a therapeutic window during which permanent damage might be prevented.

The RNFL reflectance at all wavelengths arises from light scattering by cylindrical structures, with RNFL reflectance spectra suggesting the presence of at least two scattering mechanisms [12]. The spectra can be fit by a model where the reflectance at wavelengths shorter than 570 nm comes from cylinders with diameters much smaller than the wavelength, but the reflectance at wavelengths longer than ~680 nm comes from cylinders with effective diameters of 350 nm to 900 nm. A recent report showed that the RNFL reflectance in the rat glaucoma model also demonstrated spectral changes; at short wavelengths (< 560nm) the reflectance of the affected RNFL decreased much more than at long wavelengths [13]. This result suggests that the subtle structural alteration of axons in early glaucoma produces a wavelength dependent change in the RNFL reflectance.

In conventional ophthalmic OCT, the RNFL appears as a brightly reflecting layer. Because the light sources used in ophthalmic OCT systems are in the near infrared (NIR), the detected light from the RNFL should come mainly from scattering by thick cylindrical structures, making NIR OCT intrinsically insensitive to the very early changes seen in the rat model of glaucoma. If we shift the wavelength of the OCT to the visible (shorter than 570 nm), however, the changes to the thin cylindrical structures may be detectable. Ideally, an OCT system that covers both long and short wavelengths should be able to image contributions from both the thin and thick cylindrical structures in the nerve fibers. To test this hypothesis, an OCT system working in the visible spectrum (preferably at a center wavelength < 500 nm) is needed.

In this paper we report on our first effort in building and testing an OCT system working at a center wavelength of ~ 415 nm specifically for imaging the spectral contrast of the RNFL. The current study is a technical preparation for investigations on early diagnosis and progressive assessment of glaucoma using imaging technologies.

2. Experimental system

Figure 1 shows a schematic of the experimental system. The system consists of two spectral-domain OCT systems: one working in the NIR band and the other one working in the visible (VIS) band. A broadband Ti: Sapphire laser (MICRA, center wavelength: 800 nm, bandwidth: 120 nm, Coherent Inc, USA) was used as the fundamental light source. A frequency doubling crystal (BBO, Castech, Fujian, China) was used to convert the laser light to the visible. To ensure a broadband output from the crystal the fundamental laser light was focused into the crystal by a lens of $f=14$ mm (L1). The light exiting the BBO crystal was first collimated with an achromatic lens of $f=40$ mm (L2) and then reflected by a long-pass filter (FEL0700, Thorlabs, USA) to separate the visible light from the residual fundamental light.




Fig. 1. Schematic of the dual-band OCT system. M1~M5: mirror; L1~L4: lens; BBO: BBO crystal; LPF: long-pass filter. C1~C6: Collimator; PC: Polarization Controller.

The visible light was then coupled into the source arm of a single mode optical fiber-based Michelson interferometer. In the detection arm the reflected light from the sample and reference arms was collimated and detected by a spectrometer, which consisted of an 1800 line/mm transmission grating, a multi-element imaging lens ($f=150$ mm), and a line scan CCD camera (Aviiva-SM2-CL-2010, 2048 pixels with 10 micron pixel size operating in 12-bit mode, e2V). An image acquisition board (NI IMAQ PCI 1428) acquired the image captured by the camera and transferred it to a workstation (HP xw4600, 4 GB memory) for signal processing and image display. The design of the NIR-OCT is similar to that in our previous publications [14,15] except that the residual fundamental light after the long-pass filter was used as the light source.

After exiting their corresponding optical fibers in the sample arms the visible and NIR probing light beams were collimated with collimators that can be interchangeably connected with the scanning and optical delivery system. The scanning and optical delivery system was built on a modified slit lamp. The probing light beams were scanned by an X-Y galvanometer scanner and delivered to the eye through an achromatic relay lens ($f = 75$ mm) and an objective lens ($f = 19$ mm).

3. Results and discussion

3.1 Spectral reflectance of the RNFL

We measured the spectral reflectance of the RNFL of living *ex vivo* rat retinas by using imaging micro-reflectometry (IMR) [13,16]. The result of measurement is shown in Fig. 2. Rectangular areas are defined both on a nerve fiber bundle (black boxes) and nearby gaps (white boxes), where no nerve fiber bundles exist. The average reflected light intensity of both bundle (I_b) and gap (I_g) areas are calculated. Reflectance from RNFL only is then derived as $I = I_b - I_g$. Figure 2B shows the RNFL spectral reflectance of a glaucomatous eye and its control eye in a rat glaucoma model produced by laser treatment of the trabecular meshwork [9–11]. Intra-ocular pressure (IOP) before treatment for both eyes was 24 mmHg. The treated eye was exposed to high IOP for 18 days with peak IOP of 48 mmHg. The spectra were measured 4 weeks after the start of treatment. The experiments showed that glaucoma causes decreased RNFL reflectance at short wavelengths while it remained approximately the same at longer wavelengths.




Fig. 2. Measurement of RNFL reflectance. A: typical reflection image of the RNFL; nerve fiber bundles appear as bright stripes; B: calculated spectral reflectance of the RNFL; Bar: 20 μ m.

3.2 VIS-OCT experiments

Figure 3 shows the measured spectra and the point-spread functions of the two OCT systems. The center wavelength and full width at half maximum (FWHM) of the NIR light are 808 nm and about 105 nm, respectively. The NIR-OCT achieved a depth resolution of about 4.7 μ m and an imaging depth of 2.5 mm in the air. The center wavelength and FWHM of the visible light are 415 nm and 7 nm, respectively. The VIS-OCT achieved a depth resolution of about 12.2 μ m and an imaging depth of 1.3 mm in the air.

The system was applied to imaging the normal rat retina *in vivo*. The animals (Sprague Dawley rats, body weight: 600–750 g, Charles Rivers) were anesthetized by intraperitoneal injection of a cocktail containing ketamine (54 mg/kg body weight) and xylazine (6 mg/kg body weight). In the meantime, the pupils were dilated with 10% phenylephrine solution. Drops of artificial tears were applied to the eyes every two minutes to prevent dehydration of the cornea and cataract formation. After anesthetization, the animals were restrained in a mounting tube, which was fixed on a five-axis platform. Raster scans with the fast axis along the horizontal direction were performed. All experiments were performed in compliance with the guidelines of the University of Southern California's Institutional Animal Care and Use Committee.




Fig. 3. Measured parameters of the NIR and VIS OCT systems. (a) and (b): spectra of the NIR and VIS OCT systems; (c) and (d): PSF of the NIR and VIS OCT systems.

In the experiment, we collected the NIR and VIS-OCT images sequentially. The initial alignment including finding the area of interest was done with the NIR OCT. After the acquisition of the NIR OCT image we mechanically switched the VIS OCT into position and acquired the VIS OCT images. In our current study the laser energy after the objective lens was measured to be 450 μW for the visible and 750 μW for the NIR light, which are both below the ANSI safety limits for eye imaging. Since the OCT images of the two different bands were not acquired simultaneously they are not completely registered due to movement of the animal during the switching process. The advantage of using the NIR-OCT for alignment is to reduce the retinal exposure to the visible light. The higher axial resolution and greater imaging depth of the NIR-OCT also aided the alignment.




Fig. 4. Sample images of a rat retina acquired with the NIR OCT and VIS OCT. Images consist of 2048 A-lines. Bar: 100 μm.

Figure 4 shows one pair of the acquired NIR and VIS OCT images. The images consist of 2048 A-scans and cover a length of about 1.3 mm across the retina. The differences between the NIR and VIS OCT images are significant. As expected the blood vessels blocked the VIS probing light completely with only reflections from their front boundaries, while the NIR-OCT still revealed some features beneath the vessels. This difference comes from the significant difference of the absorption coefficients of hemoglobin at the two bands: at 415 nm $\mu_a = 2801.52 \text{ (cm}^{-1}\text{)}$ for oxy-hemoglobin (HbO_2) and $\mu_a = 1891.79 \text{ (cm}^{-1}\text{)}$ for deoxy-hemoglobin (Hb); at 808 nm $\mu_a = 4.58 \text{ (cm}^{-1}\text{)}$ for HbO_2 and $\mu_a = 3.88 \text{ (cm}^{-1}\text{)}$ for Hb [17].




Fig. 5. Normalized average of the A-lines of the OCT images in Fig. 4.

In the NIR OCT image there are strong signals from deep into the choroid. In the VIS OCT image, although we can still recognize all the retinal layers there are much weaker signals from behind the RPE layer due to the higher attenuation coefficient of the RPE in the VIS band. To further demonstrate the different characteristics of the OCT signal in the two bands we averaged all the A-lines of the images in Fig. 4. Before averaging we flattened each image by first drawing their ILMs and then shifting each A-line in the depth direction accordingly. The averaged A-lines are shown in Fig. 5. From Fig. 4 and Fig. 5 we can see that the RPE signal is the strongest in the NIR OCT image. In contrast, the RNFL signal is the strongest in the VIS OCT image. Another noticeable feature is that the boundaries among the different retinal layers in the VIS OCT image are not as clear as those in the NIR OCT image. This feature may be caused by the different depth resolutions of the two OCT systems. Other factors like the differences of the optical properties, e.g. index of refraction of the different layers in the VIS and NIR bands may also play a role.




Fig. 6. Retinal images acquired with the NIR and VIS OCT systems. (a): B-scan image of the NIR OCT; (b): B-scan image of the VIS OCT; (c): NIR-OCT fundus image; (d): VIS-OCT fundus image. BV: blood vessel. Bar: 100 μm .

Figure 6 shows another typical pair of NIR and VIS OCT images from an acquired 3D data set, which consists of 512×128 A-scans. Also shown in the figure are the corresponding OCT fundus images, on which the locations of the OCT cross sectional images are marked. We can see that the contrast of the retinal blood vessels is higher in the VIS-OCT fundus

image due to the higher attenuation coefficient of the red blood cells in the VIS band. We can also see from the fundus images that the lateral resolution in the VIS band is better since we can recognize some small blood vessels that are not available in the NIR fundus image.

Owing to the depth resolving capability of OCT we can calculate the light reflection from a certain layer of the retina by summation of the OCT signals inside the layer:

$$I_{r(z_1 \sim z_2)}(x) = \sum_{z_1}^{z_2} I_{OCT}(x, z) \quad (1)$$

where I_{OCT} is the OCT intensity signal, x is the coordinates of the horizontal direction, z is the coordinates of the depth, z_1 and z_2 are the lower and upper boundaries of the layer. To calculate the reflections from the RNFL we first segmented the RNFL manually and calculated the reflection from inside the layer. We then calculated the ratio of the RNFL reflection over the fundus reflection calculated from the whole imaging depth using:

$$R_{RNFL} = \frac{I_{r(z_1 \sim z_2)}(x)}{I_{r(total)}}. \quad (2)$$

Figure 7 shows the segmentation of the RNFL boundaries in a pair of NIR and VIS OCT B-Scan images at the same location of the retina. Also shown in the figure are the calculated RNFL reflection ratios. From the calculation we can see that the reflection from the RNFL in the NIR band takes about 5% of the total fundus reflection. In the VIS band, however, the ratio increases to about 10%. The calculation shows that a higher portion of the fundus reflection comes from the RNFL reflection when imaged in the VIS band.




Fig. 7. Calculation of the ratio of RNFL reflection over the total fundus reflection for NIR (a) and VIS (b) OCT B-scans. Bar: 100 μ m.

4. Conclusion

We have for the first time built a dual-band OCT system with one band centered at 415 nm and the other at 808 nm for imaging the spectral contrast of the RNFL. The system was successfully applied to imaging the rat retina *in vivo*. Significantly different appearances between the OCT cross sectional images at the two bands were revealed in the preliminary imaging tests. The experiments showed that a higher portion of the fundus reflection comes from the RNFL reflection when imaged in the VIS band. We can conclude that the dual-band OCT system is feasible for imaging the spectral contrast of the RNFL.

Acknowledgements

This work is supported in part by the following grants: National Institutes of Health grant 7R21EB008800-02 (Jiao), 1R01EY019951-01A1 (Jiao), and R01EY19084 (Huang) and Department of Defense grant W81XWH-11-1-0197 (Jiao).