



Submitted by Stephanie M Price in partial fulfillment of the requirements for the degree of Master of Science in Oral Biology.

Accepted on behalf of the Faculty of the Graduate School by the thesis committee:



LTC Stacy Larsen, DDS
AEGD Program Director

10 June 16
Date



MAJ Jason Bullock, DMD, MS
AEGD Assistant Program Director

10 June 16
Date



LTC Manuel Peláez, DMD, MS
Research Director

10 June 16
Date

The author hereby certifies that the use of any copyrighted material in the thesis manuscript entitled:
Evaluating the effectiveness of biomaterial removal from dental implant drills
Is appropriately acknowledged and beyond brief excerpts, is with the permission of the copyright owner.



Stephanie M Price
2-Year AEGD Program, Fort Bragg
Uniformed Services University
Date: 13 JUN 2016

The author hereby certifies that the use of any copyrighted material in the thesis manuscript entitled:


Evaluating the effectiveness of biomaterial removal from dental implant drills

Is appropriately acknowledged and beyond brief excerpts, is with the permission of the copyright owner.

Stephanie M Price
2-Year AEGD Program, Fort Bragg
Uniformed Services University
Date: 13 JUN 2016


Submitted by Stephanie M Price in partial fulfillment of the requirements for the degree of Master of Science in Oral Biology.

Accepted on behalf of the Faculty of the Graduate School by the thesis committee:



LTC Stacy Larsen, DDS
AEGD Program Director

10 June 16
Date



MAJ Jason Bullock, DMD, MS
AEGD Assistant Program Director

10 June 16
Date



LTC Manuel Peláez, DMD, MS
Research Director

10 June 16
Date

Evaluating the effectiveness of biomaterial removal from dental implant drills

STEPHANIE M. PRICE, DDS

B.M.E. University of Delaware, Newark, DE 1995

D.D.S. University of North Carolina, Chapel Hill, NC 2013

MAJ, United States Army

Submitted in partial fulfillment of the requirements for the degree of Master of Science in the Department of Oral Biology in the Graduate School of The Uniformed Services University of Health Sciences

Fort Bragg, NC

2016

ABSTRACT

Multiple studies question the efficacy of cleaning and sterilizing protocols for multi-use dental burs and endodontic files. The US Army, among other institutions, implemented a single use policy for burs and endodontic files; included in this policy are implant osteotomy drills. We hypothesize that our cleaning protocols cannot guarantee complete debris removal thereby supporting the single use drill policy.

This in-vitro study examined 160 Biomet 3i™ drills of two configurations imaged with SEM at 30X magnification. Eighty drills were used to conduct osteotomies on a porcine mandible; 40 (positive control) were steam autoclaved and imaged directly and 40 were cleaned using our standardized protocol, steam autoclaved, and then imaged. Eighty drills were used as negative control; 40 were imaged directly from the package and 40 were cleaned using the same protocol, steam autoclaved and then imaged. Four independent observers scored visible debris as a percentage of total drill area on each image. The results were compared using ANOVA post-hoc and Tukey HSD tests.

None of the four groups showed an absence of debris on all drills. Unused and cleaned drills showed on average 18% less debris than used and cleaned drills; however the results were not statistically significant. The scoring system did not differentiate between surgical debris and remnants of cleaning solutions. This initial study does not support a single use policy for dental implant drills. More research is indicated to differentiate surgical debris from solution remnants, to verify results with different cleaning procedures, and to investigate pathogenicity of debris.

Key words: infection control, implant drills, single-use medical equipment

BACKGROUND

In the past 50 years, research studies evaluating the effectiveness of cleaning dental instruments focused primarily on endodontic files and dental burs because of their intricate architecture and frequent use and reuse. Most studies assessed the presence of contamination by comparing either microbial culture or microscopic imaging and concluded that debris cannot be fully removed from endodontic files and dental burs prior to sterilization.¹⁻¹² Evidence shows that burs, files and drills not labeled as “sterile” from the manufacturer contain both organic and manufacturing debris and must be sterilized prior to use.^{11,13-16} Because of this both the Department of Veterans Affairs and the US Army have implemented a single use policy for burs and endodontic files and have expanded this policy to include implant osteotomy drills.^{17,18} In 2015 the FDA published new recommendations for reusable medical devices giving guidance to manufacturers to validate cleaning protocols.¹⁹

Threats to complete sterilization include fungi, bacteria, viruses, and prions with average sizes 3-30µm, 1-3µm, 0.02-0.03 µm, and 0.005-0.01µm respectively. A principle of infection control is that instruments should be as clean as possible; we rely on sterilization to remove infectious viability from debris.²⁰⁻²² Studies of endodontic files, dental burs, and dental instruments show our current standard of practice utilizing steam sterilization kills fungi, viruses and bacteria.^{8,23-31} Prions, associated with Creutzfeldt-Jakob Disease (CJD), are not destroyed by the standard sterilization techniques.³² Destruction is only confirmed with incineration at 1000°C,³³ although recently enzymatic soak combined with sterilization or plasma sterilization were found to be effective at eliminating prion infectivity.^{34,35} The presence of inorganic and organic debris can promote thermal resistance of pathogens and interfere with complete sterilization.²¹ Finally, although it is difficult to confirm with evidence, it is possible that inert biological debris itself could potentially cause an inflammatory risk.

There are four basic methods of cleaning dental instruments: presoaking, ultrasonic, automatic washer disinfectant, thermal washer disinfectant and hand scrubbing- with benefits of each method subject to debate. The addition of an enzymatic presoak shows a statistically significant benefit, although they vary in composition.^{1,29,32,36,37} Some studies support the use of ultrasonic cleaners⁶ while others failed to find a statistically significant advantage over the washer disinfectant,³⁸ however both are considered a safe alternative to hand scrubbing as there are concerns about percutaneous blood exposure.²¹

Although, there are no published incidents of transmission of a pathogen through an instrument after standard of care sterilization in the dental office, instances of infection control failures are likely under reported.³⁹ There is one known patient to patient transfer of Hepatitis B in 2001 involving an oral surgery clinic. A subsequent investigation discovered no breakdown in infection control and sterilization protocols. The authors suspected cross-contamination from an environmental surface as the etiology of the transmission.⁴⁰ A second instance of Hepatitis B transmission occurred in 2009 associated with a portable dental clinic with both volunteers (2) and patients (3) infected. Again, an investigation was unable to determine the exact cause of the transmission, although failure to sterilize headpieces and failure to wrap sterilized instruments were cited as potential causes.⁴¹

The trend towards single use instruments in the United States was fueled by an instance in 2013 of bacterial transfer of E. Coli to 39 patients exposed to a single duodenoscope in Illinois (CDC HICPAC minutes). Since then, recommendations have been made by both the CDC (facilities) and FDA (manufacturers) for multiple use medical instruments which are critical, i.e. exposed to mucosa or vascular tissue, and have intricate architecture making them inherently difficult to clean. In addition to being cleaned, duodenoscope sterilization is sometimes verified with flushing or sampling and culturing techniques (CDC HICPAC). However limitations to culturing both human and technological in addition to expense, make this a difficult standard of care. Manufacturer procedures should be validated using quantitative test methods using extraction and DNA recovery¹⁹; however no published studies assess the cleanliness of dental instruments using this method. In the United Kingdom, single use endodontic and dental burs has been more public due to the confirmed case, although again, not dentally related, of prion transmission resulting in a variant of CJD.^{42,43} There is no evidence of implant complications as a direct result of contaminated (or reused) drill surfaces.⁴⁴

A cost analysis of single use versus reusable drills for the Veterans Administration is estimated at \$100 per implant. This is based on a government price and data that supports an average of two implants placed in a patient per surgical session.¹⁷ Army clinics worldwide place an average of 1.64 implants per visit (T Oringderff, oral communication, APR 2016); combined with the additional use of a starter drill and the possible use of a dense bone tap, the expense approaches approximately \$200 per implant for a total of \$1.2M year.

A single study similar to this one, investigated the cleansability of lumen free implant drills by exposing them to radioactively marked blood and concluded that predictable complete removal of all blood is impossible and therefore implant drills should be regarded as single use.⁴⁵ This study did not use enzymatic cleaners or a washer disinfector. In this study we replicated our current cleaning and sterilizing techniques to validate the US Army single use drill policy and as a pilot study to assess the need for single use implant drills. This study is unique in that it assesses the adequacy of a current standard protocol to determine if used drills can be returned to the same level of cleanliness as unused drills.

MATERIALS AND METHODS:

Eighty multiple use twist drills (2.0mmX10mm) and 80 multiple use quad shaping drills (3.25mmX10mm) (Biomet 3i™, Palm Beach Gardens, FL) were chosen for this investigation. They are externally cooled, surgical stainless steel and have the smallest grooves which we theorized would be the hardest to clean. All drills were received unsterile from the manufacturer, placed in the surgical trays in groups of 40 (20 twist and 20 quad shaping) with ungloved hands. The drills were divided into the following groups: Group A, positive control, (Cleaned, sterilized and used for porcine osteotomy, then sterilized and imaged); Group B, experimental (Cleaned, sterilized and used for porcine osteotomy, then cleaned again, sterilized and imaged); Group C, uncleaned negative control (Placed into tray and imaged); and Group D, cleaned negative control (Placed into the tray, cleaned, sterilized and imaged).

Eighty drills for Group A and B were used to perform 40 total osteotomies in a porcine mandible (sacrificed 12 hours prior). After each osteotomy (first twist then quad shaping), the drills were placed in a stainless steel surgical cup of sterile saline until all osteotomies were completed. Drills were returned to

two surgical trays (20 of each type in each tray) and placed on a shelf on the dirty side of the sterilization suite for 75 minutes prior to cleaning protocol as per our usual worst case clinic operational routine.

Forty drills for Group C were removed from manufacturers packaging, placed in the surgical tray, and sent immediately for imaging. Once Group C returned, the drills were relabeled as Group D. They were cleaned and sterilized and returned for imaging.

Each cleaning iteration followed the same procedure using both ultrasonic and thermal washer disinfectant. The ultrasonic (Elmasonic S300, Singen, Germany) was prepared by rinsing, filling with fresh water, and run at 80°C for 45 minutes with enzyme tablets (Maxizyme, Henry Schein, Melville, NY). The ultrasonic was then emptied, rinsed and refilled with fresh water. Each tray was placed individually in the ultrasonic at 80°C for 45 minutes with detergent solution (MPUS Plus, Dentronix, Cuyahoga Falls, OH). Ultrasonic cleaning was verified with Sonocheck™ (Healthmark Industries, Fraser, MI). Following ultrasonic, the trays were rinsed with fresh water for 2 min and placed in a 50min thermal washer disinfectant cycle (Getinge86, Rochester, NY), which automatically dispenses enzymatic, detergent and pre-sterilization lubricant solutions. The wash cycle was verified with a TOSI™ washer test (Healthmark Industries, Fraser, MI). The drills were placed in the vacuum steam autoclave for 5 minutes 273°F with 20 minute dry time (Getinge, Rochester, NY). Both the Bowie-Dick (Assure S.M.A.R.T., Getinge, Rochester NY) and biological spore test (Attest™, 3M™, Maplewood, MN) verified functionality of the steam autoclave.

The drills were packaged and sent for scanning electron imaging (SEM) at 30X. The drills were kept covered except during imaging and the microscopist did not touch any part of the surface to be imaged. Each image was divided into squares of approximately 300µm each with the number of squares varying depending on the amount of drill imaged. Squares that contained < ¼ drill surface were eliminated from view. The 160 drill images were observed and scored by four independent observers in the same randomized order. Their scores represented number of squares with debris divided by the total number of squares.

RESULTS

No drill groups were completely free of debris. The F test is highly significant indicating that at least one pairwise comparison is significant ($F(3,156) = 199.36, p < 0.001$). As expected, pairwise comparisons showed that B is significantly lower than A ($p < 0.0001$).

In other pairwise comparisons, D is significantly lower than C ($p < 0.0001$), additionally B was not significantly different from C; ($p = 0.09$). Therefore new drills that are cleaned and sterilized are statistically cleaner than used drills that are cleaned and sterilized, and cleaned and sterilized used drills are statistically no different than new drills from the package.

A t-test demonstrated statistical inconsistency between drill sets D1-D5 and drill sets D6-D8. This is likely due to a variation in the environmental conditions and the settings used in the SEM. The images from sets D6-D8 are clearer than the images from sets D1-D5 and the data from sets D6-D8 are more supportive of the statistically significant difference between groups D and B.

DISCUSSION

A limitation of visual assessment with SEM is that surface irregularities, for example the matt surface of the twist drill and the central groove of the quad shaping drills make it difficult to differentiate extent of debris. A second limitation of the SEM is that it is difficult to differentiate surgical debris from remnants of cleaning solutions. We hypothesized the milky white surface irregularity found on various drills of both groups B and D to be pre-sterilization lubrication added in the final stage of the washer disinfectant. Future research could disable this component of the washer disinfectant; rinse the drills with alcohol following the washer disinfectant, or use environmental scanning electron microscope (ESEM) to more specifically identify the chemical components of the debris. This debris interpretation could contribute to the large standard deviation found in all groups.

In this study the standard deviations were large in all four groups. Although observer interpretation could be a contributing factor, large standard deviations are also found in other

investigations with different study methodologies. The large deviation was also found in used and uncleaned drills which rules out individual variation from the cleaning as a cause. We hypothesize that manufacturing variances could result in a large variation in surface smoothness, thus a variation in debris retention.

The presence of debris on all drill groups in this study raises a question as to the presence of debris on other reused instruments with similar architecture- i.e. bone files, operative hand instruments, and periodontal scalers. It is unlikely that the debris contains viable bacteria, viruses or fungi- particularly in groups B-D which have more limited levels of debris because of the pressure/temperature levels confirmed by the spore and Bowie- Dick leak test. A critical step in cleaning is the visual observation of debris; however, even if debris is not visible, it is likely there as shown by the presence of debris on clean but unused drills. The quantity of debris which potentially interferes with sterilization is difficult to determine. Despite the discovery of debris on both new and used drills and the theoretical ubiquitous presence of debris, early failure of implants is remarkably low at 2%.⁴⁶

The magnification of the drills in this study, 30x, was not large enough to identify specific pathogens. All drills were imaged at 60x (drill tip and mid drill), and several images were obtained at 1000x; these images were not used in the study. Bacteria, the largest pathogen, are barely visible on a 1000x image, and the field of view is limited to only the drill tip. We felt that analyzing more of the drill surface, especially as this is an initial study, would provide more meaningful results. One of the experimental drill images at 30x was lost and reconstructed and reduced in size using two segments of 60x images of the same drill. The 1000x images, all taken of unused and cleaned drills, showed visible surface defects- larger in size than a potential bacterium. These defects may be polished out during initial use, or these surface defects may be responsible for the retained debris in used drills. This is an area for future research as we did not obtain any 1000x images of used drills to compare.

Dental clinics use different cleaning and sterilization protocols- all within the standard of care-based on their resources, size and mission. Institutional cleaning equipment like the thermal washer disinfectant used in this study may not be available in smaller clinics. While hand scrubbing is not recommended at our institution, it may be more feasible in smaller clinics. Given these variations, it is difficult for manufacturers to establish protocols which practically and universally guarantee the ability to return a used drill to its unused state. Our protocol does not follow the manufacturer's instructions. For example, the drills were soaked in the 0.9% sterile saline solution used in the surgery, whereas the manufacturer recommends a neutral pH. Saline can have a pH 4.6-5.5.^{47,48} This slight acidity may enhance the ability to retain debris or modify the surface making it appear to have debris. A worthy future investigation could be used to verify the cleaning protocols as defined by the manufacturer.

FDA requires manufacturers to produce procedures for cleaning and users to follow them, in this study we did not follow the published 3i directive.

CONCLUSION AND FUTURE RESEARCH

Although there was not a statistically significant difference between the experimental (used and cleaned) drills and the negative control (unused and cleaned) the average debris on the used and cleaned drills was higher than the average on the unused and cleaned. These results should be expanded with a second quantitative analysis- potentially DNA recovery.

Future research should include varying the cleaning methods; potentially including neutral pH proteolytic enzymatic soak at the surgical table, the use of alcohol following the washer disinfectant to remove pre-sterilization lubricant, and repeating the experiment using Nobel Biocare or other implant systems.

REFERENCES

1. Whitworth CL, Davies K, Palmer NO. Can protein contamination be removed from hand endodontic instruments? *Prim Dent Care*. 2009;16(1):7-12.
2. Assaf M, Mellor AC, Qualtrough AJ. Cleaning endodontic files in a washer disinfecter. *Br Dent J*. 2008;204(10):E17; discussion 562-3.
3. Van Eldik DA, Zilm PS, Rogers AH, Marin PD. A SEM evaluation of debris removal from endodontic files after cleaning and steam sterilization procedures. *Aust Dent J*. 2004;49(3):128-135.
4. Letters S, Smith AJ, McHugh S, Bagg J. A study of visual and blood contamination on reprocessed endodontic files from general dental practice. *Br Dent J*. 2005;199(8):522-5; discussion 513.
5. Smith A, Dickson M, Aitken J, Bagg J. Contaminated dental instruments. *J Hosp Infect*. 2002;51(3):233-235.
6. Perakaki K, Mellor AC, Qualtrough AJ. Comparison of an ultrasonic cleaner and a washer disinfecter in the cleaning of endodontic files. *J Hosp Infect*. 2007;67(4):355-359.
7. Ramadan AA. Removing hepatitis C virus from polytetrafluoroethylene-coated orthodontic archwires and other dental instruments. *East Mediterr Health J*. 2003;9(3):274-278.
8. Proff P, Bayerlein T, Kramer A, et al. Requirements and infection prophylaxis for internally cooled implant drills. *Folia Morphol (Warsz)*. 2006;65(1):34-36.
9. Harkness N, Davies EH. The cleaning of dental diamond burs. *Br Dent J*. 1983;154(2):42-45.
10. Sheriteh Z, Hassan T, Sherriff M, Cobourne M, Riley P. Decontamination of viable streptococcus mutans from orthodontic tungsten carbide debonding burs. an in vitro microbiological study. *J Orthod*. 2010;37(3):181-187.
11. Morrison A, Conrod S. Dental burs and endodontic files: Are routine sterilization procedures effective? *J Can Dent Assoc*. 2009;75(1):39.
12. Hogg NJ, Morrison AD. Resterilization of instruments used in a hospital-based oral and maxillofacial surgery clinic. *J Can Dent Assoc*. 2005;71(3):179-182.
13. Hauptman JM, Golberg MB, Rewkowski CA. The sterility of dental burs directly from the manufacturer. *J Esthet Restor Dent*. 2006;18(5):268-71; discussion 272.
14. Roth TP, Whitney SI, Walker SG, Friedman S. Microbial contamination of endodontic files received from the manufacturer. *J Endod*. 2006;32(7):649-651.
15. Filho MT, Leonardo MR, Bonifacio KC, Dametto FR, Silva AB. The use of ultrasound for cleaning the surface of stainless steel and nickel-titanium endodontic instruments. *Int Endod J*. 2001;34(8):581-585.
16. Linsuwanont P, Parashos P, Messer HH. Cleaning of rotary nickel-titanium endodontic instruments. *Int Endod J*. 2004;37(1):19-28.
17. Schoenhard W. Memorandum to VISN directors, VISN lead dentist and dental service chiefs subj: Policy for the single patient use of dental implant drills. 2011.
18. Carr D. DENCOM PAM 40-5-1 appendix J instrument processing and sterilization. 2015:4.
19. FDA. Reprocessing medical devices in health care settings: Validation methods and labeling guidance for industry and food and drug administration staff. 2015;UCM253010.
20. Rutala WA, Gergen MF, Jones JF, Weber DJ. Levels of microbial contamination on surgical instruments. *Am J Infect Control*. 1998;26(2):143-145.
21. Molinari J, Harte J. Chapter 16 instrument processing and recirculation. In: *Cottone's practical infection control in dentistry*. Thrid ed. Baltimore, MD: Wolters Kluwer/ Lippincott Williams and Wilkins; 2010:221.
22. Miller CH. Cleaning, sterilization and disinfection: Basics of microbial killing for infection control. *J Am Dent Assoc*. 1993;124(1):48-56.
23. Venkatasubramanian R, Jayanthi, Das UM, Bhatnagar S. Comparison of the effectiveness of sterilizing endodontic files by 4 different methods: An in vitro study. *J Indian Soc Pedod Prev Dent*. 2010;28(1):2-5.

24. Kumar KV, Kiran Kumar KS, Supreetha S, Raghu KN, Veerabhadrapa AC, Deepthi S. Pathological evaluation for sterilization of routinely used prosthodontic and endodontic instruments. *J Int Soc Prev Community Dent.* 2015;5(3):232-236.
25. Raju TB, Garapati S, Agrawal R, Reddy S, Razdan A, Kumar SK. Sterilizing endodontic files by four different sterilization methods to prevent cross-infection - an in-vitro study. *J Int Oral Health.* 2013;5(6):108-112.
26. Hurtt CA, Rossman LE. The sterilization of endodontic hand files. *J Endod.* 1996;22(6):321-322.
27. Van Eldik DA, Zilm PS, Rogers AH, Marin PD. Microbiological evaluation of endodontic files after cleaning and steam sterilization procedures. *Aust Dent J.* 2004;49(3):122-127.
28. Gnau HL, Goodell GG, Imamura GM. Rapid chairside sterilization of endodontic files using 6% sodium hypochlorite. *J Endod.* 2009;35(9):1253-1254.
29. Fulford MR, Ireland AJ, Main BG. Decontamination of tried-in orthodontic molar bands. *Eur J Orthod.* 2003;25(6):621-622.
30. Johnson MA, Primack PD, Loushine RJ, Craft DW. Cleaning of endodontic files, part I: The effect of bioburden on the sterilization of endodontic files. *J Endod.* 1997;23(1):32-34.
31. Rutala WA, Weber DJ. Guideline for disinfection and sterilization in healthcare facilities. 2008:1.
32. Brown SA, Merritt K, Woods TO, Busick DN. Effects on instruments of the world health organization--recommended protocols for decontamination after possible exposure to transmissible spongiform encephalopathy-contaminated tissue. *J Biomed Mater Res B Appl Biomater.* 2005;72(1):186-190.
33. World H. WHO infection control guidelines for transmissible spongiform encephalopathies. report of a WHO consultation, geneva, switzerland, 23-26 march 1999. World Health Organization Web site. http://www.who.int/csr/resources/publications/bse/WHO_CDS_CSR_APH_2000_3/en/. Published 2000. Updated 2006. Accessed APR 23, 2016.
34. Lawson VA, Stewart JD, Masters CL. Enzymatic detergent treatment protocol that reduces protease-resistant prion protein load and infectivity from surgical-steel monofilaments contaminated with a human-derived prion strain. *J Gen Virol.* 2007;88(Pt 10):2905-2914.
35. Banerjee KK, Kumar S, Bremmell KE, Griesser HJ. Molecular-level removal of proteinaceous contamination from model surfaces and biomedical device materials by air plasma treatment. *J Hosp Infect.* 2010;76(3):234-242.
36. Sanchez E, Macdonald G. Decontaminating dental instruments: Testing the effectiveness of selected methods. *J Am Dent Assoc.* 1995;126(3):359-62, 364, 366 passim.
37. Whitworth CL, Martin MV, Gallagher M, Worthington HV. A comparison of decontamination methods used for dental burs. *Br Dent J.* 2004;197(10):635-40; discussion 623.
38. Vassef M, Budge C, Poolman T, et al. A quantitative assessment of residual protein levels on dental instruments reprocessed by manual, ultrasonic and automated cleaning methods. *Br Dent J.* 2011;210(9):E14.
39. Laheij AM, Kistler JO, Belibasakis GN, Valimaa H, de Soet JJ, European Oral Microbiology Workshop (EOMW) 2011. Healthcare-associated viral and bacterial infections in dentistry. *J Oral Microbiol.* 2012;4:10.3402/jom.v4i0.17659. Epub 2012 Jun 12.
40. Redd JT, Baumbach J, Kohn W, Nainan O, Khristova M, Williams I. Patient-to-patient transmission of hepatitis B virus associated with oral surgery. *J Infect Dis.* 2007;195(9):1311-1314.
41. Radcliffe RA, Bixler D, Moorman A, et al. Hepatitis B virus transmissions associated with a portable dental clinic, west virginia, 2009. *J Am Dent Assoc.* 2013;144(10):1110-1118.
42. Will RG, Matthews WB. Evidence for case-to-case transmission of creutzfeldt-jakob disease. *J Neurol Neurosurg Psychiatry.* 1982;45(3):235-238.
43. Brown P, Preece M, Brandel JP, et al. Iatrogenic creutzfeldt-jakob disease at the millennium. *Neurology.* 2000;55(8):1075-1081.
44. Pye AD, Lockhart DE, Dawson MP, Murray CA, Smith AJ. A review of dental implants and infection. *J Hosp Infect.* 2009;72(2):104-110.

45. Parnia F, Hafezeqoran A, Mahmoudian B, et al. Comparison of different cleaning procedures of implant drills using TC99. *Implant Dent.* 2013;22(6):627-630.
46. Quirynen M, De Soete M, van Steenberghe D. Infectious risks for oral implants: A review of the literature. *Clin Oral Implants Res.* 2002;13(1):1-19.
47. Story DA, Thistlethwaite P, Bellomo R. The effect of PVC packaging on the acidity of 0.9% saline. *Anaesth Intensive Care.* 2000;28(3):287-292.
48. Reddi BA. Why is saline so acidic (and does it really matter?). *Int J Med Sci.* 2013;10(6):747-750.