

AD-A056 244

RADFORD ARMY AMMUNITION PLANT VA
NITROFORM RECOVERY AND ISOLATION STUDIES. (U)
MAY 78 R A MUNDY, E E GILBERT

F/G 19/1

UNCLASSIFIED

RAD-240.10


ARLCD-TR-78027

NL

| OF |
AD
A056244



END
DATE
FILMED
8 -78
DDC



MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS-1963-A

AD A 056244

12 LEVEL

AD

AD-E400 164

TECHNICAL REPORT ARLCD-TR-78027

NITROFORM RECOVERY AND ISOLATION STUDIES

R. A. MUNDY

RADFORD ARMY AMMUNITION PLANT

E. E. GILBERT

ARRADCOM PROJECT ENGINEER

AD NO. DDC FILE COPY

DDC RECEIVED JUL 17 1978

MAY 1978



US ARMY ARMAMENT RESEARCH AND DEVELOPMENT COMMAND
LARGE CALIBER
WEAPON SYSTEMS LABORATORY
DOVER, NEW JERSEY

APPROVED FOR PUBLIC RELEASE: DISTRIBUTION UNLIMITED.

18 06 13 088

The findings in this report are not to be construed as an official Department of the Army Position.

Destroy this report when no longer needed. Do not return to the originator.

The citation in this report of the names of commercial firms or commercially available products or services does not constitute official endorsement or approval of such commercial firms, products, or services by the US Government.

Unclassified

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER Technical Report ARLCD-TR-78027	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) NITROFORM RECOVERY AND ISOLATION STUDIES	5. TYPE OF REPORT & PERIOD COVERED Final Report December 1975 - January 1978	
7. AUTHOR(s) R. A. Mundy, Radford Army Ammunition Plant E.E. Gilbert, Project Engineer, ARRADCOM	6. PERFORMING ORG. REPORT NUMBER PE-557 (RAD 240.10)	
9. PERFORMING ORGANIZATION NAME AND ADDRESS Radford Army Ammunition Plant Hercules Incorporated Radford, Virginia 24141	8. CONTRACT OR GRANT NUMBER(s) 14/RAD-240.10, PE-557	
11. CONTROLLING OFFICE NAME AND ADDRESS US Army Armament Research & Development Command ATTN: DRDAR-TSS Dover, NJ 07801	10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS DA Project No. 1T762720048 Work Unit W-5	
14. MONITORING AGENCY NAME & ADDRESS (If different from Controlling Office) Energetic Materials Division, LCWSL US Army Armament Research & Development Command Dover, NJ 07801	12. REPORT DATE 11 May 1978	
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release, distribution unlimited. 18/ARLCD, SBIE 19/TR-78027 AD-E499164	13. NUMBER OF PAGES 69 12/68	
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)	15. SECURITY CLASS. (of this report) Unclassified	
18. SUPPLEMENTARY NOTES This project was funded by Edgewood Arsenal, Aberdeen, MD 21010	15a. DECLASSIFICATION/DOWNGRADING SCHEDULE	
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Sodium Nitroform Tetranitromethane Nitroform Pollution Abatement		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) An investigation was made to effectively remove tetranitromethane (TNM) from the effluent gas stream from trinitrotoluene (TNT) manufacturing operations and convert it to nitroform (NF). The TNM was efficiently removed by passing the gas through a column scrubber containing an aqueous solution of sodium carbonate (Na ₂ CO ₃) and stabilized hydrogen peroxide (H ₂ O ₂). The resulting sodium nitroform (SNF) solution can be vacuum distilled to obtain NF which is a marketable by-product.		

407 057

JOB

CONTENTS

	<u>Page No.</u>
Summary	1
Study	5
I. Introduction	5
A. Toxicity of Tetranitromethane	5
B. Present Fume Recovery System	6
C. Previous Investigations	6
II. Investigation of Operating Parameters	6
A. Description of Scrubbing Unit	6
B. Analysis of the Gas Stream	8
C. Stability of Hydrogen Peroxide	8
D. Scrubbing Efficiencies at Low Sodium Nitroform Concentrations	10
E. First Extended Column Operation	10
F. Second Extended Column Operation	12
G. Stability of Sodium Nitroform in the Scrubbing Liquid	16
III. Potential Economic Aspects	20
IV. Conclusions and Recommendations	21
A. Conclusions	21
B. Recommendations	21
Appendix A - Laboratory Analysis Procedures	22
Distribution List	67

LIST OF FIGURES

<u>Figure No.</u>		<u>Page No.</u>
1	Diagram of TNM scrubber flows	7
2	Albone DS - H ₂ O ₂ concentration vs time in 0.45 N Na ₂ CO ₃ solution	9
3	Gas scrubbing efficiency, extended run No. 1	14
4	Ratio of ml of H ₂ O ₂ added to grams of TNM through column versus time for extended run No. 1	15
5	Ratio of ml of H ₂ O ₂ added to grams of TNM through column versus time for extended run No. 2	17
6	Gas scrubbing efficiency, extended run No. 2	19

LIST OF TABLES

<u>Table No.</u>		<u>Page No.</u>
1	Sodium Carbonate/Hydrogen Peroxide Column Scrubbing	11
2	Summary of First Extended Column Scrubbing Run	13
3	Summary of Second Extended Column Scrubbing Run	18

STUDY

I. INTRODUCTION

Previous work ^{1/} established that TNM could be effectively removed by scrubbing the TNT fume recovery tail gas with Na₂SO₃ solution. However, recovery of the NF salt, which has a potential market value, was not feasible because the salt was not stable in the Na₂SO₃ solution. Other scrubbing solutions ^{2/} were evaluated to establish the feasibility of economically achieving recovery of NF. A NaOH solution combined with caustic stabilized H₂O₂ (Albone "DS" peroxygen compound from DuPont) produced stable NF which could be recovered by neutralization and distillation. The fume recovery stream contains, in addition to 750-1000 ppm TNM, 2 to 5 percent each of carbon dioxide (CO₂) and carbon monoxide (CO). The CO₂ reacts with NaOH to produce Na₂CO₃, thus depleting the NaOH. The purpose of this study was to investigate the effect of CO₂ absorption from the fume gas stream by using a Na₂CO₃-H₂O₂ scrubbing solution. These data would also establish the feasibility of using Na₂CO₃ as the feed alkali which is less expensive and easier to handle than NaOH.

A. Toxicity of Tetranitromethane

Tetranitromethane, C(NO₂)₄, is a toxic by-product produced during the manufacture of TNT. It is a heavy oily liquid (sp. gr. 1.65) which solidifies at 3°C and boils at 126°C without decomposition ^{3/}. It has a toxic hazard rating and the recommended personnel exposure level given by the American Conference of Government and Industrial Hygienist (ACGIH) is one part per million in air or 0.8 milligrams per cubic meter of air ^{4/}.

^{1/} Morris, J. A., Propellant Plant Pollution Abatement - Elimination of NO_x in Process Tail Gas from TNT Manufactured by Absorption in Sulfuric Acid, Final Engineering Report on Production Engineering Project PE-352 (Phase I), RAAP, April 1973.

^{2/} Mathes, C. D., Removal and Recovery of TNM from TNT Manufacture, Supplementary Engineering Report on Production Engineering Project PE-352, RAAP, October 1975.

^{3/} Urbanski, T., Chemistry and Technology of Explosives, Volume 1, p. 588, Pergamon Press, 1964.

^{4/} Sax, Irving N., Properties of Industrial Materials, p. 1151, 3d Edition.

B. Present Fume Recovery System

During previous operation of the TNT Plant at Radford Army Ammunition Plant (RAAP), the fumes from nitrating operations were passed through water scrubbers which removed approximately 90 percent of the NO_x but none of the TNM. This allowed 300 to 600 ppm (V/V) of TNM to reach the atmosphere. A separate stack venting the acid washer also discharged TNM and will be combined with the present fume gas stream on start up of the rebuilt lines. Emissions are estimated to be 750-1000 ppm when these two streams are combined.

C. Previous Investigations

Cursory laboratory and bench-scale studies were conducted under PE-352 to evaluate three scrubber solutions for the abatement of TNM. The three solutions were Na_2SO_3 , NaOH with stabilized H_2O_2 and Na_2CO_3 with stabilized H_2O_2 ^{5/}. All three solutions effectively removed TNM from the tail gases. The Na_2SO_3 scrubbing solution was eliminated from consideration because the SNF which formed decomposed very rapidly, i.e., NF could not be recovered from the scrubbing solution as a marketable by-product. Both the NaOH- H_2O_2 and Na_2CO_3 - H_2O_2 systems effectively removed the TNM and the SNF could be converted to NF by neutralization and vacuum distillation of the solution, but no long term scrubbing data were available to assess the effect of CO_2 absorption. Also, little information was available on total SNF salt buildup in the scrubbing solution.

II. INVESTIGATION OF OPERATING PARAMETERS

A. Description of Scrubbing Unit

A schematic of the bench-scale system is shown in Figure 1. The scrubbing unit was a heavy wall glass column 1.5 m high with an inside diameter of 102 mm and packed to a depth of 0.9 m with ceramic saddles. The 0.9 m of packing contained a free volume of five liters. The entire system was maintained under a reduced pressure which was produced by a Thomas Industries, Incorporated Model 2727CA39, two stage diaphragm stainless steel pump attached to the exit side of the column. TNM in the exit gas was monitored with a Miran II infrared (IR) gas analyzer ^{6/}.

^{5/} Gilligan, W. H., Hall, T. N., Removal of Tetranitromethane from Air Streams, Naval Surface Weapon Center, White Oak, Technical Report 75-128, October 1975.

^{6/} Murray, J. R., Monitoring and Controlling Systems for Tetranitromethane Pollutants in TNT Exhaust Gases, Final Report on Production Engineering Project PE-499, Radford Army Ammunition Plant, to be published.




Figure 1. Diagram of TNM scrubber flows

Liquid TNM was placed in a 250 ml Erlenmeyer flask nebulizer. Air was bubbled through the liquid TNM and then mixed at the column inlet with additional air controlled by a needle valve. By adjusting the two air flows, the TNM concentration could be varied over a wide range. The Erlenmeyer flask was used as a nebulizer because of its large diameter flat bottom as compared to standard air bubblers. The flat bottom permitted a supply of TNM for a day's operation without a rapid change in liquid depth. A change in liquid depth would have required numerous changes in air to TNM vapor flows to maintain a constant TNM inlet concentration. The TNM stream was then passed through a flowmeter and into the bottom of the scrubber column. A 1500 ml filtration flask was inserted in the gas line at the exit of the column just prior to the vacuum pump to prevent any liquid carryover from reaching the vacuum pump. A tee was also mounted in the exit line to allow the gas to be pulled through the IR analyzer for continuous monitoring of TNM concentration. The liquid phase in the scrubber column was constantly circulated from the bottom to the top with a variable speed Monostat varistaltic pump. A tee was placed in the inlet and outlet lines to the pump to take samples or make additions to the liquid phase.

B. Analysis of the Gas Stream

In the early portion of this project, the TNM in the inlet and exit gas streams was monitored using the IR analyzer. This was found to be unsatisfactory for the inlet stream because, at TNM concentration levels above 300 ppm, the slope of the absorbance versus TNM concentration curve is very flat and, therefore, is not sensitive to high TNM concentrations.

The inlet gas was, therefore, analyzed by several methods. One method used a Thermo Electron Corporation Model 10A Chemiluminescent NO-NO_x Gas Analyzer. See Appendix A-2 for description of the method. A Hewlett-Packard Gas Chromatograph Model 5700A equipped with a flame ionization detector was also used for TNM analysis. A description of this method is shown in Appendix A-1. Both methods were found suitable for inlet TNM analysis.

C. Stability of Hydrogen Peroxide

Hydrogen peroxide is not stable in basic solution unless a stabilizer such as magnesium sulfate is present. In order to determine the stability of DuPont Albene DS-50 stabilized H₂O₂ in a solution of Na₂CO₃, a known concentration was prepared and allowed to stand during which the H₂O₂ concentration was monitored. The original solution contained 1.04 percent H₂O₂ in a 0.45 N Na₂CO₃ aqueous solution. After 16 hours, the solution had spontaneously reduced in strength to 0.22 percent H₂O₂. Figure 2 is a graph showing H₂O₂ concentration versus time. These data indicate that if work is not being performed with the scrubbing solution, reactions continue thus consuming the residual H₂O₂.




Figure 2. Albone DS - H₂O₂ concentration vs time in 0.45 N Na₂CO₃ solution

D. Scrubbing Efficiencies at Low Sodium Nitroform Concentrations

The first phase of the investigation was to determine the optimum scrubber operating parameters. The variables considered were packing depth, Na_2CO_3 concentration, H_2O_2 concentration, TNM inlet concentration, gas flow rates and volume of liquid holdup in the column. From the data shown in Table 1, the following parameters were chosen as the most suitable for the scrubbing studies:

Na_2CO_3	2.25 to 2.50 percent
H_2O_2	2.0 percent
TNM Inlet	750 - 950 ppm
Gas Flow	24 - 26 l/min
Packing Height	0.9 m
Scrubbing Liquid	4 liters

E. First Extended Column Operation

Four liters of scrubbing solution were prepared containing 1.82 percent H_2O_2 and 2.58 percent NaCO_3 . This solution was then pumped into the column containing 0.9 m height of ceramic packing. This solution was circulated from the bottom of the column to the top at a rate of approximately one l/min using a variable speed pump.

The column was maintained at a pressure of 250 mm of mercury using a two-stage diaphragm vacuum pump. Room air was pulled into the column through the liquid TNM and the air makeup needle valve at a rate to produce a TNM inlet concentration between 700 and 1000 ppm and gas flow of 24 to 25 l/min.

During the first 12.25 hours of operation, an attempt was made to keep the level of H_2O_2 at 1.80 percent which required a total of 550 ml of additional 50 percent H_2O_2 . The removal of TNM from the air flow was excellent with exit concentrations between 10 and 50 ppm. The column was operated only during the day and was shut down overnight and week-ends. During shutdown periods, the H_2O_2 concentration dropped to zero. It was noted during startup that the first few milliliters of H_2O_2 entering into the scrubbing solution immediately decreased the TNM in the effluent indicating that a much lower concentration of H_2O_2 might be used.

During the next 28 hours of the run, the addition of H_2O_2 was based on the TNM concentration in the effluent stream rather than trying to maintain a constant percentage. Using the same operating conditions, except the inlet TNM averaged higher at 900 ppm, only 268 ml of H_2O_2 was required. The effluent TNM concentration averaged 45 ppm and the H_2O_2 level ranged between zero and 0.05 percent.

The effluent TNM was continuously monitored using the Miran II analyzer and when the concentration reached 70 - 80 ppm, five to ten ml of H_2O_2 were added, reducing the TNM to 35 - 40 ppm. The higher consumption of H_2O_2 during the first 12.25 hours appeared to be due to rapid decomposition

Table 1
Sodium carbonate/hydrogen peroxide column scrubbing

Run	Sodium Carbonate, %	Hydrogen Peroxide, %	TNM Inlet, ppm	TNM Exit, ppm	Removal Efficiency, %	Flow Rate Through Column, l/min	Packing Height in Column, mm	Packing Height in Column, inches	Scrubbing Liquid in Column, liters
4-1	1.40	0.76	325	55	83	34	559	22	4
4-2	1.40	0.76	275	75	73	34	559	22	2
4-3	1.40	0.76	275	160	42	34.5	559	22	1
5-1	2.53	1.29	540	15	97	25	914	36	4
5-2	2.53	1.29	580	15	97	25	914	36	3
5-3	2.53	1.29	640	18	97	30	914	36	2
5-4	5.24	1.25	760	21	97	24.5	914	36	4
5-5	5.24	1.25	760	16	98	25.5	914	36	2
5-6	5.28	1.80	250	12	95	23.8	914	36	4
5-7	4.23	0.95	395	12	97	24.6	914	36	4
5-8	4.32	0.55	465	18	96	24.3	914	36	4
5-9	4.32	0.49	540	26	95	24.6	914	36	4
5-10	0.52	0.78	720	15	98	24.6	914	36	4
5-11	0.26	0.81	720	18	97	24.6	914	36	4
5-12	2.22	0.25	950	24	98	24.8	914	36	4

of the H_2O_2 rather than being consumed in a reaction with TNM. The Na_2CO_3 concentration remained between 1.4 and 2.5 percent. The H_2O_2 addition was based on the TNM effluent concentration for the remainder of the run. After 49 hours of operation, the scrubbing solution color had changed from yellow to red and the consumption of H_2O_2 had increased four to five times. See Table 2 for comparison of the amount of H_2O_2 added per gram of TNM. This change was thought to have been due to the decrease in pH. A solution of 1.25 percent Na_2CO_3 and 0.1 percent H_2O_2 has a pH of 10.58 and the pH of the scrubbing solution after 53.5 hours was 9.21. The SNF concentration was 4.77 percent.

Operation of the column was continued for a total of 72 hours to determine if the consumption of H_2O_2 could be reduced. Adding Na_2CO_3 to the scrubbing solution resulted in short periods of reduced H_2O_2 consumption. Titration of the scrubbing solution after 59 hours with 0.3 N hydrochloric acid (HCl) using a pH meter showed the Na_2CO_3 concentration was not 2.45 percent as determined using the dead stop titration method but 1.21 percent. The difference between the 2.45 and 1.21 percent was due to 1.42 percent sodium bicarbonate ($NaHCO_3$) in the solution.

The run was terminated after 74 hours. The analysis of the scrubbing solution was 2.66 percent Na_2CO_3 , 1.87 percent $NaHCO_3$, 4.65 percent nitrates, 2.43 percent nitrites, and 6.75 percent SNF.

Figure 3 shows the decline in the efficiency of TNM removal from the gas stream versus hours of operation. The graph presented in Figure 4 shows the ratio of H_2O_2 consumption per gram of TNM entering the column versus hours of operation. It can be seen that the column solution will provide an operation for 40 hours and produce a SNF concentration of only 5.5 percent before additional reactants are required.

A true conversion efficiency of TNM to SNF was not obtained because there were some fluctuations in the total scrubber volume and also some variance in the test for SNF at the higher concentration. The average estimated consumption efficiency of the first extended run was 65 percent. This was calculated on the basis that one mole of TNM reacts to form one mole of SNF. It does not take into account any side reaction that may involve either the TNM or SNF.

F. Second Extended Column Operation

In order to better understand the operation of the TNM scrubber, a duplicate run was initiated. The four liters of starting scrubbing solution contained 2.66 percent Na_2CO_3 and 0.07 percent $NaHCO_3$ and no H_2O_2 . The H_2O_2 was added as required and the Na_2CO_3 , $NaHCO_3$ and pH were monitored in addition to parameters in the first run. After 4.5 hours of operation, the pH had dropped to 9.75, the Na_2CO_3 was 1.39 percent, and the $NaHCO_3$ was 0.57 percent. After 15.25 hours, the pH had again dropped to 9.45 and the solution contained 0.54 percent Na_2CO_3 . Only small amounts of H_2O_2 were needed for each gram of TNM and the conversion of TNM to SNF was 83.04 percent with 2.52 percent SNF in solution. The yield

Table 2
Summary of first extended column scrubbing run

Date	Operation Total Hours	Na ₂ CO ₃ , %	NaHCO ₃ , %	SNF, %	TNM Used per Day, g	pH	Inlet TNM, ppm	Exit TNM, ppm	TNM Removed, %	50% H ₂ O ₂ Added, ml	H ₂ O ₂ TNM, ml/g	60% NaOH Added, ml	Na ₂ CO ₃ Added, g	Vol. in Col, ml	H ₂ O ₂ , %
12-2-76	0	2.58	(1)	-	-		760	5	99.34	-		0	0	4000	1.82
12-2-76	5	1.74		0.65	39.2		698	6	99.14	1.50	3.83	0	0	4000	1.83
12-7-76	10	1.65		0.62			-	-		-		0	0	4000	0
12-8-76	12.25	1.63		-	46.0		719	10	98.61	400	8.70	0	0	4000	-
12-9-76	15.50	2.70		2.22	46.0		773	52	93.27	70	1.52	0	60.0	4000	0.03
12-13-76	22.00	2.36		3.28	31.9		856	50	94.16	45	1.41	0	32.0	4000	0
12-14-76	28.50	1.85		4.23	38.0		956	52	94.56	20	0.53	0	0	4000	0.01
12-15-76	35.50	1.90		5.12	65.4		889	64	92.80	63	0.96	0	40.0	4000	0.03
12-16-76	39.00	2.41		5.70	35.4		891	63	92.93	50	1.41	0	20.0	4000	0.07
2-3-77	39.00	3.10		5.62	-		-	-		-		0	0	3200	-
2-4-77	45.25	2.72		-	68.3		887	56	93.69	40	0.59	0	0	4000	-
2-8-77	48.75	1.97		-	37.7		878	71	91.91	65	1.72	0	0	4000	-
2-9-77	53.50	1.20		4.77	18.9	9.21	641	115	82.06	50	2.65	0	70.0	4000	0
2-11-77	57.50	2.33		-	30.0	9.30	701	88	87.45	140	4.67	0	45.0	3000	0
2-15-77	64.50	4.35	0.05	-	34.7	10.43	718	138	80.78	200	5.76	0	60.0	3000	-
2-16-77	67.50	3.63	0.77	-	17.6	9.85	811	79	90.26	50	2.84	0	0	3000	-
2-17-77	72.00	2.66	1.87	6.75	21.2	8.97	731	102	86.05	110	5.19	0	25.0	3000	-

(1) 12-2-76 through 2-11-77 analyzed for total alkalinity only and reported as Na₂CO₃.




Figure 3. Gas scrubbing efficiency, extended run No. 1





Figure 4. Ratio of ml of H₂O₂ added to grams of TNM through column versus time for extended run No. 1

was calculated based on one mole of TNM producing one mole of SNF. The number of moles entering the column was determined by weighing the TNM vaporizer, before and after a day's operation. After 28 hours of operation, the column contained 4.63 percent SNF but the consumption ratio (ml of H₂O₂ to grams of TNM) had increased by a factor of three and the pH steadily decreased.

Figure 5 shows an increase in H₂O₂ consumption through 28 hours and then a decrease through 43 hours. This condition was not noted for run No. 1 as seen in Figure 4. Throughout the remainder of the run, the H₂O₂ consumption increased steadily but could be reduced for short periods by addition of Na₂CO₃ or NaOH to increase the pH and lower the NaHCO₃ concentration. Table 3 shows the amounts of each reactant required.

As observed in the first extended run, the NaOH has a slightly greater effect in reducing the H₂O₂ consumption than the Na₂CO₃. This indicates that the main factors which increase H₂O₂ consumption are low pH and high NaHCO₃ concentration. The five percent CO₂ that is in the TNT process exit gas with TNM will lower the pH and increase the NaHCO₃ concentration, and in turn increase NaOH or Na₂CO₃ consumption.

The reactions that occur between the CO₂ in the TNT process exit gas, Na₂CO₃ and NaOH are as follows:



These equations indicate that during actual operation with two to five percent CO₂ in the exit gas, the NaHCO₃ content will increase and reduce the efficiency of TNM removal. The addition of NaOH will reduce the NaHCO₃ level, increase the pH and increase the efficiency of the TNM removal. The second extended run ended after 63 hours of operation when the ratio of H₂O₂ to TNM reached 5.85. During the run, the efficiency of TNM removal dropped from 94.88 to 88.06 percent (see Figure 6) and the overall efficiency of the conversion of TNM to SNF was 64 percent.

G. Stability of Sodium Nitroform in the Scrubbing Liquid

During both extended runs, there were periods when the column operation was shut down due to lack of materials or a malfunction of equipment. The SNF concentration was monitored periodically during these periods to determine its stability in the scrubber solution.

The first run was shut down for a period of 44.5 days. At the beginning of this interval, the SNF content was 5.70 percent and at the end, 5.64 percent. During this time the H₂O₂ concentration was zero and the combined concentrations of Na₂CO₃ and NaHCO₃ calculated as Na₂CO₃ was 2.41 percent.




Figure 5. Ratio of ml of H₂O₂ added to grams of TNM through column versus time for extended run No. 2

Table 3
Summary of second extended column scrubbing run

Date	Total Hours	Na ₂ CO ₃ , %	NaHCO ₃ , %	SNF, %	TNM Used per Day, g	pH	Inlet TNM, ppm (1)	Exit TNM, ppm (1)	TNM Removed, %	50% H ₂ O ₂ Added, ml	H ₂ O ₂ / TNM, ml/g	50% NaOH Added, ml	Na ₂ CO ₃ Added, g	Vol in Col, ml
2-18-77	0	2.66	0.07	-	-	10.95	-	-	-	-	-	-	-	4000
2-18-77	4.5	1.21	1.05	0.92	44.9	9.75	761	39	94.88	60	1.34	0	0	4000
2-25-77	9.5	1.39	0.57	-	36.5	10.05	807	59	92.69	80	2.19	17	0	4000
2-28-77	15.25	0.54	0.94	2.52	53.0	9.45	770	59	92.34	60	1.13	0	0	3950
3-2-77	17.75	0.29	0.86	2.85	24.4	9.30	770	66	91.43	40	1.64	0	0	3950
3-3-77	23.0	1.34	1.29	3.48	49.3	9.47	769	82	89.34	100	2.02	0	80	4000
3-4-77	28.25	0.97	1.18	4.63	48.9	9.45	740	85	88.51	150	3.07	0	0	4000
6-8-77	34.0	0.42	0.97	4.73	61.2	9.05	803	73	90.91	120	1.96	0	25	4000
6-9-77	39.25	0.74	0.97	5.57	61.0	9.65	880	67	92.39	90	1.48	30	20	4000
6-10-77	43.25	0.36	1.20	6.18	49.6	9.20	917	71	92.26	70	1.41	0	20	4000
6-13-77	48.25	1.75	0.77	7.42	62.6	9.65	872	92	89.45	210	3.35	30	40	3650
6-14-77	54.25	1.00	0.96	7.97	59.9	9.40	903	95	89.48	240	4.00	10	0	3350
6-21-77	60.50	0.54	0.83	9.42	68.7	9.06	889	95	89.31	245	3.57	20	0	2800
6-22-77	62.75	1.65	1.34	9.66	26.5	9.30	863	103	88.06	155	5.85	0	56	2800

(1) Average concentration during day's operation.





Figure 6. Gas scrubbing efficiency, extended run No. 2

The second extended run was shut down for 79 days. The SNF was 4.63 percent at the time of shutdown and 4.68 percent at the end of this period. The Na₂CO₃ was 1.18 percent and the NaHCO₃ was 0.97. These did not change during the shutdown period.

The above results show that the scrubbing solution can be stored for long periods of time without significant changes in the SNF content. This long-term stability is presumably attributable to the low level of H₂O₂ remaining in the scrubbing solution. Previous reports ^{5/} show that SNF is decomposed to form nitrite and presumably Na₂CO₃ in the presence of H₂O₂ for an extended period of time.

III. POTENTIAL ECONOMIC ASPECTS

A cost analysis was performed on the usage of Na₂SO₃ versus Na₂CO₃-Albone DS H₂O₂ and NaOH-Albone DS H₂O₂ for the recovery of a relatively stable nitroform salt. The equations describing the reaction are shown below:



*Further decomposition occurs with this compound.

The cost analysis was based on the manufacture of 204 kg per day (24 hours) of TNM from one nitration and purification building operating at a capacity of 454 Mg per day of TNT. Assuming entrained gases, CO, CO₂, NO, and NO_x react with the scrubbing medium and column loss through vaporization result in only 50 percent efficiency, the cost of the essential ingredients for one day (24 hours) would be approximately:

Chemical	Kilograms	Cost/kg, \$	Essential Ingredients
			Total Cost
Na ₂ SO ₃	263	0.370	\$96.69
Na ₂ CO ₃	248	0.127	\$31.51
NaOH	278	0.384	\$103.67
H ₂ O ₂ *	1021	0.620	\$630.00

*Required in addition to either Na₂CO₃ or NaOH.

The Na_2SO_3 scrubbing system would remove 204 kg of TNM at a cost of \$96.69 but no SNF would be available for conversion to NF. The $\text{Na}_2\text{CO}_3 + \text{H}_2\text{O}_2$ would remove the same amount of TNM at a cost of \$661.51 and the $\text{NaOH} - \text{H}_2\text{O}_2$ at a cost of \$733.67.

Assuming that all the TNM was converted to SNF according to the equations and that 90 percent of the SNF could be converted to NF, the cost per kg for essential ingredients would be \$5.34 per kg using $\text{Na}_2\text{CO}_3 - \text{H}_2\text{O}_2$. The cost of NF from a foreign supplier was \$26.50 per kg in 1975.

Additional processing cost would be incurred for utilities and labor but would be the same for all scrubbing systems. The cost of the conversion of SNF to NF has not been determined but the conversion would probably be performed by the purchaser.

IV. CONCLUSIONS AND RECOMMENDATIONS

A. Conclusions

1. The system $\text{NaOH}-\text{H}_2\text{O}_2$ is viable for the abatement of TNM and recovery of NF from a gas stream containing 750-1000 ppm TNM. The $\text{Na}_2\text{CO}_3-\text{H}_2\text{O}_2$ will also abate TNM but consumption of H_2O_2 increases drastically to maintain a 100 ppm level of TNM as the SNF salt concentration increases. This occurs because the high level of CO_2 in the fume gas stream reacts with Na_2CO_3 to produce NaHCO_3 which reacts slowly with TNM. Even with the NaOH system, CO_2 will be absorbed to produce Na_2CO_3 which in turn can react with additional CO_2 to form NaHCO_3 . Therefore, NaOH is the preferred base and its consumption will increase over theoretical requirements.

2. Long-term stability tests have indicated that solutions of SNF with the concentration studied, were stable with little or no deterioration. Based on these results, shipment of these solutions to a purchaser for final processing to NF appears possible.

B. Recommendations

Should the Army desire to recover NF for use in high energy propellant systems, on-line scrubbing studies would be required to determine control variables and SNF handling procedures. Further study of the neutralization and distillation areas would also be required if NF was the saleable end product.

APPENDIX A

LABORATORY ANALYSIS PROCEDURES

CONTENTS

Procedure

- A-1 Gas Chromatographic Method for the Determination of Tetranitromethane in Air
- A-2 Chemiluminescence Method for the Determination of Tetranitromethane in Air
- A-3 Miran II Infrared Spectrophotometric Method for the Determination of Tetranitromethane in Air
- A-4 No. L-102, Analysis of Sodium Nitroform by Ultraviolet Spectrophotometry
- A-5 No. L-101, Nitrate Analysis of Water by Ultraviolet Spectrophotometric Method
- A-6 Analysis of Sodium Carbonate and Hydrogen Peroxide in Sodium Nitroform Scrubber Solution
- A-7 Analysis of Tetranitromethane Scrubbing Solution for Sodium Carbonate and Sodium Bicarbonate Concentration
- A-8 Determination of Nitrites

Procedure A-1

Gas Chromatographic Method for the
Determination of Tetranitromethane in Air

Gas Chromatographic Method for the
Determination of Tetranitromethane in Air

1. SCOPE AND APPLICATION

- 1.1 This method is suitable for determining tetranitromethane (TNM) in gas streams with concentrations between 300 and 1200 ppm (v/v) TNM.

2. SUMMARY OF METHOD

- 2.1 Sample is taken directly from the gas stream using a gas syringe and injected directly into the chromatographic column.

3. INTERFERENCES

- 3.1 Any compounds in the gas stream that will elute from the gas chromatograph (GC) at the same time as the TNM will interfere with this test.

4. APPARATUS

- 4.1 1000 μ l gas tight syringe.
- 4.2 Hewlett Packard Model 5750 GC equipped with a flame ionization detector and a 12 ft 1/8" stainless steel column packed with Chromosorb WHP 80-100 mesh coated with 10 percent UCW 98 silicone rubber.
- 4.3 Hewlett Packard Model 3370B Integrator.

5. CALIBRATION

- 5.1 Weigh approximately 4 g of TNM into a 25 ml volumetric flask which is half filled with carbon disulfide. Fill the flask to the mark with carbon disulfide.
- 5.2 Using a 25 μ l syringe, inject 5 μ l into a gas sampling tube with a known volume. Calculate the concentration of TNM in the sample tube using the following equation:

$$\text{PPM TNM} = \frac{\text{g TNM}/25 \text{ ml CS}_2 \times V_1}{25,000 \times 196.03} \times \frac{0.08205 \times 293}{1} \times \frac{10^6}{V_2}$$

V_1 = μ l injected into gas sample tube

V_2 = vol of gas sample tube in liters

- 5.3 Repeat 5.2 using 10 and 15 μ l standards.
- 5.4 Inject a 500 μ l sample from each sample tube into the GC and record the peak areas.

5.5 Prepare a graph of peak area versus TNM concentration.

6. PROCEDURE

6.1 Withdraw 500 μ l of gas from the inlet side of the scrubber in a 1000 μ l syringe and inject into the GC. Record the peak area and determine TNM concentration from the calibration graph.

Procedure A-2

Chemiluminescence Method for the
Determination of Tetranitromethane in Air

Chemiluminescence Method for the
Determination of Tetranitromethane in Air

1. SCOPE AND APPLICATION

- 1.1 This method is suitable for measuring tetranitromethane (TNM) in gas streams with concentrations between 100 ppm and 1500 ppm TNM.

2. SUMMARY OF METHOD

- 2.1 A sample is removed from the gas stream and injected into the inlet side of the instrument. The TNM is converted to NO in the instrument and the resulting chemiluminescence generated with ozone is a measure of the TNM concentration.

3. INTERFERENCE

- 3.1 Any NO, NO₂ or material that can be converted to these species will interfere with this test.

4. APPARATUS

- 4.1 Model 10A self-contained chemiluminescent NO-NO_x Gas Analyzer manufactured by Thermo-Electron Corporation.
- 4.2 Suitable strip chart recorder.

5. CALIBRATION

- 5.1 Weigh approximately 4 g of TNM into a 25 ml volumetric flask which is half filled with carbon disulfide. Fill the flask to the mark with carbon disulfide.
- 5.2 Using a 25 μ l syringe, inject 5 μ l into a gas sampling tube with a known volume. Calculate the concentration of TNM in the tube using the following equation:

$$\text{PPM TNM} = \frac{\text{g TNM}/25 \text{ ml CS}_2 \times V_1}{25,000 \times 196.03} \times \frac{0.08205 \times 293}{1} \times \frac{10^6}{V_2}$$

V_1 = μ l injected into gas sampling tube

V_2 = vol of gas sample tube in liters

- 5.3 Repeat 5.2 using 10 and 15 μ l standards.
- 5.4 Inject a 500 μ l sample using a 1000 μ l syringe from each standard into the gas stream entering the instrument and record the peak height.

- 5.5 Prepare a graph of peak height versus concentration.
- 5.6 Set the instrument controls so that peak height is approximately three quarters maximum chart travel.
6. PROCEDURE
 - 6.1 Remove a 500 μ l sample using a 1000 μ l syringe from the scrubber gas stream.
 - 6.2 Inject the sample into the gas stream entering the instrument.
 - 6.3 Measure peak height and determine TNM concentration directly from the calibration curve.

Procedure A-3

Miran II Infrared Spectrophotometric Method
for the Determination of Tetranitromethane
in Air

Miran II Infrared Spectrophotometric Method
for the Determination of Tetranitromethane in Air

1. SCOPE AND APPLICATION

- 1.1 This method is suitable for determining the concentration of tetranitromethane (TNM) in air.

2. SUMMARY OF METHOD

- 2.1 An air stream containing TNM is passed through the Infrared Analyzer which absorbs radiation at a set wavelength. The absorption of radiation is proportional to the TNM concentration.

3. INTERFERENCES

- 3.1 Any material in the gas stream that absorbs at the same wavelength as the TNM will interfere with the test.

4. APPARATUS

- 4.1 Miran II infrared analyzer equipped with a variable path length 20 meter sample cell with filters transmitting at 798 cm^{-1} for TNM and 2564 cm^{-1} as a reference. Windows are silver bromide and mirrors are gold coated. Instrument is manufactured by Wilks Scientific Corporation.

- 4.2 Vacuum pump

5. CALIBRATION

- 5.1 The Miran II analyzer sample cell is evacuated to an absolute pressure of 250 mm of mercury.

- 5.2 Inject one μl of TNM into the closed system and allow one minute for TNM to vaporize.

- 5.3 Slowly open the sample cell valve to reach atmospheric pressure. Record absorbance reading.

- 5.4 Repeat 5.1 through 5.3 using 2, 3 and 4 μl of TNM.

- 5.5 Determine TNM concentration using the following equation:

$$\text{PPM TNM} = \frac{\mu\text{l} \times 1.639 \text{ g/ml}}{196.03} \times \frac{0.08205 \times (273 + T)}{1 \text{ atm}} \times \frac{10^3}{V}$$

μl = microliters of TNM injected

1.639 g/ml = density of TNM

196.03 g = molecular weight of TNM

0.08205 = gas constant

T = temperature of instrument °C

V = volume of instrument sample cell and connecting tubing in liters

6. PROCEDURE

- 6.1 Attach inlet side of instrument to sample port on exit side of column and allow a portion of exit gas to be flushed through the instrument by means of the vacuum pump. When instrument readings become steady, record absorbance.
- 6.2 Determine the TNM concentration directly from the calibration curve.

NOTE: Instrument settings must be the same for standards and samples.

Typical settings:

Path Length Dial = 9.05

Expansion = 1X

Zero Suppression = As necessary to zero instrument

Response = Fast

Meter = Diff, 0-1

Procedure A-4

Analysis of Sodium Nitroform by
Ultraviolet Spectrophotometry

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L- 102	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>2</u> of <u>6</u>	Date 12-18-74

4.9 125 microliter syringe

5. SAFETY

5.1 Safety glasses, flame resistant lab coats or coveralls, and conductive safety shoes shall be worn at all times.

5.2 Rubber gloves shall be worn when working with TNM.

5.3 All sampling of tetranitromethane shall be done in a hood with adequate ventilation.

6. OPERATION

6.1 Turn on instrument and tungsten lamp. Make sure knob is on TUNGSTEN LAMP and not on HYDROGEN LAMP.

6.2 Turn knobs beside cell holders on photomultiplier (IX).

6.3 Fill sample and reference cell with 2 % sodium hydroxide solution.

6.4 Set wavelength to 500A mμ

6.5 Open reference light and sample light.

6.6 Turn range to 0 - 1.

6.7 Turn to % T and adjust 100 % base line knob to 100 % reading.

6.8 Turn to A and adjust 0 % base line with zero adjust knob.

6.9 Set up instrument as follows:


A
Time Constant 0.1
0 % Adj:
100 % Adj:
Range 0 - 1
Do not adjust sensitivity.

6.10 Place sample in cell holder nearest front of instrument.

6.11 Place recording paper in instrument.

6.12 Set recorder pen to extreme right side and lower pen.

Approved: 
Chemical Laboratories Area Supervisor


Technical Director

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories Technical and Quality Assurance Department	Procedure No. L-102	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>3</u> of <u>6</u>	Date 12-18-74

- 6.13 Set scan time on 2X, scale expansion to 1X from N and motors from N to IN.
- 6.14 Turn switch below wavelength knob to ON.
- 6.15 Let run until top of first peak comes off (near 340).
- 6.16 Turn scan switch OFF.
- 6.17 Turn motors to OUT.
- 6.18 Turn scale expansion to N and lift recording pen.
- 6.19 Replace sample with a new sample and start at 6.12 above, but this time only lower the pen.
- 6.20 Repeat as often as required to analyze all samples.
- 6.21 Preparation of Standard Curve.
- 6.21.1 Prepare a 100 ml volumetric flask containing approximately 75 ml of a 2 % NaOH solution and 0.5 % H₂O₂. Accurately weigh to nearest milligram.
- 6.21.2 Add 60 microliters of pure TNM to the flask. Perform all operations with TNM in a well ventilated hood. The solution will assume a lemon yellow coloration.
- 6.21.3 Reweigh accurately to get the weight of TNM added.
- 6.21.4 Make up the flask to 100 ml volume with a 2 % NaOH/0.5 % H₂O₂ solution.
- 6.21.5 Pipette 10 ml of the solution prepared into a 100 ml volumetric flask to get a 1:10 volumetric dilution. Mix thoroughly.
- 6.21.6 Prepare 100 ml volumetric flasks to contain each, 1 ml, 2 ml, 3 ml, 4 ml, 5 ml, 6 ml, 7 ml, 8 ml, 9 ml and 10 ml of the solution prepared in 6.21.5 above. Make up to the mark using 2 % NaOH solution.
- 6.21.7 These flasks contain the standards for preparing the Standard curve. Proceed to 6.1 above and measure absorbance with the Beckman DK-2 ultraviolet spectrophotometer.
- 6.21.8 A Standard curve (below) is used to measure sodium nitroform content in samples.

Approved:

[Signature]
Chemical Laboratories Area Supervisor


[Signature]
Technical Director


HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-102	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>4</u> of <u>6</u>	Date 12-18-74

6.22 Analysis of Samples

- 6.22.1 Pipette and weigh 1 ml of sample into a 1000 ml volumetric flask. Make up to volume with 2 % NaOH. Use directly in sample cells.
- 6.22.2 Analyze the sample by the Beckman DK-2 ultraviolet spectrophotometer, starting in Step 6.1 above.
- 6.22.3 Measure the absorbance value obtained and determine the sodium nitroform concentration from the calibration curve.

Approved:


Chemical Laboratories Area Supervisor


Technical Director

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories Technical and Quality Assurance Department	Procedure No. L-102	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>5</u> of <u>6</u>	Date 12-18-74

CALCULATIONS

Standards 60 µl weighed accurately is approximately 0.1033 gram
 With 0.1033 g in 100 ml the concentration is 0.001033 g/ml. This sample
 is further diluted 10 ml to 100 ml to give 0.0001033 g/ml.

Aliquots are taken and diluted to 100 ml volume with 2 % NaOH to give the
 following concentrations:

<u>Aliquot</u>	<u>Concentration X 10⁻⁶ g/ml</u>
1 ml contains	1.033
2 ml contains	2.066
3 ml contains	3.099
4 ml contains	4.132
5 ml contains	5.165
6 ml contains	6.198
7 ml contains	7.231
8 ml contains	8.264
9 ml contains	9.297
10 ml contains	10.330

These concentrations are used to construct a standard curve of absorbance
 versus concentration.

Samples


A 1 ml sample is accurately weighed and diluted to 1000 ml in a volumetric
 flask using 2 % NaOH. If the sample does not fall within the concentration
 range of the standard curve, other dilutions will be required. The absorbance
 is used to determine nitroform concentration on the standard curve.

$$\% \text{ Nitroform salt} = \frac{(\text{conc. from curve}) (\text{dilution factor})(100)}{(\text{Sample wt})}$$

Approved:

Calvin...
 Chemical Laboratories Area Supervisor

W.J. Boll...
 Technical Director



Procedure A-5


Nitrate Analysis of Water by
Ultraviolet Spectrophotometric Method

L-101
November 14, 1974


NITRATE ANALYSIS OF WATER
BY ULTRAVIOLET SPECTROPHOTOMETRIC METHOD


HERCULES INCORPORATED
RADFORD ARMY AMMUNITION PLANT

Prepared By:



W. D. Wygal

Reviewed By:


R. A. Mundy


P. A. Walker, Jr.

Approved By:


T. G. Grady



HERCULES POWDER COMPANY Radford Army Ammunition Plant • Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories	Procedure No. L-101	Policy No. 5
	Technical and Quality Assurance Department	Function No. 2.1.2	Issue 1
		Page <u>1</u> of <u>14</u>	Date 11/14/74

Statement of Function:

Nitrate Analysis of Water by Ultraviolet Spectrophotometric Method

1. SCOPE

1.1 This document establishes the procedure for determining nitrates in outfall water using the Ultraviolet Spectrophotometric Method.

2. REQUIREMENTS

<u>Outfall</u>	<u>Maximum mg/l Nitrates as N</u>	<u>Maximum mg/l Nitrites as N</u>
004	18	-
005	516	-
006	38	-
007	260	6
009	-	4
012	159	6
013	-	3
018	2,036	4
019	270	-
020	271	-
026	-	-
028	-	-
Incinerator	-	-

3. APPLICABLE DOCUMENTS

3.1 Instruction Manuals for Beckman Spectrophotometers Models DU and DK

3.2 Standard Methods for the Examination of Water and Wastewater, Method 133B

3.3 Authorization to Discharge Under the National Pollutant Discharge Elimination System, Permit No. Va 0000248

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-101	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>2</u> of <u>14</u>	Date 11/14/74

3.4 DUP 3786

3.5 Laboratory Safety Handbook

3.6 Laboratory Safety Manual, L-6

4. MATERIALS

4.1 Spectrophotometer Beckman Model DU, DK or equivalent w/photomultiplier attachment and hydrogen lamp source

4.2 Cells, matched silica, 1 cm or longer light path

4.3 Outfall water sample

4.4 Analytical balance

4.5 Filter paper, Glass fiber Type A, 47mm, Gelman Instrument Co.

4.6 Demineralized water

4.7 Volumetric flasks, 50 ml, 1000 ml

4.8 Stock nitrate solution: 721.8 mg potassium nitrate (KNO₃) dissolved in 1000 ml demineralized water = 100 mg/l = 100 ppm NO₃ as N

4.9 Standard nitrate solution: 100 ml of KNO₃ stock solution diluted to 1000 ml with demineralized water. 1 ml = 10 ppm NO₃ as N

4.10 Hydrochloric acid (HCl) solution, 1 N = 10 ml concentrated HCl diluted with 110 ml demineralized water

4.11 Aluminum hydroxide (Al₂O₃·H₂O)

4.12 Graph paper (20 x 20 to the inch)

4.13 Standard nitrite solution: 303.95 mg potassium nitrite (KNO₂) dissolved in 1000 ml demineralized water = 50 mg/l = 50 ppm NO₃ as N

4.14 Lens paper

4.15 DK recorder paper

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-101	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>3</u> of <u>14</u>	Date 11/14/74

5. SAFETY

- 5.1 Safety glasses, flame resistant coats or coveralls shall be worn at all times.
- 5.2 Use only undamaged glassware.
- 5.3 If acid is spilled on skin or clothing, wash off immediately with water.
- 5.4 Allow pipets to drain freely; never force solutions from pipets.
- 5.5 Use rubber suction bulb for drawing solutions into pipets.

6. OPERATION

6.1 DU Settings

Fig. 1

Source Selector	(1)	- Hydrogen
Power Switch	(14)	- ON
Filament Temperature Switch	(15)	- 2
Photomultiplier Gain Switch	(16)	- 4
Phototube Selector	(3)	- 700 - 900
Lead Resistor Switch	(4)	- Photomultiplier
Screen Bias Switch	(18)	- 3
Slit Control	(9)	- 1,825
Wavelength Selector	(13)	- 220 mu, 275 mu
Function Switch	(12)	- 0 - 100
Tungsten Lamp Switch	(2)	- ON
Transmittance-Absorbance Control	(11)	- 0 - 100

6.2 Firing Hydrogen Lamp

- 6.2.1 Turn the Filament Temperature Switch to Warm-Up for about 30 seconds; then turn the switch to Position 1. After lamp is fired, turn the switch to the highest number at which the lamp is stable (Position 2 for this procedure). The lamp is fired when it emits a purple glow. If the lamp does not fire, turn the switch back to Warm-Up for a longer interval and try Position 1 again. Allow an hour for full stabilizing warm-up.

6.3 Preparation of Standards

6.3.1 Nitrate Standards

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-101	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>4</u> of <u>14</u>	Date 11/14/74

6.3.1.1 Prepare nitrate standards by diluting three or more volumes of the standard nitrate solution, plus 1 ml of 1 N HCl to 50 ml with demineralized H₂O. Use buret for measuring standard nitrate solution.

- (a) Blank - 0 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (b) 0.4 ppm - 2 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (c) 0.8 ppm - 4 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (d) 1.2 ppm - 6 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (e) 2.0 ppm - 10 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (f) 3.0 ppm - 15 ml KNO₃ solution, 1 ml 1 N HCl, dilute to volume with H₂O

6.3.2 Nitrite Standards

6.3.2.1 Prepare nitrite standards by diluting three or more volumes of the standard nitrite solution, plus 1 ml 1 N HCl to 50 ml with demineralized H₂O. Use buret for measuring standard nitrite solution.

- (a) Blank - 0 ml KNO₂ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (b) 1 ppm - 1 ml KNO₂ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (c) 2 ppm - 2 ml KNO₂ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (d) 3 ppm - 3 ml KNO₂ solution, 1 ml 1 N HCl, dilute to volume with H₂O
- (e) 4 ppm - 4 ml KNO₂ solution, 1 ml 1 N HCl, dilute to volume with H₂O

6.3.3 Other Standards

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL	Procedure No. L-101	Policy No. 5
	Chemical Laboratories	Function No. 2.1.2	Issue 1
	Technical and Quality Assurance Department	Page <u>5</u> of <u>14</u>	Date 11/14/74

- 6.3.3.1 When anionic surfactants or hexavalent chromium are known to be present in the sample, prepare correction curves for each of these substances at 2 mg/l intervals up to 10 mg/l. Use potassium dichromate ($K_2Cr_2O_7$), and linear alkylate sulfonate with demineralized water.
- 6.3.3.2 Measure the absorbances given by each substance at a wavelength of 220 mu against demineralized water and plot a separate curve for each.
- 6.4 Preparation of Standard Nitrate Curve
- 6.4.1 Rinse the reference cell several times with the nitrate blank; fill the cell with the blank.
- 6.4.2 Dry the cell with lens paper.
- 6.4.3 With the shutter closed, mount the cell in the sample positioner. (Always have the shutter closed when opening the sample compartment.)
- 6.4.4 Adjust the Dark Current Control to balance null meter. If the needle pegs left, turn the control clockwise; if the needle pegs right, turn the control counterclockwise. This sets the zero percent transmission level.
- 6.4.5 Turn the shutter switch to open and adjust the Sensitivity Control to balance the null meter as above. This sets the 100 percent transmission level.
- 6.4.6 With the sample positioner, place each nitrate standard, in turn, in the beam.
- 6.4.7 Turn the shutter switch to open.
- 6.4.8 Balance the null meter with the Adjust Transmittance-Absorption Control.
- 6.4.9 Read the absorbance of each nitrate standard at a wavelength of 220 mu.
- NOTE: If the Dark Current is reset, the Sensitivity Control generally will require re-adjustment, and vice versa.
- 6.4.10 Plot the ppm nitrate as N against the absorbance at 220 mu.
- 6.5 Preparation of Nitrite Curve
- 6.5.1 Repeat steps 6.4.1 through 6.4.9 using the nitrite standards and blank.
- 6.5.2 Plot the ppm nitrite as N against absorbance at 220 mu.

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-101	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>6</u> of <u>14</u>	Date 11/14/74

6.6 Analysis of Sample

- 6.6.1 To approximately 100 ml of sample, add a small amount (dip stirring rod into bottle) of aluminum hydroxide.
- 6.6.2 Mix sample and filter through Glass fiber filter paper.
- 6.6.3 Pipet the required quantity of the sample into a 50 ml volumetric flask. (Use Table I as a guide.)
- 6.6.4 Add 1 ml of 1 N HCl solution to the sample.
- 6.6.5 Make the sample to volume with demineralized water and thoroughly mix.
- 6.6.6 Measure the absorbance at a wavelength of 220 mu to obtain the nitrate reading and at a wavelength of 275 mu to obtain the interference due to dissolved organic matter.
- 6.6.7 Subtract 2 times the absorbance reading at 275 mu from the absorbance reading at 220 mu.
- 6.6.8 Using the nitrite standard curve, convert the ppm nitrite (which has been determined by another method), if present, to absorbance at 220 mu, and subtract this value from the absorbance of the sample at 220 mu. Using this corrected absorbance, read the nitrate concentration from the nitrate standard curve. Correct for other known interfering materials in the same manner using the appropriate standard curve.

$$\text{ppm Nitrate as N} = \text{net ppm Nitrate as N} \times \text{dilution factor}$$

$$\text{ppm Nitrate as NO}_3 = \text{ppm Nitrate as N} \times 4.43$$

6.7 DK-2 Settings

Main Power Switch	(1)	Fig. V	- ON
Detectors	(4)	Fig. IV	- Photomultiplier and 20X
Hydrogen Lamp	(1)	Fig. V	- ON
Operating Range Control	(13)	Fig. III	- 0 - 100
Sensitivity Control	(8)	Fig. III	- 11
Time Constant Selector	(10)	Fig. III	- .6
Wavelength Motor Control	(2)	Fig. II	- Out
Scale Expansion Selector	(11)	Fig. II	- Neutral
Scanning Time Selector	(3)	Fig. II	- 5
Wavelength Selector	(1)	Fig. II	- 220
Operation Selector	(9)	Fig. III	- Absorbance

- 6.7.1 Allow an hour for full stabilizing warm-up time.

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL	Procedure No. L-101	Policy No. 5
	Chemical Laboratories	Function No. 2.1.2	Issue 1
	Technical and Quality Assurance Department	Page <u>7</u> of <u>14</u>	Date 11/14/74

6.8 Preparation of Standard Curves

- 6.8.1 Using standard nitrate and nitrite solutions as prepared in paragraphs 6.3.1 and 6.3.2 scan at 220 mu as follows:
- 6.8.2 When other interfering materials are known to be present, prepare standards according to Paragraph 6.3.3, measure the absorbance at 220 mu and plot a separate curve for each.
- 6.8.3 Make certain both sample and reference knobs are in off position anytime the sample compartment is opened.
- 6.8.4 Fill both cells with the blank or reference solution; rinse cell three times.
- 6.8.5 Dry the cells with lens paper.
- 6.8.6 Turn the wavelength to 220 mu.
- 6.8.7 Open the reference control; set the pen at zero on the chart with the zero knob.
- 6.8.8 Then open the sample control and set the pen at 100% on the chart with the 100% knob.
- 6.8.9 Repeat the zero and 100% setting operation.
- 6.8.10 Close both the sample and reference controls.
- 6.8.11 Fill the front cell with the standard material rinsing three times. Dry cell with lens paper.
- 6.8.12 Place the cell in position (front) and close the sample compartment.
- 6.8.13 Turn the wavelength control to 245 mu.
- 6.8.14 Open both the sample and the reference controls at the same time (prevents bumping of the pen at either extreme).
- 6.8.15 Turn the Scale Expansion to 1X; Wavelength Motor Control to In.
- 6.8.16 Set the pen on the chart.
- 6.8.17 Turn the switch to scan, and scan from 245 mu to 220 mu.
- 6.8.18 Lift the pen at exactly 220 mu.

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL Chemical Laboratories <hr/> Technical and Quality Assurance Department	Procedure No. L-101	Policy No. 5
		Function No. 2.1.2	Issue 1
		Page <u>8</u> of <u>14</u>	Date 11/14/74

6.8.19 Turn scan switch to stop and lift pen.

6.8.20 Close both the sample and reference at the same time.

6.8.21 Plot the ppm nitrate as N against the absorbance for each standard solution.

6.9 Analysis of Sample

6.9.1 Prepare sample according to Paragraph 6.6.

6.9.2 Measure the absorbance at a wavelength of 220 mu to obtain the nitrate reading and at a wavelength of 275 mu to obtain the interference due to dissolved matter.

6.9.3 Subtract 2 times the absorbance reading at 275 mu from the absorbance reading at 220 mu.


6.9.4 Using the nitrite standard curve, convert the ppm nitrite (which has been determined by another method), if present, to absorbance at 220 mu, and subtract this value from the absorbance of the sample at 220 mu. Using this corrected absorbance, read the nitrate concentration from the nitrate standard curve. Correct for other known interfering materials in the same manner using the appropriate standard curve.

ppm Nitrate as N = net ppm Nitrate as N x dilution factor
ppm Nitrate as NO₃ = ppm Nitrate as N x 4.43

HERCULES POWDER COMPANY Radford Army Ammunition Plant Radford, Virginia	PROCEDURE MANUAL	Procedure No. L-101	Policy No. 5
	Chemical Laboratories	Function No. 2.1.2	Issue 1
	Technical and Quality Assurance Department	Page 9 of 14	Date 11/14/74

Table I

<u>Outfall</u>	<u>Dilution, ml of Sample</u>	<u>Dilution Factor</u>
004	25 - 50	2 - 1
005	10	5
006	25 - 50	2 - 1
007	1	50
009	25 - 50	2 - 1
012	1	50
013	1 - 5	50 - 10
018	10 - NO ₃ , 50 - NO ₂	5 - 1
019	10	5
020	1 - 5	50 - 10
026	10	5
028	25	2
Incinerator	25	2




NOTE
 To prevent noise and ensure optimum performance, always turn Photomultiplier Gain Switch to OFF position when using red-sensitive phototube.
 Beckman Instruments, Inc.
 Fullerton, California
 September 1964 14TW602

Figure 1. Operating Controls

- | | |
|--|---|
| <ol style="list-style-type: none"> 1. SOURCE SELECTOR
Selects either tungsten or hydrogen lamp source. 2. TUNGSTEN LAMP SWITCH
Turns tungsten lamp on and off. 3. PHOTOTUBE SELECTOR
Moves either blue-sensitive or photomultiplier or red-sensitive phototube into beam. 4. LOAD RESISTOR SWITCH
Selects proper load resistor for phototube. 5. SHUTTER SWITCH
Blocks beam from striking phototube. 6. SAMPLE POSITIONER
Positions sample and reference cells in beam. 7. FILTER POSITIONER
Positions filter or open aperture in beam. 8. SENSITIVITY CONTROL
Changes voltage across transmittance slidewire. 9. SLIT CONTROL
Selects spectral bandwidth. 10. DARK CURRENT CONTROL
Balances current flow in phototube when no energy is falling on cathode. Sets zero reference for transmittance-absorbance scale. | <ol style="list-style-type: none"> 11. TRANSMITTANCE-ABSORBANCE CONTROL
Provides readout of spectral analyses. Controls electrical null-balance system of instrument. 12. FUNCTION SWITCH
Selects either of three ranges full scale and circuit functions. 13. WAVELENGTH SELECTOR
Selects desired wavelength. Scroll range: 185-2000 millimicrons. Normal operating range: 190-1000 millimicrons. 14. POWER SWITCH
Turns power supply on and off. 15. FILAMENT TEMPERATURE SWITCH
Controls power and varies filament temperature on hydrogen or mercury lamps. 16. PHOTOMULTIPLIER GAIN SWITCH
Adjusts sensitivity on photomultiplier tube. 17. ZERO SUPPRESSION SWITCH
Applies bucking voltage to compensate for background energy effects in flame, fluorescence or reflectance measurements. 18. SCREEN BIAS SWITCH
Controls screen voltage on 2532 tube in spectrophotometer amplifier. 19. SOURCE CIRCUIT SWITCH
Selects correct circuit for either hydrogen or mercury lamps. |
|--|---|


1. **WAVELENGTH SELECTOR**
Permits manual selection of wavelength.
2. **WAVELENGTH MOTOR CONTROL**
Engages and disengages automatic wavelength drive.
3. **SCANNING TIME SELECTOR**
Selects different scanning times.
4. **SCAN LIMIT CONTROL**
Automatically stops scan at predetermined wavelength.
5. **WAVELENGTH DRIVE SWITCH**
Turns off and reverses wavelength drive.
6. **SLIT SERVO CONTROL**
Engages and disengages slit servo motor.
7. **MANUAL SLIT CONTROL**
Enables manual selection of slit width.
8. **CHART PAPER CLIP**
Holds right end of chart paper.
9. **PEN LIFT LEVER**
Raises pen from chart.
10. **CHART PAPER RETAINER**
Retains left end of chart paper.
11. **SCALE EXPANSION SELECTOR**
Disengages wavelength drive from pen carriage drive; permits wavelength expansion and contraction of recorded spectral regions.



Top Panel Controls

Figure II

1. **REFERENCE ENERGY METER**
Shows reference channel and slidewire current; indicates operation of slit servo system during transmittance or absorbance scan.
2. **MAIN POWER SWITCH**
Turns on and off all instrument power.
3. **POWER INDICATOR LIGHT**
Glows when instrument is receiving power.
4. **SCAN SWITCH**
Turns on and off the wavelength drive and recording system.
5. **RECORDER POWER SWITCH**
Not shown, behind door; controls power to recorder.
6. **CHART DRIVE SWITCH**
Starts and stops chart drive.
7. **CHART DRUM KNOB**
Permits manual rotation of chart drum.
8. **SENSITIVITY CONTROL**
Determines slit width for transmittance or absorbance operation; determines gain of system for energy operation.
9. **OPERATION SELECTOR**
Determines operating mode: absorbance, transmittance or energy.
10. **TIME CONSTANT SELECTOR**
Controls response time to change in signal level; settings of 0.1, 0.2, 0.6 and 2.0 are approximately equivalent to periods (in seconds) of 1.5, 2.0, 4.0 and 11.0, respectively.
11. **ZERO ADJUST CONTROL**
Permits setting recorder pen to zero.
12. **100% ADJUST CONTROL**
Permits setting recorder pen to 100%.
13. **OPERATING RANGE CONTROL**
Controls ordinate expansion of signal—transmittance or absorbance.



MODEL DK-2A


Front Panel Controls

Figure III

Procedure No. L-101
Policy No. 5
Function No. 2.1.2
Issue 1
Page 13 of 14
Date 11/14/74




1. SOURCE
SELECTOR
Controls mirror for selection of sources.
2. SAMPLE
SHUTTER CONTROL
Blocks sample beam for setting zero base line.
3. REFERENCE
SHUTTER CONTROL
Blocks reference beam for setting rotating mirror in correct position for energy operation.
4. DETECTOR
CIRCUIT SELECTOR
Switches into circuit the lead sulfide cell or photomultiplier tube (two sensitivity settings).
5. DETECTOR
BEAM SELECTOR
Controls mirror to direct beam onto lead sulfide cell or onto photomultiplier tube.
6. TUNGSTEN
LAMP SWITCH
Controls power to tungsten lamp.
7. SLIT HEIGHT MASK
AND
FILTER CONTROL
Used to adjust exit slit height and to interpose filter in exit beam.



Lamphouse and Photometer Controls

Figure IV



1. **HYDROGEN LAMP
POWER SWITCH**
*Turns on and off all power to the unit (WARM
setting of three-position switch is used when
starting lamp).*
2. **HYDROGEN LAMP
INTENSITY CONTROL**
Controls filament temperature of hydrogen lamp.
3. **HYDROGEN LAMP
STARTER BUTTON**
Starts arc in hydrogen lamp.
4. **POWER
INDICATOR LAMP**
Glowes when power is supplied to unit.

Hydrogen Lamp Power Supply Controls.
On late-model instruments, the transistorized unit re-
places the vacuum tube power supply.

Figure V

Procedure A-6

Analysis of Sodium Carbonate and Hydrogen
Peroxide in Sodium Nitroform Scrubber Solution

Analysis of Sodium Carbonate and Hydrogen Peroxide
in Sodium Nitroform Scrubber Solution

1. SOLUTIONS

1.1 HCl, 1/3 N

1.2 Hydrogen peroxide 1/10 strength. Dilute one part Albone DS 50 percent hydrogen peroxide with nine parts water (need not be exact).

1.3 Sodium arsenite, 0.25 N. Dissolve 16.24 g NaAsO₂ in water, add 20 ml 50 percent NaOH and dilute to one liter. To standardize, pipet exactly one ml of five percent hydrogen peroxide and titrate with NaAsO₂ solution as described in paragraph 3. Then pipet exactly one ml of the same hydrogen peroxide solution into a flask containing 50 ml water, 20 ml 1:1 H₂SO₄, 5 g KI, and three drops ammonium molybdate solution [2.5 g (NH₄)₆Mo₇O₂₄ · 4 H₂O in 25 ml water, filtered if cloudy]. Stopper, swirl, let stand for five minutes and titrate with 0.2 N Na₂S₂O₃ to starch endpoint.

$$N \text{ NaAsO}_2 = \frac{\text{ml Na}_2\text{S}_2\text{O}_3 \times N \text{ Na}_2\text{S}_2\text{O}_3}{\text{ml NaAsO}_2 \text{ to titrate one ml H}_2\text{O}_2}$$

1.4 10 percent NaOH. Dilute 200 ml of 50 percent NaOH to one liter.

2. EQUIPMENT

2.1 Precision Scientific Company Auto-Aquatator

3. DETERMINATION OF Na₂CO₃

3.1 Pipet two ml (or any suitable volume) of the sample into an aquatrator vessel which is about half full of water. Add two ml of five percent H₂O₂. Make the following instrument settings:

With electrodes dry, set knob to "Direct Adjust" and adjust black needle to 20 mA. (Needs to be checked only occasionally.) Turn knob to "Direct Titrate." Set end point hold for 30 seconds. Set Burette Selector to Left Burette. Fill left burette with standard HCl.

3.2 Put sample under burettes, start stirrer, and be sure electrode wires are covered. Set red needles at 0 and 20 for the beginning of the titration. As the black needle moves down scale, reset the right-hand red needle to about two mA higher than the lowest point reached. For example, if the mA readout is 20 → 1 - 4, set the right-hand red needle at three. The actual settings will be determined by experience.

when the "START" light illuminates, read the burette and calculate Na_2CO_3 percent.

$$\% \text{Na}_2\text{CO}_3 = \frac{\text{ml HCl} \times \text{N HCl} \times 5.3}{\text{ml of sample}}$$

4. DETERMINATION OF H_2O_2

- 4.1 Pipet two ml (or any suitable volume) of the sample into an Aquatrator vessel which is about half full of water. Add 10 ml 10% NaOH. Make the following instrument settings:

With electrodes dry, set knob to "Back Adjust" and adjust black needle to 20 μA . (Needs to be checked only occasionally.) Turn knob to "Back Titrate." Set Burette Selector to right Burette. Fill right burette with NaAsO_2 solution.

- 4.2 Put sample under burettes, start stirrer, be sure electrode wires are covered. Set red needles as necessary. (Try three and 14 to begin with.) The end point is reached when the black needle gives a low reading (probably below five μA) and has no tendency to revert to a higher reading between drops of titrant.
- 4.3 On back-titration the instrument adds titrant more slowly (in the region between the red needles) than it does in direct titration. The end point hold does not work in the back-titration mode; when the black needle moves to the left of the left red needle, the titration stops. The H_2O_2 percentage can be calculated as shown below:

$$\% \text{H}_2\text{O}_2 = \frac{\text{ml NaAsO}_2 \times \text{N NaAsO}_2 \times 1.70}{\text{ml of sample}}$$

Procedure A-7

Analysis of Tetranitromethane Scrubbing Solution for
Sodium Carbonate and Sodium Bicarbonate Concentration

Analysis of Tetranitromethane Scrubbing Solution for
Sodium Carbonate and Sodium Bicarbonate Concentration

1. SCOPE AND APPLICATION

- 1.1 This method is suitable for the determination of sodium carbonate (Na_2CO_3) and sodium bicarbonate (NaHCO_3) concentrations in tetranitromethane (TNM) scrubber solutions.

2. SUMMARY OF METHOD

- 2.1 A sample of scrubber solution is diluted with distilled water and titrated using 0.3 N hydrochloric acid. The acid is added in small increments and after each add the pH is recorded. A plot of ml of acid versus pH is prepared and from the Δ at the two points of inflection the concentrations of Na_2CO_3 and NaHCO_3 can be calculated.

3. REAGENTS

- 3.1 Distilled water
3.2 Standardized 0.3 N HCl

4. APPARATUS

- 4.1 pH meter
4.2 Magnetic stirrer
4.3 5 ml buret graduated in 0.02 ml

5. PROCEDURE

- 5.1 Pipet three ml of scrubber solution into 100 ml distilled water in a 250 ml beaker.
5.2 Place beaker on stirrer and position pH electrodes in sample.
5.3 Add 0.3 N HCl in 0.2 ml increments and record pH after each addition. As the end point is approached, allow more time between additions for the pH to stabilize.
5.4 When a pH of 3.5 is obtained, stop the titration.
5.5 On a standard graph paper, plot ml of HCl versus pH and draw a titration curve,
5.6 Determine the ml of HCl at the two inflection points.

6. CALCULATIONS

6.1 Calculation of Na_2CO_3

$$\frac{V_1 \times N \times 10.6}{3} = \text{Percent Na}_2\text{CO}_3$$

6.2 Calculation of NaHCO_3

$$\frac{(V_2 - 2V_1) \times N \times 8.4}{3} = \text{Percent NaHCO}_3$$

V_1 = Vol of HCl at first point of inflection (highest pH)

V_2 = Total vol of HCl at second point of inflection (lowest pH)

N = Normality of HCl solution

Procedure A-8

Determination of Nitrites

Determination of Nitrites

1. SCOPE AND APPLICATION

- 1.1 This method is applicable to the determination of nitrite in drinking, surface, and saline waters, domestic and industrial wastes.
- 1.2 The method is applicable in the range from 0.01 to 1.0 mg NO₂ as N/l.

2. SUMMARY OF METHOD

- 2.1 The diazonium compound formed by diazotation of sulfanilamide by nitrite in water under acid conditions is coupled with N-(1-naphthyl)-ethylenediamine to produce a reddish-purple color which is measured in a spectrophotometer at 540 nm.

3. SAMPLE HANDLING AND PRESERVATION

- 3.1 Samples should be analyzed as quickly as possible. They may be stored for 24 to 48 hours at 4°C.

4. INTERFERENCES

- 4.1 There are very few known interferences at concentrations less than 1,000 times that of the nitrite; however, the presence of strong oxidants or reductants to the samples will readily affect the nitrite concentrations. High alkalinity (>600 mg/l) will give low results due to a shift in pH.

5. APPARATUS

- 5.1 Spectrophotometer equipped with 1 cm or longer cells for use at 540 nm.
- 5.2 Nessler tubes, 50 ml or volumetric flasks, 50 ml.

6. REAGENTS

- 6.1 Distilled water free of nitrite and nitrate is to be used in preparation of all reagents and standards.
- 6.2 Buffer-color reagent: To 250 ml of distilled water, add 105 ml conc. hydrochloric acid, 5.0 g sulfanilamide and 0.5 g N-(1-naphthyl) ethylenediamine dihydrochloride. Stir until dissolved. Add 136 g of sodium acetate (CH₃COONa·3H₂O) and again stir until dissolved. Dilute to 500 ml with distilled water. This solution is stable for several weeks if stored in the dark.

- 6.3 Nitrite stock solution: 1.0 ml = 0.10 mg NO₂-N. Dissolve 0.4926 g of dried anhydrous sodium nitrite (24 hours in desiccator) in distilled water and dilute to 1000 ml. Preserve with 2 ml chloroform per liter.
- 6.4 Nitrite standard solution: 1.0 ml = 0.001 mg NO₂-N. Dilute 10.0 ml of the stock solution (6.3) to 1000 ml.

7. PROCEDURE

- 7.1 If the sample has a pH greater than 10 or a total alkalinity in excess of 600 mg/l, adjust to approximately pH 6 with 1:3 HCl.
- 7.2 If necessary, filter the sample through a 0.45 μ pore size filter using the first portion of filtrate to rinse the filter flask.
- 7.3 Place 50 ml of sample, or an aliquot diluted to 50 ml, in 50 ml Nessler tube; hold until preparation of standards is completed.
- 7.4 At the same time prepare a series of standards in 50 ml Nessler tubes as follows:

<u>ml of Standard Solution</u> 1.0 ml = 0.001 mg NO ₂ -N	<u>Conc., When Diluted to</u> <u>50 ml, mg/l of NO₂-N</u>
0.0 (Blank)	0.0
0.5	0.01
1.0	0.02
1.5	0.03
2.0	0.04
3.0	0.06
4.0	0.08
5.0	0.10
10.0	0.20

- 7.5 Add 2 ml of buffer-color reagent (6.2) to each standard and sample, mix and allow color to develop for at least 15 minutes. The color reaction medium should be between pH 1.5 and 2.0.
- 7.6 Read the color in the spectrophotometer at 540 nm against the blank and plot concentration of NO₂-N against absorbance.

8. CALCULATION


- 8.1 Read the concentration of NO₂-N directly from the curve.
- 8.2 If less than 50.0 ml of sample is taken, calculate mg/l as follows:

$$\text{NO}_2\text{-N, mg/l} = \frac{\text{mg/l from standard curve} \times 50}{\text{ml sample used}}$$

9. PRECISION AND ACCURACY

9.1 Precision and accuracy data are not available at this time.

Nitrate Test
Spectrmic 20, Wave Length - 540 nm



DISTRIBUTION LIST

Commander
US Army Armament Research
and Development Command
ATTN: DRDAR-LCE (15 cys)
DRDAR-TSS (5 cys)
DRDAR-LCM-E, R. Wolff
DRDAR-LCM-S, I. Forsten
Dover, NJ 07801

Commander
Naval Ammunition Depot
ATTN: R. Klausmeir
Crane, IN 47522

Commander
US Army Armament Materiel
Readiness Command
ATTN: DRSAR-RDM, G. Cowan
DRSAR-LEP-L
Rock Island, IL 61299

Commander
Radford Army Ammunition Plant
ATTN: W.T. Bolleter (15 cys)
Radford, VA 24141

Commander
Naval Surface Weapons Center
ATTN: T.N. Hall (4 cys)
White Oak Laboratory
Silver Spring, MD 20910

Commander
Naval Weapons Center
ATTN: A.T. Nielsen (2 cys)
China Lake, CA 93555

Commander
Naval Sea Systems Command
ATTN: A. Amster, SEA-0332
Washington, DC 20300

Commander
Chemical Systems Laboratory
ATTN: A. Hilsmeier (2 cys)
L. Schiff
Aberdeen Proving Ground, MD 21010

Commander
Army Medical Bioengineering
R&D Laboratory
ATTN: D. Rosenblatt
Fort Detrick, MD 21701

Commander (3 cys)
Volunteer Army Ammunition Plant
Chattanooga, TN 37401

Commander (3 cys)
Joliet Army Ammunition Plant
Joliet, IL 60436

Commander
US Army Armament Materiel Readiness Command
ATTN: DRCPM-PBM
Dover, NJ 07801

Commander
US Army Research Office
ATTN: G. Wyman (2 cys)
P.O. Box 12211
Research Triangle Park, NC 27709

John Brown Associates
ATTN: J. Brown
P.O. Box 145
Berkeley Heights, NJ 07922

Defense Documentation Center (12 cys)
Cameron Station
Alexandria, VA 22314

Director
US Army TRADOC Systems Analysis Activity
ATTN: ATAA-SL (Tech Library)
White Sands Missile Range, NM 88002

Weapon System Concept Team/CSL
ATTN: DRDAR-ACW
Aberdeen Proving Ground, MD 21010

Technical Library
ATTN: DRDAR-CLJ-L
Aberdeen Proving Ground, MD 21005

Technical Library
ATTN: DRDAR-TSB-S
Aberdeen Proving Ground, MD 21010

Technical Library
ATTN: DRDAR-LCB-TL
Benet Weapons Laboratory
Watervliet, NY 12189