

AD-A067 866

ARIZONA UNIV TUCSON DEPT OF CHEMISTRY
CHARACTERISTICS OF A 'HIGH SOLIDS' NEUBLIZER FOR FLAME ATOMIC A--ETC(U)
APR 79 R C FRY, M B DENTON
TR-21

F/G 7/4

N00014-75-C-0513

NL

UNCLASSIFIED

1 OF 1
AD A
067866



END
DATE
FILMED
10-80
DTIC

LEVEL

12
B.S.


OFFICE OF NAVAL RESEARCH
Contract N00014-75-C-0513
Task No. NR 051-549
Technical Report No. 21

AD A067866

CHARACTERISTICS OF A "HIGH SOLIDS" NEBULIZER
FOR FLAME ATOMIC ABSORPTION SPECTROMETRY

R. C. Fry and M. B. Denton

Department of Chemistry
University of Arizona
Tucson, Arizona 85721



DDC FILE COPY

Prepared for Publication

in

Applied Spectroscopy

Reproduction in whole or in part is permitted for any purpose
of the United States Government.

Approved for Public Release: Distribution Unlimited

79 04 20 017

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER 17R-21	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) CHARACTERISTICS OF A "HIGH SOLIDS" NEBULIZER FOR FLAME ATOMIC ABSORPTION SPECTROMETRY.		5. TYPE OF REPORT & PERIOD COVERED INTERIM / repty
7. AUTHOR(s) R. C. Fry M. B. Denton		6. PERFORMING ORG. REPORT NUMBER
		8. CONTRACT OR GRANT NUMBER(s) N00014-75-C-0513
9. PERFORMING ORGANIZATION NAME AND ADDRESS Department of Chemistry University of Arizona Tucson, AZ 85721		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS NR 051-549
11. CONTROLLING OFFICE NAME AND ADDRESS Office of Naval Research Arlington, Virginia 22217		12. REPORT DATE Apr 1979
		13. NUMBER OF PAGES 14
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office) 1240p.		15. SECURITY CLASS. (of this report) UNCLASSIFIED
15a. DECLASSIFICATION/DOWNGRADING SCHEDULE		
16. DISTRIBUTION STATEMENT (of this Report) Approved for Public Release; Distribution Unlimited		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Nebulizer; Direct Determination; High Solids Analysis; Atomic Absorption		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) Studies are presented to characterize a new nebulizer developed for direct atomic absorption analysis of extremely complex clinical and environmental materials. Important parameters are described concerning the design and performance of 'high solids' spectrochemical nebulizers based on the Babington principle. A new simplified design is presented and mechanisms of undesirable sample wastage are considered. Data are presented concerning the effect of impaction on conventional and 'high solids' aerosol generation.		

CREDIT

This work was supported in part by the Office of Naval Research and in part by an A. P. Sloan fellowship to the corresponding author.

This work was presented in part at the 20th Annual Rocky Mountain Conference on Analytical Chemistry, August 7, 1978, Denver, Colorado.

CHARACTERISTICS OF A "HIGH SOLIDS" NEBULIZER
FOR FLAME ATOMIC ABSORPTION SPECTROMETRY

R. C. Fry* and M. B. Denton†

Department of Chemistry
University of Arizona
Tucson, Arizona 85721

ACCESSION for	
NTIS	White Section ☒
DOC	Buff Section ☐
UNANNOUNCED	☐
SPECIFICATION	
BY	
DISTRIBUTION/AVAILABILITY CODES	
Dist.	SPECIAL
A	

ABSTRACT

Studies are presented to characterize a new nebulizer developed for direct atomic absorption analysis of extremely complex clinical and environmental materials. Important parameters are described concerning the design and performance of "high solids" spectrochemical nebulizers based on the Babington principle. A new simplified design is presented and mechanisms of undesirable sample wastage are considered. Data are presented concerning the effect of impaction on conventional and "high solids" aerosol generation.

INDEX HEADINGS: Nebulizer; Direct determination; High solids analysis:
Atomic absorption.

*Present address: Department of Chemistry, Kansas State University
Manhattan, Kansas 66506.

†Author to whom correspondence should be sent.

INTRODUCTION

"Clog free" production of high density, finely dispersed aerosols from solutions and slurries of elevated salt concentration, high solids content, and increased viscosity has recently been described for spectrochemical sample introduction by Fry and Denton (1). The Babington principle of aerosol generation (2,3) was adapted for direct atomic absorption analysis (using high solids burner heads) of samples as complex as sea water, urine, whole blood, evaporated milk concentrate, and tomato sauce (1). The superior tolerance of this approach to "high solids" slurries is a direct result of eliminating the conventional requirement that the sample must pass through a small bore capillary tube. Ability of the material to pass through the relatively unrestricted sample delivery pumping system was found to be the primary limitation on the complexity of materials that can be directly analyzed (1) by the new procedure. The improved sample handling capability is achieved without significant sacrifice of precision or concentration sensitivity when compared with conventional capillary pneumatic approaches (1). This relatively "preparation free" sample introduction technique for analytical atomic spectroscopy can presently be considered to be in the beginning stages of development. Once design and operational characteristics are well established, this convenient approach to high solids atomization holds great promise for application in agricultural, clinical, environmental,

and industrial analytical situations. The purpose of this paper is to outline design considerations and characterize the parameters that govern the performance of this new high solids sample introduction approach at its present state of development. Data leading to the simplification of design are presented, and mechanisms of undesirable sample wastage are explored. Data are presented concerning the effect of impaction on conventional and high solids aerosol generation.

I. EXPERIMENTAL

A. Apparatus. A previously described (1) atomic absorption spectrometer, 10 cm. high solids single slot air-acetylene burner head, 10 cm triple slot Boling burner head, Varian-Techtron (Springvale, Vic., Australia) and Jarrell Ash (Waltham, Mass.) capillary pneumatic nebulizers, Beckman (Fullerton, Calif.) total consumption burner operated as premix nebulizer, "HYDROSPHERE" (Babington principle) model respiratory nebulizer system (Owens Illinois - Toledo, Ohio), and specially developed Babington principle nebulizer chamber (cylindrical) with variable flow peristaltic pump sample delivery system were used in the course of these investigations. Quick interchange connections to the premix chamber were provided for the nebulizer mountings. Provision was made in the cylindrical nebulizer chamber for accommodation of a variety of configurations for aerosol generation involving the Babington principle. Fig. 1 diagrams the "HYDROSPHERE" system with a slight modification; other configurations constructed and evaluated in this study are illustrated in Figs. 2 - 6. All configurations shown represent adaptations of the Babington principle which involves flowing a thin

film of sample over the surface of a pressurized sphere or hemisphere insert. Interruption of the sample film at the point of an orifice by the high velocity exit of oxidant gas serves to generate the aerosol. Since the sample is not required to pass through this orifice, clogging does not occur even for the most complex materials (See Fig. 7) when the systems diagrammed in Figs. 2, 5, and 6 are used. Small ($\leq 5\mu\text{m}$) aerosol droplets (4) are carried by the gas stream to the premix chamber of the burner. Larger droplets and sample runoff from that portion of the sample film not flowing across the orifice either exit via the drain (Figs. 2, 5, 6, and 7) or may be collected for recirculation (Fig. 1, Fig. 3, and Fig 4). Recirculation was evaluated in several configurations for the purpose of conserving solution. The glass nebulizer spheres indicated in Figs. 1 - 4 were adapted from the Hydrosphere system formerly available from Owens Illinois (Toledo, Ohio) and presently available from McGaw Respiratory Therapy (Irvine, Calif.). The compression fitting designs of Figs. 5 and 6 represent a convenient prototype approach allowing rapid interchange of a variety of improved metal nebulizer inserts. The 6 mm insert was constructed by rounding and polishing one end of a 6mm o.d. rod using an engine lathe. Lathe polishing was done with emory cloth and fine steel wool to achieve an average surface roughness of 8 μin or better as judged by General Electric Roughness Standards, and as measured by profilometry. The rod was then cut to a length of 19 mm and drilled from the "cut" end at a diameter of 5.5 mm to a depth of 18.5 mm using care to avoid drilling through the rounded end. A conventional drill bit was used. The desired orifice was machined in the side of the rod 3 mm below the rounded end. This was done by holding the insert horizontally with a "vee" clamp insert in a machine vise. The

0.61 mm hole was machined using a RALMIKE (So. Plainfield, N.J.) sensitive hand drill feed fitted into a Bridgeport milling machine according to standard procedure. The 3 mm inserts were constructed in similar fashion utilizing a 6 mm base and 2.5 mm i.d. dimension with the orifice machined 1.5 mm below the rounded end. The inserts are seated conveniently in the modified compression or Swagelock (Solon, Ohio) fitting holder. The above nebulizer inserts have been made out of brass, and then electroplated with nickel to improve corrosion resistance. Stainless steel inserts have also been constructed. Non-wetted materials such as Teflon are not recommended for the nebulizer insert since aqueous samples tend to bead up on such surfaces rather than form a suitable film. Following initial cleaning, the flowing sample and solvent stream keeps the nebulizer insert sufficiently clean to maintain a thin film flow of material indefinitely (the nebulizer insert in this laboratory has not required cleaning in over a year for all high solids samples analyzed (1)). The spray chamber walls should, however, be periodically rinsed (with disinfectant if desired) following analysis of large numbers of biological samples such as blood to prevent buildup and spoilage. The dimensions given above are generally not critical. Tolerances of $\pm 10\%$ on the dimensions given should generally be sufficient unless specific exception is noted in the above text.

B. Procedure. The atomic absorption of cadmium at 228.8 nm was utilized to study the effect of parameters governing nebulizer performance. The Cd hollow cathode was operated at 5 mA, the spectral bandpass was 0.6 nm, and the overall instrumental response time was ~ 0.6 sec. Cadmium was chosen for its simple spectrum, the high volatility of its salts and its freedom from oxide formation and ionization

effects. These properties make the atomic absorption of this element in a 10 cm premixed air-acetylene flame relatively independent of instrumental parameters such as redox character and viewing position in the flame. Nebulizer performance under a wide range of conditions was thereby more readily and independently assessed. The nominal viewing position was 6 mm (beam center) above the burner head. Whenever oxidant (nebulizer) gas flow rates were varied, fuel adjustment was made to achieve a constant flame stoichiometry. Studies were also performed to ensure that the vertical absorbance profile had not been significantly altered. Oxidant gas flow rate measurements were corrected for pressure drop across the nebulizer orifice. Volumetric determination of aspiration rates and pump delivery rates were made when appropriate.

The sample delivery tube and corresponding inlet to the cylindrical chamber (Figs. 5, 6, and 7) were threaded to facilitate external adjustment of the sample delivery gap (distance between the end of the delivery tube and the top of the pressurized nebulizer hemisphere).

II. RESULTS AND DISCUSSION

A. Operating Range. A series of Babington nebulizer inserts including 0.13 mm, 0.20 mm, 0.34 mm, 0.61 mm, and 0.71 mm single hole orifices were constructed and the resultant operating range of oxidant flow rate evaluated. The 0.13 mm, 0.20 mm, and 0.34 mm orifices produced extended low flow rate operation (down to ≤ 1 L/min). The wide dynamic operating range of these orifices also extended to the normal operating range of slot burners (≥ 5 L/min), but pressures in excess of 100 psi. were required to drive the nebulizer in this higher range. The 0.71 mm insert orifice was found to operate at much higher flow rates (> 10 L/min) with reduced

pressure requirements, however the sensitivity was degraded and the operating range did not extend sufficiently low enough. The 0.61 mm orifice optimally covered the normal (8-15 L/min) slot burner range (Fig. 8) using pressures of 14-40 psi, so the other orifices (0.13 mm, 0.20 mm, 0.34 mm, and 0.71 mm) were not considered further in these atomic absorption studies. The smaller ones may prove useful, however, in low flow rate applications such as d.c. and high frequency electrical plasmas for emission spectroscopy.

Although Fig. 8 bears out the previously reported (1) sensitivity advantage of the Varian-Techtron nebulizer (when analyzing "clean" aqueous solutions), the metal version of high solids nebulizer diagrammed in Figs. 5 and 6 are seen in the present studies to have a superior operating range of oxidant flow rate (Fig. 8). The effects of Fig. 8 are primarily related to nebulizer performance. Aspiration rate measurements indicate that the Varian-Techtron nebulizer does not convert sample to aerosol form (aspiration ceases) below 4 L/min or above 20 L/min. However, the metal high solids nebulizer performance extended to flow rates below 3 L/min. The upper limit for both configurations (Fig. 4 and Fig. 5) of nebulizer based on the Babington principle was the high pressures required to exceed 40 L/min for the 0.61 mm orifice diameter utilized. This was also the reason for not pursuing the Beckman $O_2 - H_2$ total consumption burner (as a premix nebulizer) higher than 9 L/min. The pressure required at 35 L/min was 120 psi. for the 0.61 mm Babington orifice. Fig. 8 indicates that no sensitivity improvement is realized above 15 L/min for the high solids nebulizer due to the attainment of constant aerosol density at higher flow rates. Owens - Illinois indicates

that this constant aerosol density is maintained well beyond 100 L/min (5); however, the range above 30 L/min is generally not of great interest in spectrochemical applications. The higher flow aerosol range of the high solids nebulizer is, however, useful between 15 and 25 L/min for several slot burners such as the triple slot Boling head and Instrumentation Laboratories CAT No. 24036-03 high solids single slot nitrous oxide - acetylene burner. The extra large dimensions of the slot openings of these two burner heads make them somewhat dangerous (from a flashback standpoint) when operated in the optimum flow range (10 L/min) of the conventional Varian-Techtron nebulizer unless auxiliary oxidant is provided. An extra margin of safety is allowed without use of auxiliary oxidant for these specialized burner heads by operating the new high solids nebulizer in the higher flow aerosol region between 15 and 25 L/min where the performance is generally improved as compared to the Varian-Techtron nebulizer (Fig. 8). A safer high solids air- acetylene head (1) was, however, used in the present investigations. The nebulizers were therefore operated at 12 L/min for all subsequent studies not concerned with operational range.

At flow rates below 2 L/min, the pressure in the 0.61 mm orifice high solids nebulizers evaluated in these studies is no longer sufficient to rupture the sample film at high enough velocity to generate significant aerosol. The earlier mentioned smaller orifices will further extend the flow rate performance for applications in plasma emission spectrometers however, this was not deemed necessary for the present slot burner studies. A 0.61 mm single hole orifice was therefore used for all subsequent flame

studies. The large dynamic operating range of the Babington approach represents an advantage in terms of applicability to a wider variety of spectrochemical situations and reduced susceptibility to gas flow rate errors.

B. Nebulizer Efficiency. Fig. 9 illustrates the effect of the sample uptake rate on the steady state (unlimited sample available) concentration sensitivity of spectrochemical analysis for several nebulizer configurations. Despite the superior operating range, superior tolerance of complex materials, and adequate sensitivity realized with the new high solids approach, Fig. 9 illustrates the principal drawback (low efficiency in terms of sample consumption) of this nebulizer in comparison to the capillary pneumatic system. Although good concentration sensitivity is achieved utilizing the metal adaptation of the Babington principle, Fig. 9 shows that 20-50 mL/min of sample must be delivered to realize that sensitivity at the present state of development of the high solids nebulizer approach. Capillary (premix) nebulizers are normally operated at 5 mL/min. Fig. 9 also illustrates, however, that the non-recirculated 6 mm metal version (Figs. 5 and 6) of high solids nebulizer represents a dramatic efficiency improvement over the 16 mm glass respiratory nebulizers including even the most efficiently recirculated version (that of Fig. 4). Further reduction of the metal hemisphere diameter to 3 mm, however, did not further improve the efficiency or operating range.

Fig. 9 also illustrates a hydrodynamic effect of the metal high solids nebulizer. At solution uptake rates below 40 mL/min, a sample delivery gap of 5 mm leads to discrete "dropping" of the sample onto the nebulizer insert. Once a discrete drop is formed, its surface tension appears too great to be overcome as it strikes the top of the pressurized nebulizer hemisphere. Film formation is thereby inhibited and aerosol formation is subsequently depressed. Narrowing the sample delivery gap prevents "dropping" by favoring unbroken "streaming" of the sample onto the nebulizer hemisphere surface. Film formation is more highly favored in this case.

Fig. 10 further demonstrates that "dropping" phenomena tend to be encountered at either excessively low sample delivery rates or large sample delivery gaps. The most profitable way to avoid this problem is to maintain a sample delivery gap ≤ 2 mm which will still allow ready passage of all high solids samples discussed previously (1).

c. Minimum Sample Size Requirements. Fig. 9 is concerned with the spectrochemical response to unlimited sample as a function of uptake rate. Fig. 11, however, illustrates the effect of pipetting limited discrete quantities of solution into the sampling cone of the high solids nebulizer configuration of Figs. 5 and 6 as a function of solution uptake rate. Fig. 11 demonstrates that, although 40-50 mL/min is the optimum uptake rate when ≥ 2 mL sample are available, slower uptake rates maximize the transient spectrochemical response to smaller samples. In fact a 25-30% sacrifice of concentration sensitivity may be highly desirable with the

new high solids approach in order to operate at 20-25 mL/min. Fig. 11 indicates that a 500 μ L sample will (at this uptake rate) give a corresponding transient response (0.19 absorbance units) similar to the concentration sensitivity advertised for a Perkin Elmer nebulizer (see Fig. 8).

It should also be noted that the popular steady-state type of absorbance response may also be readily produced using 1 mL of sample and a lesser (12 mL/min) uptake rate to give a 5 sec (reduced sensitivity) steady-state reading of 0.21 absorbance units (see Fig. 11); 2 mL will give a 10 sec reading of 0.2 A, etc. These reduced steady-state concentration sensitivities, however, still compare favorably with the range (0.18-0.24 A) listed for capillary pneumatic nebulizers by Jarrell Ash and Perkin Elmer Corporation except that about 5 mL/min and 0.4 mL of sample is used with the capillary approach to give a 5 sec steady-state reading. Under these favorable conditions, the required steady-state sample size for the new high solids nebulizer is only 2.5 times more than the required amount for steady-state sampling with capillary pneumatic nebulizers.

Transient sampling, however, represents a somewhat different situation. In this case, Fig. 12 illustrates that, when all pertinent parameters are optimized, the capillary pneumatic nebulizer approaches its maximum concentration sensitivity with samples as small as 0.1 mL compared to 1 mL required for the metal adaptation of the Babington approach (non-recirculated) and 7 mL for the most efficiently recirculated respiratory glass version.

Fig. 13 represents an operational curve used in the laboratory for selection of uptake rates for the metal high solids nebulizer once a sample size is chosen *for transient sampling*. If a 1 mL sample size is chosen, then 40 mL/min will produce an optimum transient of 0.3 A/ppm Cd. The same 1 mL sample may alternatively be used at 12 mL/min (Fig. 11) to produce a 5 sec steady state at $\sim 30\%$ less sensitivity.

Table I summarizes the sample size requirements of the spectrochemical adaptations of the Babington principle developed in these studies for the production of aerosols from "high solids" materials. The table demonstrates that a 500-fold overall improvement in sample size requirement has been achieved in these studies in comparison with the original respiratory nebulizer approach.

D. Mechanisms of Sample Wastage. The difference in sample size requirements (Table I) for the configurations of Fig. 3 and 4 suggest that recirculation is only significantly beneficial if it is carried out in such a way as to collect both the larger spray droplets (that settle out in the chamber) and the nebulizer sphere runoff. This suggests that, although sphere or hemisphere runoff (back and sides) plays a role, a major mechanism of sample waste is through the large spray droplets produced in addition to small ($< 5 \mu\text{m}$) aerosol droplets. These larger spray droplets settle out and are either recirculated (Figs. 1 and 4) or exit via the drain (Figs. 2, 3, 5, and 6). In the interest of simplifying the design, reducing memory and sample cross contamination effects, and in consideration of the earlier report (1) of particle size fractionation that may occur in some high solids samples, recirculation has now been

abandoned in favor of the approach shown in Figs. 5, 6, and 7. The real improvement of the smaller metal hemisphere with non-recessed orifice is seen in the 30-fold difference in results (Table I) for the configurations of Figures 2, 5 and 6. Although the 1 mL sample size of the high solids transient or 5 sec steady-state approach at its present stage of development (Figs. 5, 6 and 7) is larger than the 0.1 mL required for transient sampling (0.4 mL for a 5 sec steady state) by the capillary pneumatic nebulizer, a 1 mL sample size requirement is still reasonable for many analytical situations where sample complexity (e.g. Fig. 7) precludes the use of more conventional nebulizers.

E. Impactor Effects. Fig. 14 illustrates the effect of impactor gap on spectrochemical concentration sensitivity. The Varian-Techtron capillary nebulizer is seen to require very critical adjustment (± 0.2 mm) in comparison to the high solids nebulizer. In fact, the Varian-Techtron (adjustable impactor model) capillary pneumatic nebulizer seems to derive its sensitivity advantage (compared to the other commercial capillary nebulizers evaluated) from the precise nature of adjustment provided for the impactor bead gap. Many commercial designs do not include the capability of making such precision adjustments. It may, therefore, not be possible to readily achieve the peak sensitivity shown in Fig. 14 on a routine daily basis for capillary pneumatic nebulizers unless a precision externally adjustable model is available. In contrast, the high solids nebulizer functions almost as well without any impactor at all as it does when the impactor is adjusted with the greatest of care (see Fig. 14).

F. New Design. On the basis of the above data, the latest nebulizer design (Fig. 6) has been simplified. The earlier (1) externally adjustable impactor (based on an internal rotary-translational motion converter) has been eliminated in favor of the simple impactor mount (of Fig. 6) which is much easier to construct and is only adjustable when the chamber is dismantled.

III. CONCLUSIONS

The high solids spectrochemical nebulizer based on the Babington principle represents a system of similar concentration sensitivity, similar precision, superior operating range, simpler design, and far superior sample handling (high solids) capability as compared to the capillary pneumatic approach. High density, finely dispersed aerosols may be produced from solutions and slurries of elevated salt concentration, high solids content, and increased viscosity (1). The parameter to be improved on, however, is the required sample size. Although the present 1 mL transient signal represents a considerable improvement over earlier high solids adaptations of the Babington principle, further improvements are needed to reduce the sample size requirements. Evidence has been presented to indicate that a primary mechanism of sample wastage is production of an increased percentage of larger spray droplets that settle out in the premix chamber. At its present state of development, the high solids approach may be optimally used as an auxiliary nebulizer for analysis of those materials (≥ 1 mL) which are too complex to pass through the capillary orifice of a standard pneumatic nebulizer. The

high solids approach diagrammed in Fig. 5 and implemented in Fig. 7 is proving to be a simple and convenient "clog-free" approach to spectrochemical sample introduction of highly complex clinical and environmental materials when sufficient sample is available. A newer simplified design has been presented (Fig. 6) which is easier to implement without any sacrifice in performance.

LITERATURE CITED

- (1) R. C. Fry and M. B. Denton, Anal Chem. 49, 1413 (1977).
- (2) R. S. Babington, U. S. Patents: 3,421,692; 3,421,699; 3,425,058; 3,425,059; and 3,504,859.
- (3) R. S. Babington, Popular Science, May, 102 (1973).
- (4) M. Litt and D. E. Swift, Am. Rev. Respir. Dis. 105, 308 (1972).
- (5) Owens - Illinois, technical data sheet, Consumer and Technical Products Group, Toledo, Ohio (1974).

Table I. Sample Size Requirements of Several Spectrochemical Adaptations of the Babington Principle.

Configuration	Figure No.	Minimum Sample Size (mL)
3 mm metal hemisphere	(configured as in 6)	1
6 mm metal hemisphere	6	1
6 mm metal hemisphere	5	1
16 mm glass (recirculated) sphere	4	7
16 mm glass (recirculated) sphere	3	22
16 mm glass sphere	2	30
dual 16 mm glass ("HYDROSPHERE") spheres	1	500

Fig. 1. Complete "HYDROSPHERE" Respiratory Nebulizer Configuration (Babington Principle).

a. single post aerosol generator (16 mm glass sphere with recessed orifice -- 0.61 mm hole); b. double post aerosol generator (two 0.43 mm recessed holes); c,d. oxidant gas inlets (rear entry -- not shown); e. modified sample change port.

Fig. 2. Flow-Through Adaptation of "HYDROSPHERE" (Babington Principle Nebulizer for Spectrochemical Analysis.

a. single post aerosol generator mount from "HYDROSPHERE" system; b. 16 mm pressurized glass sphere with recessed orifice (0.61 mm hole); c. 3 mm plastic impaction sphere, fixed position; d. cylindrical Plexiglas housing for nebulizer inserts based on the Babington principle.

Fig. 3. Recirculated Flow-Through Adaptation of "HYDROSPHERE" (Babington Principle) Nebulizer.

Simplified cross-sectional view: a. single post aerosol generator from "HYDROSPHERE" system; b. conical reservoir (nylon) for sample entry and nebulizer runoff recirculation; c. sample entry (fill) channel; d. nebulizer delivery channel (from reservoir); e. two-way valve; f. sample exit (empty) after reading obtained.

Fig. 4. Increased Recirculation Efficiency Flow-Through Adaptation of "HYDROSPHERE" (Babington Principle) Nebulizer.

Simplified cross-sectional view: a. single post aerosol generator from "HYDROSPHERE" system; b. "half cone" (nylon) for runoff collection (drains into reservoir, c); c. conical reservoir (Plexiglas; sealed to container side) for sample entry and runoff collection from chamber walls (condensed large droplets) as well as nebulizer sphere; d. conical Plexiglas ceiling to minimize large droplet hangup. Conical shape allows condensed droplets to flow more readily down to recirculation reservoir (c); e. sample entry (fill) channel; f. nebulizer delivery channel (from reservoir); g. two-way valve; h. sample exit (empty) after reading obtained.

Fig. 5. Improved Flow-Through Adaptation of the Babington Principle for Spectrochemical Analysis (No Recirculation).

a. interchangeable 6 mm o.d. or 3 mm o.d. pressurized closed-end polished metal tube nebulizer insert with non-recessed orifice (0.61 mm hole); b. 3 mm plastic impaction sphere (translational movement possible); c. mounting block with internal rotary-translational motion converter for impactor movement; d. sample delivery tube (2.38 mm i.d.); e. modified compression "elbow" fitting; f. external adjustment for impaction gap; g. set screw to hold compression elbow in place.

Fig. 6. Improved Flow-Through Adaptation with Simplified Impactor and Mounting and Safety "Pop Off" Plug.

a. Nebulizer insert (same as in Fig. 5); b. simplified impactor mount ("on stream" external adjustment no longer necessary); c. newly simplified nebulizer insert mount utilizing Swagelock "Tee" fitting; d. solid rod (structural support); and e. newly added "pop off" safety plug in case of burner flashback.

Fig. 7. Finely Dispersed Tomato Sauce Aerosol Emanating from 10 cm Slot Burner.

Direct aerosol production (using system diagrammed in Fig. 5) is seen to be easily accomplished without system clogging. Burner is shown without flame burning to demonstrate high density finely dispersed aerosol. The same result is achieved with the system diagrammed in Fig. 6.

Fig. 8. Nebulizer Operating Range.

- | | | | |
|---|---|---|-------------|
| ○ | Varian-Techtron capillary nebulizer; | □ | High solids |
| | nebulizer, 6 mm metal (from Fig. 5); | ▲ | High solids |
| | nebulizer, 16 mm glass respiratory (from Fig. 4); | | |
| △ | Beckman total consumption burner as premix | | |
| | nebulizer. | | |

Fig. 9. Nebulizer Efficiency Comparison.

- Varian-Techtron capillary nebulizer; △ High solids nebulizer, 6 mm metal, non-recirculated, 5 mm gap;
- High solids nebulizer, 6 mm metal, non recirculated, 0.6 mm gap; ▴ High solids nebulizer, 16 mm glass respiratory, recirculation of all excess (from Fig. 4).

Fig. 10. Sample Delivery Gap Effects.

Sample Delivery Rate: △ 53 mL/min; ○ 21 mL/min; □ 10 mL/min;
▴ 4.5 mL/min.

Fig. 11. Sample Size and Uptake Rate Requirements.

Sample Size: -- unlimited; ○ 2 mL; ▽ 1 mL; △ 0.5 mL; □ 0.2 mL;
▴ 0.1 mL.

Fig. 12. Comparison of Minimum Required Sample Volume.


- Varian Techtron capillary nebulizer;
- △ Jarrell Ash (Hetco) capillary nebulizer;
- High solids nebulizer, 6 mm metal, non-recirculated (Figs. 5 and 6);
- ▴ High solids nebulizer, 16 mm glass respiratory, recirculation of all excess (from Fig. 4).


Fig. 13. Operational Curve of Sample Volume-Uptake Rate (Optimum).


Note that the ordinate represents flow rate. This figure does not reflect on the sensitivity at a given volume.


Fig. 14. Impactor Effects.


- Varian-Techtron capillary nebulizer;
- High solids nebulizer (Fig. 5)






































TECHNICAL REPORT DISTRIBUTION LIST, GEN

	<u>No.</u> <u>Copies</u>		<u>No.</u> <u>Copies</u>
Office of Naval Research 800 North Quincy Street Arlington, Virginia 22217 Attn: Code 472	2	Defense Documentation Center Building 5, Cameron Station Alexandria, Virginia 22314	12
ONR Branch Office 536 S. Clark Street Chicago, Illinois 60605 Attn: Dr. George Sandoz	1	U.S. Army Research Office P.O. Box 1211 Research Triangle Park, N.C. 27709 Attn: CRD-AA-IP	1
ONR Branch Office 715 Broadway New York, New York 10003 Attn: Scientific Dept.	1	Naval Ocean Systems Center San Diego, California 92152 Attn: Mr. Joe McCartney	1
ONR Branch Office 1030 East Green Street Pasadena, California 91106 Attn: Dr. R. J. Marcus	1	Naval Weapons Center China Lake, California 93555 Attn: Dr. A. B. Amster Chemistry Division	1
ONR Area Office One Hallidie Plaza, Suite 601 San Francisco, California 94102 Attn: Dr. P. A. Miller	1	Naval Civil Engineering Laboratory Port Hueneme, California 93401 Attn: Dr. R. W. Drisko	1
ONR Branch Office Building 114, Section D 666 Summer Street Boston, Massachusetts 02210 Attn: Dr. L. H. Peebles	1	Professor K. E. Woehler Department of Physics & Chemistry Naval Postgraduate School Monterey, California 93940	1
Director, Naval Research Laboratory Washington, D.C. 20390 Attn: Code 6100	1	Dr. A. L. Slafkosky Scientific Advisor Commandant of the Marine Corps (Code RD-1) Washington, D.C. 20380	1
The Assistant Secretary of the Navy (R,E&S) Department of the Navy Room 4E736, Pentagon Washington, D.C. 20350	1	Office of Naval Research 800 N. Quincy Street Arlington, Virginia 22217 Attn: Dr. Richard S. Miller	1
Commander, Naval Air Systems Command Department of the Navy Washington, D.C. 20360 Attn: Code 310C (H. Rosenwasser)	1	Naval Ship Research and Development Center Annapolis, Maryland 21401 Attn: Dr. G. Bosmajian Applied Chemistry Division	1
		Naval Ocean Systems Center San Diego, California 91232 Attn: Dr. S. Yamamoto, Marine Sciences Division	1

Encl 1

TECHNICAL REPORT DISTRIBUTION LIST, 051C

	<u>No.</u> <u>Copies</u>		<u>No.</u> <u>Copies</u>
Dr. M. B. Denton University of Arizona Department of Chemistry Tucson, Arizona 85721	1	Dr. K. Wilson University of California, San Diego Department of Chemistry La Jolla, California	1
Dr. R. A. Osteryoung Colorado State University Department of Chemistry Fort Collins, Colorado 80521	1	Dr. A. Zirino Naval Undersea Center San Diego, California 92132	1
Dr. B. R. Kowalski University of Washington Department of Chemistry Seattle, Washington 98105	1	Dr. John Duffin United States Naval Postgraduate School Monterey, California 93940	1
Dr. S. P. Perone Purdue University Department of Chemistry Lafayette, Indiana 47907	1	Dr. G. M. Hieftje Department of Chemistry Indiana University Bloomington, Indiana 47401	1
		Dr. Victor L. Rehn Naval Weapons Center Code 3813 China Lake, California 93555	1
Dr. D. L. Venezky Naval Research Laboratory Code 6130 Washington, D.C. 20375	1	Dr. Christie G. Enke Michigan State University Department of Chemistry East Lansing, Michigan 48824	1
Dr. H. Freiser University of Arizona Department of Chemistry Tucson, Arizona 85721		Dr. Kent Eisentraut, MBT Air Force Materials Laboratory Wright-Patterson AFB, Ohio 45433	1
Dr. Fred Saalfeld Naval Research Laboratory Code 6110 Washington, D.C. 20375	1	Walter G. Cox, Code 3632 Naval Underwater Systems Center Building 148 Newport, Rhode Island 02840	1
Dr. E. Chernoff Massachusetts Institute of Technology Department of Mathematics Cambridge, Massachusetts 02139	1		

DAT
ILMI