

AD-A081 092

AIR FORCE OCCUPATIONAL AND ENVIRONMENTAL HEALTH LAB --ETC F/6 6/5  
A SURVEY OF WASTE ANESTHETIC GAS LEVELS IN SELECTED USAF VETERI--ETC(U)  
FEB 79 J E WILLIGAN, J L SABLAN  
OEHL-TR-79-33

UNCLASSIFIED

NL

1 of 1  
ADDITIONAL



END  
DATE  
SERIAL  
3-80  
doc

MA081092

## NOTICES

When US Government drawings, specifications, or other data are used for any purpose other than a definitely related Government procurement operation, the Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

The mention of trade names or commercial products in this publication is for illustration purposes and does not constitute endorsement or recommendation for use by the United States Air Force.

Do not return this copy. Retain or destroy.

Please do not request copies of this report from the USAF Occupational and Environmental Health Laboratory. Additional copies may be purchased from:

National Technical Information Service  
5285 Port Royal Road  
Springfield, Virginia 22161


Federal Government Agencies and their contractors registered with the (DDC) should direct requests for copies of this report to:

Defense Documentation Center (DDC)  
Cameron Station  
Alexandria, Virginia 22314

This report has been reviewed by the Public Affairs Office and is releasable to the National Technical Information Service (NTIS). At NTIS, it will be available to the general public, including foreign nations.

This technical report has been reviewed and is approved for publication.

*William E. Mabson*  
WILLIAM E. MABSON, Colonel, USAF, BSC  
Commander



UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM	
1. REPORT NUMBER 14 USAF OEHL-TR-79-33	2. GOVT ACCESSION NO.	3. RECIPIENT'S CATALOG NUMBER	
4. TITLE (and Subtitle) 6 A Survey of Waste Anesthetic Gas Levels in Selected <u>USAF</u> Veterinary Surgeries.	5. TYPE OF REPORT & PERIOD COVERED 9 Technical Report Final rept.		
7. AUTHOR(s) 10 Joseph E. Milligan, Major, USAF, MSC James L. Sablan, TSgt, USAF	8. CONTRACT OR GRANT NUMBER(s)		
9. PERFORMING ORGANIZATION NAME AND ADDRESS USAF Occupational and Environmental Health Laboratory Brooks AFB TX 78235	10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS		
11. CONTROLLING OFFICE NAME AND ADDRESS	11 Feb 1979	12. NUMBER OF PAGES 28	
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office)	15. SECURITY CLASS. (of this report) Unclassified		
12 27	15a. DECLASSIFICATION/DOWNGRADING SCHEDULE		
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited.			
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)			
18. SUPPLEMENTARY NOTES			
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Anesthetic Gas Scavenging      Halothane Environmental Exposure        Inhalation Anesthesia Environmental Health          Nitrous Oxide Environmental Pollution       Occupational Diseases Halogenated Anesthetics       Occupational Exposure			
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) Veterinarians and their assistants should consider the potential environmental hazards of chronic exposure to gaseous anesthetics. In humans, epidemiologic surveys have circumstantially shown adverse health effects following exposure to anesthetic gases. These effects include abortion, congenital abnormalities, hepatic and renal diseases, CNS disturbances, increased risk of cancer and decreased fertility. In animals, similar effects due to occupational exposure levels of halothane and nitrous oxide have been experimentally established.			

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

410936

JSM

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

Item 19. Key Words (cont):

Occupational Health	Veterinary Anesthetists
Trace Anesthetic Gases	Veterinary Operating Room Air
Veterinary Anesthesia	Veterinary Surgeons
Veterinary Anesthesiologists	Veterinary Surgeries
Veterinary Anesthetics	Veterinary Technicians
Waste Anesthetic Gases	

Item 20. Abstract (cont):

Previous investigations regarding this personal and personnel health hazard have all concerned medical and dental personnel. There are no documented reports of the potential hazard to veterinary personnel. Therefore, the present survey was conducted to determine the exposure levels of halothane and nitrous oxide in selected USAF veterinary surgeries. In this survey, 11 of 35 halothane samples and 15 of 20 nitrous oxide samples exceeded the maximum exposure levels recommended by NIOSH for those compounds. A potential hazard to the health of personnel working in USAF veterinary surgeries exists. A complete waste anesthetic gas management program and a periodic monitoring program should be maintained. (32 References).

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)

TR-79-33

A SURVEY OF WASTE ANESTHETIC GAS LEVELS  
IN SELECTED USAF VETERINARY SURGERIES

By

Major Joseph E. Milligan  
TSgt James L. Sablan

February 1979

USAF Occupational and Environmental Health Laboratory  
Aerospace Medical Division (AFSC)  
Brooks Air Force Base, Texas 78235

## TABLE OF CONTENTS

<u>SECTION</u>	<u>PAGE</u>
TABLE OF CONTENTS. . . . .	i
LIST OF FIGURES. . . . .	11
LIST OF TABLES . . . . .	111
INTRODUCTION . . . . .	1
METHODS AND MATERIALS. . . . .	3
Sampling Dates and Protocol . . . . .	3
Infrared Spectrophotometry. . . . .	3
Bag Sampling. . . . .	4
Charcoal Tube Sampling. . . . .	4
RESULTS AND DISCUSSION . . . . .	12
BIBLIOGRAPHY . . . . .	17

## LIST OF FIGURES

<u>FIGURE</u>		<u>PAGE</u>
1	Miran-IA Calibration Curve of Standard Concentrations of Halothane. . . . .	5
2	Miran-IA Calibration Curve of Standard Concentrations of N <sub>2</sub> O. . . . .	6
3	Simultaneous Sampling of Halothane and N <sub>2</sub> O at the Breathing Zone of the Surgeon. . . . .	7
4	Simultaneous Sampling of Halothane and N <sub>2</sub> O at the Breathing Zone of the Assistant Surgeon. . . . .	7
5	Simultaneous Sampling of Halothane and N <sub>2</sub> O at the Breathing Zone of the Anesthetist. . . . .	8
6	Simultaneous Recording of Halothane and N <sub>2</sub> O Levels During Sampling. . . . .	8
7	The Miran-IA Portable Gas Analyzer. . . . .	9
8	Calibration of the Miran-IA Infrared Spectrophotometer . . . . .	9
9	Charcoal Tube . . . . .	11
10	Attaching Charcoal Tubes and Pumps to the Anesthetist. . . . .	11

LIST OF TABLES

<u>TABLE</u>		<u>PAGE</u>
1	Levels of Halothane (ppm) in Operating Room Air Sampled at USAFSAM/SGV - OR #1, Brooks Air Force Base. Samples Collected December 1978. . . . .	13
2	Levels of Halothane (ppm) and Nitrous Oxide (ppm) in Operating Room Air Sampled at USAFSAM/SGV - OR #2, Brooks Air Force Base. Samples Collected February 1979. .	14
3	Levels of Halothane (ppm) and/or Nitrous Oxide (ppm) in Operating Room Air Sampled at the Clinical Research Lab and the Military Dog Veterinary Service of Wilford Hall, Lackland Air Force Base. Samples Collected December 1977 and May 1977, Respectively. . . . .	15
4	Levels of Halothane (ppm) and Nitrous Oxide (ppm) in Operating Room Air Sampled at AMRL. Samples Collected May 1978. . . . .	16

## INTRODUCTION

We live in an age of increasing awareness and concern about occupational health and diseases associated with environmental pollution. Attention has recently been called to the possible adverse effects of trace anesthetic gases in the operating room (19). Veterinarians and their assistants who use gaseous anesthetics on a regular basis should consider the potential environmental hazards posed by these chemicals.

Interest in trace anesthetic gases was stimulated in 1967 by a survey of 303 Russian anesthesiologists who reported a high incidence of nausea, irritability, headache, fatigue and pruritus, and an alarming rate of spontaneous abortions (31). Subsequent to this report, a number of epidemiologic studies of humans, and laboratory experiments with animals have been conducted in an attempt to identify any prominent health effects associated with chronic exposure to waste anesthetic gases.

In humans, the major adverse health effects identified by the epidemiologic surveys were abortion and congenital abnormalities. An increased incidence of spontaneous abortion in exposed female workers (1,12-14,21-23) and of congenital abnormalities among their children (13,14,16,22,23) were found. The same increases among the unexposed wives and children of exposed men were also noticed (1,13,14,23). Furthermore, effects on fertility (22), hepatic and renal diseases (13,14,30), the CNS (15,30,31) and the risk of cancer (3,13,18) have been described following exposure to anesthetic gases.

Evidence that occupational exposure to gas anesthetics constitutes a health hazard, however, is only circumstantial. A cause and effect relationship in humans has not been documented. Other factors affecting medical personnel, such as stress, fatigue, and unfavorable working hours may be involved. Nevertheless, the potential harm of the long term exposure to anesthetic gases should not be ignored. Since the lag time between exposure and disease may be as long as 20 years, protection is necessary long before clinical manifestations of disease appear (33).

Recent experimental evidence for laboratory animal studies is supportive of the suggestion that health hazards are actually caused by anesthetic gas exposure. Fetal resorption has occurred in rats following exposure to nitrous oxide at levels within the range of occupational exposure (17). The teratogenic potential of halothane and nitrous oxide at occupational exposure levels has also been established in rats (4,8,9). Moderate liver and kidney damage has been observed in animals following chronic exposure to volatile halogenated anesthetics at levels associated with an occupational environment (6,7,10,27). Cerebral damage and permanent learning deficits have been seen in adult rats exposed to halothane at occupational levels (5,24). The same effect was seen in offspring of rats chronically exposed to occupational levels of halothane during pregnancy (4,26). Furthermore, it has been suggested that halothane and nitrous oxide, when administered simultaneously to rats at

occupational levels, might result in a synergistic cytologic or reproductive effect (20). Such a regimen has been shown to result in dose-dependent increases in cytogenetic aberrations in both bone marrow and spermatogonial cell populations, as well as a decrease in ovulation and implantation efficiency and retardation in fetal development (11).

Where do we stand in USAF veterinary anesthesia? Although all studies to date on this problem have been performed in human operating rooms, there is no reason to think that the problem is less important in veterinary facilities, where gas anesthetics are in frequent use. Obviously, veterinary anesthesiologists, surgeons, technicians and assistants breathe the same gases as their medical and dental counterparts. Some would argue that because of smaller patients, lower flow rates are used (25). However, levels of trace anesthetic gases in veterinary operating areas may well be greater than those reported for human facilities. The difficulties of high concentrations of halothane and high flow rates of oxygen and nitrous oxide for large animals, mask induction and maintenance of laboratory animals with potential leakage problems, along with frequent use of non-rebreathing systems all may contribute to high levels of these potentially hazardous gases (2).

Is breathing trace amounts of waste anesthetic gases in USAF veterinary surgeries a personal or personnel hazard? The evidence for or against is not available. There have been no scientifically documented reports of measurements of trace gases in veterinary hospitals. Only a limited study, sponsored by NIOSH (conducted by Witcher and Hart at the Veterinary Medical Teaching Hospital, University of California at Davis, 1976) addresses the occupational exposure of veterinary personnel to inhalation anesthetics (19). In that study, the average breathing zone concentration of nitrous oxide, in the absence of control measures, was 310 ppm for two anesthetists during two different small animal surgical procedures. The average breathing zone concentration of halothane was 7.1 ppm for the anesthetists during surgery in a gelding. In both cases measurements were obtained using an infrared gas analyzer. Using appropriate scavenging techniques and control measures, the average breathing zone concentrations for anesthetists were reduced to 7.8 ppm for nitrous oxide and 0.35 ppm for halothane.

Until the present, we could only look at data gathered on human operating and dental treatment room personnel and conclude that efforts be made to depollute veterinary preparation, operating and recovery rooms. As a result, the following survey of waste anesthetic gases in selected USAF veterinary surgeries was conducted. The primary emphasis of this study was to establish analytically determined values of exposures levels to two commonly used inhalational anesthetics (halothane and nitrous oxide) at the breathing level of veterinary personnel in selected USAF veterinary surgeries.

## METHODS AND MATERIALS

### Sampling Dates and Protocol

Halothane and/or nitrous oxide levels were measured at five USAF veterinary surgical facilities, as follows:

<u>FACILITY</u>	<u>DATE</u>
1. USAFSAM/SGV Brooks AFB TX Operating Room #1	Dec 78
2. USAFSAM/SGV Brooks AFB TX Operating Room #2	Feb 79
3. Wilford Hall Medical Center Clinical Research Lab Lackland AFB TX	Dec 77
4. Wilford Hall Medical Center Military Dog Veterinary Service Lackland AFB TX	May 77
5. Aeromedical Research Lab Wright-Patterson AFB OH	May 78

The methods of anesthetic gas sampling included direct sampling using an infrared spectrophotometer, grab-bag sampling using an evacuated cylinder, and adsorbent trapping using an activated charcoal tube.

### Infrared Spectrophotometry

Many compounds have characteristic absorption in the infrared region strong enough for both identification and quantitative determination by means of an infrared spectrophotometer (29). All of the anesthetic gases have one or more absorption bands in the infrared spectral region. By passing an infrared light beam of the proper wavelength through a cell containing the sample to be analyzed, the concentration of agent present can be determined by measuring the amount of infrared light absorbed. A specific measurement can be made by selecting a wavelength where other gases do not absorb. Halothane can be measured at wavelengths of 8.45  $\mu\text{m}$ , 8.8  $\mu\text{m}$  and 12.3  $\mu\text{m}$ . A wavelength of 12.3  $\mu\text{m}$  is recommended when halothane is used in combination with nitrous oxide, to avoid interference from the latter. Nitrous oxide can be measured at wavelengths of 4.5  $\mu\text{m}$ , 7.7  $\mu\text{m}$  and 7.9  $\mu\text{m}$ . The best sensitivity and least interference occurs in the 4.5  $\mu\text{m}$  band for nitrous oxide. Excellent sensitivity is achieved by using a multi-pass, long pathlength cell, where the light beam is reflected back and forth through the sample many times (28).

Halothane and nitrous oxide levels were sampled directly at USAFSAM/SGV operating room number 2 (Figures 3-6) and at AMRL using the Wilks Miran-IA General Purpose Gas Analyzer. The Miran-IA portable gas analyzer (Figure 7) is a single-beam variable wavelength analyzer equipped with a 20 meter gas cell. The instrument is small enough (30 pounds) for use in the field. The variable pathlength feature allows the light beam to traverse through the sample many times, so a high enough absorbance value is obtained to achieve excellent sensitivity (28).

The parameters for sampling using the Miran-IA were:

<u>PARAMETER</u>	<u>HALOTHANE</u>	<u>N<sub>2</sub>O</u>
Wavelength	12.3 $\mu$ m	4.5 $\mu$ m
Slit Setting	1.0 mm	1.0 mm
Absorbance Range	0-0.25 units	0-1.0 units
Pathlength	20.25 m	14.25 m
Volume	5.6 l	5.6 l
Windows	NaCl	NaCl

Calibration of the infrared spectrophotometer (Figure 8), construction of standard curves and calculations were accomplished using USAF OEHL methodology based on the Miran-IA Operation, Maintenance and Service Manual (Wilks/Foxboro Analytical, Norwalk CT, 1978). The Miran-IA calibration curves for halothane and nitrous oxide are shown in Figure 1 and Figure 2, respectively.

#### Bag Sampling

Sampling bags of Tedlar, Mylar, or Saran, or evacuated containers may be used for collecting air samples to be analyzed for anesthetic gas content. Samples collected in this manner may be analyzed for halogenated gases using gas chromatography and for nitrous oxide using the infrared spectrophotometric method (19).

A grab sample of air, to be analyzed for N<sub>2</sub>O, was collected from the Military Dog Veterinary Service surgical facility. Sampling was accomplished by opening the valve on a 6 liter evacuated container and allowing it to reach atmospheric pressure. After sampling, all parts were sealed to minimize leakage in or out of the container. The container was transported to the laboratory for analysis of N<sub>2</sub>O by the infrared spectrophotometric method.

#### Charcoal Tube Sampling

Charcoal tube sampling for halothane was accomplished at the following USAF veterinary surgical facilities using the indicated parameters:

<u>FACILITY</u>	<u>FLOW RATE</u>	<u>SAMPLING TIME</u>
USAFSAM/SGV - OR #1	0.2 l/min.	12 min.
Wilford Hall/Clinical Research Lab	0.5 l/min.	60 min.
Military Dog Veterinary Service	0.45 l/min.	44 min.




FIGURE 1. MIRAN-1A CALIBRATION CURVE OF STANDARD CONCENTRATIONS OF HALOTHANE.




FIGURE 2. MIRAN-1A CALIBRATION CURVE OF STANDARD CONCENTRATIONS OF N<sub>2</sub>O.



FIGURE 3. SIMULTANEOUS SAMPLING OF HALOTHANE AND N<sub>2</sub>O AT THE BREATHING ZONE OF THE SURGEON.



FIGURE 4. SIMULTANEOUS SAMPLING OF HALOTHANE AND N<sub>2</sub>O AT THE BREATHING ZONE OF THE ASSISTANT SURGEON.



FIGURE 5. SIMULTANEOUS SAMPLING OF HALOTHANE AND N<sub>2</sub>O AT THE BREATHING ZONE OF THE ANESTHETIST.




FIGURE 6. SIMULTANEOUS RECORDING OF HALOTHANE AND N<sub>2</sub>O LEVELS DURING SAMPLING.




FIGURE 7. THE MIRAN-IA PORTABLE GAS ANALYZER.



FIGURE 8. CALIBRATION OF THE MIRAN-IA INFRARED SPECTROPHOTOMETER.

Samples were also collected at USAFSAM/SGV - OR #2 and at AMRL as backups for the infrared spectrophotometric analyses.

The charcoal tubes used were glass tubes with both ends flame-sealed, 7 cm long, 6 mm O. D. and 4 mm I. D., containing two sections of 20/40 mesh activated coconut shell charcoal separated by a 2 mm portion of urethane foam (Figure 9). Calibration, collection using a portable pump (Figure 10), and gas chromatographic analysis were accomplished using the procedure outlined by NIOSH (19).



FIGURE 9. CHARCOAL TUBE.



FIGURE 10. ATTACHING CHARCOAL TUBES AND PUMPS TO THE ANESTHETISTS.

## RESULTS AND DISCUSSION

The results of the survey of halothane and/or nitrous oxide levels in the operating room air of USAF veterinary surgical facilities are summarized in Tables 1-4. Animal species utilized during the surgical procedures were as follows:

<u>FACILITY</u>	<u>SPECIES</u>
1. USAFSAM/SGV - OR #1	Monkey
2. USAFSAM/SGV - OR #2	Dog
3. Wilford Hall/Clinical Research Lab	Dog
4. Military Dog Veterinary Service	Dog
5. AMRL	Rat

Except where indicated in the Tables, scavenging techniques and control measures recommended by USAF OEHL (32) were in use during the air sampling.

To insure personnel safety, NIOSH has recommended in its recent Criteria Document on Waste Anesthetic Gases and Vapors (19) that certain concentration levels should not be exceeded. These levels are:

(1) 2 ppm for halothane and methoxyflurane.

(2) 25 ppm for nitrous oxide when used alone or in combination with the halogenated agents. NIOSH recommends monitoring to insure that concentration levels are kept within safe limits.

In the present survey of selected USAF veterinary surgeries, halothane concentration levels ranged from 0.06 ppm to 37.2 ppm and nitrous oxide concentration levels ranged from 6 ppm to 270 ppm. In 11 of 35 samples collected, halothane concentration exceeded the maximum levels recommended by NIOSH. In 15 of 20 samples collected, nitrous oxide concentration exceeded the maximum levels recommended by NIOSH.

It is therefore obvious that a potential hazard to the health of personnel working in USAF veterinary surgical facilities exists. A complete waste anesthetic gas management program along with a periodic monitoring program as previously outlined by USAF OEHL (32) should be maintained in order to decrease personal chronic anesthetic exposure to the absolute minimum possible.

TABLE 1. LEVELS OF HALOTHANE (PPM) IN OPERATING ROOM AIR SAMPLED AT USAFSAM/SGV - OR #1, BROOKS AIR FORCE BASE. SAMPLES COLLECTED DECEMBER 1978.

<u>COLLECTION SITE</u>	<u>HALOTHANE CONCENTRATION<sup>a</sup></u>	<u>TIME<sup>b</sup></u>
Breathing Zone - Surgeon	0.6	1.5
	0.6	3.5
	1.5	4.5
Breathing Zone - Assistant Surgeon	0.2	1.5
	0.6	3.5
	0.7	4.5
Breathing Zone - Anesthetist	21.4 <sup>c</sup>	1.5
	37.2 <sup>c</sup>	3.5
	8.4 <sup>c</sup>	4.5
Breathing Zone - Technician	0.7	1.5
	1.5	3.5
	1.1	4.5

<sup>a</sup>Time weighted average (ppm)

<sup>b</sup>Hours after start of surgery

<sup>c</sup>A leak was detected around the mask of the patient

TABLE 2. LEVELS OF HALOTHANE (PPM) AND NITROUS OXIDE (PPM) IN OPERATING ROOM AIR SAMPLED AT USAFSAM/SGV - OR #2, BROOKS AIR FORCE BASE. SAMPLES COLLECTED FEBRUARY 1979.

<u>COLLECTION SITE</u>	<u>ANESTHETIC GAS CONCENTRATION<sup>a</sup></u>		<u>TIME<sup>b</sup></u>
	<u>HALOTHANE</u>	<u>NITROUS OXIDE</u>	
Breathing Zone - Surgeon	0.6	86	0.5
	0.6	120	1.0
	0.6	120	1.5
Breathing Zone - Assistant Surgeon	0.6	49	0.5
	0.4	62	1.0
	0.6	86	1.5
Breathing Zone - Anesthetist	18.0	27	0.5
	0.2	27	1.0
	2.4	270	1.5
Breathing Zone - Technician	0.4	14	0.5
	0.2	11	1.0
	0.2	11	1.5
Animals Head	3.4	223	0.5
	2.4	210	1.0
	1.2	167	1.5

<sup>a</sup>Time weighted average (ppm)

<sup>b</sup>Hours after start of surgery

TABLE 3. LEVELS OF HALOTHANE (PPM) AND/OR NITROUS OXIDE (PPM) IN OPERATING ROOM AIR SAMPLED AT THE CLINICAL RESEARCH LAB AND THE MILITARY DOG VETERINARY SERVICE OF WILFORD HALL, LACKLAND AIR FORCE BASE. SAMPLES COLLECTED DECEMBER 1977 AND MAY 1977, RESPECTIVELY.

<u>COLLECTION SITE</u>	<u>ANESTHETIC GAS CONCENTRATION HALOTHANE<sup>a</sup></u>	<u>NITROUS OXIDE<sup>b</sup></u>
Clinical Research Lab Breathing Zone - Anesthetist	0.06	-
Military Dog Veterinary Service Breathing Zone - Anesthetist	5.4	6 28

<sup>a</sup>Time weighted average (ppm)

<sup>b</sup>Grab samples (ppm)

TABLE 4. LEVELS OF HALOTHANE (PPM) AND NITROUS OXIDE (PPM) IN OPERATING ROOM AIR SAMPLED AT AMRL. SAMPLES COLLECTED MAY 1978.

<u>COLLECTION SITE</u>	<u>ANESTHETIC GAS CONCENTRATION<sup>a</sup></u>	
	<u>HALOTHANE</u>	<u>NITROUS OXIDE</u>
Floor Level	1.7	14
Animals Head	2.7	-
Breathing Zone - Surgeon	1.5	-
Breathing Zone - Anesthetist (Expelling Bag)	2.0 5.0	- -
Breathing Zone - Anesthetist (Manual Ventilation)	4.0	100 120 <sup>b</sup>

<sup>a</sup>Time weighted average (ppm)

All samples collected with scavenger on, except where noted

<sup>b</sup>Sample collected with scavenger off

## BIBLIOGRAPHY

1. Askrog, V. and Harvald, B. 1970. Teratogenic effect of inhalation anesthetics. *Nord. Med.* 833:498-504.
2. Best, J. L. and Mc Grath, C. J. 1977. Trace anesthetic gases: An overview. *J. Am. Vet. Med. Assoc.* 171:1268-1269.
3. Bruce, D. L., Eide, K. A., Linde, H. W. and Eckenhoff, J. E. 1968. Causes of death among anesthesiologists. A 20-year study. *Anesthesiology* 29:565-569.
4. Chang, L.W., Dudley, A. W., Jr., Katz, J. and Martin, A. H. 1974. Nervous system development following in utero exposure to trace amounts of halothane. *Teratology* 9:A-15.
5. Chang, L. W., Dudley, A. W., Jr., Lee, Y. K. and Katz, J. 1974. Ultrastructural changes in the nervous system after chronic exposure to halothane. *Exp. Neurol.* 45:209-219.
6. Chang, L. W., Dudley, A. W., Jr., Lee, Y. K. and Katz, J. 1975. Ultrastructural changes in the kidney following chronic exposure to low levels of halothane. *Am. J. Pathol.* 78:225-232.
7. Chang, L. W., Dudley, A. W., Jr., Lee, Y. K. and Katz, J. 1975. Ultrastructural studies of the hepatocytes after chronic exposure to low levels of halothane. *Exp. Mol. Pathol.* 23:35-42.
8. Chang, L. W., Dudley, A. W., Jr., Lee, Y. K. and Katz, J. 1975. Ultrastructural studies on the pathological changes in the neonatal kidney following in-utero exposure to halothane. *Environ. Res.* 10:174-189.
9. Chang, L. W., Lee, Y. K., Dudley, A. W., Jr. and Katz, J. 1975. Ultrastructural evidence of the hepatotoxic effect of halothane in rats following in-utero exposure. *Can. Anaesth. Soc. J.* 22:330-337.
10. Chenoweth, M. B., Leong, B. K. J., Sparschu, G. L. and Torkelson, T. R. 1972. Toxicities of methoxyflurane, halothane and diethyl ether in laboratory animals on repeated inhalation of subanesthetic concentrations, "In Cellular Biology and Toxicity of Anesthetics." Williams and Wilkins CO., Baltimore, pp. 275-285.
11. Coate, W. B., Kapp, R. W., Jr. and Ulland, B. M. 1974. Toxicity of low concentration, long term exposure, to an airborne mixture of nitrous oxide and halothane. Final Report NIOSH Contract No. CDC-99-74-46.

12. Cohen, E. N., Belvill, J. W. and Brown, B. W., Jr. 1971. Anesthesia, pregnancy, and miscarriage: A study of operating room nurses and anesthetists. *Anesthesiology* 35:343-347.
13. Cohen, E. N. Brown, B. W., Bruce, D. L., Cascorbi, H. F., Corbett, W., Jones, T. H. and Whitcher, C. E. 1974. Occupational disease among operating room personnel: A national study. *Anesthesiology* 41:321-340.
14. Cohen, E. N. Brown, B. W., Bruce, D. L., Cascorbi, H. F., Corbett, W., Jones, T. H. and Whitcher, C. E. 1975. A survey of anesthetic health hazards among dentists. *J. A. Dent. Assoc.* 90:1291-1296.
15. Cook, T. L., Smith, M., Starkweather, J. A., Winter, P. M. and Eger, E. I. 1978. Behavioral effects of trace and subanesthetic halothane and nitrous oxide in man. *Anesthesiology* 49:419-424.
16. Corbett, T. H., Cornell, R. G., Endres, J. L. and Lieding K. 1974. Birth defects among children of nurse-anesthetists. *Anesthesiology* 41:341-344.
17. Corbett, T. H., Cornell, R. G., Endres, J. L. and Millard, R. I. 1973. Effects of low concentrations of nitrous oxide on rat pregnancy. *Anesthesiology* 39:299-301.
18. Corbett, T. H., Cornell, R. G., Lieding, K. and Endres, J. L. 1973. Incidence of cancer among Michigan nurse-anesthetists. *Anesthesiology* 38:260-263.
19. Criteria for a Recommended Standard - Occupational Exposure to Waste Anesthetic Gases and Vapors. 1977. HEW Publication No. (NIOSH) 77-140. US Dept. of Health, Education and Welfare, Public Health Service, Center for Disease Control, National Institute for Occupational Safety and Health, 194 pp.
20. Garro, A. J., Phillips, R. A., Milliken, R. A. and Rendell-Baker, L. 1979. Chronic anesthetic exposure - what is the margin of safety. *Anesthesiology* 50:77-78.
21. Garstka, K., Wagner, K. L., Hamacher, M. 1974. Pregnancy complications in anesthesiologists. Proc. 159th Meeting, Lower Rhine-Westphalia Gyn Obs. Assoc., The Women's Clinic, Bonn University, Bonn, West Germany.
22. Knill-Jones, R. P., Moir, D. D., Rodrigues, L. V. and Spence, A. A. 1972. Anesthetic practice and pregnancy: Controlled survey of women anesthetists in the United Kingdom. *Lancet* 1:1326-1328.
23. Knill-Jones, R. P., Newman, B. J. and Spence, A. A. 1975. Anesthetic practice and pregnancy: Controlled survey of male anesthetists in the United Kingdom. *Lancet* 2:807-809.

24. Quimby, K. L., Aschkenase, L. J., Bowman, R. E., Katz, J. and Chang, L. W. 1974. Enduring learning deficits and cerebral synaptic malformation from exposure to 10 parts of halothane per million. *Science* 185:625-627.
25. Sawyer, D. C. 1976. A personnel hazard: Operating room environmental pollution. *J. Amer. An. Hosp. Assoc.* 12:214-217.
26. Smith, R. F., Bowman, R. E. and Katz, J. 1978. Behavioral effects of exposure to halothane during early development in the rat: Sensitive period during pregnancy. *Anesthesiology* 49:319-323.
27. Stevens, W. C., Eger, E. I., II, White, A., Halsey, M. J., Munger, W., Gibbons, R. D., Dolan, W. and Shargel, R. 1975. Comparative toxicities of halothane, isoflurane, and diethyl ether at subanesthetic concentrations in laboratory animals. *Anesthesiology* 42:408-419.
28. Syrjala, R. J. 1977. The analysis of anesthetic gases with Miran infrared analyzers. Application Report: Infrared Methods of Analysis, No. 10, Wilks, Norwalk CT, 21 pp.
29. Thompson, B. 1974. Hazardous gases and vapors: Infrared spectra and physical constants. Industrial Technical Report TR-595, Beckman Instruments, Fullerton CA, 323 pp.
30. Uhlirova, A. and Pokorny, J. 1974. Results of questionnaire survey of health damage to anesthesiologists. *Rozhl. Chir.* 53:761-770.
31. Vaisman, A. I. 1967. Working conditions in surgery and their effect on the health of anesthesiologists. *Eksp. Khir. Anesteziol.* 3:44-49.
32. Witcher, C. E. and Rock, J. C. 1978. Control of occupational exposure to inhalational anesthetics-current status. OEHL Report TP-78-1, USAF Occupational and Environmental Health Laboratory, Aerospace Medical Division, Brooks AFB TX, 21 pp.
33. White, R. J. 1978. Pollution control in the operating room. *Vet. Clinics N. Amer.* 8:337-342.