

AD-A102 703

NAVAL AVIONICS CENTER INDIANAPOLIS IN  
REVIEW OF DATA GENERATED WITH INSTRUMENTS USED TO DETECT AND ME--ETC(U)  
FEB 81


F/G 7/4

UNCLASSIFIED

NAC-MRR-3-78-REV

NL

[ OF ]  
AD  
800000



END  
DATE  
FILMED  
9-81  
DTIC

LEVEL II

2

Materials Research Report 3-78

29 AUGUST 1978

6

REVIEW OF DATA GENERATED WITH INSTRUMENTS  
USED TO DETECT AND MEASURE IONIC CONTAMINANTS  
ON PRINTED-WIRING ASSEMBLIES. Revision 1

AD A102703



11 17 Feb 81

12 36

APPROVED FOR PUBLIC RELEASE;  
DISTRIBUTION UNLIMITED.

Materials  
Laboratory and  
Consultants  
Division



NAVAL AVIONICS CENTER  
Indianapolis, Indiana 46218

DTIC  
ELECTE  
AUG 11 1981

D

14 NAC-MRR-3-78-REV

DTIC FILE COPY

410501

81 8 10 109

76

llm



DEPARTMENT OF THE NAVY  
 NAVAL AVIONICS CENTER  
 INDIANAPOLIS, INDIANA 46218

IN REPLY REFER TO:  
 712:WTH:RJD:mjr  
 5213/3  
 29 August 1978

(A) Revised 17 Feb 1981

MATERIALS RESEARCH REPORT NO. 3-78

SUBJECT: Review of Data Generated with Instruments Used to Detect and Measure Ionic Contaminants on Printed-Wiring Assemblies

ABSTRACT:

The comparison of test methods; the manufacture, cleaning, and controlled ionic contamination of the printed wiring assemblies; the calibration of each test instrument; and the test data are presented. The steps of each method are detailed. Interim instrument factors are elucidated as well as the rationale for their development.

INTRODUCTION:

An investigation was undertaken to measure various contaminant levels on equivalent printed wiring assemblies. The results obtained utilizing the procedure in a Naval Avionics Center (NAC) Modified MIL-P-28809 test were compared with those obtained using commercially available test instruments. Measurements in the NAC-Modified MIL-P-28809 test were made with a Beckman Conductivity Bridge (Model RC-16C) and a digital read-out Markson ElectroMark Conductivity Analyzer (Model J-4405). Comparison measurements were made with a NAC-Modified Alpha Ionograph <sup>®</sup>, a Kenco Omega Meter <sup>®</sup> (Model 200), and a NAC-Modified DuPont Ion Chaser <sup>®</sup>. Uncomplicated printed wiring assemblies were used in studies documented in Materials Research Report 3-72. However, printed wiring assemblies with quite different assembly components were designed for this study. These varied assemblies were designed to detect any difference in response of the various test procedures.

For constant levels of contamination, the amount of contamination indicated by each instrument invariably fell in the same order. The highest to the lowest indications of ionic contaminants always fell in this order:

Highest - NAC-Modified DuPont Ion Chaser <sup>®</sup>  
 NAC-Modified Alpha Ionograph <sup>®</sup>  
 Kenco Omega Meter <sup>®</sup>

Lowest - NAC-Modified MIL-P-28809

Inspection of the data indicated that each test instrument had good

Accession For	
NTIS GRA&I	<input checked="" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By	
Distribution/	
Availability Codes	
Avail and/or	
Dist	Special
A	

DTIC  
 ELECTE  
 AUG 11 1981  
 S D  
 D

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

precision. The proposed interim instrument factors are the consensus judgment of representatives from industry and NAC present at a review meeting\* held at NAC on 9-10 February 1978.

PROCEDURE:

In order to determine the effect of the printed wiring assemblies configuration on the response of the various test instruments, nine different printed wiring assemblies were designed and eighteen of each kind were fabricated (Appendix A-2 - A-10). The configurations of these assemblies show the probability of a wide range of possible entrapment of ionic contaminants.

Verification of the cleaning procedure was accomplished in the following manner. Eighteen AV21921 jumper-wire-assembly boards (Appendix A-10) were first cleaned in the vapor area of a degreaser using inhibited 1-1-1 trichloroethane. Next, all eighteen printed wiring assemblies were transferred to a cleaning system containing a wash solution composed of 75 volume-% isopropyl alcohol and 25 volume-% deionized water (Appendix A-11). The resistivity of this wash solution was originally greater than  $25 \times 10^6$  ohm-cm. After introduction of the assemblies, the resistivity of the wash solution fell to  $13.8 \times 10^6$  ohm-cm. The wash solution was circulated through the cleaning system until its resistivity reached  $20 \times 10^6$  ohm-cm. At that time, the assemblies were removed from the cleaning system and spray rinsed using a plastic squeeze wash bottle containing wash solution with a resistivity greater than  $25 \times 10^6$  ohm-cm. Handled with clean forceps, the printed wiring assemblies were blown dry with water pumped nitrogen and placed in individual known-clean plastic bags. The bags were closed and sealed with plastic tape.

Proof of cleaning was established by testing all of the above printed wiring assemblies by the NAC-Modified MIL-P-28809 Method using a Beckman Conductivity Bridge. The volume of test solution collected was equal to 10-ml per square inch of board plus component area. Statistical analysis of the resistivity of the collected test solution of each individual printed wiring assembly gave the following results:

$$\begin{aligned}\bar{X} &= 11.468 \times 10^6 \text{ ohm-cm}; \sigma = 3.59 \times 10^6 \text{ ohm-cm} \\ \bar{X} + 3\sigma &= 22.22 \times 10^6 \text{ ohm-cm} \\ \bar{X} - 3\sigma &= 5.92 \times 10^6 \text{ ohm-cm}\end{aligned}$$

Since the lowest resistivity of washings measured was  $9.0 \times 10^6$  ohm-cm, it is evident that the printed wiring assemblies were adequately cleaned.

\* See list of attendees, PWB Ionic Contaminants Meeting, Naval Avionics Center, 9-10 Feb 1978 (Appendix A-1).

Each unique group of 18 printed wiring assemblies was cleaned in this manner. The lot cleaning was deemed to be adequate when the measured resistivity of test solution from any specimen in the sample of three printed wiring assemblies was not lower than  $9 \times 10^6$  ohm-cm when tested by the NAC-Modified MIL-P-28809 method. These three proof-of-cleaning printed wiring assemblies were not used subsequently. The remaining fifteen printed wiring assemblies were then contaminated with MIL-F-14256D, Type RA Flux.

Contamination of the printed wiring assemblies was accomplished by dipping the assemblies into the flux and withdrawing them at the rate of eight inches per minute by a synchronous motor. The flux-coated assemblies were hung vertically and allowed to drain for five minutes; hung in an air circulating oven for thirty minutes at  $107^\circ\text{C}$  ( $225^\circ\text{F}$ ); removed, allowed to cool; and placed in individual known-clean plastic bags. The top of each bag was folded over to form a flap which was secured with plastic tape.

The above flux was composed of 50 g/l WW gum rosin, purchased to meet requirements of LL-R-626; 1 g/l hexadecyltrimethyl ammonium chloride (Eastman Chemical Co., Catalog #9536); and the remainder ACS Reagent Grade Isopropyl Alcohol. The flux had a specific gravity of 0.807 at  $23^\circ\text{C}$  ( $73^\circ\text{F}$ ). The resistivity of water extract, determined in accordance with MIL-F-14256D, was  $5.8 \times 10^4$  ohm-cm. This value meets the water extract resistivity requirements for MIL-F-14256D, type RA Flux.

#### NAC TEST APPARATUS:

It became evident that the procedure in para 4.8.3 of MIL-P-28809 is most satisfactory for small test assemblies. Since larger assemblies required more test solution<sup>1</sup> than was convenient to deliver from a plastic squeeze wash-bottle, MIL-P-28809 was modified. A five-gallon plastic rectangular carboy containing the test solution was elevated so that the distance from the top of the contained solution to the top of the work bench was 4 feet 8 inches. The test solution passed through Tygon<sup>®</sup> tubing to a mixed-bed deionizer column, through an in-line conductivity cell, thence to a stainless steel needle valve, and through additional Tygon<sup>®</sup> tubing terminating in a stainless steel nozzle 2.75 inches (70 mm) long with a nominal inside diameter of 0.0625 inches (1.59 mm). The test solution was sprayed over both sides of the specimens which were suspended over an 8" polyethylene funnel. The test solution was collected in a TPX-polymethyl pentene beaker which was pre-marked to a computed volume (10 ml/sq.in.-area computed from both board and component dimensions). Resistivity and conductivity

<sup>1</sup> Throughout this report, "test solution" denotes a solution composed of 75 volume-percent ACS Reagent Grade 2-Propanol (Isopropyl Alcohol) and 25 volume-% deionized water.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

readings were made with a CEL-A001 dip cell ( $K=0.0100/\text{cm}$ ) attached to a Beckman Model RC-16C Conductivity Bridge. Conductivity measurements were made subsequently on the same solution with a Model J-4405 Markson ElectroMark Conductivity Analyzer. The test solution reservoir level was maintained to standardize head pressure and consequently the velocity of the test solution at the delivery tip. The Conductivity Bridge was connected to a double-pole double-throw switch. Thus, resistivity measurements of the test solution were made before delivery and after spraying the test specimens.

New plastic ware was cleaned and reserved exclusively for this test. It is imperative that this laboratory ware be used for this test only. Experimentation proved that rinsing the dip probe, funnel and beaker with fresh solution provided sufficient cleaning between test runs on the same day.

This modified apparatus improved the precision of measurements because the solution was fresh from the ion-exchange column and had no opportunity to absorb carbon dioxide from the air before being ejected at the delivery tip. Typical resistivity of the test solution before use is  $25 \times 10^6$  ohm-cm. Also, for continuous operation, elimination of the squeeze bottle significantly reduced operator fatigue.

ADEQUATE VENTILATION SHOULD BE PROVIDED WHEN PERFORMING IONIC CONTAMINANT TESTING IN ACCORDANCE WITH MIL-P-28809. PREFERABLY, TESTING WITH PORTABLE INSTRUMENTATION SHOULD BE PERFORMED IN A FUME HOOD!

#### BECKMAN CONTINUITY BRIDGE:

The Beckman, Model RC-16C, Conductivity Bridge (Beckman Instruments, Cedar Grove, NJ 07009) is a compact, portable, completely solid state, manually balanced general purpose conductivity bridge. It is a wide-range instrument of high sensitivity and accuracy operating on self-contained batteries or direct from an AC Line. It may be used for measurements in grounded or isolated containers. It is provided with a dual color-coded scale for simultaneous resistance and conductance measurements over the ranges of 0.2 to 2,500,000 ohms-cm and 0.4 to 5,000,000 micromhos/cm. This range is modified with the cell constant of the conductivity cell. The test solution in this study is poorly conductive. The use of a conductivity dip cell with a low constant (CEL-A001;  $K=0.0100/\text{CM}$ ) is required. Measurements are made utilizing the null indicating meter which also functions as a battery condition indicator.

MARKSON CONDUCTIVITY ANALYZER:

The Markson<sup>®</sup>, Model J-4405, ElectroMark Conductivity Analyzer (Markson Science, Inc., Del Mar, CA 92014) is a digital display conductivity meter with an operating range of 0 - 199,900 micromhos/cm. The sensing probe is a cylindrical tube with gold-plated electrodes inside. A thermistor inside the tube provides automatic temperature compensation from 0 - 100°C.

KENCO OMEGA METER

The Kenco, Model 200, Omega Meter<sup>®</sup>, (Kenco Alloy and Chemical Co., Inc., Addison, IL 60101) is an instrument specifically designed for measuring ionic contaminants on printed wiring assemblies. It consists of a reservoir of test solution, mixed-bed deionizer columns in parallel, a pump, a test chamber with conductivity cell, a timer, a plotter and a control console. It has a programmable electronic interlock system. In use, the pump forces the test solution out of the reservoir, through the deionizer columns into the test chamber and back into the reservoir. When the resistivity of the test solution in the test chamber has reached a preselected value, the test chamber is isolated from the loop. The volume of test solution in the test chamber is adjusted for the surface area of the specimens being examined. The printed wiring assemblies are immersed in the chamber and a magnetic stirrer agitates the test solution. A timer is set and, as the test proceeds, the resistivity is continuously measured and recorded on chart paper.

ALPHA IONOGRAPH<sup>®</sup>, NAC-MODIFIED:

The Ionograph<sup>®</sup> (Alpha Metals, Jersey City, NJ 07304) is an instrument designed for measuring ionic contaminants on printed wiring assemblies. It consists of a metering pump, a mixed-bed deionizer column, a test chamber, an in-line conductivity cell, a meter, an integrator, a plotter and a control console. It is a closed loop system in which the test solution continuously circulates throughout the test. When a test is underway, the metering pump forces the test solution through the deionizer column, through the test chamber in which the specimen is suspended, through the in-line conductivity cell and back to the pump. The sodium chloride equivalent of the ionic contaminants is continuously measured, integrated, and plotted on chart paper. NAC modified the instrument by installing a larger mixed-bed deionizer (Barnstead Hose Nipple cartridge, D8902HN, Ultra-pure), a relay to stop the pump when a pre-selected conductivity base-line is reached, and a test chamber with a capacity of four liters. A magnetic stirrer agitates this large volume of test solution.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

DUPONT ION CHASER<sup>®</sup>, NAC-MODIFIED:

The Ion Chaser ("Freon" Products Laboratory, E. I. du Pont de Nemours & Co., Wilmington, DE 19898) is an instrument used for measuring the ionic contaminants level on printed wiring assemblies. It consists of a metering pump, a mixed-bed deionizer column, six test chambers connected with Tygon<sup>®</sup> tubing and twelve 3-way valves (two per test chamber), an in-line conductivity cell, a multiple magnetic stirrer, a meter, an integrator, a plotter, and suitable consoles. Prior to a test, the solution in all six chambers is circulated sequentially until a preselected conductivity base line is obtained. The test chambers are isolated from the loop. A specimen is placed in each chamber and allowed to extract for one hour. The specimens are removed from the chambers. Test solution from each chamber is then consecutively pumped through the measuring section until the preselected conductivity base line is reached. The integrator then displays the sodium chloride equivalent of the ionic contaminants. The instrument was modified by installing a larger mixed bed deionizer column (Barnstead Hose-Nipple Cartridge, D8902HN Ultra-pure).

NAC-MODIFIED MIL-P-28809 METHOD, CALIBRATION CURVE FOR:

STANDARD SODIUM CHLORIDE SOLUTIONS:

SOLUTION TO BE ADDED FROM A MICROLITER SYRINGE:

1. Dry reagent grade sodium chloride at 105°C for 4 hours minimum or overnight. Cool in a dessicator.
2. Weigh  $10.0000 \pm 0.0002$  g of the dried salt and carefully transfer to a clean 1-liter volumetric flask.
3. Dissolve with high quality distilled or deionized water and dilute to 1-liter. Stopper tightly.
4. Label: NaCl-S; 10 g/l NaCl ( $30 \mu\text{l} = 300 \mu\text{g NaCl}$ ).

SOLUTION TO BE ADDED FROM A 50-ml BURET:

1. Dry reagent grade sodium chloride at 105°C for 4 hours minimum or overnight. Cool in a desiccator.
2. Weigh  $6.0000 \pm 0.0002$  g of the dried salt and carefully transfer to a clean 1-liter volumetric flask.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

3. Dissolve with high quality distilled or deionized water and dilute to 1-liter. Label NaCl-A; 6 g/l NaCl.
4. Pipet 10 ml of the NaCl-A solution to a clean 1-liter volumetric flask. Dilute to 1-liter with test solution<sup>2</sup>. Stopper and mix well.
5. Label: NaCl-B; 0.06 g/l NaCl (5-ml = 300 $\mu$  NaCl).

#### ACQUISITION OF CALIBRATION DATA:

Operate the conductivity and/or resistivity bridge, with appropriate conductivity cell, according to the manufacturer's instructions. It should measure accurately resistivity in the range between  $1 \times 10^6$  ohm-cm and  $25 \times 10^6$  ohm-cm (conductivity in the range between  $0.04 \times 10^{-6}$  mho/cm and  $1 \times 10^{-6}$  mho/cm).

Add 2-liters of test solution to a clean TPC-Polymethyl pentene beaker. Stir magnetically for two minutes. Measure the resistivity/conductivity of the solution with the bridge that is to be calibrated. If resistivity measures less than  $6 \times 10^6$  ohm-cm (conductivity measures more than  $0.167 \times 10^{-6}$  mho/cm), reclean the apparatus and repeat procedure. A valid calibration cannot be made unless the virgin test solution has a resistivity of at least  $6 \times 10^6$  ohm-cm (conductivity -  $0.167 \times 10^{-6}$  mho/cm).

#### ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A MICROLITER SYRINGE:

1. Measure and record the temperature of the test solution above.
2. Add 30 $\mu$ l of NaCl-S (30 $\mu$ l = 300 $\mu$ g NaCl) with a microliter syringe.
3. Mix test solution for 2 minutes. Measure and record the resistivity/conductivity.
4. Repeat steps 2 and 3 until a total of ten 30 $\mu$ l additions of NaCl-S (3000 $\mu$ g NaCl) has been made.

#### ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A 50-ml BURET:

Proceed as above, adding 5-ml portions of NaCl-B, until a total of ten 5-ml additions of NaCl-B (3000 $\mu$ g NaCl) has been made.

<sup>2</sup> 75-volume-percent 2-propanol solution (minimum resistivity -  $25 \times 10^6$  ohm-cm) was used to dilute this standard. If water was used for this dilution, error would be introduced because the final alcohol-water ratio of the test solution would not be 75% 2-propanol, 25% water.

#### PLOTTING CALIBRATION CURVE:

This method is valid for additions made with either NaCl-S or NaCl-B solutions.

Study the Calibration Curve for MIL-P-28809 Method using Beckman Bridge (Appendix A-12) and the Calibration Curve for MIL-P-28809 using Markson Meter (Appendix A-13).

Plot the calibration curve on 2-cycle log-log graph paper. Plot this carefully because it will be used to calculate the amount of ionic contaminants.

#### IONOGRAPH<sup>®</sup>, CALIBRATION CURVE FOR:

NaCl Standard Solutions: Prepare appropriate NaCl Standard Solutions as described on pages 6 and 7.

#### PREPARATION OF TEST INSTRUMENT:

1. Fill the Ionograph<sup>®</sup> with test solution according to the manufacturer's instructions. NAC's test chamber for this study had a volume of four liters.
2. Turn on the electrical power switch.
3. Set the pump rate adjustment to 9.
4. Set the integrator count rate at 40 counts/sec.
5. Set the meter range to X1.
6. Turn on the pump.
7. Turn on the recorder.
8. Run the pump until the conductivity meter reads 0.025  $\mu$ mho/cm.
9. Adjust the integrator to turn off the counter and stop the pump. (On unmodified Ionographs, the pump may be stopped manually when the meter falls to 0.025  $\mu$ mho/cm. Null and balance the integrator against this level as a base line.)

#### ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A MICROLITER SYRINGE:

1. Measure and record the temperature of the test solution.
2. Add 30  $\mu$ l of NaCl-S (30  $\mu$ l = 300  $\mu$ g NaCl).
3. Turn on the electrical power switch.
4. Turn on the pump.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1979

5. Turn on the integrator counter.
6. Turn on the recorder.
7. Allow the pump to run until the conductivity meter reads  $0.025 \mu \text{ mho/cm}$ .
8. Turn off the pump. (The NAC-Modified Ionograph automatically turns off the pump upon reaching the set base line.)
9. Record the integrator count on the strip chart.
10. Repeat steps 2-9 increasing the amount of NaCl-S by  $30 \mu \text{ l}$  each time until the final addition is  $300 \mu \text{ l}$  ( $3000 \mu \text{ g NaCl}$ ).

#### ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A 50-ml BURET:

Proceed as above. Add NaCl-B in 5 ml increments until the final addition is 50 ml ( $3000 \mu \text{ g NaCl}$ ).

#### PLOTTING CALIBRATION CURVE:

This method is valid for additions made with either NaCl-S or NaCl-B Standard Solutions.

Study the Calibration Curve for the Ionograph<sup>®</sup> (Appendix A-14).

Plot the calibration curve on 2-cycle log-log paper. Plot this carefully because it will be used to calculate the amount of ionic contaminants.

#### ION CHASER<sup>®</sup>, CALIBRATION CURVE FOR:

NaCl Standard Solutions: Prepare appropriate NaCl Standard Solutions as described on pages 6 and 7.

#### PREPARATION OF TEST INSTRUMENT:

1. Fill the Ion Chaser<sup>®</sup> with test solution according to manufacturer's instructions.
2. Turn on the electrical power.
3. Set pump rate setting on 9.
4. Check valving to make sure solution is being pumped to and from cell number 1.
5. Make sure that conductivity meter is operating on the green scale.
6. Turn on the integrator power.
7. Adjust integrator counter to zero.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1979

8. Allow the pump to run until the conductivity meter reads  $0.025\ \mu\text{ mho/cm}$ .
9. Adjust the integrator to turn off the counter.

ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A MICROLITER SYRINGE:

1. Measure and record the temperature of the solution.
2. Add  $30\ \mu\text{ l}$  of NaCl-S solution ( $30\ \mu\text{ l} = 300\ \mu\text{ g NaCl}$ ) to cell number 1.
3. Turn on the electrical power.
4. Turn on the pump.
5. Turn on integrator counter.
6. Turn on recorder.
7. Run the pump until the conductivity meter reads  $0.025\ \mu\text{ mho/cm}$ . (The indicator light should go out at this reading.)
8. Turn off the pump.
9. Record the integrator count.
10. Repeat steps 2-9 increasing the amount of NaCl-S by  $30\ \mu\text{ l}$  each time until the final addition is  $300\ \mu\text{ l}$  ( $3000\ \mu\text{ g NaCl}$ ).

ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A 50 ml-BURET:

Proceed as above. Add NaCl-B in 5 ml increments until the final addition is 50 ml ( $3000\ \mu\text{ g NaCl}$ ).

PLOTTING CALIBRATION CURVE:

This method is valid for either NaCl-S or NaCl-B Standard Solutions.

Study the Calibration Curve for the Ion Chaser (Appendix A-15).

Plot the calibration curve on 2 cycle log-log paper. Plot this carefully because it will be used to calculate the amount of ionic contaminants.

OMEGA METER<sup>®</sup>, CALIBRATION CURVE FOR:

NaCl Standard Solutions: Prepare appropriate NaCl Standard Solutions as described on pages 6 and 7.

PREPARATION OF TEST INSTRUMENT:

1. Fill the Omega Meter<sup>®</sup> with test solution.

2. Turn on the electrical power.
3. Adjust the high limit set point past 20.
4. Turn on the Clean Cycle.
5. Fill and empty the test chamber until the meter needle reaches the maximum resistivity setting.
6. Adjust the resistivity high limit set point to 20. This illuminates the Clean Cycle COMPLETE button and activates an audible alarm.
7. Manipulate drain and fill push-switches until the test chamber contains exactly 2 liters of test solution (50 sq.in. line on test chamber). The instrument is now ready for calibration.

ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH MICROLITER SYRINGE:

1. Measure and record the temperature of the test solution in the test chamber.
2. Set the timer for 3 minutes.
3. Add  $30\mu\text{l}$  of NaCl-S solution ( $30\mu\text{l} = 300\mu\text{g NaCl}$ ).
4. Push on the Recorder ON switch.
5. Push on the Test Cycle ON switch. (Unless Clean Cycle COMPLETE switch is illuminated, no calibration or test can be performed.)
6. At the end of three minutes, an alarm will sound, the ACCEPT indicator will be illuminated, and the timer will reset automatically. Read and record the resistivity of the solution.
7. Adjust the high limit set point on the meter to a value somewhat less than the reading in Step 6.
8. Push Clean Cycle ON button. Clean Cycle COMPLETE button should illuminate immediately. An alarm will sound. Instrument is now ready for further calibration.
9. Repeat steps 3-8 until a total of ten  $30\mu\text{l}$  additions of NaCl-S solution has been made. ( $300\mu\text{l} = 3000\mu\text{g NaCl}$ ).

ADDITION OF STANDARD SODIUM CHLORIDE SOLUTION WITH A 50-ml BURET:

Proceed as above, adding 5-ml portions of NaCl-B, until a total of ten 5-ml additions of NaCl-B ( $3000\mu\text{g NaCl}$ ) has been made.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1979

#### PLOTTING CALIBRATION CURVE:

This method is valid for additions made with either NaCl-S or NaCl-B Standard Solution.

Study the Calibration Curve for the Omega Meter (Appendix A-16).

Plot the calibration curve on 2 cycle log-log paper. Plot this carefully because it will be used to calculate the amount of ionic contaminants.

#### IONIC CONTAMINANT TESTING:

##### GENERAL:

The results of all assemblies tested with each method were entered as raw data in a log book. To maintain as much objectivity as possible, the raw data was not converted until the entire series of comparison tests was completed.

The assemblies were always handled with clean forceps or clean hemostats. These assemblies were always stored individually in known-clean plastic bags during any in-process delay. The continued accuracy of the test instruments was assured by periodic standardization with standard salt solutions. Three printed wiring assemblies of each drawing number of MIL-F-14256 Type RA flux contaminated assemblies were tested by each test method.

#### NAC-MODIFIED MIL-P-28809 TEST METHOD (See pages 3 and 4):

A scrupulously clean TPX-Polymethyl pentene beaker was premarked to the computed volume (10 ml/sq.in. - area computed from the board and attached component dimensions). A scrupulously clean polyethylene funnel was used to direct the  $25 \times 10^6$  ohm-cm test solution into the beaker. This test solution was directed from a nozzle at a constant head in such a manner as to uniformly wash all surfaces of the printed wiring assembly. Washing was stopped when the premarked volume was reached.

BECKMAN CONDUCTIVITY BRIDGE, Model RC-16C, Serial No. 5503;  
CEL-A001 DIP Cell (K - 0.0100/cm):

Essentially, the Beckman Instrument was used in accordance with the manufacturer's instructions to measure the resistivity of the test solution collected by the NAC-Modified MIL-P-28809 Test Method.

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

MARKSON ELECTROMARK<sup>®</sup> CONDUCTIVITY ANALYZER, Model J4005,  
Serial No. 7211:

Essentially, the Markson ElectroMark<sup>®</sup> Conductivity Analyzer was used in accordance with the manufacturer's instructions to measure the conductivity of the wash solution that was first measured with the Beckman Conductivity Bridge. This measurement was made immediately following the Beckman Conductivity Bridge measurement.

KENCO OMEGA<sup>®</sup> METER, Model 200, Serial No. 1008:

All printed wiring assemblies were tested according to the manufacturer's operating instructions except for assemblies with so much surface area that 40 ml/in<sup>2</sup> of surface would exceed the test chamber's capacity. A 15 minute extraction time was used for each printed wiring assembly. Those printed wiring assemblies with so much surface area that they could not be tested with 40 ml/in<sup>2</sup> without overflowing the test chamber were tested with 20 ml/in<sup>2</sup> of test solution. It should be noted that the calibration curve generated for the Kenco Omega<sup>®</sup> Meter was used to convert the measured resistivities to  $\mu$ g NaCl equivalent/in<sup>2</sup>.

ALPHA IONOGRAPH<sup>®</sup>, NAC-MODIFIED, Serial No. 207-WI:

The NAC-Modified Alpha Ionograph<sup>®</sup> was filled with test solution. It was operated per the manufacturer's instructions except 0.050  $\mu$  mho/cm was used as the conductivity baseline or cutoff, rather than the so-called natural conductivity base line (the lowest conductivity that the system could attain). Generally, a system factor,  $\mu$ g of NaCl added divided by the number of integrator counts accumulated in returning to the baseline, was determined before each set of three printed wiring assemblies of a given drawing number were tested. This system factor multiplied by the integrator count determined the  $\mu$ g of NaCl equivalents on the printed wiring assembly tested.

DUPONT ION CHASER<sup>®</sup>, NAC-MODIFIED:

The NAC-Modified DuPont Ion Chaser<sup>®</sup> was filled with test solution. It was used according to DuPont's Instructions except 0.050  $\mu$  mho/cm was used as the conductivity baseline or cutoff rather than the so-called natural conductivity (the lowest conductivity that the system could attain). Generally, a system factor,  $\mu$ g of NaCl added divided by the number of integrator counts accumulated in returning to the baseline, was determined for each of three test chambers before each set of three printed wiring assemblies of a given drawing number were

tested. In the testing mode, each printed wiring assembly was allowed to extract for one hour; then the assembly was removed. The contents of each individual tank were circulated through the system until the system reached the set base line ( $0.050 \mu \text{ mho/cm}$ ). The integrator count multiplied by the system factor for the tank used gave the  $\mu \text{ g}$  of NaCl equivalents on the printed wiring assembly tested.

#### DISCUSSION:

The NAC test values found by each method are tabulated in Table I (Appendix A-17). The calibration curves (Appendix A-12 through A-16) were used to convert the raw test data to  $\mu \text{ g}$  of NaCl equivalents per square inch of the total board and component area for each printed wiring assembly tested by the NAC-Modified MIL-P-28809 Test Method and the Kenco Omega<sup>®</sup> Meter. As a systems factor was found for both the NAC-Modified Alpha Ionograph<sup>®</sup> and the NAC-modified DuPont Ion Chaser<sup>®</sup>, it was deemed superfluous to use calibration curves. However, where the printed wiring assembly being tested has less than  $400 \mu \text{ g}$  salt equivalents of soil, it would be well to determine the system factor with an amount of NaCl which approximates the expected amount of ionic contaminants to be found on the printed wiring assembly.

The calibration curves for the NAC-Modified MIL-P-28809 Method (Beckman Bridge and Markson Meter) and the Omega Meter<sup>®</sup> were used as follows to calculate the  $\mu \text{ g}$  of NaCl equivalents per square inch of printed wiring assembly.

1. Extend the resistivity or conductivity, as appropriate, horizontally from its value on the y-axis until it intersects the calibration curve.
2. Extend a vertical from the point of intersection to the x-axis. Read  $\mu \text{ g/l}$  NaCl.
3. Multiply the concentration in  $\mu \text{ g NaCl/l}$  by the total liters of test solution used. This result indicates the total  $\mu \text{ g}$  of NaCl equivalents removed from the printed wiring board.
4. Divide the  $\mu \text{ g}$  of NaCl equivalents determined in Step 3 by the area (board plus components) of the printed wiring assembly expressed as square inches or square centimeters. This yields the micrograms of NaCl equivalents per square inch or square centimeter.

The data generated is in Table I (Appendix A-17). The average amount of ionic contaminants expressed a  $\mu \text{ g}$  NaCl equivalents per

square inch found for each group of three printed-wiring assemblies of the same drawing number tested by each method was plotted (Appendix A-18). It is evident that each test method has reasonable precision. For a given level of ionic contamination, it is apparent the NAC-Modified MIL-P-28809 Method gave the lowest indication of ionic contaminants. In ascending order, the NAC-Modified MIL-P-28809 Method, the Kenco Omega Meter<sup>®</sup>, the NAC-Modified Alpha Ionograph<sup>®</sup>, and the NAC-Modified DuPont Ion Chaser<sup>®</sup>, gave higher indications of that level of ionic residue.

MANUFACTURERS - MILITARY

The test instrument manufacturers and the involved military personnel met at Naval Avionics Center in Indianapolis on 9-10 February 1978. A representative of one of these test instrument manufacturers commented that the NAC-developed-and-modified Ionic Contaminant Test using  $2 \times 10^6$  ohm-cm resistivity ( $10.06 \mu\text{g}/\text{in}^2$  or  $1.56 \mu\text{g}/\text{cm}^2$  sodium chloride equivalents) as the limit of ionic contaminants for a printed wiring assembly has been used successfully for over 5 years in producing high reliability electronic gear. Therefore, it was proposed that the grand average ( $\bar{x}$ ) be computed for the ionic contaminants found by each test method. An interim "equivalence factor" for each test would be calculated by dividing its grand average ( $\bar{x}$ ) by the grand average of the NAC-Modified MIL-P-28809 Method. This met the approval of the attendees.

The following were computed and will be acceptable as interim "equivalence factors".

METHOD	$\bar{x}$ $\mu\text{g NaCl}/\text{in}^2$	Equivalence Factor	INSTRUMENT "ACCEPTANCE LIMIT"	
			$\mu\text{gNaCl}/\text{Cm}^2$	$\mu\text{gNaCl}/\text{in}^2$
MIL-P-28809-Beckman	7.47	$\frac{7.545}{7.545} = 1$	1.56	10.06
MIL-P-28809-Markson	7.62	$\frac{7.545}{7.545} = 1$	1.56	10.06
Omega Meter	10.51	$\frac{10.51}{7.545} = 1.39$	2.2	14
Ionograph	15.20	$\frac{15.20}{7.545} = 2.01$	3.1	20
Ion Chaser	24.50	$\frac{24.50}{7.545} = 3.25$	5.1	32

(A)

MRR No. 3-78  
712:WTH:RJD:mjr  
29 August 1978

It is hoped that the IPC Round Robin of Ionic Contaminant Testing will elucidate our findings and make plain a reasonable and lasting solution.

CONCLUSION:

All of the instruments tested are suitable for determining ionic contaminants remaining after solvent cleaning of post-soldering rosin-flux residues as described in MIL-P-28809 providing the interim equivalence factors are employed. **THIS CENTER IN NO WAY ENDORSES OR RECOMMENDS A PREFERRED INSTRUMENT. Each prospective buyer must evaluate and purchase what he feels to be most acceptable to him.**

REPORT PREPARED BY:

W. T. Hobson  
W. T. HOBSON, Materials Engineer  
R. J. DeNoon  
R. J. DENOON, Chemist

APPROVED BY:

R. D. Hott  
R. D. HOTT, Head, Metallurgical  
Materials Branch  
B. C. Vaughn  
B. C. VAUGHN, Director, Materials  
Laboratory and Consultants Division

MRR 3-78

APPENDIX

## APPENDIX

Attendees at the PWB Ionic Contaminants Meeting, Naval Avionics Center, 9-10 February 1978:\*

<u>NAME</u>	<u>AFFILIATION</u>	<u>TELEPHONE</u>
John W. McCormick	RADC	315-330-4029
Jack Brous	Alpha Metals 57 Freeman Street Newark, NJ 07105	201-589-2007
William G. Kenyon	E. I. DuPont Freon <sup>®</sup> Products Lab Wilmington, DE 19898	302-999-4167
Ed Wolfgram	Kenco 418 W. Belden Ave. Addison, IL 60101	312-543-2510
Richard E. Martz	NAC, D/714	317-353-3274
David O. Pond	NAC, D/714	317-353-3274
Michael Cowart	NAC, D/714	317-353-3274
William T. Hobson	NAC, D/712	317-353-3266
R. J. DeNoon	NAC, D/712	317-353-3266

### ADDITIONAL DISTRIBUTION:


Jerry Inman, Code 5541  
Naval Weapons Center  
China Lake, CA 93555

John Kerr, Code 50452  
NESC

700  
710  
711  
712  
713  
714

\*NOTE: All attendees listed are to receive one copy each.

MRR 3-78



4	AS ROD	MIL-F-14256	FLUX	
3	AS ROD	QQ-S-571	SOLDER	
2	108	40 C 501	CAPACITOR, SPRAGUE ELECTRIC CO.	
1	1	AV 21902	RESISTOR, CAPACITOR BD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.

DRAWN BY  
*J. Washon*

CAPACITOR ASSY. BOARD

PROJ. ENGR.  
*R.E. Martz*

SIZE  
A

CODE IDENT NO.  
02387

DRAWING NO.

AV 21923

A-2


DATE 8-3-77

SCALE ~

REV

SHEET 1 of 1

MRR 3-78



5	96	~	SOLUBLE MOUNTING PADS	
4	AS RQD	MIL-F-14256	FLUX	
3	AS RQD	QQ-S-571	SOLDER	
2	32	TO-5 CASE	TRANSISTOR	
1	1	AV 21906	TRANSISTOR BOARD	
ITEM	QTY.	PART No,	DESCRIPTION	REF.

DRAWN BY  
*J. Washon*

TRANSISTOR ASSY. BOARD

PROJ. ENGR  
*Richard E. May*

SIZE  
A

CODE IDENT NO.  
02387

DRAWING NO.

AV 21924


A-3

DATE 8-4-77

SCALE ~

REV


SHEET 1 of 1



4	AS ROD	MIL-F-14256	FLUX	
3	AS ROD	QQ-S-57	SOLDER	
2	108	RCR05G	RESISTOR, 1/8 WATT	
1	1	AV 21902	RESISTOR, CAPACITOR BOARD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.

DRAWN BY <i>J. Washon</i>		RESISTOR ASSEMBLY BOARD		
PROJ. ENGR. <i>R. E. Martz</i>	SIZE A	CODE IDENT NO. 02387	DRAWING NO. AV 22100	
DATE 8-8-77	SCALE	REV	SHEET 1 of 1	

A-4



7	AS ROD	MIL-F-14256	FLUX	
6	AS ROD	QQ-S-571	SOLDER	
5	8	6-32 FIL HD. SCREW, CRES 3/8 LONG		
4	8	6-32 HX. HD NUT, CRES		
3	4	TO-66 CASE	TRANSISTOR	
2	4	690-66-P	HEAT SINK, WAKEFIELD ENGINEERING INC	
1	1	AV 21903	POWER TRANSISTOR WITH HEAT SINK BOARD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.


DRAWN BY *J. Washin* POWER TRANSISTOR W/HT SK ASYBC

PROJ. ENGR. *RE Martz* SIZE A CODE IDENT NO. 02387 DRAWING NO. AV 21926

A-5

DATE 8-5-77 SCALE *~* REV SHEET 1 of 1

MRR 3-78



4	AS RQD	MIL-F-14256	FLUX	
3	AS RQD	QQ-S-571	SOLDER	
2	361	RCR 20G	RESISTOR, 1/2 WATT	
1	2	AV 21904	CORDWOOD PACK BD.	
ITEM	QTY.	PART No.	DESCRIPTION	REF.


DRAWN BY *J. Washon* CORDWOOD PACK ASSEMBLY BD.

PROJ. ENGR. *RE Marty* SIZE A CODE IDENT NO. 02387 DRAWING NO. AV 22102

A-6

DATE 8-9-77 SCALE \_\_\_\_\_ REV \_\_\_\_\_ SHEET 1 of 1

MRR 3-78



5	AS RQD	MIL-F-14256	FLUX	
4	AS RQD	QQ-S-571	SOLDER	
3	32	7717-86N	TRANSISTOR PADS, THERMALLOY COMPANY	
2	32	TO-5 CASE	TRANSISTOR	
1	1	AV 21906	TRANSISTOR BOARD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.


DRAWN BY *J. Hachow* TRANSISTOR W/PADS ASSY. BD.

PROJ. ENGR. *R.E. Martz* SIZE A CODE IDENT NO. 02387 DRAWING NO. AV 22103

A-7

DATE 8-9-77 SCALE REV SHEET 1 of 1


MRR 3-78



6	AS ROD	MILF-14256	FLUX	
5	AS ROD	00-S-571	SOLDER	
4	16	6-32 HEX HEAD NUT, CRES		
3	16	6-32 FIL HD. SCR., CRES $\frac{5}{8}$ LONG		
2	8	00-5007-018-163-001	18 PIN EDGE CARD CONN	ELCO CORP.
1	1	AV 21907	EDGE CARD CONNECTOR BOARD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.

DRAWN BY <i>J. Wash</i>		EDGE CARD CONNECTOR ASSY. BD.		
PROJ. ENGR. <i>RE Martz</i>	SIZE A	CODE IDENT NO. 02387	DRAWING NO. AV 22101	
DATE 8-9-77	SCALE	REV	SHEET 1 of 1	

A-8



4	AS RQD	MIL-F-14256	FLUX	
3	AS RQD	QQ-S-571	SOLDER	
2	50	MIL-M-38510C D-1 CASE	DIP, 14 LEAD, 1/4" X 3/4" CASE	
1	1	AV 21909	DIP BOARD	
ITEM	QTY.	PART No.	DESCRIPTION	REF.

DRAWN BY  
*J. Washon*

DIP ASSEMBLY BOARD

PROJ. ENGR  
*R.E. Martz*

SIZE  
A

CODE IDENT NO.  
02387

DRAWING NO.

AV 21925

A-9


DATE 8-3-77

SCALE ~

REV

SHEET 1 of 1

MRR 3-78




ITEM	QTY.	PART No.	DESCRIPTION
1	1	AV 21910	JUMPER WIRE BOARD
2	96	57A5A52-1	TERMINAL, STUD
3	48	200AS108-9	WIRE, 3 1/8" LONG, STRIPPED 3/8" BOTH ENDS
4	AS REQD	00-S-571	SOLDER
5	AS REQD	MIL-F-14256	FLUX

DRAWN BY <i>J. Washen</i>		JUMPER WIRE ASSY. BOARD		
PROJ. ENGR <i>R.E. Martz</i>	SIZE A	CODE IDENT NO. 02387	DRAWING NO. AV 21921	
DATE 8-2-77	SCALE	REV	SHEET 1 of 1	

A-10

BARNSTEAD HOSE NIPPLE  
 CARTRIDGE D 8902 ULTRA-PURE  
 (MIXED BED) MAXIMUM FLOW RATE  
 PER CARTRIDGE 12 GALLONS PER HOUR.

MRR 3-78





SYSTEM SOLUTION


75% 2-PROPANOL  
 25% DEIONIZED WATER  
 APPROXIMATELY 12 GALLONS


A-11

DRAWN BY <i>J. Washow</i>		CLEANING SYSTEM SCHEMATIC	
PROJ. ENGR. <i>R.E. Martz</i>	SIZE A	CODE IDENT NO. 02387	DRAWING NO. AV 21922
DATE 8-2-77	SCALE	REV	SHEET 1 of 1










µg NaCl/LITER



Contaminants Expressed as  
Micrograms NaCl Equivalents  
per in<sup>2</sup> as Found by Noted  
Test Procedures

Instrument	Factor	AV 21923 Capacitor (56.2 in. <sup>2</sup> , 363 cm <sup>2</sup> )	AV 21924 Transistor (65 in. <sup>2</sup> , 419 cm <sup>2</sup> )	AV 22100 Resistor (55 in. <sup>2</sup> , 355 cm <sup>2</sup> )	AV 21926 Power Trans- istor w/heat sink (157 in. <sup>2</sup> , 1010 cm <sup>2</sup> )	AV 22102 Cordwood Pack (162 in. <sup>2</sup> , 1050 cm <sup>2</sup> )	AV 22103 Transistor (72 in. <sup>2</sup> , 465 cm <sup>2</sup> )	AV 22101 Edge Card Connector (139 in. <sup>2</sup> , 897 cm <sup>2</sup> )	AV 21925 Dip 103 in. <sup>2</sup> (665 cm <sup>2</sup> )	AV 21921 Jumper Wire (66 in. <sup>2</sup> , 426 cm <sup>2</sup> )
MIL-P-28809 Beckman Conductivity Bridge RA Flux X	2	6.74	6.75	5.55	7.07	7.65	6.75	10.8	13.1	5.92
		6.30	5.40	6.07	5.61	6.15	6.07	12.1	12.8	5.92
		7.25	6.37	5.85	6.08	6.07	6.56	12.5	8.55	5.70
		6.76	6.17	5.82	6.25	6.63	6.46	11.8	11.5	5.85
MIL-P-28809 Markson Meter Bridge RA Flux X	4	6.44	6.22	4.65	6.30	8.35	6.44	12.3	13.8	6.21
		5.70	5.69	5.18	5.47	6.75	6.00	13.1	13.8	6.00
		7.26	6.49	5.18	5.77	6.75	8.10	13.5	8.85	5.70
		6.47	6.14	5.00	5.84	7.45	6.85	13.0	11.9	5.97
Kenco Omega Meter RA Flux X	6	10.9	10.2	10.5	7.76	8.40	10.2	11.6	14.5	10.2
		10.8	9.72	10.5	6.22	7.80	11.4	10.4	17.1	9.61
		11.5	9.95	9.87	8.21	8.56	10.7	11.6	16.5	9.00
		11.1	9.96	10.29	7.40	8.25	10.8	11.2	16.0	9.60
Alpha Iono- graph RA Flux X	8	13.8	11.6	12.6	7.89	--	14.8	18.5	26.1	16.1
		13.7	11.8	11.6	8.18	--	13.4	20.7	25.0	12.8
		15.4	13.4	12.9	6.91	--	15.2	24.1	23.1	15.9
		14.3	12.3	12.3	7.49	--	14.5	21.1	24.7	14.9
DuPont Ion Chaser RA Flux X	10	24.1	22.7	22.7	11.1	18.5	20.8	36.5	40.9	--
		25.2	22.2	20.3	14.1	16.0	22.1	35.8	50.0	22.3
		28.5	16.5	16.9	11.5	17.9	23.8	40.0	36.3	24.4
		25.9	20.5	20.0	12.2	17.5	22.2	37.4	42.4	22.35

$$\bar{X} = \left\{ \begin{array}{l} 7.55 \\ 7.55 \\ 7.55 \end{array} \right. = 1$$

$$\bar{X} = 10.5 \quad \frac{10.5}{7.55} = 1.39$$

$$\bar{X} = 15.20 \quad \frac{15.20}{7.55} = 2.01$$

$$\bar{X} = 24.5 \quad \frac{24.5}{7.55} = 3.25$$

TABLE I.

DATA POINTS

MRR 3-78

MIL-P-28809, Beckman Conductivity Bridge  
RA Flux  $\bar{X}^2$

45 - MIL-P-28809, Markson Meter Bridge  
RA Flux  $\bar{X}^4$

Kenco Omega Meter  
RA Flux  $\bar{X}^6$

40 - Alpha Ionograph  
RA Flux  $\bar{X}^8$

DuPont Ion Chaser  
RA Flux  $\bar{X}^{10}$

$\mu\text{g NaCl Equivalents}/\text{in}^2$

10.2 megohm-cm

5

