

2

THE USE OF NON-BARBITURATE BUFFERS IN COUNTERIMMUNOELECTROPHORESIS

AD A120036

E.A. EDWARDS
W.C. SUITER

REPORT NO. 82-11

SDTIC
ELECTED
OCT 7 1982
H



NAVAL HEALTH RESEARCH CENTER

P. O. BOX 85122
SAN DIEGO, CALIFORNIA 92138

NAVAL MEDICAL RESEARCH AND DEVELOPMENT COMMAND

82 10 07 049 BETHESDA, MARYLAND

This document has been approved
for public release and sale; its
distribution is unlimited.

DIG FILE COPY

**Best
Available
Copy**

2

The Use of Non-Barbiturate Buffers in Counterimmunoelectrophoresis

Earl A. Edwards

William C. Suiter

Biological Sciences Department

Naval Health Research Center

P. O. Box 85122

San Diego, CA 92138

Report No. 82-11, supported by Naval Medical Research and Development Command, Bethesda, Maryland, Department of the Navy, under research Work Unit M0095PN, M0095PN002-5044. The views presented in this paper are those of the authors. No endorsement by the Department of the Navy has been given or should be inferred.

SUMMARY

Counterimmunoelectrophoresis (CIE) has become a widely used test for the rapid identification of bacterial infections. It is less often used for identifying viral infections except for hepatitis B infections, in which CIE was the earliest test used for its detection. CIE testing has traditionally been performed using a barbital buffer as the electrolyte of choice. This is most likely due to the fact that barbital buffers were used for paper electrophoresis in serum protein studies for many years. The barbiturates used in these buffers have now become controlled substances due to their potential for drug abuse. The continued use of barbiturate buffers for electrophoresis then becomes an unsafe laboratory practice.

We report here the compositions of four non-barbiturate buffers, each appears to have the potential of replacing the barbital buffer in the CIE test. We used these buffers to rapidly detect 4 bacterial and one viral antigen in the CIE test. The sensitivity achieved with these non-barbital buffers with these five antigens was comparable to that achieved with the standard barbital buffer. Further evaluation of the use of non-barbiturate buffers in the CIE test are being conducted.

Accession For	
NTIS GRA&I	☒
DTIC TAB	☐
Unannounced	☐
Justification	
By	
Distribution/	
Availability Codes	
Dist	Avail and/or Special
A	



INTRODUCTION

Counterimmunoelectrophoresis (CIE) has traditionally been performed utilizing barbital buffers, a carryover from the serum protein electrophoresis technique used the past 4 decades. However, the barbiturates, with a potential for drug abuse, are now controlled chemicals. The use of barbital buffers, therefore, creates a potential laboratory hazard which non-barbital alternative(s) would eliminate. A recent description of two non-barbiturate buffers for electrophoresis of serum proteins on cellulose membrane suggested their usefulness for CIE testing (1).

This report compares CIE identification of 1 viral and 4 bacterial antigens using 4 non-barbiturate buffers and 1 barbiturate buffer.

MATERIALS AND METHODS

Agarose: Indubiose A45, lot #3162 (Fisher Scientific Co)

Antigens: Three spinal fluids from patients with cultures positive for *H. influenzae* type b, (supplied by Dr. John Sipple, Naval Biomedical Labs, Oakland); meningococcus group A antigen, purified polysaccharide (Lot #M1072, Merieux Institute, Lyon, France); pneumococcus antigen, (vaccine polysaccharide, Pneumocoque Tetradeccavalent Lot #5088-AOUT 80, Merieux Institute, Lyon, France) Streptococcus group A, (In-House produced extract by autoclaving (2)).

Antisera: Meningococcus group A (Lot #1-80, a gift of the Environmental Preventive Medicine Unit #5, San Diego); *H. influenzae* type b (kindly supplied by CDC, Atlanta); pneumococcus (OMNI anti-serum, Lot #49-79, Statens Seruminstitut, Copenhagen) and streptococcus group A (Lot #K8291, Wellcome Reagent Ltd., England).

Buffers: Buffer A, 13.0 gr Tris; 5.4 g Tricine; 0.424 gr calcium lactate; 0.8 g sodium azide per liter distilled water.

Buffer B. 8.0 g Tris; 3.75 g Tricine; 2.25 g sodium chloride, 0.75 g sodium salicylate per liter distilled water (1).

Buffer C. 7.5 g Tris; 10.0 gr sodium hippurate; 3.0 g hippuric acid per liter distilled water (1).

Buffer D. 8.0 g Tris; 3.75 Tricine g per liter distilled water.

Buffer E. Barbital Buffer (Lot #900M240 Kallestad Labs)

CIE Test: A 1% suspension of agarose was made in each of the 5 buffers. Twelve ml of each buffered suspension were pipetted onto separate lantern slides (80 x 100 mm) (3). The agarose was allowed to gel in a moist chamber at room temperature. Two parallel rows of wells, 3 mm in diameter, were cut 1.5 mm apart. Serial dilutions of antigen were placed in the wells nearest the cathode and undiluted antisera was placed in those wells nearest the anode. Electrophoresis with the appropriate buffers in the electrode vessels was for 60 minutes with 12-15 mA per slide. Examination for a precipitin reaction between the wells was made immediately after electrophoresis and after 24 hours.

RESULTS

The antigen-antibody precipitin reactions for the 5 antigens tested appeared similar in all 5 buffers and the clarity of the precipitin was comparable in all buffers. The reciprocal of the highest dilution of each antigen yielding a positive precipitin for each buffer system is shown in Table I. Except for pneumococcal antigen in buffer A and *H. influenzae* type b antigen in buffer C, the highest dilution of each antigen producing a positive test was similar for each of the non-barbital buffers and the barbital buffer.

TABLE I.

A Quantitative Study of Five Bacterial Antigens by CIE

Using Five Different Buffer Electrolytes

Buffers	Strep A*	ANTIGEN *			
		Pneumococcus	Meningococcus A	Adenovirus Type 4	Hemophilus Influenzae Type b
Buffer A	8 **	1024	512	8	8
Buffer B	16	4096	1024	8	4
Buffer C	16	4096	512	8	0
Buffer D	8	4096	1024	8	ND
Buffer E	8	4096	1024	8	8

* Antigens diluted $\log_2$ with physiological saline. Results represent the reciprocal of the highest dilution giving a precipitin reaction.

** Titer from 3 trial runs performed on different days.

ND - Not tested

DISCUSSION

Barbital buffers, long the standard for electrophoresis, are now controlled chemicals and their use in a laboratory creates a legal and administrative burden. Alternative buffers must demonstrate similar specificity and sensitivity.

This study compared results when 4 bacterial antigens and 1 viral antigen were identified by CIE using 4 non-barbital buffers and 1 barbital buffer. Non-barbital buffers B and D demonstrated comparable sensitivity to the barbital buffer for each antigen tested and non-barbital buffers A and C were comparable to the barbital buffer in 4 of 5 antigens tested.

The sensitivity achieved with these non-barbital buffers, in an admittedly small range of antigens, encourages their further evaluation. Buffers B and D particularly appear to be deserving of further testing. Studies with these buffers on a broader range of bacterial and viral antigens as well as on other parameters (pH, molarity, endosmosis, color distortion, storability, current stability, agarose concentration, time for the reaction to optimize, etc) are currently being conducted.

REFERENCES

1. Ambler, J. and Rodgers, M. Two new non-barbiturate buffers for electrophoresis of serum proteins on cellulose acetate membranes. J. Clin. Chem. 26:1221-1223, 1980.
2. Rantz, L.A. and F. Randall. Use of autoclaved extracts of Hemolytic streptococci for serological grouping. Stanford Med. Bull. 13:290-291, 1955.
3. Edwards, E.A. Immunologic Investigations of Meningococcal Disease. 1. Group specific Neisseria meningitidis antigens present in the serum of patients with fulminant meningococcemia. J. Immunol. 106:314-317, 1971.

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)


REPORT DOCUMENTATION PAGE		READ INSTRUCTIONS BEFORE COMPLETING FORM
1. REPORT NUMBER 82-11	2. GOVT ACCESSION NO. AD-820000	3. RECIPIENT'S CATALOG NUMBER
4. TITLE (and Subtitle) The Use of Non-Barbiturate Buffers in Counter-immunoelectrophoresis		5. TYPE OF REPORT & PERIOD COVERED FINAL
7. AUTHOR(s) E.A. Edwards W.C. Suiter		6. PERFORMING ORG. REPORT NUMBER
9. PERFORMING ORGANIZATION NAME AND ADDRESS Naval Health Research Center P.O. Box 85122 San Diego, CA 92138		8. CONTRACT OR GRANT NUMBER(s)
11. CONTROLLING OFFICE NAME AND ADDRESS Naval Medical Research & Development Command Bethesda, MD 20814		10. PROGRAM ELEMENT, PROJECT, TASK AREA & WORK UNIT NUMBERS M0095-PN.002-5044
14. MONITORING AGENCY NAME & ADDRESS (if different from Controlling Office) Bureau of Medicine & Surgery Department of the Navy Washington, DC 20372		12. REPORT DATE July/1982
		13. NUMBER OF PAGES 5
		15. SECURITY CLASS. (of this report) UNCLASSIFIED
		15a. DECLASSIFICATION DOWNGRADING SCHEDULE
16. DISTRIBUTION STATEMENT (of this Report) Approved for public release; distribution unlimited		
17. DISTRIBUTION STATEMENT (of the abstract entered in Block 20, if different from Report)		
18. SUPPLEMENTARY NOTES		
19. KEY WORDS (Continue on reverse side if necessary and identify by block number) Counterimmunoelectrophoresis Barbital buffers Microbial antigen Non-barbiturate buffers		
20. ABSTRACT (Continue on reverse side if necessary and identify by block number) Counterimmunoelectrophoresis (CIE) and cellulose electrophoresis have traditionally been run by using barbital buffers as the electrolyte of choice. We compared results of CIE identification of 1 viral and 4 bacterial antigens using four non-barbiturate buffers and a barbital buffer. One of the non-barbiturate buffers gave results equal to the barbital buffer for all 5 antigen systems tested. These data show that buffers other than barbital buffers can be used in the CIE test to identify bacterial and viral antigens.		

DD FORM 1473
1 JAN 73EDITION OF 1 NOV 65 IS OBSOLETE
S/N 0102 LF 014 6601

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE (When Data Entered)

SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)



SECURITY CLASSIFICATION OF THIS PAGE(When Data Entered)