

AD-A189 345

DEVELOPMENT AND EVALUATION OF INSTRUMENTATION AND
METHODS FOR MULTI-ELEMENT ANALYSIS(U) INDIANA UNIV AT
BLOOMINGTON DEPT OF CHEMISTRY G M HIEFTJE 13 JAN 88

1/1


UNCLASSIFIED

NO0014-76-C-0038

F/G 20/9

NL





AD-A189 345

ACTION PAGE

RESTRICTIVE MARKINGS

DTIC FILE COPY

1a REPORT SECURITY CLASSIFICATION

NA

2a SECURITY CLASSIFICATION AUTHORITY

NA

2b DECLASSIFICATION/DOWNGRADING SCHEDULE

NA

4 PERFORMING ORGANIZATION REPORT NUMBER(S)

3 DISTRIBUTION/AVAILABILITY OF REPORT

Distribution Unlimited; Approved for
Public Release

5 MONITORING ORGANIZATION REPORT NUMBER(S)

NA

6a NAME OF PERFORMING ORGANIZATION

Indiana University

6b OFFICE SYMBOL
(If applicable)

7a NAME OF MONITORING ORGANIZATION

ONR

6c ADDRESS (City, State, and ZIP Code)

Department of Chemistry
Bloomington, IN 47405

7b ADDRESS (City, State, and ZIP Code)

800 N. Quincy Street
Arlington, VA 222178a NAME OF FUNDING/SPONSORING
ORGANIZATION8b OFFICE SYMBOL
(If applicable)

9 PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER

Contract N00014-76-C-0838

8c ADDRESS (City, State, and ZIP Code)

10 SOURCE OF FUNDING NUMBERS

PROGRAM
ELEMENT NOPROJECT
NOTASK
NOWORK UNIT
ACCESSION NO

NR 051-622

11 TITLE (Include Security Classification)

Development and Evaluation of Instrumentation
and Methods for Multi-Element Analysis

12 PERSONAL AUTHOR(S)

Gary M. Hieftje

13a TYPE OF REPORT
FINAL13b TIME COVERED
FROM 5/1/76 TO 4/30/8614 DATE OF REPORT (Year, Month, Day)
1988 January 1315 PAGE COUNT
21

16 SUPPLEMENTARY NOTATION

17 COSATI CODES

FIELD

GROUP

SUB-GROUP

18 SUBJECT TERMS (Continue on reverse if necessary and identify by block number)

19 ABSTRACT (Continue on reverse if necessary and identify by block number)

FINAL report on Contract N00014-76-C-0838

DTIC
ELECTE
JAN 27 1988
S H D

20 DISTRIBUTION/AVAILABILITY OF ABSTRACT

[X] UNCLASSIFIED UNLIMITED [] SAME AS REPORT [] FOR USE ONLY

21 ABSTRACT SECURITY CLASSIFICATION

Distribution Unlimited

22a NAME OF RESPONSIBLE INDIVIDUAL

Gary M. Hieftje

22b TELEPHONE (Include Area Code)
(812) 335-2189

22c OFFICE SYMBOL

OFFICE OF NAVAL RESEARCH

FINAL REPORT

for

1 May 1976 through 30 April 1986

for

Contract N00014-76-C-0838

Task No. NR 051-622

DEVELOPMENT AND EVALUATION OF INSTRUMENTATION
AND METHODS FOR MULTI-ELEMENT ANALYSIS

Gary M. Hieftje

Department of Chemistry
Indiana University
Bloomington, IN 47405



Reproduction in whole, or in part, is permitted for any purpose of the United States Government.

* This document has been approved for public release and sale; its distribution is unlimited.

ion For	
GRA&I	☒
AR	☐
anced	☐
ocation	
By	
Distribution/	
Availability Codes	
Avail and/or	
Dist	Special
A-1	

Technical reports submitted to ONR during contract term:

<u>Technical Report #</u>	<u>Title</u>
1	A Selective Spectral-Line Modulation Technique for High Sensitivity Continuum Source Atomic Absorption Spectrometry
2	Automatic Background-Correcting Spectrometer
3	Saturation of Energy Levels in Analytical Atomic Fluorescence Spectrometry - I. Theory
4	Spectral and Noise Characteristics of a 300-W Eimac Arc Lamp
5	A Modified Microwave-Induced Plasma (MIP) Discharge Chamber Exhibiting High Stability and Immunity from Sample Solvent Extinguishment
6	Physical Characterization of a New Atom Cell for Atomic Spectrometry: The Helium-Oxygen-Acetylene Flame
7	Report on Combined Meeting of the XX Colloquium Spectroscopicum Internationale and 7th International Conference on Atomic Spectroscopy
8	Simplified Apparatus for Routine Selective Spectral-Line Modulation Atomic Absorption Spectrometry
9	Enhancement of Pneumatic Nebulization Efficiency Through Application of an Electric Field
10	A New, Accurate Method for the Measurement of Rise Velocities in Laminar Flames
11	Atomic Emission Spectrometry of Trace Metals Using a New Kind of Microwave-Induced Helium Plasma at Atmospheric Pressure
12	Biennial Review on Flame Emission, Atomic Absorption, and Atomic Fluorescence Spectrometry for <i>Analytical Chemistry</i>
13	Development and Characterization of a Miniature Inductively Coupled Plasma Source for Atomic Emission Spectrometry
14	Investigation into the Operating Characteristics of a Microarc Atmospheric-Pressure Glow Discharge
15	Microwave-Supported Discharges

- 16 Theoretical Study of the Spatial Distribution of Atoms Surrounding an Individual Solute Particle Vaporizing in an Analytical Flame
- 17 Measurement and Comparison of Relative Free-Atom Fractions in the Helium-Oxygen-Acetylene Flames
- 18 Comparison of Atomic Fluorescence Power Efficiencies for the Helium-Oxygen-Acetylene and Air-Acetylene Flames
- 19 Wide-Bandwidth Analog Correlator and Its Application to Mode-Locked Laser Measurements
- 20 Experimental Studies on Spatial Distributions of Atoms Surrounding an Individual Solute Particle Vaporizing in an Analytical Flame
- 21 Hydrodynamic Flow Patterns as a Simple Aid to Effective ICP Torch Design
- 22 Characteristics of the Background Emission Spectrums from a Miniature ICP
- 23 Evaluation of Classical Vaporization and Ionization Interferences in a Miniature ICP
- 24 An Accurate Model for Sample Droplet Acceleration in Analytical Flames
- 25 Correction of Quenching Errors in Analytical Fluorimetry Through Use of Time Resolution
- 26 A Linear Response Theory Approach to Time-Resolved Fluorimetry
- 27 Reduction of Spectral Interferences in Flame Emission Spectrometry by Selective Spectral-Line Modulation
- 28 Development and Characterization of a 9-mm Inductively Coupled Argon Plasma (ICP) Source for Atomic Emission Spectrometry
- 29 Effects of Gas Composition and Flame Sheathing on the Spatial Velocity Profiles of Laminar Analytical Acetylene Flames
- 30 Microwave-Supported Discharges
- 31 An Approximate Model for the Liberation and Ionization of Atoms from Individual Solute Particles in Flame Spectrometry

- 32 A Novel Device for the Accurate Dispensing of Small
 Volumes of Liquid Samples
- 33 Power Stabilization Method for a Synchronously Pumped
 Dye Laser System
- 34 Saturation of Energy Levels in Analytical Atomic
 Fluorescence Spectrometry II. Experimental
- 35 Correlation-Based Approaches to Time-Resolved
 Fluorimetry
- 36 Correlation Methods in the Chemistry Laboratory
- 37 Simple and Inexpensive Design for an Isolated Droplet
 Generator Useful in Studies of Atomization in Flames
- 38 A New Method for Measuring Rates of Photosensitive
 Reactions
- 39 A New Instrument for Time-Resolved Reduction of
 Scattered Radiation in Fluorescence Measurements
- 40 On the Significance of Radiation Trapping in the
 Inductively Coupled Plasma
- 41 Determination of Atomic and Molecular Excited-State
 Lifetimes Using an Opto-Electronic Cross-Correlation
 Method
- 42 Advances in Sample Introduction for Elemental
 Analysis
- 43 A New, Directly Computer-Controlled pH Stat
- 44 Microdroplet Sample Application in Electrothermal
 Atomization for Atomic Absorption Spectrometry
- 45 Reduction of Spectral Interferences in Inductively
 Coupled Plasma-Atomic Emission Spectrometry by
 Selective Spectral-Line Modulation
- 46 Design and Construction of a Low-Flow, Low-Power
 Torch for Inductively Coupled Plasma Spectrometry
- 47 The Use of a Microarc Atomizer for Sample
 Introduction into an Inductively Coupled Plasma
- 48 Replacement Ion Chromatography with Flame Photometric
 Detection. I. Cations
- 49 Approaching the Limit in Atomic Spectrochemical
 Analysis

- 50 A Separated Impedance Matcher/Load Coil Assembly for
Convenient Spatial Translation of an ICP Torch
- 51 A New Background-Correction Method for Atomic
Absorption Spectrometry
- 52 Signal-to-Noise Considerations in Fluctuation
Analysis Spectroscopic Techniques (FAST)
- 53 Near Infrared Reflectance Analysis by Gauss-Jordan
Linear Algebra
- 54 Modification of a Plasma-Therm Inductively Coupled
Plasma Supply to Enable RF Power Modulation
- 55 Sample Introduction into the Inductively Coupled
Plasma by a Radio-Frequency Arc
- 56 New Techniques and Tools for Clinical Chemistry
- 57 Jet-Impact Nebulization for Sample Introduction in
Inductively Coupled Plasma Spectrometry
- 58 A Detailed Consideration of Resonance Radiation
Trapping in the Argon Inductively Coupled Plasma
- 59 A Comparison of Signal-to-Noise Ratios for Near-
Infrared Detectors
- 60 Mini, Micro, and High-Efficiency Torches
- 61 A Detailed Consideration of Resonance Radiation
Trapping in the Argon Inductively Coupled Plasma
- 62 A New Method for Obtaining Individual Component
Spectra from Those of Complex Mixtures
- 63 Vaporization in the Absence of Ionization of Alkali
Chloride Particles in a Laminar Air-Acetylene Flame
- 64 Development of a Microwave Induced Nitrogen Discharge
at Atmospheric Pressure (MINDAP)
- 65 Analytical Characteristics of the Microwave Induced
Nitrogen Discharge at Atmospheric Pressure (MINDAP)
- 66 Identification of Limiting Noise Sources in The
Microwave Induced Nitrogen Discharge at Atmospheric
Pressure
- 67 Microsample Introduction into the MINDAP Using a
Microarc Atomizer

- 68 A Steady-State Approach to Excitation Mechanisms in
 the ICP
- 69 Interferences in a Low-Flow, Low-Power Inductively
 Coupled Plasma
- 70 Number of Samples and Wavelengths Required for the
 Training Set in Near-Infrared Reflectance
 Spectroscopy
- 71 Micro-Droplet Titration Apparatus for Analyzing
 Small Sample Volumes
- 72 Comparison of Atomic Fluorescence Power Efficiencies
 for the Helium-Oxygen-Acetylene and Air-Acetylene
 Flames -- Resubmission of Technical Report #18
- 73 Burnt-Gas Composition of the Helium-Oxygen-Acetylene
 Flame
- 74 Analytical Characteristics of Near-Infrared Nonmetal
 Atomic Emission from a Helium Microwave-Induced
 Plasma
- 75 Near-Infrared Nonmetal Atomic Emission from a Helium
 Microwave-Induced Plasma: Element Ratio
 Determinations
- 76 Interferometric Detection of Near-Infrared Nonmetal
 Atomic Emission from a Microwave-Induced Plasma
- 77 Optimization of Sample Absorbance for Quantitative
 Analysis in the Presence of Pathlength Error in the
 IR and NIR Regions
- 78 An Optimization of Detector-Noise-Limited Infrared
 Multi-Wavelength Determinations
- 79 Unique-Sample Selection via Near-Infrared Spectral
 Subtraction
- 80 Near-Infrared Determination of Some Physical
 Properties of Hydrocarbons
- 81 Analytical Characteristics of a Low-Flow, Low-Power
 Inductively Coupled Plasma
- 82 The Two Sides of Analytical Chemistry
- 83 Analytical Characteristics of an Optimized Miniature
 Inductively Coupled Plasma Source
- 84 Optical Imaging Spectrometers

- 85 An Apparatus for Thomson Scattering in an Inductively Coupled Plasma
- 86 A Comparison of the Helium-Oxygen-Acetylene and Air-Acetylene Flames as Atom Sources for Continuum-Source Atomic Fluorescence
- 87 The Use of Sample Additives in Flame Emission Spectroscopy
- 88 Near-Infrared Spectrophotometric Methods Development with a Limited Number of Samples. Application to Carbonate in Geological Samples
- 89 Modification of an ICP Radio-Frequency Supply for Amplitude Modulation with Complex Waveforms
- 90 Energy Transport and Analyte Excitation in the ICP
- 91 Novel Techniques for the Determination of Fluorescence Lifetimes
- 92 Fundamental Studies in the ICP Using a Sinusoidally Modulated Power Input
- 93 Time-Resolved Fluorescence Using an Optical Fiber Probe
- 94 Thomson Scattering from an ICP
- 95 Electron Temperatures and Electron Number Densities Measured by Thomson Scattering in the ICP
- 96 A Fiber-Optic Absorption Cell for Remote Determination of Copper in Industrial Electroplating Baths
- 97 A Modulated Sample Introduction System for Atomic Emission Spectrometry
- 98 Twenty-five Years of Analytical Atomic Spectrometry
- 99 A Steady-State Approach to Evaluation of Proposed Excitation Mechanisms in the Analytical ICP
- 100 Use of Time-Resolution to Eliminate Bilirubin Interference in the Determination of Fluorescein
- 101 A Computerized Tool for Studying Two-Dimensional Images of Spectroscopic Sources
- 102 A Novel Light-Scattering Technique to Study Gas-Flow Dynamics in an ICP Torch

- 103 Determination of Subnanosecond Fluorescence Lifetimes
 with a UHF Television Tuner and a CW or Mode-Locked
 Laser
- 104 An Improved Uniform Size Droplet Generator
- 105 A New Technique for the Elimination of Fluorescence
 Interference in Raman Spectroscopy
- 106 Temporally and Spatially Resolved Studies in an
 Amplitude Modulated ICP
- 107 Taming the Surfatron
- 108 Analytical Characteristics of a New, Versatile
 Inductively Coupled Plasma-Mass Spectrometer
- 109 Morphological Investigation of Solute Particles
 Trapped from an Air-Acetylene Flame
- 110 The Use of Sample Modulation to Simplify the
 Experimental Arrangement of Selective Line Modulation
 in Atomic Emission Spectrometry
- 111 A Detailed Study of Noise Sources in Selective Line
 Modulation Atomic Spectrometry
- 112 A Modulated Sample-Introduction System for ICP-Atomic
 Emission Spectrometry

List of Publications Emanating from Contract

1. "Automatic Background-Correcting Spectrometer", Anal. Chem. 48, 2030-3 (1976) (with R. J. Sydor and J. R. Sinnamon).
2. "A Selective Spectral-Line Modulation Technique for High Sensitivity Continuum-Source Atomic Absorption Spectrometry", Anal. Chem. 49, 98-105 (1977) (with R. L. Cochran).
3. "Influence of Droplet Desolvation, Particle Vaporization and Atomic Diffusion on the Distribution of Atoms in Analytical Flames and Plasmas", Bulletin of the Technical University, Prague, Czechoslovakia, 113-23 (1977) (with C. B. Boss).
4. "Physical Characterization of the Helium-Oxygen-Acetylene Flame as an Atom Cell for Atomic Spectrometry", Anal. Chem. 49, 2013-8 (1977) (with K. A. Saturday).
5. "Spectral and Noise Characteristics of a 300-W Eimac Arc Lamp", Anal. Chem. 49, 2040-3 (1977) (with R. N. Cochran).
6. "Saturation of Energy Levels in Analytical Atomic Fluorescence Spectrometry--I. Theory", Spectrochim. Acta 33B, 79-99 (1978) (with D. R. Olivares).
7. "Determination of Trace Metals by Microwave Plasma Spectrometry with an Atmospheric Pressure Helium Discharge", Anal. Chem. 50, 1257-60 (1978) (with A. T. Zander).
8. "A New, Accurate Method for the Measurement of Rise Velocities in Laminar Flames", Appl. Spectrosc. 32, 377-80 (1978) (with C. B. Boss).
9. "Enhancement of Pneumatic Nebulization Efficiency Through Application of an Electric Field", Rev. Sci. Instrum. 49, (10) 1418-24 (1978) (with R. N. Savage).
10. "Development and Characterization of a Miniature Inductively Coupled Plasma Source for Atomic Emission Spectrometry", Anal. Chem. 51, 408-13 (1979) (with R. N. Savage).
11. "Investigation into the Operating Characteristics of a 'Microarc' Atmospheric-Pressure Glow Discharge", Appl. Spectrosc. 33, (3) 230-40 (1979) (with R. I. Bystroff, and L. R. Layman).
12. "Theoretical Study of the Spatial Distribution of Atoms Surrounding an Individual Solute Particle Vaporizing in an Analytical Flame", Anal. Chem. 51, 895-901 (1979) (with C. B. Boss).
13. "Wide-Bandwidth Analog Correlator and Its Application to Mode-Locked Laser Measurements", Rev. Sci. Instrum. 50, (8) 997-1001 (1979) (with J. M. Ramsey and G. R. Haugen).

14. "Hydrodynamic Flow Patterns as a Simple Aid to Effective ICP Torch Design", Appl. Spectrosc. 33, (6) 643-6 (1979) (with E. Sexton and R. N. Savage).
15. "Vaporization and Ionization Interferences in a Miniature Inductively Coupled Plasma", Anal. Chem. 52, 1267-72 (1980) (with R. N. Savage).
16. "Characteristics of the Background Emission Spectrum from a Miniature ICP", Anal. Chim. Acta 123, 319-24 (1981) (with R. N. Savage).
17. "Experimental Studies on Spatial Distributions of Atoms Surrounding an Individual Solute Particle Vaporizing in an Analytical Flame", Anal. Chem. 51, 1897-1905 (1979) (with C. B. Boss).
18. "Comparison of Atomic Fluorescence Power Efficiencies for the Helium-Oxygen-Acetylene and Air-Acetylene Flames", Anal. Chim. Acta 164, 51-66 (1984) (with K. A. Saturday and A. Yuen).
19. "Measurement and Comparison of Relative Free-Atom Fractions in the Helium-Oxygen-Acetylene Flame and Air-Acetylene Flames", Anal. Chem. 52, 786-92 (1980) (with K. A. Saturday).
20. "An Accurate Model for Sample Droplet Acceleration in an Air-Acetylene Flame", Anal. Chim. Acta 118, 293-9 (1980) (with R. E. Russo).
21. "Correction of Quenching Errors in Analytical Fluorimetry Through Use of Time Resolution", Anal. Chim. Acta 123, 255-61 (1981) (with G. R. Haugen).
22. "Reduction of Spectral Interferences in Flame Emission Spectrometry by Selective Spectral-Line Modulation", Anal. Chim. Acta 121, 165-74 (1980) (with S. W. Downey and J. G. Shabushnig).
23. "Effects of Gas Composition and Flame Sheathing on the Spatial Velocity Profiles of Laminar Analytical Acetylene Flames", Spectrochim. Acta 36B, 231-41 (1981) (with R. E. Russo).
24. "Development and Characterization of a 9-mm Inductively Coupled Argon Plasma (ICP) Source for Atomic Emission Spectrometry", Anal. Chim. Acta 124, 245-58 (1981) (with A. D. Weiss and R. N. Savage).
25. "An Approximate Model for the Liberation and Ionization of Atoms from Individual Solute Particles in Flame Spectrometry", Spectrochim. Acta 36B, 205-13 (1981) (with B. D. Bleasdel and E. P. Wittig).
26. "Microwave-Supported Discharges", Appl. Spectrosc. 35, 357-71 (1981) (invited review paper with A. T. Zander).
27. "Power Stabilization Method for a Synchronously Pumped Dye Laser System", Rev. Sci. Instrum. 52, 172-3 (1981) (with R. E. Russo and R. Withnell).

28. "A Novel Device for the Accurate Dispensing of Small Volumes of Liquid Samples", Anal. Chim. Acta 126, 167-74 (1981) (with J. G. Shabushnig).
29. "Correlation Methods in the Chemistry Laboratory", American Laboratory 13, (3) 76-83 (1981) (with G. Horlick).
30. "Correlation-Based Approaches to Time-Resolved Fluorimetry", Anal. Chem. 53, 755-66A (1981) (with G. R. Haugen).
31. "Saturation of Energy Levels in Analytical Atomic Fluorescence Spectrometry--II. Experimental", Spectrochim. Acta 36B, 1059-79 (1981) (with Dorys Rojas de Olivares).
32. "Simple and Inexpensive Design for an Isolated Droplet Generator Useful in Studies of Atomization in Flames", Appl. Spectrosc. 35, 531-6 (1981) (with R. E. Russo and R. Withnell).
33. "Stochastic Photolysis: A New Method for Measuring Rates of Photo-sensitive Reactions", Appl. Spectrosc. 36, 203-9 (1982) (with G. R. Haugen, L. L. Steinmetz, and R. E. Russo).
34. "A New Instrument for Time-Resolved Reduction of Scattered Radiation in Fluorescence Measurements", Anal. Chim. Acta 134, 13-9 (1982) (with R. E. Russo).
35. "On the Significance of Radiation Trapping in the Inductively Coupled Plasma", Spectrochim. Acta 37B, 191-7 (1982) (with M. W. Blades).
36. "Determination of Atomic and Molecular Excited-State Lifetimes Using an Opto-Electronic Cross-Correlation Method", Appl. Spectrosc. 36, 92-9 (1982) (with R. E. Russo).
37. "Advances in Sample Introduction for Elemental Analysis", Industrial Chemical News 3, (1) 1 (1982).
38. "A New, Directly Computer-Controlled pH Stat", Anal. Chim. Acta 141, 173-86 (1982) (with R. E. Lemke).
39. "Microdroplet Sample Application in Electrothermal Atomization for Atomic Absorption Spectrometry", Anal. Chim. Acta 148, 181-92 (1983) (with J. G. Shabushnig).
40. "Reduction of Spectral Interferences in Inductively Coupled Plasma-Atomic Emission Spectrometry by Selective Spectral-Line Modulation", Anal. Chim. Acta 141, 193-205 (1982) (with S. W. Downey).
41. "Approaching the Limit in Atomic Spectrochemical Analysis", J. Chem. Educ. 50, 900-9 (1982).
42. "A Separated Impedance Matcher/Load Coil Assembly for Convenient Spatial Translation of an ICP Torch", Appl. Spectrosc. 36, 689-91 (1982) (with J. W. Carr and M. W. Blades).

43. "Modification of a Plasma-Therm Inductively Coupled Plasma Supply to Enable RF Power Modulation", Appl. Spectrosc. 37, 571-573 (1983) (with R. E. Ensman and J. W. Carr).
44. "Design and Construction of a Low-Flow, Low-Power Torch for Inductively Coupled Plasma Spectrometry", Appl. Spectrosc. 36, 627-31 (1982) (with R. Rezaaiyaan, H. Anderson, H. Kaiser, and B. Meddings).
45. "Signal-to-Noise Considerations in Fluctuation Analysis Spectroscopic Techniques", in New Directions in Molecular Luminescence, D. Eastwood, ed., ASTM Publication 822, Philadelphia, 1983, pp. 82-100 (with J. M. Ramsey).
46. "The Use of a Microarc Atomizer for Sample Introduction into an Inductively Coupled Plasma", Appl. Spectrosc. 37, 101-5 (1983) (with J. P. Keilsohn and R. D. Deutsch).
47. "Replacement Ion Chromatography with Flame Photometric Detection", Anal. Chim. Acta 153, 1-13 (1983) (with S. W. Downey).
48. "Near Infrared Reflectance Analysis by Gauss-Jordan Linear Algebra", Appl. Spectrosc. 37, 491-497 (1983) (with D. E. Honigs, J. M. Freelin, and T. Hirschfeld).
49. "Sample Introduction into the ICP by a Radio-Frequency Arc", Anal. Chem. 55, 1414-17 (1983) (with P. B. Farnsworth).
50. "A New Method for Obtaining Individual Component Spectra from Those of Complex Mixtures", Appl. Spectrosc. 38, 317-322 (1984) (with D. E. Honigs and T. Hirschfeld).
51. "A Detailed Consideration of Resonance Radiation Trapping in the Argon Inductively Coupled Plasma", Spectrochim. Acta, Part B 39, 859-866 (1984) (with J. W. Mills).
52. "Jet-Impact Nebulization for Sample Introduction in Inductively Coupled Plasma Spectrometry", Appl. Spectrosc. 38, 405-412 (1984) (with M. P. Doherty).
53. "New Techniques and Tools for Clinical Chemistry", Clin. Chem. 29, (9) 1659-1664 (1983).
54. "Mini, Micro, and High-Efficiency Torches for the ICP -- Toys or Tools?", Spectrochim. Acta 38B, 1465-1481 (1983).
55. "A Comparison of Signal-to-Noise Ratios for Near-Infrared Detectors", Appl. Spectrosc. 38, 837-843 (1984) (with J. E. Freeman).
56. "Vaporization in the Absence of Ionization of Alkali Chloride Particles in a Laminar Air-Acetylene Flame", Spectrochim. Acta, 40B, 209-216 (1985) (with Y. Pak).

57. "Development of a Microwave Induced Nitrogen Discharge at Atmospheric Pressure (MINDAP)", Appl. Spectrosc. 39, 214-222 (1985) (with R. D. Deutsch).
58. "Analytical Characteristics of the Microwave Induced Nitrogen Discharge at Atmospheric Pressure (MINDAP)", Appl. Spectrosc. 39, 531-534 (1985) (with R. D. Deutsch and J. P. Keilsohn).
59. "Identification of Limiting Noise Sources in the Microwave Induced Nitrogen Discharge at Atmospheric Pressure", Appl. Spectrosc. 39, 19-24 (1985) (with R. D. Deutsch).
60. "Microsample Introduction into the MINDAP Using a Microarc Atomizer", Anal. Chem. 56, 1923-1927 (1984) (with R. D. Deutsch).
61. "A Steady-state Approach to Excitation Mechanisms in the ICP", Spectrochim. Acta 40B, 167-176 (1985) (with G. D. Rayson and J. W. Olesik).
62. "Interferences in a Low-Flow, Low-Power Inductively Coupled Plasma", Spectrochim. Acta 40B, 73-83 (1985) (with R. Rezaaiyaan and J. W. Olesik).
63. "Number of Samples and Wavelengths required for the Training set in Near-Infrared Reflectance Spectroscopy", Appl. Spectrosc. 38, 844-847 (1984) (with D. E. Honigs and T. Hirschfeld).
64. "Micro-Droplet Titration Apparatus for Analyzing Small Sample Volumes", Anal. Chem. 56, 2884-2888 (1984) (with A. W. Steele).
65. "Burnt-Gas Composition of the Helium-Oxygen-Acetylene Flame", Anal. Chim. Acta 162, 403-407 (1984) (with A. Yuen).
66. "Analytical Characteristics of Near-Infrared Nonmetal Atomic Emission from a Helium Microwave-Induced Plasma", Spectrochim. Acta 40B, 475-492 (1985) (with J. E. Freeman).
67. "Near-Infrared Nonmetal Atomic Emission from a Helium Microwave-Induced Plasma: Element Ratio Determinations", Spectrochim. Acta, 40B, 653-664 (1985) (with J. E. Freeman).
68. "Interferometric Detection of Near-Infrared Nonmetal Atomic Emission from a Microwave-Induced Plasma", Appl. Spectrosc. 39, 211-214 (1985) (with J. E. Freeman).
69. "Optimization of Sample Absorbance for Quantitative Analysis in the Presence of Pathlength Error in the IR and NIR Regions" Appl. Spectrosc. 39, 430-433 (1985) (with D. Honigs and T. Hirschfeld).
70. "An Optimization of Detector-Noise-Limited Infrared Multi-Wavelength Determinations" Appl. Spectrosc. 39, 253-256 (1985) (with T. Hirschfeld and D. Honigs).

71. "Unique-Sample Selection via Near-Infrared Spectral Subtraction" Anal. Chem. 57, 2299-2303 (1985) (with D. E. Honigs, H. Mark, and T. B. Hirschfeld).
72. "Near-Infrared Determination of Several Physical Properties of Hydrocarbons", Anal. Chem. 57, 443-445 (1985) (with D. E. Honigs and T. B. Hirschfeld).
73. "Analytical Characteristics of a Low-Flow, Low-Power Inductively Coupled Plasma", Anal. Chem. 57, 412-415 (1985) (with R. Rezaaiyaan).
74. "The Two Sides of Analytical Chemistry", Anal. Chem. 57, 256A-267A (1985).
75. "Analytical Characteristics of an Optimized Miniature Inductively Coupled Plasma Source for Atomic Emission Spectrometry", Anal. Chim. Acta 173, 63-75 (1985) (with R. Rezaaiyaan).
76. "Optical Imaging Spectrometers", Anal. Chem. 57, 2049-2055 (1985) (with J. W. Olesik).
77. "An Apparatus for Thomson Scattering in an Inductively Coupled Plasma", Spectrochim. Acta 40B, 1211-1217 (1985) (with M. Huang and K. A. Marshall).
78. "Comparison of the Helium-Oxygen-Acetylene and Air-Acetylene Flames as Atom Sources for Continuum-Source Atomic Fluorescence Spectrometry", Anal. Chim. Acta 171, 241-249 (1985) (with D. A. Wilson and A. Yuen).
79. "The Use of Sample Additives in Flame Emission Spectroscopy", Anal. Chim. Acta 181 195-201 (1986) (with R. Rezaaiyaan and T. Hirschfeld).
80. "Near-Infrared Spectrophotometric Methods Development with a Limited Number of Samples. Application to Carbonate in Geological Samples", Appl. Spectrosc. 39, 1062-1065 (1985) (with D. E. Honigs and T. Hirschfeld).
81. "Modification of an Inductively Coupled Plasma Radio-Frequency Supply for Amplitude Modulation with Complex Waveforms", Anal. Chem. 57, 2414-2417 (1985) (with R. Withnell, G. D. Rayson, and A. F. Parisi).
82. "Energy Transport and Analyte Excitation in the ICP", in Analytical Chemistry in the Exploration, Mining and Processing of Materials, L.R.P. Butler, ed., Blackwell Scientific Publications, 1986 (with J. W. Mills, J. W. Carr, G. D. Rayson, M. Huang, and K. A. Marshall).
83. "Novel Techniques for the Determination of Fluorescence Lifetimes", Anal. Instrum. 14, 483-502 (1985) (with D. A. Wilson, and G. H. Vickers).
84. "Fundamental Studies in the ICP Using a Sinusoidally Modulated Power Input", Appl. Spectrosc. 40, 181-185 (1986) (with A. F. Parisi).

85. "Time-Resolved Fluorescence With an Optical-Fiber Probe", Anal. Chim. Acta 192, 145-153 (1987) (with G. H. Vickers and R. M. Miller).
86. "Thomson Scattering from An ICP", Spectrochim. Acta 40B, 1387-1400 (1985) (with M. Huang).
87. "Electron Temperatures and Electron Number Densities Measured by Thomson Scattering in the Inductively Coupled Plasma", Anal. Chem. 58, 207-210 (1986) (with M. Huang and K. A. Marshall).
88. "A Fiber-Optic Absorption Cell for Remote Determination of Copper in Industrial Electroplating Baths", Anal. Chim. Acta 177, 121-128 (1985) (with J. Freeman, A. Childers, and A. Steele).
89. "A Modulated Sample Introduction System for Atomic Emission Spectrometry", Appl. Spectrosc. 40, 357-363 (1986) (with A. Steele).
90. "Twenty-five Years of Analytical Atomic Spectrometry", Anal. Chim. Acta 180, 69-98 (1986) (with S. Greenfield, N. Omenetto, A. Scheeline, and W. Slavin).
91. "A Steady-State Approach to Evaluation of Proposed Excitation Mechanisms in the Analytical ICP", Spectrochim. Acta, 41B, 683-697 (1986) (with G. D. Rayson).
92. "Use of Time-Resolution to Eliminate Bilirubin Interference in the Determination of Fluorescein", Anal. Chem. 58, 1225-1227 (1986) (with F. V. Bright and G. H. Vickers).
93. "A Computerized Tool for Studying Two-Dimensional Images of Spectroscopic Sources", Computer Enhanced Spectrosc. 3, 91-94 (1986) (with B. R. Hites, J. W. Olesik, and A. F. Parisi).
94. "A Novel Light-Scattering Technique to Study Gas-Flow Dynamics in an ICP Torch", Appl. Spectrosc. 40, 680-683 (1986) (with A. G. Childers).
95. "Determination of Subnanosecond Fluorescence Lifetimes with a UHF Television Tuner and a CW or Mode-Locked Laser", Appl. Spectrosc. 40, 683-687 (1986) (with F. V. Bright and D. A. Wilson).
96. "An Improved Uniform Size Droplet Generator", Appl. Spectrosc. 40, 688-691 (1986) (with A. G. Childers).
97. "Temporally and Spatially Resolved Studies in an Amplitude Modulated Inductively Coupled Plasma", Spectrochim. Acta 42B, 361-376 (1987) (with A. F. Parisi, G. D. Rayson, and J. W. Olesik).
98. "Taming the Surfatron", Spectrochim. Acta 42B, 285-298 (1987) (with M. Selby).
99. "A New Technique for the Elimination of Fluorescence Interference in Raman Spectroscopy", Appl. Spectrosc. 40, 583-587 (1986) (with F. V. Bright).

100. "Analytical Characteristics of an Inductively Coupled Plasma-Mass Spectrometer", *Spectrochim. Acta* 42B, 29-38 (1987) (with D. A. Wilson, G. H. Vickers, and A. T. Zander).
101. "Morphological Investigation of Solute Particles Trapped from an Air Acetylene Flame", *Appl. Spectrosc.* 40, 939-944 (1986) (with A. G. Childers).
102. "The Use of Sample Modulation to Simplify the Experimental Arrangement of Selective Line Modulation in Atomic Emission Spectrometry", *Appl. Spectrosc.* 40, 1110-1117 (1986) (with A. W. Steele).
103. "A Detailed Study of Noise Sources in Selective Line Modulation Atomic Spectrometry", *Appl. Spectrosc.* 40, 1117-1126 (1986) (with A. W. Steele).
104. "A Modulated Sample-Introduction System for ICP-Atomic Emission Spectrometry", *Appl. Spectrosc.* 40, 1127-1131 (1986) (with A. W. Steele).
105. "Torches for Inductively Coupled Plasmas" in Inductively Coupled Plasma Emission Spectrometry--Part 1: Methodology, Instrumentation and Performance, P.W.J.M. Boumans, ed., Wiley, 1987, 258-295 (with P.W.J.M. Boumans).

List of all people paid from contract N00014-76-C-0838 during contract period .

Bartelt, Joan
Blades, Michael
Bright, Frank
Carr, Jeffrey
Deutsch, Robert
Doherty, Michael
Downey, Stephen
Farnsworth, Paul
Freeman, James
Galante, Leonard
Hieftje, Gary
Hites, Bonnie
Lemke, Ralph
Lodder, Robert
Merrill, Dawn
Mitchell, Joel
Monnig, Curtis
Pak, Yongnam
Ramsey, J. Michael
Rezaaiyaan, Ruhangiz
Saturday, Kathy
Savage, Richard
Selby, Mark
Shabushnig, John
Steele, Andrew
Swarthout, Mary
Vickers, George
Weiss, Aaron
Wilson, Daniel
Wyatt, Wayde
Yuen, Anna
Zander, Andrew

This constitutes the final report of work accomplished under ONR contract N00014-76-C-0838. Over the ten-year period during which this contract was in force, a great deal was accomplished. As revealed earlier in this document, 112 technical reports were prepared, out of which has stemmed already 104 publications in refereed journals. Additional publications are still appearing and will be forwarded as they are received.

In addition, a substantial number of research personnel have been supported under this contract. A total of 9 Master's theses resulted from these studies, as did 14 Ph.D. theses. Also, 5 post-doctoral associates were funded, as were a number of part-time, clerical, and technical staff members.

Because of the wide range of activities that were pursued, it is difficult to summarize them here in a brief way. Instead, this report will concentrate on several highlights of the studies that have been completed to date. Importantly, a portion of this work is being continued under a new ONR contract, N00014-86-K-0366. In the following narrative, the papers cited are those listed elsewhere in this report under "List of Publications Emanating from Contract".

The principal focus of work pursued under the ONR contract just ending is to develop and characterize new methods for multielement analysis. Three such methods were originally emphasized and included the development of low-flow, high-efficiency torches for use in inductively coupled plasma (ICP) spectrometry, the exploration of novel microwave-induced plasmas (MIP) for elemental analysis, and the exploration of a new technique for the spectral selection of atomic or ionic emission lines, termed selective spectral line modulation (SLM).

Our studies to develop high-efficiency torches for ICP use have been highly successful and have resulted in systems that are now widely used in commercial instruments. This work has been reviewed in a journal (54) and a book chapter (105).

Our work with high-efficiency torches began with the development of miniaturized systems (10, 14-16, 24). In those studies, we progressed from a "standard" 18-mm torch, to 13-mm systems, and finally to a 9-mm device. By careful design, we were able to reduce thereby the gas-flow and power requirements for sustaining a usable plasma, and without loss of operating performance.

Later, we turned to a careful optimization of torch dimensions (11) and showed that it was possible to achieve the same kinds of reduction in operating power and gas flows but with a standard-size (18-mm) torch. This approach has the advantage of permitting the new high-efficiency torch to be employed with older ICP instruments. Performance of this kind of system was later evaluated (62, 73) and its application to miniaturized torches examined (75). It was found that there is little additional advantage to miniaturizing a fully optimized torch. Such optimized torches are now available commercially and at least two

vendors of ICP instruments offer reduced-size torches patterned after our design.

Our studies into microwave-induced plasmas began with atmospheric pressure systems in helium (7) that have strong advantages over alternative argon-supported discharges or those that operate at reduced pressure. The characteristics of microwave plasmas were also reviewed (26). Because the key to successful operation of an MIP was found to be the delivery of a pre-vaporized form of the sample material, we turned our attention in part to the development and exploration of alternative sample-introduction techniques. The "microarc" system (11, 46) served nicely and was evaluated further (see below). Because of the high excitation energy available in the helium MIP, we found that it was useful not only for the excitation of metal atoms but also nonmetallic species (66-68). A more recent design (98), supported by a surface wave structure termed a "surfatron" offers even greater capability than our earlier MIP, supported by a resonant cavity device. We continue the exploration of this novel system under the new ONR contract.

In the course of our MIP explorations, we pursued the use of nitrogen as an inexpensive and conveniently available support gas. The resulting device, which we termed the "microwave-induced nitrogen discharge at atmospheric pressure" (MINDAP) (57, 60) offers a number of attractive alternatives over the atmospheric pressure helium system. In particular, the MINDAP is a thermally hotter source, so that it can accept sample aerosols directly, without the need for pre-vaporization. In addition, the MINDAP produces a tail flame, much like the ICP. As a consequence, emission from it can be viewed in a side-on fashion rather than end-on, and background radiation from the plasma is minimized. In more recent investigations, we have found that the MINDAP is suitable also for use as an ionization source with mass spectrometry.

Our studies into the use of selective spectral line modulation (SLM) began with applying the technique to atomic absorption spectrometry (2) and to the selection of an appropriate light source to be used with the method (5). Later, we turned to the application of the method to atomic-emission spectrometry (22, 40, 103) and to improving the instrumentation used with the technique (102). An even more recent development, supported under the current ONR contract, makes SLM an extremely attractive alternative for eliminating spectral interference such as those which frequently occur in ICP emission work.

Because of the importance of sample introduction to both ICP and MIP studies, and to the potential application of modulated sample introduction to the SLM procedure, we occupied a portion of our time pursuing novel sample introduction approaches. Early studies involved the enhancement of pneumatic nebulization efficiency by application of an electric field (9) and by using the microarc (11) described above. In addition, we constructed several microsampling devices (28, 32, 99, 100) and showed how sample could be introduced into spectroscopic sources in the form of uniform, discrete droplets that are less than one nanoliter in volume. A number of these studies have been reviewed (25, 101). We learned later that samples could be introduced directly into an ICP by extracting a portion of the plasma energy and drawing it to a grounded

sample (49). Recent investigations have shown that this "RF arc" technique is highly attractive for solid, conducting samples. Another potentially attractive approach to sample introduction involves the jet-impact nebulizer (52), which functions by directing a sample solution onto a solid surface. The kinetic energy of the jet is thereby converted into internal energy within the jet, so it is disrupted into a fine aerosol.

Lastly, we examined modulated sample-introduction systems for use with emission spectroscopic sources like the ICP (89, 104). Such systems served to modulate a sample and its signal but not a potentially interfering background level. As a consequence, the signal attributable to the sample can be extracted with the use of simple, frequency-selective amplifiers. Such systems are useful also in implementing the SIM approach (102).

END

DATE

FILMED

MARCH

1988

DTIC