

NO-A192 190

A DISPOSABLE LIQUID MICROCELL FOR NEAR-INFRARED
REFLECTANCE ANALYSIS(U) INDIANA UNIV AT BLOOMINGTON
DEPT OF CHEMISTRY R A LODDER ET AL. 22 FEB 88

1/1


UNCLASSIFIED

INDU/DC/GMH/TR-88-22 N00014-86-K-0366

F/G 28/6

NL





MICROCOPY RESOLUTION TEST CHART
NATIONAL BUREAU OF STANDARDS 1963-A

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE

DTIC FILE COPY

4

REPORT DOCUMENTATION PAGE

1a. REPORT SECURITY CLASSIFICATION NA		1b. RESTRICTIVE MARKINGS NA	
2a. SECURITY CLASSIFICATION AUTHORITY NA		3. DISTRIBUTION/AVAILABILITY OF REPORT Distribution Unlimited; Approved for Public Release	
2b. DECLASSIFICATION/DOWNGRADING SCHEDULE NA			
4. PERFORMING ORGANIZATION REPORT NUMBER(S) INDU/DC/GMII/TR-88-22		5. MONITORING ORGANIZATION REPORT NUMBER(S) NA	
6a. NAME OF PERFORMING ORGANIZATION Indiana University	6b. OFFICE SYMBOL (if applicable) NA	7a. NAME OF MONITORING ORGANIZATION ONR	
6c. ADDRESS (City, State, and ZIP Code) Department of Chemistry Bloomington, IN 47405		7b. ADDRESS (City, State, and ZIP Code) 800 N. Quincy Street Arlington, VA 22217	
8a. NAME OF FUNDING/SPONSORING ORGANIZATION	8b. OFFICE SYMBOL (if applicable)	9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER Contract N00014-86-K-0366	
8c. ADDRESS (City, State, and ZIP Code)		10. SOURCE OF FUNDING NUMBERS	
		PROGRAM ELEMENT NO.	PROJECT NO.
		TASK NO.	R&T Code
			4134006
11. TITLE (Include Security Classification) A Disposable Liquid Microcell for Near-Infrared Reflectance Analysis			
12. PERSONAL AUTHOR(S) Robert A. Lodder and Gary M. Hieftje			
13a. TYPE OF REPORT Technical	13b. TIME COVERED FROM TO	14. DATE OF REPORT (Year, Month, Day) 22 February 1988	15. PAGE COUNT 12
SUPPLEMENTARY NOTATION Accepted for publication in APPLIED SPECTROSCOPY			
COSATI CODES		18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)	
FIELD	GROUP	SUB-GROUP	
18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)			
		NIRA, Solution analysis; Microsampling; Near infrared reflectance analysis.	
ABSTRACT (Continue on reverse if necessary and identify by block number)			
A disposable, 70 microliter, liquid analysis transmission cell is described for use in near-infrared reflectance instruments. The small sample volume relative to the overall cell size allows many analyses to be performed without a thermostatically controlled heating cycle. Complex purge/fill and wash cycles are also unnecessary. The cell is ideally suited for the analysis of potentially hazardous liquid samples. <i>Keywords:</i>			
20. DISTRIBUTION/AVAILABILITY OF ABSTRACT ☒ UNCLASSIFIED/UNLIMITED ☐ SAME AS RPT ☐ DTIC USERS		21. ABSTRACT SECURITY CLASSIFICATION Distribution Unlimited	
22a. NAME OF RESPONSIBLE INDIVIDUAL Gary M. Hieftje		22b. TELEPHONE (Include Area Code) (812) 335-2189	22c. OFFICE SYMBOL

AD-A192 190

DTIC
ELECTE
MAR 01 1988

OFFICE OF NAVAL RESEARCH

Contract N14-86-K-0366

R&T Code 4134006

TECHNICAL REPORT NO. 22

A DISPOSABLE LIQUID MICROCELL FOR
NEAR-INFRARED REFLECTANCE ANALYSIS

by

Robert A. Lodder and Gary M. Hieftje

Prepared for Publication

in

APPLIED SPECTROSCOPY

Indiana University
Department of Chemistry
Bloomington, Indiana 47405

22 February 1988

Reproduction in whole or in part is permitted for
any purpose of the United States Government

This document has been approved for public release
and sale; its distribution is unlimited

88 2 26 007

While near-infrared reflectance instruments are best known for their ability to rapidly analyze powdered solid samples¹⁻², some manufacturers have provided the means to measure liquid samples as well. Unfortunately, liquid-analysis accessories are often cumbersome and expensive. A typical accessory requires a relatively large volume of the sample (on the order of milliliters) and complex purge/fill and wash cycles to prevent clogging. If clogging does occur, cleaning can be difficult. The thermostatically controlled heating cycle only adds to a total analysis time already extended by the other operations.

Figure 1 depicts an alternative liquid cell and cell holder for the Technicon InfraAlyzer 400, constructed from solid aluminum and a single-cavity microscope slide with a cover slip. The main body of the cell holder has machined into it a 90 degree conical reflector and is based on a design used to analyze intact pharmaceutical capsules³. This main body fits into the solid-sample drawer of the spectrometer in place of the standard closed sample cup. The reflector is a polished curved surface of a right-circular cone with a height and a base radius of 13 mm. A hole 2 mm in diameter located at the vertex of this cone serves to stabilize a polished aluminum insert. This insert is essentially a cylinder capped with a second right-circular cone whose base is oriented in the direction opposite to that of the main body cone. The cone on the insert is machined with a vertex of 135 degrees. A standard single-cavity



n For	
AI	☒
ed	☐
tion	☐
tion/	
Availability Codes	
Dist	Avail and/or Special
A-1	

microscope slide (25 x 76 mm) is centered with its 22 x 22 mm cover slip over the insert. The position of the slide is made stable and reproducible by resting it against two screws fastened into the main body.

The cavity slide has some distinct advantages over a conventional flat microscope slide in this application: (1) it provides a longer and more reproducible optical pathlength, (2) the cover slip acts as a lid on the cavity and lowers the liquid-sample evaporation rate, and (3) the cavity shape acts as a lens to scatter transmitted light into the integrating sphere of the spectrophotometer. A fully filled liquid cell based on a single-cavity slide (Dickinson and Company, Parsippany, NJ, #3720) and an ordinary cover slip (American Scientific Products, McGaw Park, IL, #M6045-2) contains 110 microliters of sample. However, different slides and cover slips with different masses can be used to vary the optical pathlength and the sample cell volume. The cover slip actually floats on the sample, and heavier cover slips tend to squeeze the sample and reduce the cell volume. When the cover slip is resting against the slide the cell volume is 70 microliters.

The 135-degree aluminum insert returns collimated light that passes through the slide back through the slide parallel to the walls of the main-body reflector cone. This design allows the bulk of the light that passes through the liquid in the cavity to be reflected directly into the instrument's integrating sphere at a 45 degree angle from the source light.

In this design the 135-degree conical portion of the insert is placed atop a small cylinder because the sample is actually below the integrating sphere; if the insert cone were to be lowered to the bottom (vertex end) of the main-body reflector cone much of the reflected light would miss the window of the integrating sphere.

Initial tests of this liquid cell included an analysis of a set of aqueous sodium chloride solutions. The determination of sodium chloride in water can be difficult for several reasons⁴, including: (1) sodium chloride has no absorption bands in the near-infrared, (2) water has very strong absorption bands in the near-infrared, and (3) these water absorption bands are very temperature-dependent. Nevertheless, successful determinations of aqueous sodium chloride in concentrations from 30-38 grams per liter have been reported⁴ by using four wavelengths selected in a standard multiple linear regression procedure.

Twenty aqueous solutions of reagent-grade sodium chloride (ten for the training set and ten for the validation set) were prepared for analysis in the new liquid cell. Solutions ranged in concentration from 5 to 38 grams per liter. Each solution was loaded into a single-cavity slide two times, and four spectra were taken from each sample loading. Spectra were recorded at 16 wavelengths and the data were transformed to their principal axes to avoid the need for a time-consuming all-possible-combinations of wavelengths regression. In order to

demonstrate that one need not be very particular about the initial selection of analytical wavelengths, the wavelength data near water absorption peaks were deliberately deleted from the recorded spectra (which contained data from 19 wavelengths). This also shows that relatively complex instruments, utilizing scanning monochromators to collect data at hundreds of wavelengths, are often unnecessary in NIRA.

Multiple linear regression was then carried out on the 80 training spectra using only the data along the first five principal axes (these axes accounted for over 99.9% of the total spectral variation). Data from five axes (rather than the four used in reference 4) were required because evaporative loss from the cell produced pathlength variations that called for an additional degree of freedom in the system. The results of the training process are summarized in the calibration line in Figure 2. The r^2 for the training set that produced the line is 0.97, and the r^2 value for the 80 validation spectra (shown superimposed on the calibration line, with error bars) is also 0.97. The detection limit for sodium chloride, calculated from both the error in the validation spectra and from four solvent blanks, is 1 gram per liter (1000 ppm). This value corresponds to an absolute detection limit of approximately 100 micrograms in the 110 microliter sample cell.

The liquid microcell that has been described here has a number of practical advantages. It is faster and easier to use than an ordinary liquid accessory. No heating or thermostating

is required because 110 microliters of liquid rapidly reaches thermal equilibrium. No purging/filling or wash cycles are required. Any number of cells can be rapidly filled with a precision pipette if desired, and the cells can be easily cleaned or simply discarded afterward (an advantage for potentially dangerous and toxic samples). The configuration of the cell permits sensitive detection by enhancing transmission through the sample in a reflectance instrument. The apparent lack of pathlength reproducibility for volatile samples is compensated simply by using a random selection of pathlengths when the training-set spectra are recorded and by letting the calibration process take care of the rest. This microcell design adds a versatility to liquid analysis in near-infrared reflectance instruments that complements the flexibility of the near-infrared calibration procedure.

ACKNOWLEDGEMENTS

Supported in part by Technicon Industrial Systems, the Office of Naval Research, The Upjohn Company, and by the National Science Foundation through Grant CHE 83-20053.

REFERENCES

1. C. A. Watson, *Anal.Chem.* 49(9), 835A (1977).
2. D. L. Wetzel, *Anal.Chem.* 55(12), 1165A (1983).
3. R. A. Lodder and G. M. Hieftje, *Anal.Chem.* (in press, 1987).
4. T. Hirschfeld, *Appl.Spectrosc.* 39(4), 740 (1985).

FIGURE CAPTIONS

Figure 1. A cross-section of the microsample cell.

Figure 2. NIR calibration for NaCl in H₂O, obtained with the new microcell.



*Fig. 1*

Fig. 2



TECHNICAL REPORT DISTRIBUTION LIST, GEN

	<u>No. Copies</u>		<u>No. Copies</u>
Office of Naval Research Attn: Code 1113 800 N. Quincy Street Arlington, Virginia 22217-5000	2	Dr. David Young Code 334 NORDA NSTL, Mississippi 39529	1
Dr. Bernard Douda Naval Weapons Support Center Code 50C Crane, Indiana 47522-5050	1	Naval Weapons Center Attn: Dr. Ron Atkins Chemistry Division China Lake, California 93555	1
Naval Civil Engineering Laboratory Attn: Dr. R. W. Drisko, Code L52 Port Hueneme, California 93401	1	Scientific Advisor Commandant of the Marine Corps Code RD-1 Washington, D.C. 20380	1
Defense Technical Information Center Building 5, Cameron Station Alexandria, Virginia 22314	12 high quality	U.S. Army Research Office Attn: CRD-AA-IP P.O. Box 12211 Research Triangle Park, NC 27709	1
DTNSRDC Attn: Dr. H. Singerman Applied Chemistry Division Annapolis, Maryland 21401	1	Mr. John Boyle Materials Branch Naval Ship Engineering Center Philadelphia, Pennsylvania 19112	1
Dr. William Tolles Superintendent Chemistry Division, Code 6100 Naval Research Laboratory Washington, D.C. 20375-5000	1	Naval Ocean Systems Center Attn: Dr. S. Yamamoto Marine Sciences Division San Diego, California 91232	1

END

DATE

FILMED

5-88
DTIC