

AD-A200 306

2

DTIC FILE COPY

OFFICE OF NAVAL RESEARCH

Contract N00014-83-K-0470-P00003

R&T Code NR 33359-718

Technical Report No. 103

Ultramicroelectrode Sensors and Detectors. Considerations of the  
Stability, Sensitivity, Reproducibility, and Mechanism of Ion Transport in  
Gas Phase Chromatography and in High Performance Liquid Phase Chromatography

by

R. Brina, S. Pons and M. Fleischmann

Prepared for publication in J. Electroanal. Chem.

Department of Chemistry  
University of Utah  
Salt Lake City, UT 84112

July 15, 1988

DTIC  
ELECTE  
NOV 14 1988  
S D  
CH

Reproduction in whole, or in part, is permitted for  
any purpose of the United States Government

DISTRIBUTION STATEMENT A  
Approved for public release;  
Distribution Unlimited

88 11 10 059

100-2-206

**REPORT DOCUMENTATION PAGE**

1a. REPORT SECURITY CLASSIFICATION Unclassified		1b. RESTRICTIVE MARKINGS	
2a. SECURITY CLASSIFICATION AUTHORITY		3. DISTRIBUTION / AVAILABILITY OF REPORT Approved for public release and sale. Distribution unlimited.	
2b. DECLASSIFICATION / DOWNGRADING SCHEDULE		4. PERFORMING ORGANIZATION REPORT NUMBER(S) ONR Technical Report No. 103	
5. MONITORING ORGANIZATION REPORT NUMBER(S)		6a. NAME OF PERFORMING ORGANIZATION University of Utah	
6b. OFFICE SYMBOL (if applicable)		7a. NAME OF MONITORING ORGANIZATION	
6c. ADDRESS (City, State, and ZIP Code) Department of Chemistry Henry Eyring Building Salt Lake City, UT 84112		7b. ADDRESS (City, State, and ZIP Code)	
8a. NAME OF FUNDING / SPONSORING ORGANIZATION Office of Naval Research		8b. OFFICE SYMBOL (if applicable)	
9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER N00014-83-K-0470-P00003		8c. ADDRESS (City, State, and ZIP Code) Chemistry Program, Code 1113 800 N. Quincy Street Arlington, VA 22217	
10. SOURCE OF FUNDING NUMBERS		11. TITLE (Include Security Classification) Considerations of the Stability, Sensitivity, Reproducibility, and Mechanism of Ion Transport in Gas Phase Chromatography and in High Performance Liquid Phase Chromatography	
PROGRAM ELEMENT NO. 4	PROJECT NO.	TASK NO.	WORK UNIT ACCESSION NO.
12. PERSONAL AUTHOR(S) R. Brina, S. Pons and M. Fleischmann			
13a. TYPE OF REPORT Technical	13b. TIME COVERED FROM 9/87 TO 7/88	14. DATE OF REPORT (Year, Month, Day) July 15, 1988	15. PAGE COUNT 24
16. SUPPLEMENTARY NOTATION			
17. COSATI CODES		18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)	
FIELD	GROUP	SUB-GROUP	microelectrodes, detectors
19. ABSTRACT (Continue on reverse if necessary and identify by block number)  Attached.			
20. DISTRIBUTION / AVAILABILITY OF ABSTRACT <input checked="" type="checkbox"/> UNCLASSIFIED/UNLIMITED <input type="checkbox"/> SAME AS RPT <input type="checkbox"/> DTIC USERS		21. ABSTRACT SECURITY CLASSIFICATION Unclassified	
22a. NAME OF RESPONSIBLE INDIVIDUAL Stanley Pons		22b. TELEPHONE (Include Area Code) (801)581-4760	22c. OFFICE SYMBOL

**ABSTRACT**

We describe in this paper the use of ultramicroelectrodes as sensitive sensors and detectors in the gas phase and in liquid solutions of high resistance; particularly aspects regarding the sensitivity, stability, and reproducibility of the devices. The advantages of these detectors can be improved notably by appropriate choice of method of pretreatment of the electrode surface. The surface separating the working ultramicroelectrode from the secondary/reference electrode provides the medium for ionic conduction for the cell. The mechanism for ion conduction involves proton migration and diffusion across the surface of the separator.



Accession For	
NTIS GRA&I	<input checked="" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By _____	
Distribution/ _____	
Availability Codes	
Dist	Avail and/or Special
A-1	

J. Electroanal. Chem. 00 (1988) JEC09532  
Elsevier Sequoia S.A., Lausanne - Printed in The Netherlands

## ULTRAMICROELECTRODE SENSORS AND DETECTORS

### CONSIDERATIONS OF THE STABILITY, SENSITIVITY, REPRODUCIBILITY, AND MECHANISM OF ION TRANSPORT IN GAS PHASE CHROMATOGRAPHY AND IN HIGH PERFORMANCE LIQUID PHASE CHROMATOGRAPHY

ROSSELLA BRINA and STANLEY PONS \*

*Department of Chemistry, University of Utah, Salt Lake City UT 84112 (U.S.A.)*

MARTIN FLEISCHMANN

*Department of Chemistry, The University, Southampton, Hants. SO9 5NH (Great Britain)*

(Received 31st August 1987; in revised form 20th November 1987)

#### ABSTRACT

We describe in this paper the use of ultramicroelectrodes as sensitive sensors and detectors in the gas phase and in liquid solutions of high resistance; particularly aspects regarding the sensitivity, stability, and reproducibility of the devices. The advantages of these detectors can be improved notably by appropriate choice of method of pretreatment of the electrode surface. The surface separating the working ultramicroelectrode from the secondary/reference electrode provides the medium for ionic conduction for the cell. The mechanism for ion conduction involves proton migration and diffusion across the surface of the separator. (10)

#### INTRODUCTION

Ultramicroelectrodes have found new and interesting applications as sensitive electrochemical sensors and detectors in a variety of media [1-4]. The improved analytical sensitivity observed at ultramicroelectrodes (compared to electrodes with characteristic dimensions larger than about 25  $\mu\text{m}$ ) is due to the increase in the relative contribution of faradaic to nonfaradaic currents as the characteristic size of the electrode is decreased [5]. At very small electrodes, electrochemical measurements in the absence of added supporting electrolyte is possible [6-8]; the number

\* To whom correspondence should be addressed.

of ions/dipoles required to charge the double layer of a low area electrode is small so that impurities or ions resulting from to autoprotolysis of the solvent are sufficient. Further, the small currents ( $< 10^{-10}$  A typically in dilute solutions) give rise to negligible ohmic corrections to the applied potential; this correction can be predicted analytically in some circumstances. Electrochemical measurements in highly resistive media are therefore straightforward at ultramicroelectrodes [9], and can be made at higher potentials that are possible in conventional systems where the limiting potential is due to electrolysis of the added supporting electrolyte [10,11]. The small currents flowing in the cell also eliminate the need for potentiostats since the counter electrode potential remains essentially constant during the experiment. System electronic noise is therefore reduced. Typical measurements are the made with two electrodes, a wave form generator, and a picoammeter. [12].

The present work concerns the behavior of ultramicroelectrodes in gas phase measurements, particularly in gas chromatographic applications, and in high performance liquid chromatography. It was suggested in an earlier report [12] that the electrochemical reactions of species in the gas phase at ultramicroelectrodes are coupled to ion migration across the surface of the insulator separating the two electrodes. Evidence for this process and the role of protons in the ion transport mechanism are discussed.

The use of electrochemical detection in high performance liquid chromatography is expanding rapidly [13]. The requirement of supporting electrolytes in the liquid phase has generally precluded the use of electrochemical detection in pure liquid chromatographic phases; the solvent often must be a non-polar organic solvent. The same experimental requirements are generally required for flow injection analysis experiments. We demonstrate herein the utility of ultramicroelectrodes for chemical sensing and detection under these circumstances.

#### EXPERIMENTAL

An ultramicroelectrode gas phase detector assembly is shown in Fig. 1. A detailed line drawing of the electrode assembly itself is presented in ref. 1. The working electrode used in this work for both the gas phase and high performance liquid phase chromatography experiments consisted of a 25  $\mu$ m diameter platinum wire (Johnson Matthey, Inc.), lead soldered to the end of an 18 ga. copper wire. The assembly was then inserted into a quartz tube (0.5-0.7 mm i.d. at the tip) until a few mm of wire protruded from the tip of the tube. The tube was previously coated with a thick film of palladium metal on the outside surface to a length of about 2 cm from the tip. This coating was prepared by thermal reduction of Liquid Bright Palladium (Englehard Industries, Inc.) organometallic at 625°C for 1 h in a tube furnace. Electrical contact to the palladium coating was made by lead soldering to it a fine insulated copper wire. The platinum ultramicroelectrode sealed in the glass tube with epoxy resin, and when dry was polished with alumina. A wave form generator (Hi Tek Instruments, Model PPR1) was used to control the potential. The currents were measured with a digital picoammeter (Keithley model 617), placed in




Fig. 1. Schematic diagram of the ultramicroelectrode chromatographic detector system.

series with the wave form generator and the electrochemical cell.

To reduce capacitively coupled noise, the ultramicroelectrode assembly was positioned inside an aluminum Faraday cage. The gas chromatograph was a Hewlett-Packard Model 700. The gas outlet port was connected to the ultramicroelectrode detector by means of a section of Teflon tube. The tip of the ultramicroelectrode was inserted inside the Teflon tube which was vented near the ultramicroelectrode tip. A dual pen strip chart recorder (Omniscribe, Houston Instruments, Inc.) was used for the simultaneous recording of the response of the ultramicroelectrode and the thermal conductivity detector (Hewlett-Packard Model 4-4188). The reverse phase column used in all of the experiments described herein was Silicone OV-17 10% on ChromasorbW. The carrier gas was helium which was maintained at a constant flow rate of 30 ml/min. Tetra-*n*-butylammonium tetrafluoroborate (Chem.-Biochem. Research, Inc.) was used received. Reagent grade solvents were used.

For high performance liquid chromatographic studies, the DNA bases thymine, adenine, cytosine, uracil, and guanine (Aldrich Chemical) were used. The chromatograph consisted of a Valco 6-way electric valve, a Waters 6000A pump, and a 3.5 cm  $\times$  0.5 mm octadecylsilane reverse phase capillary column. The ultramicroelectrode detector was mounted on the end of the column so that the effluent passed across the tip.

#### RESULTS AND DISCUSSION

The initial experiments were intended to demonstrate the selectivity and relative sensitivity of the ultramicroelectrode in gas phase detection. The injection port was at a temperature of 170°C and the column at 80°C. The applied potential was +3.0 V vs. the palladium pseudoreference electrode. Table 1 shows the responses measured for injections of 1  $\mu$ l of the various analytes. The first column in the table lists the ultramicroelectrode responses for the electrode prepared as described in the

TABLE 1

Response of the platinum ultramicroelectrode detector <sup>a</sup>

Compound	Response for electrode as prepared pA	Response after treatment w/TBABF <sub>4</sub> pA	Ratio $\frac{\text{column 2}}{\text{column 1}}$
Acetone	0.1180	12.66	107.0
Acetonitrile	10.19	13.940	1368
Aniline <sup>c</sup>	0.2000	11.350	56.750
Benzonitrile <sup>c</sup>	0.8670	191.0	220.0
Carbon tetrachloride	0.2500	0.3120	1.205
Cyclohexanone	16.51	1.125	68.57
Ethanol	565.0	89.540	158.5
Ethyl Acetate	0.5300	136.6	258.3
Isopropylamine	0.1620	8.340	51.33
Methylene chloride	3.650	620.0	169.5
Methylamine (40% in water)	51.000	168.700	3.394
<i>o</i> -Xylene	0.1010	16.13	160.1
Toluene	0.1620	0.3402	2.102
Triethylamine	0.5080	2.140	4.215
1,2,4-Trimethylbenzene	0.0120	2.590	216.3

<sup>a</sup> Injection of 1  $\mu$ l; solution approx. 0.01 M H<sub>2</sub>O.<sup>b</sup> The ultramicroelectrode was pretreated in a 0.3 M in a solution of TBABF<sub>4</sub> in CH<sub>3</sub>CN for 2 min.<sup>c</sup> Injection of 0.5  $\mu$ l; solution approx. 0.01 M H<sub>2</sub>O.

experimental section, while the second column gives the measured responses after the electrode had been treated with a 0.3 M tetra-n-butylammonium tetrafluoroborate (TBABF<sub>4</sub>) solution in acetonitrile for 2 min. The electrode was allowed to equilibrate at the exit port of the chromatograph for about 20 min. Generally, an increase in the sensitivity of the electrode was observed when treated with TBABF<sub>4</sub>, which for the epoxy resin used was superior to other common electrolytes. The response of this detector is quite reproducible over a period of 12 h of continuous use. If the electrode was not treated with TBABF<sub>4</sub> for two consecutive days, a few percent decrease in sensitivity occurs, and will continue to degrade. Excellent reproducibility is obtained (better than 1%) if the electrode is treated daily by the procedure described. The absolute value of the increase depended on the solvent used, and varied from 100 to 1000 in over the untreated electrode. If the electrode was treated with 0.2 M H<sub>2</sub>SO<sub>4</sub> solution for 1 min, a further increase in sensitivity could be observed, but was not reproducible or constant, presumably due to evaporation and oxidation of the epoxy resin.

The following experiment was carried out in order to compare the sensitivity of the ultramicroelectrode detector to the thermal conductivity detector. The effluent was first detected by the thermal conductivity detector and then the ultramicroelectrode detector. The ultramicroelectrode was located 5 mm downstream of the thermal conductivity detector. The gas chromatograph and the ultramicroelectrode detector were operated under the same conditions as described in the last experiment.

1  $\mu$ l injections of dilute cyclohexanone solutions in acetone were analyzed with both detectors. The results are given in Table 2. Correlation coefficients and slopes of the response/injected concentration plots for the ultramicroelectrode and for the thermal conductivity detector are also listed. The linearity of the two detectors can be compared better by plotting the logarithm of the response vs. the logarithm of the injected concentration [14], as shown in Fig. 2. We measure correlation coefficients of 0.998 and 0.999 for the ultramicroelectrode and the thermal conductivity detector, respectively, in the concentration range from 15% to 0.025% (v/v) of cyclohexanone. The slope of these plots, defined as the linearity of the detector [14], are 1.18 for the ultramicroelectrode and 0.84 for the thermal conductivity detector. The ultramicroelectrode used in these measurements has a ripple fluctuation of 3 fA peak to peak. The background noise fluctuation of the thermal conductivity detector was ca. 2  $\mu$ V peak to peak. Defining the detection limits as the concentration that gives an average signal three times larger than the background noise level, and using the values of the slope in Table 2, we obtain detection limits for cyclohexanone in acetone of 160 ppm when measured at the ultramicroelectrode and 600 ppm at the thermal conductivity detector. A similar analysis for aniline solutions in acetone indicated detection limits of 82 ppm at the ultramicroelectrode detector and 900 ppm for the thermal conductivity detector.

The dependence of the ultramicroelectrode response on the applied potential and the nature of the analyte was investigated. The ultramicroelectrode response for an injection of a solution of 50% toluene (v/v) in triethylamine was determined in the potential region between +0.6 and +3.0 V vs. the palladium reference electrode. The column and the carrier gas used were the same as described in the experimental

TABLE 2  
Ultramicroelectrode and thermal conductivity detector responses for cyclohexanone solutions <sup>a</sup>

Injected concentration % cyclohexanone	Ultramicroelectrode response pA <sup>b</sup>	Thermal conductivity detector response mV <sup>b</sup>
8	4.47	0.8
5	2.6	0.54
3	1.351	0.36
1	0.345	0.126
0.5	0.166	0.07
0.25	0.095	0.046
0.1	0.028	0.0186
0.05	0.0153	-
Corr. coefficient <sup>c</sup>	0.997	0.998
Slope <sup>c</sup>	0.556 pA/%	0.100 mV/%
Detection limits <sup>d</sup>	160 ppm	600 ppm

<sup>a</sup> 1  $\mu$ l solution injected.

<sup>b</sup> Peak height.

<sup>c</sup> From the plot of the response vs. the injected concentration for the two detectors.

<sup>d</sup> Three times the electronic noise level.




Fig. 2. Logarithm of ultramicroelectrode response vs. logarithm of cyclohexanone concentration (O), and log of thermal conductivity response vs. log of cyclohexanone concentration (□), for the injection of  $1 \mu\text{l}$  of solution.

section. The injection port temperature was maintained at  $120^\circ\text{C}$  and the column at  $85^\circ\text{C}$ . Figure 3 shows the polarization curves obtained for the two compounds  $1 \mu\text{l}$  injections of the solution. There is a greater dependence on potential of the toluene response than for the triethylamine response. This leads to a crossing of the two voltammograms at lower potentials. The same experiment was performed with a 30% cyclohexanone solution (v/v) in triethylamine, between  $+0.8$  and  $+7.0$  V. The




Fig. 3. Ultramicroelectrode response vs. potential applied for the injection of  $1 \mu\text{l}$  of a solution 50% in toluene and 50% in triethylamine. (□) Toluene response, (O) triethylamine response.





Fig. 4. Ultramicroelectrode response vs. potential applied for the injection of  $1 \mu\text{l}$  of a solution of 30% cyclohexanone in triethylamine. (□) triethylamine response.

injection port temperature was  $170^\circ\text{C}$ , and the column temperature was  $110^\circ\text{C}$ . The polarization curves of the two species are shown in Fig. 4. A discontinuity in the response for both compounds is observed at about  $+6.0 \text{ V}$ . An analogous feature between  $+4.0$  and  $+6.0 \text{ V}$  was observed in other binary mixtures. The same phenomena was observed in previous work [1] in the polarization curve for methylamine at potentials greater than about  $+2.5 \text{ V}$ . The discontinuity is avoided if the ultramicroelectrode is mounted in the glass tube with a glass or glass solder instead of epoxy resin. This result suggests that the discontinuity is due to oxidative breakdown of the epoxy seal. While the degradation in response is eliminated when pure glass seals are used, the glass mounted ultramicroelectrodes are generally less sensitive, and their sensitivity is not increased with the pretreatment methods to any appreciable extent; other methods of pretreatment are under investigation.

Figures 3 and 4 demonstrate that analysis of an injected mixture is possible from the voltammetric gas phase response; more complex mixtures can be similarly analyzed. Figure 5 shows a chromatogram obtained at the thermal conductivity detector and the ultramicroelectrode detector for an injection of  $1 \mu\text{l}$  of a solution containing 12% acetone, 12% cyclohexanone, and 30% toluene in *o*-xylene (v/v). The ultramicroelectrode was maintained at  $+3.0 \text{ V}$  vs the palladium pseudoreference electrode. The injection port temperature was  $170^\circ\text{C}$  and the column at  $90^\circ\text{C}$ .

We have studied the conduction mechanism across the insulating area between the two electrodes. Table 3 shows the response of the ultramicroelectrode detector to  $1 \mu\text{l}$  injections of a series of substituted benzenes. The temperature of the injection port in each case was  $170^\circ\text{C}$  and the temperature of the column was  $120^\circ\text{C}$ . The potential of the ultramicroelectrodes was maintained at  $+3.0 \text{ V}$ . As the number of methyl groups on the benzene ring is increased, the ultramicroelectrode response



CONTAINING

Fig. 5. (a) Ultramicroelectrode detector, (b) thermal conductivity detector responses after the injection of a solution containing 30% toluene, 46% (v/v) *o*-xylene. The peaks from left to right correspond to acetone, toluene, *o*-xylene, and cyclohexanone, respectively. Applied potential was +3.0 V. The total time is 16 min.

CONTAINING

Fig. 6. HPLC of a mixture of DNA and RNA bases. Lower chromatogram is the ultramicroelectrode response with potential difference of  $-3.0$  V vs. the palladium secondary reference. Upper chromatogram is under the same conditions at  $-2.50$  V. The injected concentrations of the bases were guanine:  $11 \text{ pg ml}^{-1}$ , adenine:  $9 \text{ pg ml}^{-1}$ , cytosine  $10 \text{ pg ml}^{-1}$  uracil:  $19 \text{ pg ml}^{-1}$ , and thymine  $19 \text{ pg ml}^{-1}$ . The peaks correspond to (a) guanine, (b) uracil, (c) cytosine, (d) adenine, and (e) thymine respectively.

decreases rapidly. This behavior is enhanced when the substituents on the benzene are ethyl groups. In general, methyl-substituted benzenes are anodically oxidized [15,16] to the cation radicals which lose a proton rapidly, to give the free radical. The free radical is oxidized further to and the following reactions lead to a variety of products. In general, increasing the number of alkyl substituents on the benzene ring leads to a decrease in the oxidation potential of the hydrocarbon [17] and an increase in the stability of the cation radical. As a result, increasing the number of methyl groups leads to a decrease in the population of protons at the interface, although not in the total number of charged species. The response is observed to decrease with increased number of substituents (fewer free protons). Ethyl group

TABLE 3

Ultramicroelectrode detector response to methyl- and ethyl-substituted benzenes

Compound	Ultramicroelectrode response/ $\mu\text{A}$ <sup>a</sup>
Toluene (99%)	57.33
1,2-Dimethylbenzene (95%)	16.13
1,2,4-Trimethylbenzene (99%)	2.590
Pentamethylbenzene <sup>b</sup>	0.0000
Hexamethylbenzene <sup>c</sup>	0.0000
Ethylbenzene	0.0760
1,3,5-Triethylbenzene <sup>c</sup>	0.0000
Hexaethylbenzene <sup>c</sup>	0.0000

<sup>a</sup> Injection of 1  $\mu\text{l}$ .<sup>b</sup> 7 M solution in acetone.<sup>c</sup> Saturated acetone solution.

substitution leads to much greater stabilization of the cation radical (smaller population of free protons). As is seen in the table, the response drops markedly for ethyl substitution. The conclusion is that protons, and not necessarily other charged species produced in the redox reaction, are responsible for the ionic conductivity across the surface of the insulator separating the two electrodes.

The use of ultramicroelectrodes as detectors in resistive liquid solutions in high performance liquid chromatography is shown in Fig. 6. The ultramicroelectrode was not coated in this experiment. The DNA bases guanine, adenine, cytosine, uracil, and thymine were detected for a gradient elution of a mixture of the bases in a solution of acetonitrile in hexane (0 to 50%, v/v) in the chromatograph described in the experimental section. The small unlabeled peaks in the figure are due to impurities. The lower plot was the response of the ultramicroelectrode at  $-3.0\text{ V}$  while the upper plot was that obtained at  $-2.5\text{ V}$ . It is seen that the compound with the highest reduction potential, uracil, shows a much lower response at the lower potential, as expected. Further decrease in the potential results in the progressive and sequential disappearance of the responses in the order adenine, uracil, thymine, cytosine, and guanine, respectively, as expected from their reduction potentials. As the potential is made more positive, the response appear again, but with positive currents flowing. The sensitivity of the ultramicroelectrode detector is higher in the liquid system than in the gas phase experiments due to the increased conductivity of this medium.

It is clear that these types of detectors ar both quantity and species sensitive for chromatographic applications. Their use in a variety of new applications is evident.

## ACKNOWLEDGEMENTS

We thank the Office of Naval Research for support of this work. We express appreciation to Mr. Joseph Pons for design and preparation of the ultramicroelectrodes used in this work.

## REFERENCES

- 1 J. Ghoroghchian, F. Sarfarazi, T. Dibble, J. Cassidy, J.J. Smith, A. Russell, G. Dunmore, M. Fleischmann and S. Pons, *Anal. Chem.*, 58 (1986) 2278.
- 2 L.A. Knecht, E.J. Guthrie and J.W. Jorgenson, *Anal. Chem.*, 56 (1984) 479.
- 3 W.L. Cardill, J.O. Howell and R.M. Wightman, *Anal. Chem.*, 55 (1983) 1877.
- 4 R.M. Wightman, *Anal. Chem.*, 53 (1981) 1125A.
- 5 Southampton Electrochemistry Group, *Instrumental Methods in Electrochemistry*, Ellis Horwood, Chichester, 1985, p. 217.
- 6 A.M. Bond, M. Fleischmann and J. Robinson, *J. Electroanal. Chem.*, 172 (1984) 11.
- 7 A.M. Bond, M. Fleischmann and J. Robinson, *J. Electroanal. Chem.*, 168 (1984) 299.
- 8 J.O. Howell and R.M. Wightman, *Anal. Chem.*, 53 (1984) 524.
- 9 A.M. Bond, M. Fleischmann and J. Robinson, *J. Electroanal. Chem.*, 180 (1984) 257.
- 10 J. Cassidy, S.B. Khoo, S. Pons and M. Fleischmann, *J. Phys. Chem.* 89 (1985) 3933.
- 11 T. Dibble, S. Bandyopadhyay, J. Ghoroghchian, J.J. Smith, F. Sarfarazi, M. Fleischmann and S. Pons, *J. Phys. Chem.*, 90 (1986) 5275.
- 12 J.W. Bixler, A.M. Bond, P.A. Lay, W. Thormann, P. van den Bosch, M. Fleischmann and S. Pons, *Anal. Chim. Acta*, 187 (1986) 67.
- 13 R.E. Majors, H. G. Barth and C.H. Lochmuller, *Anal. Chem.*, 56 (1984) 300R.
- 14 J. Ševčík, *Detectors in Gas Chromatography*, Journal of Chromatography Library, Vol. 4, Elsevier, Amsterdam, Oxford, New York, 1976, p. 32.
- 15 E.S. Pysh and N.C. Yang, *J. Am. Chem. Soc.*, 85 (1963) 2124.
- 16 W.C. Niekam and M.M. Desmond, *J. Am. Chem. Soc.*, (1964) 4811.
- 17 J.W. Loveland and G.R. Dimeler, *Anal. Chem.*, 33 (1961) 1196.

56 (1964) 4811

TECHNICAL REPORT DISTRIBUTION LIST, GEN

	<u>No. Copies</u>		<u>No. Copies</u>
Office of Naval Research Attn: Code 1113 800 N. Quincy Street Arlington, Virginia 22217-5000	2	Dr. David Young Code 334 NORDA NSTL, Mississippi 39529	1
Dr. Bernard Douda Naval Weapons Support Center Code 50C Crane, Indiana 47522-5050	1	Naval Weapons Center Attn: Dr. Ron Atkins Chemistry Division China Lake, California 93555	1
Naval Civil Engineering Laboratory Attn: Dr. R. W. Drisko, Code L52 Port Hueneme, California 93401	1	Scientific Advisor Commandant of the Marine Corps Code RD-1 Washington, D.C. 20380	1
Defense Technical Information Center Building 5, Cameron Station Alexandria, Virginia 22314	12 high quality	U.S. Army Research Office Attn: CRD-AA-IP P.O. Box 12211 Research Triangle Park, NC 27709	1
DTNSRDC Attn: Dr. H. Singerman Applied Chemistry Division Annapolis, Maryland 21401	1	Mr. John Boyle Materials Branch Naval Ship Engineering Center Philadelphia, Pennsylvania 19112	1
Dr. William Tolles Superintendent Chemistry Division, Code 6100 Naval Research Laboratory Washington, D.C. 20375-5000	1	Naval Ocean Systems Center Attn: Dr. S. Yamamoto Marine Sciences Division San Diego, California 91232	1

ABSTRACTS DISTRIBUTION LIST, SDIO/IST

Dr. Robert A. Osteryoung  
Department of Chemistry  
State University of New York  
Buffalo, NY 14214

Dr. Douglas N. Bennion  
Department of Chemical Engineering  
Brigham Young University  
Provo, UT 84602

Dr. Stanley Pons  
~~Department of Chemistry~~  
University of Utah  
Salt Lake City, UT 84112

Dr. H. V. Venkatesetty  
Honeywell, Inc.  
10701 Lyndale Avenue South  
Bloomington, MN 55420

Dr. J. Foos  
EIC Labs Inc.  
111 Downey St.  
Norwood, MA 02062

Dr. Neill Weber  
Ceramatec, Inc.  
163 West 1700 South  
Salt Lake City, UT 84115

Dr. Subhash C. Narang  
SRI International  
333 Ravenswood Ave.  
Menlo Park, CA 94025

Dr. J. Paul Pemsler  
Castle Technology Corporation  
52 Dragon Ct.  
Woburn, MA 01801

Dr. R. David Rauh  
EIC Laboratory Inc.  
111 Downey Street  
Norwood, MA 02062

Dr. Joseph S. Foos  
EIC Laboratories, Inc.  
111 Downey Street  
Norwood, Massachusetts 02062

Dr. Donald M. Schleich  
Department of Chemistry  
Polytechnic Institute of New York  
333 Jay Street  
Brooklyn, New York 01

Dr. Stan Szpak  
Code 633  
Naval Ocean Systems Center  
San Diego, CA 92152-5000

Dr. George Blomgren  
Battery Products Division  
Union Carbide Corporation  
25225 Detroit Rd.  
Westlake, OH 44145

Dr. Ernest Yeager  
Case Center for Electrochemical  
Science  
Case Western Reserve University  
Cleveland, OH 44106

Dr. Mel Miles  
Code 3852  
Naval Weapons Center  
China Lake, CA 93555

Dr. Ashok V. Joshi  
Ceramatec, Inc.  
2425 South 900 West  
Salt Lake City, Utah 84119

Dr. W. Anderson  
Department of Electrical &  
Computer Engineering  
SUNY - Buffalo  
Amherst, Massachusetts 14260

Dr. M. L. Gopikanth  
Chemtech Systems, Inc.  
P.O. Box 1067  
Burlington, MA 01803

Dr. H. F. Gibbard  
Power Conversion, Inc.  
495 Boulevard  
Elmwood Park, New Jersey 07407

D/1113/87/2

ABSTRACTS DISTRIBUTION LIST, SDIO/IST

Dr. V. R. Koch  
Covalent Associates  
52 Dragon Court  
Woburn, MA 01801

Dr. Randall B. Olsen  
Chronos Research Laboratories, Inc.  
4186 Sorrento Valley Blvd.  
Suite H  
San Diego, CA 92121

Dr. Alan Hooper  
Applied Electrochemistry Centre  
Harwell Laboratory  
Oxfordshire, OX11 0RA UK

Dr. John S. Wilkes  
Department of the Air Force  
The Frank J. Seiler Research Lab.  
United States Air Force Academy  
Colorado Springs, CO 80840-6528

Dr. Gary Bullard  
Pinnacle Research Institute, Inc.  
10432 N. Tantan Avenue  
Cupertino, CA 95014

Dr. J. O'M. Bockris  
Ementech, Inc.  
Route 5, Box 946  
College Station, TX 77840

Dr. Michael Binder  
Electrochemical Research Branch  
Power Sources Division  
U.S. Army Laboratory Command  
Fort Monmouth, New Jersey 07703-5000

Professor Martin Fleischmann  
Department of Chemistry  
University of Southampton  
Southampton, Hants, SO9 5NH UK