

MMIC FILE COPY


USAARL Report No. 88-8

2



AD-A200 431

**A Comparison of Two Computer Implemented
Psychophysical Procedures Applied to Real-ear
Attenuation Testing (ANSI S12.6-1984)**



By
William R. Nelson
and
Ben T. Mozo

Sensory Research Division

June 1988

Approved for public release; distribution unlimited

88 11 09 068

**United States Army Aeromedical Research Laboratory
Fort Rucker, Alabama 36362-5292**

Notice

Qualified requesters

Qualified requesters may obtain copies from the Defense Technical Information Center (DTIC), Cameron Station, Alexandria, Virginia 22314. Orders will be expedited if placed through the librarian or other person designated to request documents from DTIC.

Change of address

Organizations receiving reports from the U.S. Army Aeromedical Research Laboratory on automatic mailing lists should confirm correct address when corresponding about laboratory reports.

Disposition

Destroy this report when it is no longer needed. Do not return it to the originator.


Disclaimer

The views, opinions, and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy, or decision, unless so designated by other official documentation. Citation of trade names in this report does not constitute an official Department of the Army endorsement or approval of the use of such commercial items.

Human use

Human subjects participated in these studies after giving their free and informed voluntary consent. Investigators adhered to AR 70-25 and USAMRDC Reg 70-25 on Use of Volunteers in Research.

Reviewed:




BRUCE C. LEIBRECHT, Ph.D.
LTC, MS
Director, Sensory Research
Division



J.D. LaMOthe, Ph.D.
COL, MS
Chairman, Scientific
Review Committee

Released for publication:



DAVID H. KARNEY
Colonel, MC
Commanding

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE

REPORT DOCUMENTATION PAGE				Form Approved OMB No. 0704-0188	
1a. REPORT SECURITY CLASSIFICATION Unclassified			1b. RESTRICTIVE MARKINGS		
2a. SECURITY CLASSIFICATION AUTHORITY			3. DISTRIBUTION/AVAILABILITY OF REPORT Approved for public release; distribution unlimited		
2b. DECLASSIFICATION/DOWNGRADING SCHEDULE			5. MONITORING ORGANIZATION REPORT NUMBER(S)		
4. PERFORMING ORGANIZATION REPORT NUMBER(S) USAARL Report No. 88-8					
6a. NAME OF PERFORMING ORGANIZATION Sensory Research Division U.S. Army Aeromedical Resch Lab		6b. OFFICE SYMBOL (If applicable) SGRD-UAS-AS		7a. NAME OF MONITORING ORGANIZATION U.S. Army Medical Research and Development Command	
6c. ADDRESS (City, State, and ZIP Code) P.O. Box 577 Fort Rucker, AL 36362-5292		7b. ADDRESS (City, State, and ZIP Code) Fort Detrick, Frederick, MD 21701-5012			
8a. NAME OF FUNDING/SPONSORING ORGANIZATION		8b. OFFICE SYMBOL (If applicable)		9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER	
8c. ADDRESS (City, State, and ZIP Code)		10. SOURCE OF FUNDING NUMBERS			
		PROGRAM ELEMENT NO. 135		PROJECT NO. 3E162773A819	WORK UNIT ACCESSION NO. 050
11. TITLE (Include Security Classification) A comparison of computer implementations of two psychophysical procedures applied to real-ear attenuation testing (ANSI S12.6-1984)					
12. PERSONAL AUTHOR(S) William R. Nelson, Ben T. Mozo					
13a. TYPE OF REPORT Final		13b. TIME COVERED FROM _____ TO _____		14. DATE OF REPORT (Year, Month, Day) 1988 July	
15. PAGE COUNT 23					
16. SUPPLEMENTARY NOTATION					
17. COSATI CODES			18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)		
FIELD	GROUP	SUB-GROUP	Acoustics, audiometry, hearing, threshold, psychophysical procedure, auditory aquity.		
05	08				
06	04				
19. ABSTRACT (Continue on reverse if necessary and identify by block number) The application of computer technology to acoustical instrumentation has significantly increased the capability and flexibility of modern acoustical laboratories. The need to replace old recording attenuators used in real-ear sound attenuation testing with state-of-the-art instrumentation prompted the combination of a CMOS multiplying D/A converter chip (which can accurately and reliably attenuate an analog signal) and a table-top computer to control the D/A chip. The computer was also used to record the measurement of auditory threshold, perform statistical analysis, and permanently store data. The flexibility of computer technology allowed the choice of psychophysical procedure. Consideration was given to two such procedures, tracking and method of adjustment. This study was undertaken to determine if one of these procedures would produce faster, more accurate results. <i>Ker...</i>					
20. DISTRIBUTION/AVAILABILITY OF ABSTRACT ☒ UNCLASSIFIED/UNLIMITED ☐ SAME AS RPT. ☐ DTIC USERS			21. ABSTRACT SECURITY CLASSIFICATION		
22a. NAME OF RESPONSIBLE INDIVIDUAL Chief, Scientific Information Center			22b. TELEPHONE (Include Area Code) (205) 255-6907		22c. OFFICE SYMBOL SGRD-UAX-SI

DD Form 1473, JUN 86

Previous editions are obsolete.

SECURITY CLASSIFICATION OF THIS PAGE

UNCLASSIFIED

Table of contents

Introduction.....	3
Methods.....	3
Results.....	5
Conclusions.....	12
References.....	14
Manufacturers' list.....	15

List of tables

1. Means and standard deviations of attenuator dB settings for each psychophysical procedure for each frequency band.....	6
2. Mean, standard deviation, and correlation coefficient of attenuator dB settings obtained using two psychophysical procedures for all subjects, across frequencies and days.....	6
3. Absolute differences in average threshold measurements obtained from 4 vs 5 and 4 vs 6 dB criteria of acceptable range, by frequency and days (measured in dB).....	7
4. Cumulative percentage of successful completion of the threshold determination for day 1 for the indicated number of trials using a range criteria of 4, 5, and 6 dB.....	9
5. Cumulative percentage of successful completion of the threshold determination for day 2 for the indicated number of trials using a range criteria of 4, 5, and 6 dB.....	10
6. Cumulative percentage of successful completion of the threshold determination for day 3 for the indicated number of trials using a range criteria of 4, 5, and 6 dB.....	11
7. Average dB of real-ear attenuation obtained from the David-Clark 9AN/2 earmuff as measured using each psychophysical test procedure.....	12
8. Cumulative percentage of successful completion of the threshold determination for day 4 for the indicated number of trials using range criteria of 4, 5, and 6 dB.....	13



=====

This page left blank intentionally.

=====

Introduction

The use of computer technology in clinical audiometric equipment has become wide-spread. Microelectronics has revolutionized screening, clinical, and immittance audiometers and made clinical brain stem audiometry commonplace. However, the electromechanical recording attenuator used in real-ear attenuation testing of hearing protective devices (ANSI S12.6-1984) has not benefitted from these recent advances in instrument technology. Therefore, a CMOS Logarithmic D/A Converter chip which could be computer controlled and used as a programmable audio attenuator was used to replace the obsolete recording attenuator. The D/A chip was installed on a circuit board and interfaced to a tabletop computer via a parallel interface for control.

Since the new audio circuitry was controlled by a computer system, a choice of psychophysical procedures for threshold testing was possible. The tracking method described by Bekesy (1947) has been used at the U.S. Army Aeromedical Research Laboratory (USAARL), Fort Rucker, Alabama, in the past; however, Hirsh (1952) has observed, "Perhaps the easiest and quickest way to obtain a threshold measurement with an intelligent observer involves the method of adjustment." The relative merits of tracking and adjustment were assessed and it was decided to take advantage of the flexibility of computer technology by developing software programs for both. This study reports the comparison of results obtained from those two procedures.

Methods

Subjects

Ten college students with normal hearing were selected as subjects. They were required to have hearing thresholds for both ears no greater than 10 dB at test frequencies from 250 to 1000 Hertz and no greater than 20 dB at the higher test frequencies as measured on a standard audiometer (ANSI S3.6-1969).

Instrumentation

The auditory threshold and real-ear attenuation tests were accomplished in a custom-built audiometric examination room measuring 10' X 9'4" X 6'6" (l X w X h) located at USAARL. This room was modified to meet the reverberation characteristics specified in ANSI S12.6-1984. All tests were accomplished in a sound field consistent with that standard. No tests were made under earphones.

Signal intensity and linearity were calibrated to the test space as required by ANSI S12.6-1984. A plumb-bob was used to maintain the subject's head position in the calibrated test space.

The signals used in the test were one-third octave bands of noise with center frequencies at 125, 250, 500, 1K, 2K, 3.15K, 4K, 6.3K, and 8K Hz. The test signals were generated and controlled by the instrumentation shown in Figure 1. The noise generator (Bruel and Kjaer (B&K) Type 1405*) was set to deliver white noise to the band pass filter, B&K type 1618. The selected band of noise was input to the electronic switch, Grason-Stadler Type 1287B,* which was pulsed with a 1 Hz symmetrical square wave control signal. The rise and fall times of the electronic switch were adjusted to 30 milliseconds to exclude audible transients during on-off and off-on transitions of the test signal. Signal intensity was controlled with an Analog Devices CMOS Logarithmic D/A Converter, Model AD7111LN* and a B&K power amplifier, type 2706. Both the D/A converter and the filter were under program control of a Hewlett-Packard (HP) Table Top Computer, Model 9845B* via an HP model 98032-A* 16-bit parallel interface.

A multikey touch pad was interfaced to the computer and used by the subjects to control signal intensity. During the tracking sessions, only one key was required to indicate when the signal was heard. For the method of adjustment sessions, five keys were used. Four were used to control signal intensity as follows: fast increase, slow increase, fast decrease, slow decrease; and the last key to indicate the subject was at threshold. Data points were recorded in terms of attenuator settings.

Procedures

The design of this study follows the general case of repeated measures as discussed by Keppel (1973). To preclude any procedural bias, subjects with no experience in real-ear attenuation testing were selected to participate. All procedures used in this study comply with paragraph 3 of ANSI S12.6-1984. The same listeners were used for both the tracking task and the method of adjustment. Half of the subjects accomplished the tracking procedure first while the other half completed the method of adjustment procedure first.

The study was divided into two parts; first, the comparison of soundfield auditory threshold measurements using

* See manufacturers' list

the two procedures. For this part of the experiment, 6 threshold measurements were obtained on 3 different occasions for each procedure for a total of 18 threshold measurements for each subject for each method. The second part of the study involved the standard measurement of real-ear attenuation for a circumaural hearing protector (a David-Clark model 9AN/2 earmuff*) using each procedure. Compliant with the standard, three free-field and three attenuated threshold were measured for each procedure. These data also were collected on two separate occasions for a total of eight data collection sessions for each subject.

For the tracking method, the subject controlled the signal level as described by Bekesy (1947). The computer recorded 10 reversal points. The threshold level for each test frequency was calculated as the average of the attenuation settings at these 10 reversal points.

For the method of adjustment, the test signal was presented to the subject at a random intensity. The subject used the keypad to control signal intensity as described above and to indicate to the computer when his threshold was reached. Four threshold responses were recorded and tested against a range criterion of no more than 4 dB. If the four responses failed to meet this criterion, additional trials were administered until four successive responses fell within the 4 dB range. When the criterion was met, an average was calculated for the four accepted responses and that average was taken as the threshold for that subject at that frequency. The same procedure was followed for each test frequency.

The data acquired by both methods were stored on magnetic tape. Anecdotal comments made by the subjects about each procedure were noted.

Results

The means and standard deviations of the sound field threshold data for all subjects by frequency for each psychophysical procedure are summarized in Table 1. It should be noted that attenuator dB settings are arbitrary values which are dependent on the specific associated instrumentation. The thresholds are not adjusted to audiometric zero, but values are consistent between the two methods because the same instrumentation is used for both.

Table 1

Means and standard deviations* of attenuator dB settings
for each psychophysical procedure for each frequency band

One-third octave center frequencies	Method of adjustment		Tracking method	
	Mean	S.D.	Mean	S.D.
125 Hz	57.43	4.93	54.86	5.62
250 Hz	61.74	5.69	61.03	5.36
500 Hz	74.23	6.04	74.03	6.14
1000 Hz	76.28	4.94	77.23	5.20
2000 Hz	79.71	4.50	80.17	4.47
3150 Hz	81.90	2.82	82.74	4.20
4000 Hz	81.22	4.06	81.75	4.10
6300 Hz	74.75	3.70	74.90	4.63
8000 Hz	72.45	4.55	71.69	5.32

* Based on 180 threshold determinations per frequency.

A linear regression analysis of the threshold data for each procedure for each test session and for all sessions across subjects was completed and the results are in Table 2. These data were recorded in attenuator dB settings with no adjustment made for the between frequency differences in the sensitivity of the human ear. Had this been accomplished, the variance across frequencies would have been reduced and the standard deviations would have been substantially smaller. The high correlation between the two procedures is as expected.

Table 2

Mean, standard deviation, and correlation coefficient
of attenuator dB settings obtained using two psychophysical
procedures for all subjects, across frequencies and days

Day	Method of adjustment		Tracking method		Correlation coefficient
	Mean	S.D.	Mean	S.D.	
1	69.94	11.48	71.05	12.55	.97
2	71.40	11.48	70.56	11.83	.90
3	71.24	11.81	70.60	13.02	.94
1-3	70.93	11.57	70.71	12.43	.93

The raw data were reanalyzed to determine the effect of a less stringent range criterion for the method of adjustment. A comparison of the average differences in thresholds obtained when 5 dB or 6 dB criteria were used rather than the 4 dB criterion is contained in Table 3. The differences between the threshold averages obtained using the 5 dB and 6 dB vs 4 dB range criteria are well within the range of acceptable variability for auditory threshold determination (Hirsh, 1952).

Table 3

Absolute differences in average threshold measurements obtained from 4 vs 5 and 4 vs 6 dB criteria of acceptable range, by frequency and days (measured in dB)									
Third-octave test center frequency in Hertz									
	125	250	500	1000	2000	3150	4000	6300	8000
Day 1									
4 vs 5 dB									
Mean	.050	.088	.125	.000	.125	.200	.038	.113	.050
S.D.	.120	.181	.219	.151	.128	.267	.074	.083	.093
4 vs 6 dB									
Mean	.075	.125	.188	.000	.163	.263	.113	.175	.050
S.D.	.175	.183	.398	.169	.160	.297	.203	.175	.093
Day 2									
4 vs 5 dB									
Mean	.080	.030	.080	.010	.090	.090	.060	.000	.030
S.D.	.155	.116	.132	.099	.137	.137	.108	.047	.048
4 vs 6 dB									
Mean	.070	.100	.130	.010	.110	.140	.070	.090	.000
S.D.	.170	.189	.157	.129	.166	.158	.106	.185	.067
Day 3									
4 vs 5 dB									
Mean	.010	.020	.020	.020	.000	.040	.060	.030	.100
S.D.	.160	.114	.220	.103	.094	.097	.165	.116	.105
4 vs 6 dB									
Mean	.010	.000	.070	.010	.010	.030	.090	.040	.110
S.D.	.173	.133	.289	.152	.110	.134	.173	.126	.185
Days 1-3									
4 vs 5 dB									
Mean	.039	.043	.071	.011	.068	.104	.054	.043	.039
S.D.	.147	.135	.190	.113	.128	.179	.120	.096	.099
4 vs 6 dB									
Mean	.043	.071	.125	.007	.089	.136	.089	.096	.054
S.D.	.171	.172	.281	.144	.155	.215	.157	.167	.132

Tables 4, 5, and 6 demonstrate the relative efficiency of the 4 dB, 5 dB, and 6 dB criteria for acceptable ranges in terms of cumulative proportion of subjects who were able to complete the task in a given number of trials. The maximum number of trials required by any subject also is reported. As expected, the larger the criterion, the more quickly the task could be completed. The 6 dB criterion allowed completion of the task with many fewer trials per frequency while maintaining accuracy and reducing test time.

For the second part of the experiment, both procedures were used to test the real-ear attenuation of the same circumaural device, a David-Clark model 9AN/2 earmuff. Table 7 contains the mean attenuation and standard deviation values for each test frequency obtained from the two procedures. The contents of this table were compared using a t-test of significance at the .05 level of confidence. No significant difference was discovered between real-ear attenuation results measured with the two procedures at any frequency.

Table 4

Cumulative percentage of successful completion of the threshold determination for day 1
for the indicated number of trials using a range criteria of 4, 5, and 6 dB

Test frequency in hertz	Range criteria	4	5	6	7	8	9	10	Total trials required
125	4 dB	.62	.79	.82	.84	.89	.89	.92	13
	5 dB	.78	.88	.90	.92	.99	.99	.99	11
	6 dB	.83	.90	.93	.96	1.00			8
250	4 dB	.68	.80	.82	.87	.90	.93	.93	12
	5 dB	.85	.92	.94	.96	.99	.99	.99	11
	6 dB	.92	.99	.99	.99	.99	.99	.99	11
500	4 dB	.53	.70	.78	.86	.89	.94	.94	11
	5 dB	.78	.86	.89	.96	.98	1.00		9
	6 dB	.85	.93	.96	1.00				7
1000	4 dB	.58	.68	.80	.87	.87	.90	.95	20
	5 dB	.80	.87	.89	.92	.95	.97	.99	11
	6 dB	.87	.94	.96	.98	.98	.98	1.00	10
2000	4 dB	.53	.70	.80	.88	.91	.91	.94	20
	5 dB	.73	.86	.91	.98	1.00			8
	6 dB	.80	.90	.93	.98	1.00			8
3150	4 dB	.52	.65	.72	.85	.92	.94	.94	16
	5 dB	.72	.84	.91	.94	.99	.99	.99	13
	6 dB	.83	.93	.95	.98	.98	.98	.98	13
4000	4 dB	.62	.72	.80	.92	.97	.99	.99	12
	5 dB	.75	.83	.90	.97	.99	1.00		9
	6 dB	.88	.95	.98	1.00				7
6300	4 dB	.58	.73	.88	.95	.98	.98	.98	12
	5 dB	.80	.87	.95	.98	1.00			8
	6 dB	.83	.93	.96	1.00				7
8000	4 dB	.80	.92	.95	.98	1.00			8
	5 dB	.93	.98	1.00					6
	6 dB	.95	.98	1.00					6

Table 5

Cumulative percentage of successful completion of the threshold determination for day 2
for the indicated number of trials using a range criteria of 4, 5, and 6 dB

Test frequency in hertz	Range criteria	4	5	6	7	8	9	10	Total trials required
125	4 dB	.58	.71	.79	.84	.91	.94	1.00	10
	5 dB	.73	.88	.91	.96	.98	1.00		9
	6 dB	.92	.97	.97	.99	1.00			8
250	4 dB	.62	.72	.80	.90	.95	.98	.98	13
	5 dB	.77	.85	.90	.95	.98	1.00		9
	6 dB	.87	.95	.97	1.00				7
500	4 dB	.55	.70	.80	.87	.94	.97	.97	12
	5 dB	.77	.90	.95	1.00				7
	6 dB	.93	1.00						5
1000	4 dB	.50	.72	.77	.89	.94	.99	.99	11
	5 dB	.77	.85	.90	.95	.95	.98	.98	11
	6 dB	.85	.93	.96	1.00				7
2000	4 dB	.68	.81	.93	.96	.96	.96	.98	13
	5 dB	.87	.95	.98	.98	.98	.98	1.00	10
	6 dB	.95	.98	1.00					6
3150	4 dB	.62	.77	.84	.89	.94	.97	.99	12
	5 dB	.78	.88	.95	.97	1.00			8
	6 dB	.88	.95	.98	1.00				7
4000	4 dB	.73	.86	.91	.94	1.00			8
	5 dB	.93	.96	1.00					6
	6 dB	.97	.99	1.00					6
6300	4 dB	.65	.82	.90	.93	.93	.95	.97	13
	5 dB	.77	.92	.95	.97	.99	.99	.99	13
	6 dB	.88	.95	.98	1.00				7
8000	4 dB	.77	.94	.97	1.00				7
	5 dB	.88	.96	.96	1.00				7
	6 dB	.98	1.00						5

Table 6

Cumulative percentage of successful completion of the threshold determination for day 3
for the indicated number of trials using a range criteria of 4, 5, and 6 dB

Test frequency in hertz	Range criteria	4	5	6	7	8	9	10	Total trials required
125	4 dB	.72	.79	.92	.94	.96	.96	.98	14
	5 dB	.73	.88	.91	.96	.98	1.00		9
	6 dB	.93	.95	.98	.98	1.00			8
250	4 dB	.78	.85	.92	.92	.99	1.00		9
	5 dB	.95	.98	.98	.98	1.00			8
	6 dB	.97	1.00						5
500	4 dB	.70	.80	.88	.93	.95	.97	.97	13
	5 dB	.82	.90	.95	.97	1.00			8
	6 dB	.88	.95	.97	.99	1.00			8
1000	4 dB	.65	.77	.89	.91	.91	.93	.96	15
	5 dB	.82	.89	.96	.96	.96	.96	.98	13
	6 dB	.93	.96	.98	.98	.98	.98	.98	12
2000	4 dB	.65	.78	.86	.93	.96	.96	1.00	10
	5 dB	.87	.97	.99	1.00				7
	6 dB	.93	.98	1.00					6
3150	4 dB	.65	.73	.91	.93	.95	.97	.97	13
	5 dB	.77	.87	.94	.99	1.00			8
	6 dB	.87	.95	.98	1.00				7
4000	4 dB	.77	.87	.89	.97	.99	.99	.99	13
	5 dB	.92	.97	1.00					6
	6 dB	.95	.98	1.00					6
6300	4 dB	.73	.86	.86	.91	.94	.97	.97	15
	5 dB	.88	.95	.95	1.00				7
	6 dB	.90	.95	.95	1.00				7
8000	4 dB	.52	.69	.74	.86	.93	.98	1.00	14
	5 dB	.78	.85	.88	.93	.98	1.00		9
	6 dB	.92	.92	.94	.97	1.00			8

Table 7

Average dB of real-ear attenuation obtained from
the David-Clark 9AN/2 earmuff as measured
using each psychophysical test procedure

Third-octave test center frequencies	Method of adjustment		Tracking method	
	Mean	S.D.	Mean	S.D.
125 Hz	16.04	2.95	15.69	3.10
250 Hz	20.84	4.59	22.35	4.05
500 Hz	27.50	4.17	28.99	4.25
1000 Hz	29.22	4.39	30.87	3.25
2000 Hz	26.87	2.90	28.60	3.81
3150 Hz	24.60	2.65	25.25	3.82
4000 Hz	26.71	2.28	28.17	3.31
6300 Hz	27.62	3.54	29.40	3.46
8000 Hz	27.66	4.49	28.56	4.11

Table 8 contains cumulative percentages of successful trials for each criterion obtained during testing of the David-Clark 9AN/2 earmuff. Again, the greater efficiency of the 6 dB criterion is demonstrated clearly.

Conclusions

It can be concluded from the results of this study that: 1) computer implementation of both procedures is practical; 2) microcircuits can be adapted for laboratory applications; 3) the psychophysical procedures of tracking and method of adjustment yield similar results for threshold tasks; 4) range criteria of 4, 5, or 6 dB all yield similar threshold and attenuation results; 5) the 6 dB criterion is more efficient since fewer trials are required to complete the test; and 6) subjects report a preference for the method of adjustment.

This is consistent with Hirsh's (1952) observations.

Table 8

Cumulative percentage of successful completion of the threshold determination for day 4
for the indicated number of trials using a range criteria of 4, 5, and 6 dB

Test frequency in hertz	Range criteria	4	5	6	7	8	9	10	Total trials required
125	4 dB	.60	.78	.79	.85	.95	.95	.95	16
	5 dB	.88	.91	.94	.96	.98	1.00		9
	6 dB	.95	.97	.97	.99	.99	1.00		9
250	4 dB	.73	.85	.92	.95	.97	.97	.99	11
	5 dB	.83	.93	.95	.97	.97	.97	.99	11
	6 dB	.97	1.00						5
500	4 dB	.67	.80	.92	.94	.96	.99	.99	11
	5 dB	.83	.96	.96	.96	.96	1.00		9
	6 dB	.97	.99	.99	.99	.99	1.00		9
1000	4 dB	.70	.80	.83	.88	.91	.93	.95	14
	5 dB	.90	.92	.92	.97	.97	.99	1.00	10
	6 dB	.93	.98	1.00					6
2000	4 dB	.77	.79	.81	.84	.94	.96	.98	15
	5 dB	.90	.90	.93	.93	.98	1.00		9
	6 dB	.95	.95	.97	.97	1.00			3
3150	4 dB	.60	.68	.85	.87	.92	.94	.94	15
	5 dB	.77	.89	.92	.94	.96	.98	1.00	10
	6 dB	.90	.97	.97	.99	1.00			8
4000	4 dB	.80	.82	.89	.89	.94	.97	1.00	10
	5 dB	.95	.97	.99	.99	1.00			8
	6 dB	.98	.98	1.00					6
6300	4 dB	.72	.79	.86	.89	.92	.92	.97	15
	5 dB	.85	.92	.99	1.00				7
	6 dB	.90	.95	.98	1.00				7
8000	4 dB	.78	.86	.88	.88	.93	.96	.96	18
	5 dB	.95	.97	.97	.99	1.00			8
	6 dB	.98	.98	.98	.98	1.00			8

References

- American National Standards Institute. 1969. American National Standard Specification for Audiometers. New York: American National Standards Institute. ANSI S3.6-1969.
- American National Standards Institute. 1984. American National Standard Method for the Measurement of Real-Ear Attenuation of Hearing Protectors. New York: American Institute of Physics. ANSI S12.6-1984.
- Von Békésy, G. 1947. A new audiometer. Acta otolaryngologica. Volume 35: 411-422.
- Hirsh, I. J. 1952. The measurement of hearing. New York: McGraw-Hill.
- Keppel, G. 1973. Design and analysis: a researcher's handbook. Englewood Cliffs, New Jersey: Prentice-Hall, Inc.

Manufacturer's list

Analog Devices
One Technology Way
Norwood, MA 02062-9106

Bruel and Kjaer Instruments Incorporated
185 Forest Street
Marborough, MA 017752

David-Clark Co., Inc.
P.O. Box 155
Worcester, MA 01613

Grason-Stadler
56 Winthrop Street
Concord, MA 01742

Hewlett-Packard Company
2000 South Park Place
Atlanta, GA 30348

Initial distribution

Commander
U.S. Army Natick Research
and Development Center
ATTN: Documents Librarian
Natick, MA 01760

Naval Submarine Medical
Research Laboratory
Medical Library, Naval Sub Base
Box 900
Groton, CT 05340

Commander/Director
U.S. Army Combat Surveillance
& Target Acquisition Lab
ATTN: DELCS-D
Fort Monmouth, NJ 07703-5304

Commander
10th Medical Laboratory
ATTN: Audiologist
APO NEW YORK 09180

Commander
Naval Air Development Center
Biophysics Lab
ATTN: G. Kydd
Code 60B1
Warminster, PA 18974

Naval Air Development Center
Technical Information Division
Technical Support Detachment
Warminster, PA 18974

Dr. E. Hendler
Human Factors Applications, Inc.
295 West Street Road
Warminster, PA 18974

Under Secretary of Defense
for Research and Engineering
ATTN: Military Assistant
for Medical and Life Sciences
Washington, DC 20301

Commander
U.S. Army Research Institute
of Environmental Medicine
Natick, MA 01760

U.S. Army Avionics Research
and Development Activity
ATTN: SAVAA-P-TP
Fort Monmouth, NJ 07703-5401

U.S. Army Research and Development
Support Activity
Fort Monmouth, NJ 07703

Chief, Benet Weapons Laboratory
LCWSL, USA ARRADCOM
ATTN: DRDAR-LCB-TL
Watervliet Arsenal, NY 12189

Commander
Man-Machine Integration System
Code 602
Naval Air Development Center
Warminster, PA 18974

Commander
Naval Air Development Center
ATTN: Code 6021 (Mr. Brindle)
Warminster, PA 18974

Commanding Officer
Naval Medical Research
and Development Command
National Naval Medical Center
Bethesda, MD 20014

Director
Army Audiology and Speech Center
Walter Reed Army Medical Center
Washington, DC 20307-5001

COL Franklin H. Top, Jr., MD
Walter Reed Army Institute
of Research
Washington, DC 20307-5100

HQ DA (DASG-PSP-0)
Washington, DC 20310

Naval Research
Laboratory Library
Code 1433
Washington, DC 20375

Harry Diamond Laboratories
ATTN: Technical Infor-
mation Branch
2800 Powder Mill Road
Adelphi, MD 20783-1197

U.S. Army Materiel Systems
Analysis Agency
ATTN: Reports Processing
Aberdeen Proving Ground
MD 21005-5017

U.S. Army Ordnance Center
and School Library
Building 3071
Aberdeen Proving Ground,
MD 21005-5201

U.S. Army Environmental Hygiene
Agency Laboratory
Building E2100
Aberdeen Proving Ground,
MD 21010

Technical Library
Chemical Research
and Development Center
Aberdeen Proving Ground,
MD 21010-5423

Commander
U.S. Army Institute
of Dental Research
Walter Reed Army Medical Center
Washington, DC 20307-5300

Naval Air Systems Command
Technical Air Library 950D
Rm 278, Jefferson Plaza II
Department of the Navy
Washington, DC 20361

Naval Research Laboratory Library
Shock and Vibration Infor-
mation Center, Code 5804
Washington, DC 20375

Director
U.S. Army Human Engineer-
ing Laboratory
ATTN: Technical Library
Aberdeen Proving Ground,
MD 21005-5001

Commander
U.S. Army Test
and Evaluation Command
ATTN: AMSTE-AD-H
Aberdeen Proving Ground,
MD 21005-5055

Director
U.S. Army Ballistic
Research Laboratory
ATTN: DRXBR-OD-ST Tech Reports
Aberdeen Proving Ground,
MD 21005-5066

Commander
U.S. Army Medical Research
Institute of Chemical Defense
ATTN: SGRD-UV-AO
Aberdeen Proving Ground,
MD 21010-5425

Commander
U.S. Army Medical Research
and Development Command
ATTN: SGRD-RMS (Ms. Madigan)
Fort Detrick, Frederick, MD 21701

Commander
U.S. Army Medical Research
Institute of Infectious Diseases
Fort Detrick, Frederick,
MD 21701

Director, Biological
Sciences Division
Office of Naval Research
600 North Quincy Street
Arlington, VA 22217

Commander
U.S. Army Materiel Command
ATTN: AMCDE-S (CPT Broadwater)
5001 Eisenhower Avenue
Alexandria, VA 22333

Commandant
U.S. Army Aviation
Logistics School
ATTN: ATSQ-TDN
Fort Eustis, VA 23604

U.S. Army Training
and Doctrine Command
ATTN: ATCD-ZX
Fort Monroe, VA 23651

Structures Laboratory Library
USARTL-AVSCOM
NASA Langley Research Center
Mail Stop 266
Hampton, VA 23665

Naval Aerospace Medical
Institute Library
Bldg 1953, Code 102
Pensacola, FL 32508

Command Surgeon
U.S. Central Command
MacDill Air Force Base
FL 33608

Air University Library
(AUL/LSE)
Maxwell AFB, AL 36112

Commander
U.S. Army Medical Bioengineering
Research and Development Lab
ATTN: SGRD-UBZ-I
Fort Detrick, Frederick,
MD 21701

Defense Technical
Information Center
Cameron Station
Alexandria, VA 22313

U.S. Army Foreign Science
and Technology Center
ATTN: MTZ
220 7th Street, NE
Charlottesville, VA 22901-5396

Director,
Applied Technology Laboratory
USARTL-AVSCOM
ATTN: Library, Building 401
Fort Eustis, VA 23604

U.S. Army Training
and Doctrine Command
ATTN: Surgeon
Fort Monroe, VA 23651-5000

Aviation Medicine Clinic
TMC #22, SAAF
Fort Bragg, NC 28305

U.S. Air Force Armament
Development and Test Center
Eglin Air Force Base, FL 32542

U.S. Army Missile Command
Redstone Scientific
Information Center
ATTN: Documents Section
Redstone Arsenal, AL 35898-5241

U.S. Army Research and Technology
Laboratories (AVSCOM)
Propulsion Laboratory MS 302-2
NASA Lewis Research Center
Cleveland, OH 44135

AFAMRL/HEX
Wright-Patterson AFB, OH 45433

University of Michigan
NASA Center of Excellence
in Man-Systems Research
ATTN: R. G. Snyder, Director
Ann Arbor, MI 48109

John A. Dellinger, MS, ATP
University of Illinois-
Willard Airport
Savoy, IL 61874

Project Officer
Aviation Life Support Equipment
ATTN: AMCPO-ALSE
4300 Goodfellow Blvd.
St. Louis, MO 63120-1798

Commander
U.S. Army Aviation
Aviation Systems Command
ATTN: DRSAB-ED
4300 Goodfellow Blvd
St. Louis, MO 63120

Commanding Officer
Naval Biodynamics Laboratory
P.O. Box 24907
New Orleans, LA 70189

U.S. Army Field Artillery School
ATTN: Library
Snow Hall, Room 14
Fort Sill, OK 73503

Commander
U.S. Army Health Services Command
ATTN: HSOP-SO
Fort Sam Houston, TX 78234-6000

U.S. Air Force Institute
of Technology (AFIT/LDEE)
Building 640, Area B
Wright-Patterson AFB, OH 45433

Henry L. Taylor
Director, Institute of Aviation
University of Illinois-
Willard Airport
Savoy, IL 61874

Commander
U.S. Army Aviation
Systems Command
ATTN: DRSAB-WS
4300 Goodfellow Blvd
St. Louis, MO 63120-1798

Commander
U.S. Army Aviation
Systems Command
ATTN: SGRD-UAX-AL (MAJ Lacy)
4300 Goodfellow Blvd., Bldg 105
St. Louis, MO 63120

U.S. Army Aviation
Systems Command
Library and Information
Center Branch
ATTN: DRSAB-DIL
4300 Goodfellow Blvd
St. Louis, MO 63120

Federal Aviation Administration
Civil Aeromedical Institute
CAMI Library AAC 64D1
P.O. Box 25082
Oklahoma City, OK 73125

Commander
U.S. Army Academy
of Health Sciences
ATTN: Library
Fort Sam Houston, TX 78234

Commander
U.S. Army Institute
of Surgical Research
ATTN: SGRD-USM (Jan Duke)
Fort Sam Houston, TX 78234-6200

Director of Professional Services
AFMSC/GSP
Brooks Air Force Base, TX 78235

U.S. Army Dugway Proving Ground
Technical Library
Bldg 5330
Dugway, UT 84022

U.S. Army Yuma Proving Ground
Technical Library
Technical Library
Yuma, AZ 85364

AFFTC Technical Library
6520 TESTG/ENXL
Edwards Air Force Base,
CAL 93523-5000

Commander
Code 3431
Naval Weapons Center
China Lake, CA 93555

Aeromechanics Laboratory
U.S. Army Research
and Technical Labs
Ames Research Center,
M/S 215-1
Moffett Field, CA 94035

Sixth U.S. Army
ATTN: SMA
Presidio of San Francisco,
CA 94129

Commander
U.S. Army Aeromedical Center
Fort Rucker, AL 36362

Directorate
of Combat Developments
Bldg 507
Fort Rucker, AL 36362

U.S. Air Force School
of Aerospace Medicine
Strughold Aeromedical Library
Documents Section, USAFSAM/TSK-4
Brooks Air Force Base, TX 78235

Dr. Diane Damos
Department of Human Factors
ISSM, USC
Los Angeles, CA 90089-0021

U.S. Army White Sands
Missile Range
Technical Library Division
White Sands Missile Range,
NM 88002

U.S. Army Aviation Engineering
Flight Activity
ATTN: SAVTE-M (Tech Lib)
Stop 217
Edwards Air Force Base,
CA 93523-5000

U.S. Army Combat Developments
Experimental Center
Technical Information Center
Bldg 2925
Fort Ord, CA 93941-5000

Commander
Letterman Army Institute
of Research
ATTN: Medical Research Library
Presidio of San Francisco,
CA 94129

Director
Naval Biosciences Laboratory
Naval Supply Center, Bldg 844
Oakland, CA 94625

Commander
U.S. Army Aviation Center
and Fort Rucker
ATTN: ATZQ-CDR
Fort Rucker, AL 36362

Directorate
of Training Development
Bldg 502
Fort Rucker, AL 36362

Chief
Army Research Institute
Field Unit
Fort Rucker, AL 36362

Commander
U.S. Army Safety Center
Fort Rucker, AL 36362

U.S. Army Aircraft Development
Test Activity
ATTN: STEBG-MP-QA
Cairns AAF
Fort Rucker, AL 36362

Chief
Defence and Civil Institute
of Environmental Medicine
P.O. Box 2000
ATTN: Director MLSD
Downsview, Ontario Canada M3M 3B9

Staff Officer, Aerospace Medicine
RAF Staff, British Embassy
3100 Massachusetts Avenue, NW
Washington, DC 20008

Canadian Society
of Aviation Medicine
c/o Academy of Medicine, Toronto
ATTN: Ms. Carment King
288 Bloor Street West
Toronto, Canada M5S 1V8

Canadian Forces
Medical Liaison Officer
Canadian Defence Liaison Staff
2450 Massachusetts Avenue, NW
Washington, DC 20008

Officer Commanding
School of Operational
and Aerospace Medicine
DCIEM P.O. Box 2000
1133 Sheppard Avenue West
Downsview, Ontario, Canada M3M 3B9

Chief
Human Engineering Laboratory
Field Unit
Fort Rucker, AL 36362

Commander
U.S. Army Aviation Center
and Fort Rucker
ATTN: ATZQ-T-ATL
Fort Rucker, AL 36362

President
U.S. Army Aviation Board
Cairns AAF
Fort Rucker, AL 36362

USA Medical Liaison Officer
U.S. Embassy Box 54
ATTN: USADO-AMLO
FPO New York 09509

HQ, Department of the Army
Office of The Surgeon General
British Medical Liaison Officer
DASG-ZX/COL M. Daly
5109 Leesburg Pike
Falls Church, VA 22401-3258

Canadian Airline Pilot's
Association
MAJ (Retired) J. Soutendam
1300 Steeles Avenue East
Brampton, Ontario, Canada L6T 1A2

Commanding Officer
404 Squadron CFB Greenwood
Greenwood, NS, Canada B0P 1N0

National Defence Headquarters
101 Colonel By Drive
ATTN: DPM
Ottawa, Ontario, Canada K1A 0K2

Commanding Officer
Headquarters, RAAF Base
Point Cook Victoria,
Australia 3029

Netherlands Army Liaison Office
Buildingg 602
Fort Rucker, AL 36362

British Army Liaison Office
Building 602
Fort Rucker, AL 36362

Canadian Army Liaison Office
Building 602
Fort Rucker, AL 36362

German Army Liaison Office
Buildingg 602
Fort Rucker, AL 36362

French Army Liaison Office
Building 602
Fort Rucker, AL 36362