

4

THIS IS A COPY

AD _____

REPORT NO. T4-90

RELIABILITY OF PATTERN REVERSAL EVOKED POTENTIALS (PREP) USING THE NICOLET PATHFINDER II

AD-A219 664

U S ARMY RESEARCH INSTITUTE
OF
ENVIRONMENTAL MEDICINE
Natick, Massachusetts

FEBRUARY 1990

DTIC
ELECTE
MAR 26 1990
S B D



Approved for public release distribution unlimited

UNITED STATES ARMY
MEDICAL RESEARCH & DEVELOPMENT COMMAND

The findings in this report are not to be construed as an official Department of the Army position, unless so designated by other authorized documents.

DISPOSITION INSTRUCTIONS

Destroy this report when no longer needed.

Do not return to the originator.

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

1a. REPORT SECURITY CLASSIFICATION Unclassified		1b. RESTRICTIVE MARKINGS	
2a. SECURITY CLASSIFICATION AUTHORITY		3. DISTRIBUTION/AVAILABILITY OF REPORT Approved for public release; distribution is unlimited	
2b. DECLASSIFICATION/DOWNGRADING SCHEDULE			
4. PERFORMING ORGANIZATION REPORT NUMBER(S) T4-90		5. MONITORING ORGANIZATION REPORT NUMBER(S)	
6a. NAME OF PERFORMING ORGANIZATION US Army Research Institute of Environmental Medicine	6b. OFFICE SYMBOL (if applicable) SGRD-UE-HP	7a. NAME OF MONITORING ORGANIZATION US Army Medical Research & Development Command	
6c. ADDRESS (City, State, and ZIP Code) Natick, MA 01760-5007		7b. ADDRESS (City, State, and ZIP Code) Ft. Detrick, Frederick, MD 21701-5012	
8a. NAME OF FUNDING/SPONSORING ORGANIZATION Same as 6a.	8b. OFFICE SYMBOL (if applicable)	9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER	
8c. ADDRESS (City, State, and ZIP Code) Same as 6c.		10. SOURCE OF FUNDING NUMBERS	
		PROGRAM ELEMENT NO. 63002A	PROJECT NO. 3M2630- 02D995
		TASK NO. AE	WORK UNIT ACCESSION NO. DA305221
11. TITLE (Include Security Classification) Reliability of Pattern Reversal Evoked Potentials (PREP) Using the NICOLET Pathfinder II			
12. PERSONAL AUTHOR(S) W.J. Tharion, D.J. McMenemy, and T.M. Rauch			
13a. TYPE OF REPORT Technical Report	13b. TIME COVERED FROM 11/88 TO 2/90	14. DATE OF REPORT (Year, Month, Day) 1990 February 21	15. PAGE COUNT 40
16. SUPPLEMENTARY NOTATION			
17. COSATI CODES		18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)	
FIELD	GROUP	Evoked Potential, Pattern Reversal Evoked Potential, Evoked Response, Nicolet Pathfinder, JWGD ³ MILPERF Tag Level 1, PREP Baseline Norms	
19. ABSTRACT (Continue on reverse if necessary and identify by block number)			
<p>Evoked potentials (EPs) are emerging as a useful tool to determine the functional integrity of the central and peripheral nervous systems. The Pattern Reversal Evoked Potential (PREP) provides an assessment of the functional integrity of the visual nervous system pathway.</p> <p>Signal averaging systems such as the Nicolet Pathfinder II extract the EP from the background electroencephalogram. Because different laboratories may have slightly different techniques and environmental conditions, each laboratory first must establish normative values for the system before research studies can be undertaken.</p> <p>The purposes of this study were to assess the reliability of our system for the collection of PREPs and to establish laboratory norms. Twenty-four males and eleven females served as volunteer research subjects. Subjects were administered two trials for each eye on each</p>			
20. DISTRIBUTION/AVAILABILITY OF ABSTRACT <input checked="" type="checkbox"/> UNCLASSIFIED/UNLIMITED <input type="checkbox"/> SAME AS RPT. <input type="checkbox"/> DTIC USERS		21. ABSTRACT SECURITY CLASSIFICATION Unclassified	
22a. NAME OF RESPONSIBLE INDIVIDUAL Terry M. Rauch, MAJ, MS		22b. TELEPHONE (Include Area Code) (508) 651-4195	22c. OFFICE SYMBOL SGRD-UE-HP

19. ABSTRACT (Continued)

of two days. Conditions remained constant for all trials. Using the International Electrode Placement System, surface electrodes were placed at C_2 , O_2 , O_1 , O_2 and F_{p_2} . Two sets of 100 pattern reversals were presented at the rate of 1.7/sec. A 32 square checkerboard pattern on a 12 inch TV monitor was the stimulus used.

The latencies of the two negative waveforms $N75$ and $N145$ as well as the positive peak $P100$ were assessed for replicability. Amplitudes from $N75-P100$ and from $P100-N145$ were also examined. Gender differences were noted. Means were calculated and established to use as norms for the new laboratory.

A repeated measures analysis of variance (ANOVA) and an intraclass reliability analysis were performed. The ANOVA assessed differences over days and between males and females. Intraclass reliability assessed error variances due to trials, days, and subjects. Latencies and amplitudes proved reliable and replicable for recording sites $F_{p_2}-O_2$ and $F_{p_2}-O_2$. Females had significantly shorter $P100$ latencies and greater amplitudes than males.

The results from this study are comparable to other laboratories engaged in EP research. The Nicolet Pathfinder II is considered to be a reliable system in our laboratory for collecting PREPs.

Brain evoked responses have traditionally been used for neurodiagnostic purposes in the clinical setting. A novel utilization of evoked potential technology enables the systematic correlation of changes in the performance of cognitive and/or sensorimotor tasks to changes in the amplitude and latency of neuronal signals in specific pathways within the brain. Such performance assessment technology represents the only window to the human central nervous system with direct application to the performance of military skills.

The views, opinions and/or findings contained in this report are those of the authors and should not be construed as an official Department of the Army position, policy or decision, unless so designated by other official documentation.

Human subjects participated in this study after giving their free and informed voluntary consent. Investigators adhered to AR 70-25 and USAMRDC Regulation 70-25 on Use of Volunteers in Research.

Accession For	
NTIS GRA&I	<input checked="" type="checkbox"/>
DTIC TAB	<input type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification	
By	
Distribution/	
Availability Codes	
Dist	Avail and/or Special
A-1	



ACKNOWLEDGEMENTS

We would like to thank Lynn Sleeper of Harvard University for providing valuable statistical assistance, Rebecca Kittredge for her assistance in the data reduction, and Everett Harman for review of the manuscript.

Approved for public release
Distribution unlimited

AD

TECHNICAL REPORT
NO.

RELIABILITY OF PATTERN REVERSAL EVOKED POTENTIALS (PREP)
USING THE NICOLET PATHFINDER II

W. J. Tharion, D. J. McMenemy, T.M. Rauch

February 1990

US Army Research Institute of Environmental Medicine
Natick, Massachusetts 01760-5007

FOREWORD

The following project was conducted under the guidance of the Tri-Service Joint Working Group on Drug Dependent Degradation in Military Performance (JWGD3 MILPERF) as part of the Task Area Group (TAG) Level I. The main purpose of the TAG Level I is to identify adverse drug effects on neurological functions in order to provide guidance to other performance related TAG Levels. One goal of TAG Level I is the development of an automated, standardized and clinically relevant assessment of nervous system integrity. This will be achieved through the creation of the Neurophysiological Performance Assessment Battery (NP-PAB), consisting of a set of eight evoked potential protocols. Before the NP-PAB can be fully implemented, standardization of the test procedures must be accomplished. Then validation of the NP-PAB with two classes of antihistamines will proceed, using the standardization procedures, by a network of laboratories. This will result in a common archive for JWGD3 MILPERF related data.

Several different evoked potential assessment systems are in use by the laboratories in the JWGD3 MILPERF network. Standardization of the procedures will insure that similar results can be produced by different systems in different laboratory settings. The Health and Performance Division at the US Army Research Institute of Environmental Medicine was requested to participate in this validation effort by assessing three of the standardized procedures on the Nicolet Pathfinder II. Testing the ability of this system and laboratory setting to validate previous findings is the first step in the effort towards the standardization of the NP-PAB. A database of normal values for the

evoked potential laboratory at USARIEM will also be established. The second step, validation of the NP-PAB with two classes of antihistamines, may then proceed.

TABLE OF CONTENTS

	<u>Page</u>
List of Figures	vi
List of Tables	vii
Abstract	viii
Introduction	1
Method	5
Subjects	5
Procedure	5
Analysis	7
Results	11
Stability of Measures	11
Consistency of Measures	12
Gender Effects for Measures	12
Discussion	21
Summary	23
References	24
Distribution List	26

LIST OF FIGURES

	<u>Page</u>
Figure 1 PREP waveform with the standard latencies N75, P100 and N145	4
Figure 2 Replicability of PREP waveform on Day 1	9
Figure 3 Replicability of PREP waveform on Day 2	10

LIST OF TABLES

	<u>Page</u>	
Table 1	Latencies and amplitudes from recording site $F_{pz}-O_1$ for the left eye	13
Table 2	Latencies and amplitudes from recording site $F_{pz}-O_2$ for the left eye	14
Table 3	Latencies and amplitudes from recording site $F_{pz}-O_2$ for the left eye	15
Table 4	Latencies and amplitudes from recording site $F_{pz}-O_1$ for the right eye	16
Table 5	Latencies and amplitudes from recording site $F_{pz}-O_2$ for the right eye	17
Table 6	Latencies and amplitudes from recording site $F_{pz}-O_2$ for the right eye	18
Table 7	Intraclass reliability analysis for PREP latencies for left and right eyes	19
Table 8	Intraclass reliability analysis for PREP amplitudes for left and right eyes	20

ABSTRACT

Evoked potentials (EPs) are emerging as a useful tool to determine the functional integrity of the central and peripheral nervous systems. The Pattern Reversal Evoked Potential (PREP) provides an assessment of the functioning of the visual nervous system pathway.

Signal averaging systems such as the Nicolet Pathfinder II extract the EP from the background electroencephalogram. Because different laboratories may have slightly different techniques and environmental conditions, each laboratory first must establish normative values for the system before research studies can be undertaken.


The purposes of this study were to assess the reliability of our system for collection of PREPs and to establish laboratory norms. Twenty-four males and eleven females served as volunteer research subjects. Subjects were administered two trials for each eye on each of two days. Conditions remained constant for all trials. Using the International Electrode Placement System, surface electrodes were placed at C₂, O₂, O₁, O₂, and F_{pz}. Two sets of 100 pattern reversals were presented at the rate of 1.7/sec. A 32 square checkerboard pattern on a 12 inch TV monitor was used as the stimulus source.

The latencies of the two negative waveforms N75 and NI45 as well as the positive peak P100 were assessed for replicability. Amplitudes from N75-P100 and from P100-NI45 were also examined. Gender differences were noted. Means were calculated and established to use as norms for the EP laboratory.

A repeated measures analysis of variance (ANOVA) and an intraclass reliability analysis were performed. The ANOVA assessed differences over days and between males and females. Intraclass reliability assessed error variances due to trials, days, and subjects. Latencies and amplitudes proved reliable and replicable for recording sites $F_{pz}-O_2$, and $F_{pz}-O_2$. Females had significantly shorter $P100$ latencies and greater amplitudes than males.

The results from this study are comparable to other laboratories engaged in EP research. The Nicolet Pathfinder II is considered to be a reliable system in our laboratory for collecting PREPs.

Brain evoked responses have traditionally been used for neurodiagnostic purposes in the clinical setting. A novel utilization of evoked potential technology enables the systematic correlation of changes in the performance of cognitive and/or sensorimotor tasks to changes in the amplitude and latency of neuronal signals in specific pathways within the brain. Such performance assessment technology represents the only window to the human central nervous system with direct application to the performance of military skills. (EP)



Introduction

Electroencephalograms (EEG's) have long been used to record spontaneous and/or background activity of the brain. Differences in electrical potentials resulting from the ionic current flow across membranes, produce the different waves found in EEG's. Evoked responses obtained directly from the eye were first reported by Holmgren in 1865 (Celesia, 1985). The stimuli for these responses were bright flashes and the resultant evoked potentials (EPs) were several hundred microvolts (μV) in amplitude. This type of EP is called an electroretinogram (ERG).

Recently, EPs have emerged as a useful tool in assessing functional integrity of the central and peripheral nervous systems. Since the voltage from the visual EP ([flash] visual evoked response (VER) and pattern reversal evoked potential (PREP)) obtained from recordings on the scalp is very small (.5 to 20 μV) compared to background EEG voltage (up to 200 μV depending on the EEG rhythm) it is necessary to extract the EP from the background EEG (Best & Taylor, 1973). The evoked potential is "time-locked" to the stimulus, occurring at a specific time after stimulus presentation. It has been found that latencies of responses to these stimuli are nearly constant not only with the same subject across time, but also between different non-impaired subjects. Since the EP is the only aspect of the EEG which is stimulus dependent, the EP can be extracted from other brain activity through signal averaging. Successive evoked responses are digitized and added to the previous responses. After each addition, the sums are divided by the number of responses collected to produce a running average until the desired number of

responses have been collected. Successive averages are used to obtain a clean, well defined signal.

Several signal-averaging systems have been designed specifically for the collection of EPs. Attention to the technical specifications and proper calibration of the system help to insure that the system is reliable. However, slight variations may still exist between different types of systems, and even between different units of the same system. Further variations can occur in procedure or laboratory locations. Many researchers (Chiappa, 1983; Colon, Visser, deWeerd and Zonnerveldt, 1983; and Spehlman, 1985) strongly recommend that a new EP laboratory establish a normative database for each procedure to be used. Evoked potential data from other laboratories may be used initially as a reference standard. However, once normative data from the laboratory have been established, that data should be used in future studies conducted in that laboratory. To make comparisons with other laboratories, the normative values should replicate those of the reference laboratory. Spehlman (1985) recommends that 95 percent of the subjects tested in the new laboratory fall within the limits derived from the reference laboratory before the results should be considered replicated.


The visual EP provides a measure of integrity of the visual pathway (optic nerve function) and neuronal conduction velocity within the central nervous system structures concerned with visual information processing (AEEGS, 1984). Originally, visual EPs were obtained by stimulating with bright light flashes. A flash stimulus is still used with certain populations (e.g. infants, mentally impaired and coma patients) because these individuals cannot fixate on a TV monitor which

is necessary when using a pattern reversal stimulus. Using a flash does not require patient cooperation. It is not the stimulus of choice, however, because as Ciganek (1975) states: "though most intensively studied, the [flash] visual evoked response (VER) is very difficult to interpret." In addition, differences between what is normal and what is not has been difficult to establish (Celesia, 1985). Hubel and Wiesel (1959, 1974) found the visual system responds with a high specificity of the cortical neurons when stimulated by shapes. Over the past decade these findings have led to stimulation of the visual system using specific patterns. Most frequently used is the visual presentation of a reversing pattern of black and white checks. This type of stimulation is referred to as the pattern reversal evoked potential (PREP). The PREP wave has three major components. The most important being a positive peak occurring at approximately 100 msec (known as P1 or P100) after stimulus presentation. The latencies of the two negative peaks or "valleys" of normal subjects are N75, (N1), and N145, (N2), (AEEGS, 1984). Figure 1 is an example of a PREP waveform whose origin is in the visual cortex. As is the the case with most EPs the exact origin is undetermined. The primary visual cortex on the medial surface as well as the secondary visual cortex on the lateral surface are thought to contribute but to an unknown extent (Spehlman, 1985). In general, cortical EPs are due largely to the spatial and temporal summation of post synaptic potentials generated at the membrane of nerve cell bodies and dendrites in response to the stimulus (Spehlman, 1985). It is believed that P100 results from primary activation of the neurons in the striate and prestriate regions of the occipital cortex (Broadman's

FIGURE 1

PREP WAVEFORM KEY COMPONENTS

$\overline{N75}$, $\overline{P100}$, $\overline{N145}$



cortical area 19) and also from subsequent thalamocortical volleys (Chippa, 1983).

The purpose of this study was twofold. The test-retest reliability of a newly acquired Nicolet Pathfinder II (Nicolet Biomedical Instruments: Madison WI) in the environment of USARIEM's Evoked Potential Laboratory was assessed. Secondly, a normative PREP database for the laboratory was established.

METHOD

Subjects

Subjects were 24 male and 11 female USARIEM civilian and military personnel, between the ages of 20 and 39 years with normal or correctable vision (20/20 Snellen). The P100 latency of subjects accepted for the database was required to fall within 2.5 standard deviation units of the norms determined by our reference laboratories (Colon, et al., 1983; and Chiappa, 1983).

Procedure

Testing occurred over two days following the same procedure. The PREP does not vary significantly over a few hours, days, or even several months in normal subjects. Stringent scheduling of the two test days was not necessary, however subjects were tested with two to seven days in between sessions. Each test session took approximately 20 minutes. The basic procedure used to collect PREP data in our laboratory was drawn directly from the procedure protocol published by Nicolet Biomedical

Instruments (1987) and is also the procedure standardized by the JWGD3 MILPERF Level I TAG (Reeves et al., 1990). All data was collected using the Nicolet Pathfinder II (Madison, WI), a neurodiagnostic system for the collection of evoked potentials. The Nicolet Pathfinder II is a self-contained system, including the electrode board, amplifier, stimuli and stimuli controls, memory blocks for recording and displaying waveforms, and waveform analysis.

The 10-20 International Electrode Placement System was utilized to insure that electrodes were placed in the same locations over repeated trials (Jasper, 1958). Surface electrodes were applied to the scalp at five sites: the vertex C_z ; three occipital sites; O_1 , O_2 , and O_z ; and at F_{pz} . F_{pz} served as the reference electrode with O_1 , O_2 , and O_z as the active electrodes. C_z was the ground electrode.

To minimize interference in the recorded signal the electrode site was prepared using an abrasive solution, Omni-Prep (D.O. Weaver & Co.; Aurora, CO) to remove oils and dead skin. Medi-Trace EEG Sol (Graphic Controls Corp.; Buffalo, NY) electrode cream was then used to adhere the electrode to the prepared site and provide clear electrical conduction.

The resistance to current flow, known as impedance, is in part a measure of the quality of the electrode-scalp interface. Impedance of the electrode-scalp connection, measured with an impedance meter on the Nicolet Pathfinder II, was required to be between one and five kilohms to avoid excessive artifact.

Two trials were performed monocularly on each eye. The left eye was always recorded first. The non-stimulated eye was covered with an eye patch (covered with a sterile gauze pad) while data were being

collected on the stimulated eye. Stimulation was presented to the eye in the form of a reversing checkerboard pattern on a Nicolet NIC 1015 Pattern Reversal Stimulator with a 12 inch screen monitor positioned at eye level one meter in front of the subject. If the subject used corrective lenses, the lenses were worn during testing. The stimulus pattern consisted of a 32 square black and white checkerboard. The pattern was set to reverse 1.7 times per second. Two sets of 100 responses each were collected from the left eye, with the subject closing his/her eyes for two minutes between sets for rest. The patch was then switched to the left eye and the procedure repeated for two more sets of 100 responses from the right eye. Sensitivity was set to 50 uV and bandpass filters set at 1 HZ (low bandpass) and 100 HZ (high bandpass), to filter out undesired signals. The Pathfinder intrinsically filters out signals which appear to be muscle artifacts, obtaining 100 acceptable responses per trial. Subjects relaxed in an easy chair in an upright sitting position during data collection. They were instructed to fixate on a paper adhesive dot in the center of the video screen and refrain from excessive blinking. All data collection occurred in the dimly lit EP laboratory at USARIEM. Environmental conditions such as lighting, temperature, and stimulus intensity remained similar between trials.

Analysis

For PREP waveform analysis, the most important measure is peak latency of P_{100} . Also important is the amplitude from valley to peak for the preceding N_{75} wave to the P_{100} peak (AEEGS, 1984). Abnormal waves are considered to be those where a response is absent or there is a

prolonged P100 peak latency, i.e. greater than 2.5 standard deviation units from the mean. No wave amplitude or latencies of the other waves, N75 and NI45, need be considered for a normal wave if P100 is normal. The latter measures are only examined if the P100 latency is greater than 2.5 standard deviation units from the mean of our reference laboratory. Values for all waves were obtained directly from the Pathfinder using Nicolet software and a moving cursor which allows for marking waves on the Pathfinder CRT screen. Reliability of the various latency and amplitude measures is necessary before further analysis is appropriate. Proven reliability allows for future analysis of the effects of various treatments. Reliability estimates deal with both stability and consistency of measures (Kroll, 1966). Stability is a measure of resistance to influence from repeated testing and can be assessed with a repeated measures analysis of variance (ANOVA). Consistency shows that the same factor is being measured on different occasions. Intraclass reliability is a technique that assesses consistency via analysis of variance (Lindquist, 1953). The mean square values for the different sources of error are used to calculate the error terms which is described by Lindquist (1953) in detail. This technique is preferred over interclass Pearson product-moment correlation in its ability to distinguish the various sources of error within an observed score (Kroll, 1966). Intraclass reliability is expressed as follows:

$$R = \frac{\sigma^2 \text{ True Score}}{\sigma^2 \text{ True Score} + \sigma^2 \text{ Error Due to Trials} + \sigma^2 \text{ Error Due to Days}}$$

FIGURE 2

REPLICABILITY OF PREP WAVEFORM ON DAY 1





FIGURE 3

REPLICABILITY OF PREP WAVEFORM ON DAY 2



Results

Figure 1 shows a standard PREP wave recorded during this study. Key wave components (N75, P100, N145) are noted. Amplitude N75-P100 is from the lowest point of the N75 valley to the apex of the P100 peak. The same peak to valley measurement is used for amplitude P100-N145. Figures 2 and 3 illustrate a clear, representative example of the reproducibility of the PREP waveform of one subject (female) on two different days. Trials 1 and 2 in Figures 2 and 3 are each shown separately as well as superimposed on one another to demonstrate the reproducibility of the waves.

Stability of Measures

The ANOVA analysis for stability showed only one significant treatment effect on latency. For wave P100 from recording site $F_{p_2}-O_1$ on the left eye; ($F(1,33) = 5.94, p < .02$), Day 2's latency was significantly longer than Day 1's. There were significant day effects for both amplitude measures from both eyes N75-P100 and P100-N145 for recording site $F_{p_2}-O_1$ only; left eye; (N75-P100, $F(1,33) = 5.35, p < .03$; P100-N145, $F(1,33) = 7.32, p < .01$) right eye; (N75-P100, $F(1,33) = 7.54, p < .01$; P100-N145, $F(1,33) = 6.01, p < .02$). Amplitudes were also greater on Day 2. These were the only treatment effects observed for amplitude. Actual latencies and amplitudes for the different waves are listed in Tables 1 through 6 with respect to stimulated eye (left or right) and recording site ($F_{p_2}-O_1, O_2,$ or O_2').

Consistency of Measures

Consistency for PREP latencies are listed in Table 7. Reliability scores for PREP latencies for waves P100 and N145 range from .44 to .79. Reliability for wave N75 was much lower ranging from .12 to .45.

Reliability scores for amplitudes (Table 8) ranged from .48 to .85 for the different recording sites.

Gender Effects

There were significant differences between males and females in the P100 wave for both eyes and for all recording sites: left eye; (F_{Pz-O_1} , $F(1,33) = 8.81$, $p < .01$; F_{Pz-O_2} , $F(1,33) = 6.54$, $p < .01$; and F_{Pz-O_2} , $F(1,33) = 7.78$, $p < .01$), and right eye; (F_{Pz-O_1} , $F(1,33) = 4.61$, $p < .04$; F_{Pz-O_2} , $F(1,33) = 7.70$, $p < .01$; and F_{Pz-O_2} , $F(1,33) = 7.60$, $p < .01$). Males had longer latencies in all cases than females. Latency values for the P100 wave are listed in Tables 1 through 6. There were no gender differences for either N75 or N145 latencies for any recording sites for either eye.

Gender effects existed on amplitude for both eyes at all three recording sites for both waves (N75-P100 and P100-N145). Where significant differences existed between males and females, females always exhibited greater amplitude. There were significant differences for wave N75-P100 for right eye; (F_{Pz-O_1} , $F(1,33) = 5.55$, $p < .02$; F_{Pz-O_2} , $F(1,33) = 12.62$, $p < .01$; F_{Pz-O_2} , $F(1,33) = 7.57$, $p < .01$), and left eye; (F_{Pz-O_1} , $F(1,33) = 5.28$, $p < .01$; F_{Pz-O_2} , $F(1,33) = 6.61$, $p < .01$). The only significant difference for wave P100-N145 was for right eye; (F_{Pz-O_1} , $F(1,33) = 4.91$, $p < .03$).

TABLE 1
 Latencies and Amplitudes for Recording
 Site F_{p2}-O₁ for the Left Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	74.94	±	7.48	71.61	±	3.90
** Wave P100	103.71	±	7.45	97.18	±	4.88
Wave N145	141.88	±	11.64	138.19	±	13.92
AMPLITUDE						
* AMP N75-P100	5.31	±	2.64	7.32	±	2.52
AMP P100-N145	6.68	±	3.38	8.50	±	3.26

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	72.42	±	5.18	75.36	±	6.20
* Wave P100	100.24	±	5.41	103.06	±	6.96
Wave N145	140.84	±	11.69	140.61	±	13.88
AMPLITUDE						
* AMP N75-P100	5.47	±	2.17	6.41	±	3.02
** AMP P100-N145	6.79	±	2.94	7.72	±	3.87

* p < .05

** p < .01

TABLE 2
Latencies and Amplitudes for Recording
Site F_{P2}-O₂ for the Left Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	74.95	±	5.22	73.13	±	2.67
** Wave P100	102.15	±	6.46	97.08	±	5.29
Wave NI45	140.72	±	10.09	138.51	±	11.49
AMPLITUDE						
** AMP N75-P100	7.47	±	3.01	10.12	±	3.05
AMP P100-NI45	9.23	±	4.39	10.67	±	3.80

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	72.81	±	4.41	74.94	±	3.48
Wave P100	99.35	±	5.31	101.75	±	6.43
Wave NI45	137.66	±	9.49	139.25	±	12.09
AMPLITUDE						
AMP N75-P100	8.11	±	2.96	8.49	±	3.08
AMP P100-NI45	9.74	±	4.57	9.62	±	3.83

* p < .05

** p < .01

TABLE 3
 Latencies and Amplitudes for Recording
 Site F_{P2}-O₂ for the Left Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	75.13	±	5.64	72.16	±	3.39
** Wave P100	103.50	±	7.35	97.43	±	5.75
Wave N145	143.09	±	10.39	138.92	±	15.36
AMPLITUDE						
AMP N75-P100	5.88	±	2.78	7.84	±	3.47
AMP P100-N145	7.60	±	4.14	9.39	±	4.67

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	73.86	±	5.06	74.54	±	3.98
Wave P100	100.61	±	6.26	102.57	±	6.89
Wave N145	140.38	±	12.57	141.30	±	13.67
AMPLITUDE						
AMP N75-P100	6.48	±	2.92	6.51	±	3.07
AMP P100-N145	8.14	±	4.14	8.18	±	4.55

* p < .05

** p < .01

TABLE 4
Latencies and Amplitudes for Recording
Site F_{p2}-O₁ for the Right Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	74.74	±	7.75	72.43	±	6.08
* Wave P100	103.89	±	7.39	98.94	±	7.26
Wave N145	140.31	±	10.24	139.49	±	12.65
AMPLITUDE						
* AMP N75-P100	5.12	±	2.06	6.81	±	2.61
* AMP P100-N145	6.18	±	2.65	8.49	±	3.91

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	72.46	±	7.91	75.58	±	6.54
Wave P100	102.12	±	7.78	102.54	±	6.91
Wave N145	140.24	±	10.24	139.87	±	11.57
AMPLITUDE						
** AMP N75-P100	5.17	±	1.84	6.14	±	2.63
* AMP P100-N145	6.50	±	2.66	7.32	±	3.44

* p < .05

** p < .01

TABLE 5
Latencies and Amplitudes for Recording
Site F_{pz}-O₂ for the Right Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	74.23	±	5.60	73.80	±	3.33
** Wave P100	102.01	±	5.19	97.55	±	4.33
Wave NI45	141.43	±	8.51	138.10	±	7.97
AMPLITUDE						
** AMP N75-P100	6.99	±	2.57	10.44	±	3.72
AMP P100-NI45	8.77	±	3.69	11.62	±	5.35

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	73.14	±	5.42	75.04	±	4.34
Wave P100	100.17	±	5.40	101.04	±	4.44
Wave NI45	140.39	±	8.66	140.38	±	8.01
AMPLITUDE						
AMP N75-P100	8.01	±	3.20	8.13	±	2.65
AMP P100-NI45	9.87	±	4.49	9.46	±	3.95

* p < .05

** p < .01

TABLE 6
Latencies and Amplitudes for Recording
Site F_{z2}-O₂ for the Right Eye

	GENDER					
	MALES			FEMALES		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	74.00	±	5.83	71.74	±	3.99
** Wave P100	103.57	±	7.10	97.40	±	5.70
Wave N145	143.01	±	9.61	139.77	±	13.02
AMPLITUDE						
** AMP N75-P100	5.71	±	2.38	8.40	±	3.75
AMP P100-N145	7.19	±	3.53	10.38	±	6.19

	DAY					
	DAY 1			DAY 2		
	MEAN	±	S.D.	MEAN	±	S.D.
LATENCY						
Wave N75	72.72	±	5.91	78.86	±	4.59
Wave P100	100.33	±	7.41	101.94	±	6.08
Wave N145	141.46	±	10.30	142.51	±	11.06
AMPLITUDE						
AMP N75-P100	8.01	±	3.20	6.46	±	2.69
AMP P100-N145	8.14	±	4.27	8.24	±	4.45

* p < .05

** p < .01

TABLE 7
Intraclass Reliability Analysis for PREP Latencies
for Left and Right Eyes

LATENCIES	σ^2 Trials	σ^2 Days	σ^2 True Score	R
LEFT EYE				
Wave F _{PZ} -O ₁				
N75	20.60	38.00	7.98	.12
PI00	21.00	11.65	36.99	.53
NI45	55.90	34.15	99.39	.52
Wave F _{PZ} -O ₂				
N75	10.02	2.85	15.32	.54
PI00	8.86	16.33	28.17	.53
NI45	20.10	7.95	104.99	.79
Wave F _{PZ} -O ₂				
N75	23.90	4.20	12.05	.30
PI00	10.21	20.86	34.61	.53
NI45	36.70	49.80	102.58	.54
RIGHT EYE				
Wave F _{PZ} -O ₁				
N75	31.20	37.70	11.70	.15
PI00	15.00	22.50	34.83	.48
NI45	52.60	0.15	91.79	.64
Wave F _{PZ} -O ₂				
N75	17.52	4.44	14.18	.39
PI00	20.62	0.00	18.23	.47
NI45	34.70	6.40	48.20	.44
Wave F _{PZ} -O ₂				
N75	19.61	2.51	18.13	.45
PI00	12.52	8.45	39.93	.66
NI45	30.60	10.50	91.90	.69

TABLE 8
 Intraclass Reliability Analysis for PREP Amplitudes
 For Left and Right Eyes

AMPLITUDES	σ^2 TRIALS	σ^2 DAYS	σ^2 TRUE SCORE	R
LEFT EYE				
Wave $F_{pz}-O_1$				
N75-PI00	1.65	1.89	5.60	.61
PI00-NI45	2.15	1.34	10.42	.75
Wave $F_{pz}-O_2$				
N75-PI00	2.31	2.45	8.12	.63
PI00-NI45	2.39	3.05	14.49	.73
Wave $F_{pz}-O_2$				
N75-PI00	1.64	.89	8.15	.76
PI00-NI45	4.02	.12	16.49	.80
RIGHT EYE				
Wave $F_{pz}-O_1$				
N75-PI00	1.31	2.41	3.50	.48
PI00-NI45	1.36	3.59	7.93	.61
Wave $F_{pz}-O_2$				
N75-PI00	1.70	1.93	8.89	.71
PI00-NI45	2.71	1.78	17.06	.79
Wave $F_{pz}-O_2$				
N75-PI00	2.38	1.03	19.98	.85
PI00-NI45	1.52	1.18	7.94	.75

Discussion

Both latencies and amplitudes have been shown to be stable over time with the exception of recording site $F_{pz}-O_1$. Latencies of waves P100 and N145 have been shown to be consistent measures from recording sites $F_{pz}-O_2$, and $F_{pz}-O_2$. Error variances due to trials and days were relatively small compared to true score variance. Because all variances are small, (standard deviations were less than 10% of the means), and there were only two trials each per day on two separate days, the R values are very reasonable and the system can be considered reliable. Furthermore, the demonstrated stability over time for recording sites $F_{pz}-O_2$, and $F_{pz}-O_2$, indicate these measures may be held in confidence when testing for PREPs in this laboratory. Most of the variance was due to inter-subject differences. While amplitudes vary much more than latencies, amplitudes were reliable at all recording sites except $F_{pz}-O_1$, where instability was evidenced over days for amplitude as well as for latencies.

The following reasons have been suggested as causes for small amounts of variability in waves: 1) small changes in arousal that were unnoticed during testing, 2) subjects closing their eyes or losing focus on the dot before being alerted by the technician, 3) pupil size, 4) body temperature, or 5) small differences in electrode placement (Osken, Chiappa, and Gill, 1987).

Females have previously been reported to have shorter PREP wave latencies probably due to smaller brain and head sizes (Celesia, 1985; Halliday and McDonald, 1981; Shearer and Dustman, 1980; and Stockard, Hughes, and Sharbrough, 1979). In agreement with these findings, our

laboratory found significant differences for the P_{100} latencies between males and females regardless of recording site or eye tested. Females always had shorter latencies. However, males tended to show a much greater variability in their latency measures. The differences may be because of greater variability in head sizes of males compared to females. In addition, amplitudes for females were significantly greater, probably due to a shorter distance from the electrical potential generator to the recording site.

Finally, all our values on waves P_{100} fall within the 2.5 standard deviation units of our reference laboratory 89.55 - 115.08 msec (Colon, et al., 1983 and Chiappa, 1983). Therefore, the values for the different recording sites in Tables 1-6 will serve as the normative values in future studies conducted in USARIEM's evoked potential laboratory. Separate normative values for males and females will be used because of the significant differences observed on wave P_{100} .

Recording site $F_{pz}-O_z$ has been shown to be the most reliable PREP site. It is recommended that PREP research in this laboratory utilize this site using Grand Averages (Nicolet Biomedical Instruments, 1987). Grand Averages combine two or more PREP recordings into one wave, to give the equivalent of 200 averages. This procedure is done internally by the Pathfinder to produce a cleaner, more representative wave. Recordings should only be combined into Grand Averages when there is high replication in the separate waveform components. Recording site $F_{pz}-O_z$ is the site reported most frequently in the literature regarding PREP research.

Summary

The PREP is emerging as a useful tool to assess functioning of the visual pathway. The Nicolet Pathfinder II is one signal averaging system which can be used to extract the EP from the EEG. This study assessed the ability of our laboratory procedures using the Nicolet Pathfinder II to validate results from other laboratories, measured reliability and provided normative values as reference points for future PREP research to be performed by USARIEM. The system was found to provide stable and consistent recordings of both PREP latencies and amplitudes for sites $F_{Pz}-O_2$, and $F_{Pz}-O_1$. Recording site $F_{Pz}-O_1$ showed some test-retest variability. The study has established PREP norms by gender for the USARIEM laboratory which may be found in Tables 1 through 6.

References

1. Allison, T., C.C. Wood, and W.P. Goff (1983). Brain stem auditory, pattern reversal visual and short-latency somatosensory evoked potentials: latencies in relation to age, sex, and brain and body size. American Journal of Electroencephalography Technology, 55, 619-636.
2. American Journal of Electroencephalographic Society (1984). Guidelines for clinical evoked potential studies. Journal of Clinical Neurophysiology, 1, 3-53.
3. Best, C.H. and N.B. Taylor (1973). Best and Taylor's physiological basis of medical practice. J.R. Brobeck (Editor), Baltimore: Williams and Wilkins Company.
4. Celesia, G.G. (1985). Visual evoked responses. In J. Owens and H. Davis (Editors). Evoked potential testing: clinical applications. Orlando: Grune and Stratton Inc. pp. 1-54.
5. Chiappa, K.H. (1983). Evoked potentials in clinical medicine. New York: Raven Press.
6. Ciganek, L. (1975). Visual evoked responses. In W. Storm Van Leeuwen, F.M. Lopes da Silva, and A. Kamp (Editors). Evoked responses volume 8 part A, handbook of EEG and clinical neurophysiology. Amsterdam: Elsevier, pp. 35-59.
7. Colon, E., S. Visser, J. deWeerd, and A. Zannerveldt (1983). Evoked potential manual: a practical guide to clinical applications. Boston: Martinus Nijhoff Publishers.
8. Halliday, A.M. and W.I. McDonald (1981). Visual evoked potentials. In: E. Stalberg and R.R. Young (Editors), Neurology I: Clinical Neurophysiology. London: Butterworth Publishing, pp. 228-258.
9. Hubel, D.M. and T.N. Wiesel (1959). Receptive fields of single neurons in the cat's striate cortex. Journal of Physiology (London). 148, 574-591.
10. Hubel, D.M. and T.N. Wiesel (1974). Uniformity of monkey striate cortex, a parallel relationship between field size, scatter and magnification factor, Journal of Comparative Neurology, 158, 295-306.
11. Jasper, H.H. (1958). Report of the committee on methods of clinical examinations in electroencephalography. Electroencephalography and Clinical Neurology, 10, 370-375.
12. Kroll, W. (1966). Reliability theory and research decision in selection of a criterion score. Research Quarterly, 38, 412-419.
13. Lindquist, E.F. (1953). Design and Analysis of Experiments in Psychology and Education. Boston: Houghton Mifflin, Co.

14. Osken, B.S., K.H. Chiappa and E. Gill (1987). Normal temporal variability of the P100. Electroencephalography and Clinical Neurophysiology, 68, 153-156.

15. Reeves, D.L., R.R. Stanny, D.F. Wilson, R.I. Herning, W. Pickworth, K.F. VanOrden, and J.A. Caldwell. The JWGD3 MILPERF Level I Neurophysiological Performance Assessment Battery: NP-PAB. Technical Report in Progress, 1990.

16. Shearer, D.E. and Dustman (1980). The pattern reversal evoked potential: The need for laboratory norms. American Journal of Electroencephalography and Technology, 20, 185-200.

17. Spehlman, R. (1985). Evoked potential primer: visual, auditory, and somatosensory evoked potentials in clinical diagnosis. Boston: Butterworth Publishing.

18. Stockard, J.J., J.F. Hughes, and F.W. Sharbrough (1979). Visually evoked potentials to electronic pattern reversal: latency variations with gender, age, and technical factors. American Journal of Electroencephalography and Technology, 19, 171-204.

DISTRIBUTION LIST

2 Copies to:

Commander
U.S. Army Medical Research and Development Command
ATTN: SGRD-OP
Fort Detrick
Frederick, MD 21701-5012

2 Copies to:

Commander
U.S. Army Medical Research and Development Command
ATTN: SGRD-PLE
Fort Detrick
Frederick, MD 21701-5012

2 Copies to:

Commander
U.S. Army Medical Research and Development Command
ATTN: SGRD-PLC
Fort Detrick
Frederick, MD 21701-5012

1 Copy to:

Commandant
Academy of Health Sciences, U.S. Army
ATTN: AHS-COM
Fort Sam Houston, TX 78234-6100

1 Copy to:

Stimson Library
Academy of Health Sciences, U.S. Army
ATTN: Chief Librarian
Bldg. 2840, Room 106
Fort Sam Houston, TX 78234-6100

1 Copy to:

Director, Biological Sciences Division
Office of Naval Research - Code 141
800 N. Quincy Street
Arlington, VA 22217

1 Copy to:

Commanding Officer
Naval Medical Research and Development Command
NMC-NMR/Bldg. 1
Bethesda, MD 20814-5044

1 Copy to:

Office of Undersecretary of Defense for Acquisition
ATTN: Director, Defense Research and Engineering
Deputy Undersecretary for Research & Advanced
Technology
(Environmental and Life Sciences)
Pentagon, Rm. 3D129
Washington, D.C. 20301-3100

1 Copy to:

Dean
School of Medicine
Uniformed Services University of The Health Sciences
4301 Jones Bridge Road
Bethesda, MD 20814-4799

2 Copies to:

Commander
U.S. Army Medical Research Institute of Chemical
Defense
Aberdeen Proving Ground, MD 21010-5425

2 Copies to:

Commander
U.S. Army Chemical Research, Development and
Engineering Center
Aberdeen Proving Ground, MD 21010-5423

2 Copies to:

Commandant
U.S. Army Chemical School
Fort McClellan, AL 36205-5020

2 Copies to:

Commander

U.S. Air Force School of Aerospace Medicine
Brooks Air Force Base, TX 78235-5000

2 Copies to:

Commander
Naval Health Research Center
P.O. Box 85122
San Diego, CA 92138-9174

2 Copies to:

Commander
U.S. Army Biomedical Research and Development
Laboratory
Fort Detrick
Frederick, MD 21701-5010

2 Copies to:

Commander
U.S. Army Medical Materiel Development Activity
Fort Detrick
Frederick, MD 21701-5009

2 Copies to:

U.S. Army Military Liaison Officer to DCIEM
1133 Sheppard Avenue W.
P.O. Box 2000
Downsview, Ontario
CANADA M3M 3B9

1 Copy to:

Commandant
Walter Reed Army Institute of Research
Walter Reed Army Medical Center
Acting Director for Research Management
ATTN: SGRD-UWZ-C
Washington, D.C. 20307-5100

1 Copy to:

Commander
U.S. Army Environmental Hygiene Agency
Aberdeen Proving Ground, MD 21010-5422

1 Copy to:

Commander
Walter Reed Army Institute of Research
ATTN: SGRD-UWI/Dr. Hegge
Washington, D.C. 20307-5100