

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.				
1. AGENCY USE ONLY (Leave blank)	2. REPORT DATE 1-3-95	3. REPORT TYPE AND DATES COVERED Final		
4. TITLE AND SUBTITLE A Kilohertz Femtosecond Ti:Sapphire Regenerative Amplifier			5. FUNDING NUMBERS	
6. AUTHOR(S) N. Peyghambarian and B. McGinnis			8. PERFORMING ORGANIZATION REPORT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Optical Sciences Center University of Arizona Tucson, AZ 85721				
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) U. S. Army Research Office P. O. Box 12211 Research Triangle Park, NC 27709-2211			10. SPONSORING/MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES The view, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy, or decision, unless so designated by other documentation.				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited.			12b. DISTRIBUTION CODE	
13. ABSTRACT (Maximum 200 words) To have capabilities in the near-infrared, we proposed to develop a Ti:S regenerative amplifier operating at one kilohertz. This amplifier was developed to amplify the pulses from our femtosecond Ti:S laser oscillator to the 2-3 μJ /pulse energy level. This type of energy in a femtosecond pulse allowed us to generate a broadband ($\approx 1000 \text{ \AA}$) in the near-infrared for spectroscopic studies of III-V-based semiconductors. DTIC QUALITY INSPECTED 4				
14. SUBJECT TERMS Ti:S regenerative amplifier			15. NUMBER OF PAGES 2	
			16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT UNCLASSIFIED	18. SECURITY CLASSIFICATION OF THIS PAGE UNCLASSIFIED	19. SECURITY CLASSIFICATION OF ABSTRACT UNCLASSIFIED	20. LIMITATION OF ABSTRACT UL	

19950216 017

We proposed to build a kilohertz femtosecond titanium-doped sapphire (Ti:S) regenerative amplifier system. The recent development of femtosecond laser pulses from a Ti:S laser oscillator provided us with a new stable source for femtosecond spectroscopy of near-infrared semiconductors. While visible laser dyes have provided good results for femtosecond laser systems and broadband continuum generation, infrared pulse generation from laser dyes has proven less reliable, with a limited tuning range. We built a system capable of producing modest energy pulses ($2-3 \mu\text{J}$) at a kilohertz repetition rate for higher average power and better signal-to-noise statistics. We incorporated the latest developments in all solid-state pumped lasers to produce an efficient compact system with improved stability, reliability and longevity. This amplified energy was sufficient to generate a broadband source ($\approx 1000 \text{ \AA}$) of femtosecond duration, which is necessary for studying the absorption features of III-V semiconductor structures. We performed pulse propagation studies in MQW waveguide structures. The tunability of the Ti:S laser system allowed us to study the effects of propagation of femtosecond pulses in waveguides over a broad spectral range. We intend to further study the gain dynamics of a variety of bulk and multiple quantum well semiconductor laser diodes. Since the range of gain spectra varies tremendously with growth composition and structure of the laser diode, it is necessary to have a wavelength flexible femtosecond source to continue these studies.

Diagram illustrating the optical layout of a Q-switched Nd:YAG laser resonator. The main resonator consists of a flat high reflector and a $\text{Ti:Al}_2\text{O}_3$ saturable absorber. A pump beam enters from the top right, passing through a mode-matching telescope and a $f = 10\text{ cm}$ focussing lens. A regenerative loop is formed by a glass flat mirror, a mode-matching telescope, an optical isolator, and a variable $\lambda/2$ plate, all connected to the main resonator via a Pockels cell and a polarizer. The regenerative output is shown at the bottom left.

Figure 1. Schematic of Ti:S regenerative amplifier.

Accession For	
NTIS CRA&I	☒
DTIC TAB	☐
Unannounced	☐
Justification _____	
By _____	
Distribution /	
Availability Codes	
Dist	Avail and/or Special
A-1	

