


ST1  
TECHNICAL MEMORANDUM REPORT WCRD 53-46

APPROVAL VIBRATION TESTING OF LOW PRESSURE, OXYGEN  
CHECK VALVES FOR USE IN LOW PRESSURE, AIRCRAFT, OXYGEN SYSTEM

Prepared By  
Aero Medical Laboratory  
Directorate of Research



29 May 1953

Project No. S-2

"DTIC USERS ONLY"

Accession For	
NTIS CRA&I	<input type="checkbox"/>
DTIC TAB	<input checked="" type="checkbox"/>
Unannounced	<input type="checkbox"/>
Justification .....	
By .....	
Distribution /	
Availability Codes	
Dist	Avail and/or Special
12	


Wright Air Development Center  
Air Research and Development Command  
United States Air Force  
Wright-Patterson Air Force Base, Ohio

19950824 033

DTIC QUALITY INSPECTED 8

ATI No. 200803 ST  
ASTIA FILE COPY

# DISCLAIMER NOTICE



THIS DOCUMENT IS BEST  
QUALITY AVAILABLE. THE COPY  
FURNISHED TO DTIC CONTAINED  
A SIGNIFICANT NUMBER OF  
PAGES WHICH DO NOT  
REPRODUCE LEGIBLY.

## REPORT NOTICES

When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

The information furnished herewith is made available for study upon the understanding that the Government's proprietary interests in and relating thereto shall not be impaired. It is desired that the Judge Advocate Office, WCJ, Wright Air Development Center, Wright-Patterson Air Force Base, Ohio, be promptly notified of any apparent conflict between the Government's proprietary interests and those of others.

Wright Air Development Center  
Air Research and Development Command  
United States Air Force  
Wright-Patterson Air Force Base, Ohio

Technical Memorandum Report  
WCRD 53-46  
29 May 1953

Aero Medical Laboratory  
Directorate of Research  
Project No. S-2

APPROVAL VIBRATION TESTING OF LOW PRESSURE, OXYGEN,  
CHECK VALVES FOR USE IN LOW PRESSURE, AIRCRAFT, OXYGEN SYSTEM

I. PURPOSE

To report the results of vibration tests conducted on six low pressure oxygen check valves submitted to the Aero Medical Laboratory by A. Schrader's Son, Brooklyn, New York. The above valves were new and from an approved source.

II. FACTUAL DATA

1. The check valves were submitted by A. Schrader's Son, manufacturer of the subject valves. The valves were from an approved source and were approved valves under Specification AN-V-15a. The valves were submitted for qualification testing in accordance with the vibration paragraph of Specification MIL-V-7908. This test is not required in the old AN-V-15a Specification.

2. The following low pressure, oxygen, check valves were subjected to the vibration test:

AN6030-1 - Style A  
AN6031-1 - Style B  
AN6032-1 - Style C  
AN6033-1 - Style D  
AN6036-1 - Style G  
AN6037-1 - Style H

Since the low pressure, oxygen, check valves have been approved under Specification AN-V-15a, the tests listed in the vibration paragraph of Specification MIL-V-7908 were conducted in order to qualify the valves under Specification MIL-V-7908.

3. The valves were vibrated for a period of six hours at a frequency of from 300 to 3000 cycles per minute. A constant pressure of 450 psig was used on the valves.

4. After completion of the vibration test, the valves were checked for leaks with the pressure varying from 30 psig to 500 psig. The valves were immersed in clean water and no bubbles indicating leaks were evident at any pressure.

Technical Memorandum Report

WCRD 53-46

29 May 1953

5. The check valves were then disassembled and the valve seat and valve were checked. The valve seat material was of an approved silicone type. The valve poppet was of the approved design. These parts showed no wear.

6. Water-pumped nitrogen conforming to Specification MIL-N-6011 was used as the source of pressure in the above tests.

III. CONCLUSIONS

Since the valves were from an approved source and were qualified under Specification AN-V-15a, the only test required was the vibration test under Specification MIL-V-7908. The valves passed this test.

IV. RECOMMENDATION

That A. Schrader's Son, Brooklyn, New York, be continued on the Qualified Products List as an approved source for the manufacture of low pressure, oxygen, check valves.

COORDINATION:

*Donald R. Good*

DONALD R. GOOD, WCRDE

*RJA*

IRVING H. LANTZ, WCRDE-3

*Wm E. Bills, Lt. Col. USAF (MC)*

CLARENCE C. CAIN, Lt. Col., USAF  
Chief, Aero Medical Operations

PUBLICATION REVIEW

This report has been reviewed and is approved.

ROBERT H. BLOUNT, Colonel, USAF (MC)  
Chief, Aero Medical Laboratory  
Directorate of Research

Technical Memorandum Report  
WCRD 53-46  
29 May 1953

DISTRIBUTION LIST

a.	Publications and Forms Branch	
	Publishing Division	
	Air Adjutant General, WCAPP	2 cys
	ASTIA Document Service Center	1 cy
	BAGR-CD - Miss Martin	6 cys
	WCRDA - Library	1 cy
	WCRDE	8 cys
	WCRDS-2	1 cy
	MCPPAB - Capt. Bergemann	2 cys
d.	A. Schrader's Son	
	Brooklyn, New York	