

AD_____

AWARD NUMBER DAMD17-96-1-6165

TITLE: Comparative Genomic Hybridization Onto Dense Arrays of DNA Clones: Development and Application to Breast Cancer Genomes

PRINCIPAL INVESTIGATOR: Steven M. Clark, Ph.D.

CONTRACTING ORGANIZATION: University of California
San Francisco, California 94143-0962

REPORT DATE: September 1998

TYPE OF REPORT: Annual Summary

PREPARED FOR: Commanding General
U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for Public Release;
Distribution Unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

1. AGENCY USE ONLY <i>(Leave blank)</i>	2. REPORT DATE September 1998	3. REPORT TYPE AND DATES COVERED Annual Summary (1 Sep 96 - 31 Aug 98)	
4. TITLE AND SUBTITLE Comparative Genomic Hybridization Onto Dense Arrays of DNA Clones: Development and Application to Breast Cancer Genomes		5. FUNDING NUMBERS DAMD17-96-1-6165	
6. AUTHOR(S) Steven M. Clark, Ph.D.			
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) University of California San Francisco, California 94143-0962		8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012		10. SPONSORING / MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES		19981218 024	
12a. DISTRIBUTION / AVAILABILITY STATEMENT Approved for Public Release; Distribution Unlimited		12b. DISTRIBUTION CODE	
13. ABSTRACT <i>(Maximum 200 words)</i> The work under the final portion of US Army grant DAMD17-96-1-6165 focused on the technical aspects to produce high density arrays in automated fashion. A printing robot was constructed that used a capillary print head developed in the first year of this grant. This printer was shown to produce arrays that performed as well as our previous hand-spotted arrays. Much less target DNA is necessary, permitting faster, cheaper automated clone purification procedures, such as the Autogen machine. Finally, I developed an improved narrow bore capillary print head capable of printing arrays of densities $10^4/\text{cm}^2$.			
14. SUBJECT TERMS Breast Cancer		15. NUMBER OF PAGES 8	
		16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT Unlimited

FOREWORD

Opinions, interpretations, conclusions and recommendations are those of the author and are not necessarily endorsed by the U.S. Army.

Where copyrighted material is quoted, permission has been obtained to use such material.

Where material from documents designated for limited distribution is quoted, permission has been obtained to use the material.

Citations of commercial organizations and trade names in this report do not constitute an official Department of Army endorsement or approval of the products or services of these organizations.

In conducting research using animals, the investigator(s) adhered to the "Guide for the Care and Use of Laboratory Animals," prepared by the Committee on Care and Use of Laboratory Animals of the Institute of Laboratory Resources, National Research Council (NIH Publication No. 86-23, Revised 1985).

✓ For the protection of human subjects, the investigator(s) adhered to policies of applicable Federal Law 45 CFR 46.

In conducting research utilizing recombinant DNA technology, the investigator(s) adhered to current guidelines promulgated by the National Institutes of Health.

In the conduct of research utilizing recombinant DNA, the investigator(s) adhered to the NIH Guidelines for Research Involving Recombinant DNA Molecules.

In the conduct of research involving hazardous organisms, the investigator(s) adhered to the CDC-NIH Guide for Biosafety in Microbiological and Biomedical Laboratories.



PI - Signature

Date

TABLE OF CONTENTS

Cover Page	
Report Documentation Page	
Foreword	
Table of Contents	
Recommendations in Relation to the Statement of Work	5
Method Validation for Printed Assay	5
Automation of the Array Fabrication Process	6
Defining the Minimum Amount of Target DNA Placed on the Arrays	6
Production of Dense Arrays	7
Conclusions	8

Recommendations in Relation to the Statement of Work

For the second and final year of my postdoctoral work I developed an automated array printing robot (Figure 1) in collaboration with the Engineering Division at Lawrence Berkeley Laboratory (LBL). The capillary print head developed in the first year of this work was attached to a gantry linking a translation table holding four 864-well microtitre plates and an xy translation table holding 8x6 microscope slides. A good deal of time (1/98-4/98) was spent at LBL overseeing the assembly and working out printing the parameters (print head pressures, wash conditions, stage velocities, etc.) necessary to reliably print arrays in a high throughput manner. The robot is now incorporated into a core facility at the UCSF Cancer Center.




Figure 1. The array printing robot.

Method validation for printed arrays. It is possible that this printer may introduce artifacts, so results were compared to previous results obtained from manual capillary printed arrays. An array was prepared

using 23 clones spaced along chromosome 20 (the 'chromosome 20 scanning array' mentioned in the previous report) and probe was prepared from the cell line BT474. Results were in excellent agreement (data not shown) indicating that the printing process did not introduce artifacts.

Automation of the array fabrication process. Large genome scanning arrays (~3,000 clones, resolution 1 Mbp) are planned for preparation in the upcoming years, so automated methods for clone production were evaluated. We obtained an Autogen machine which is capable of preparing ~180 clones from 1.5 ml bacterial cultures in a ten hour period. These clones were dissolved in our nitrocellulose/DMSO target solution at 1 ug/ul and compared with clones prepared using manual Quiagen columns (~6-12 clones per day from 100-500 ml cultures). The Autogen spots were quite weaker in the DAPI (counterstain) channel than the Quiagen clones were. A comparison of absorption, fluorescence, nuclease digestion and gel data for these clones indicated that the Autogen preps contained a significant amount of contaminating low molecular weight nucleic acids. Nonetheless, clones prepared in these two ways displayed identical fluorescence ratios. (data not shown)

Defining the minimum amount of target DNA placed on the arrays. From above discussion, it is clear that more targets can be prepared in a more rapid, inexpensive manner if lower DNA target concentrations are used. Also, the printing process is expected to be more reproducible for the current 2 ug/ul BAC or P1 target solutions are quite viscous. Calculations and 'double spotting' experiments indicate that only the top layer of DNA is available for hybridization. To test how little DNA was needed in the target solution a dilution series of lambda DNA was printed on a slide and hybridized with 50, 500 or 5,000 pg of FITC and

Texas Red labeled lambda probe. these experiments showed that the fluorescence ratios did not change from 2-0.06 ug/ul of target DNA concentration (data not shown). The spot shapes, however, did change, with the lower concentrations showing a 'coffee stain' appearance, an effect due to less viscosity during drying.

Production of dense arrays. The print head I developed (described in detail in the previous Progress Report) uses 375 um o.d. fused silica capillaries to deliver the target solution to the slide. I showed that by choosing smaller inside diameters, such as 25 um, extremely dense ($10^4/\text{cm}^2$) arrays can be printed. However, such narrow bore posses extremely high hydrodynamic impedance (9-fold higher than the currently used 75 um i.d. capillaries) making their use slow during the loading and washing steps. It is for these reasons that I developed a dual bore capillary, which has a short length of 25 um i.d. 150 um o.d. capillary glued inside a 4 cm length of 375 um o.d. 150 um i.d. capillary. These capillaries have not been tested yet, however, they are expected to produce dense arrays in a more rapid, reproducible manner.




Figure 2. Closeup of the dual bore capillary consisting of a 0.4 mm long 25 μm i.d. section glued into a 4 cm long 150 μm i.d. 375 μm o.d. capillary.

Conclusions

The work under the final portion of US Army grant DAMD17-96-1-6165 focused on the technical aspects to produce high density arrays in automated fashion. A printing robot was constructed that used a capillary print head developed in the first year of this grant. This printer was shown to produce arrays that performed as well as our previous hand-spotted arrays. Much less target DNA is necessary, permitting faster, cheaper automated clone purification procedures, such as the Autogen machine. Finally, I developed an improved narrow bore capillary print head capable of printing arrays of densities $10^4/\text{cm}^2$.