

REPORT DOCUMENTATION PAGE

AFRL-SR-BL-TR-01-

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments, including suggestions for reducing this burden to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503.

ces,
the
son

1. AGENCY USE ONLY (Leave blank)		2. REPORT DATE 19 September 2001	3. REPORT TYPE AND DATE COVERED Final Report 01 Apr - 31 August 2001	
4. TITLE AND SUBTITLE Wavelength-Tunable, Sub-picosecond Pulsed Laser Diagnostic Tool for III-Nitride Semiconductors			5. FUNDING NUMBERS F49620-01-1-0230	
6. AUTHOR(S) Stephen Teitworth, H. Craig Casey, and Henry Everitt				
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Duke University Department of Physics Box 90305 Durham, NC 27708-0305			8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) Department of the Air Force Air Force Office of Scientific Research (AFRL) 801 North Randolph Street, Room 732 Arlington, VA 22203-1977			10. SPONSORING/MONITORING AGENCY REPORT NUMBER	
11. SUPPLEMENTARY NOTES				
12a. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited.			<p style="text-align: center;"> AIR FORCE OFFICE OF SCIENTIFIC RESEARCH (AFOSR) NOTICE OF TRANSMITTAL DTIC. THIS TECHNICAL REPORT HAS BEEN REVIEWED AND IS APPROVED FOR PUBLIC RELEASE LAW AFR 100-12. DISTRIBUTION IS UNLIMITED. </p>	
13. ABSTRACT (Maximum 200 Words) The requested equipment would fill a tremendous void in Duke's capabilities, namely the ability to generate wavelength tunable sub-picosecond pulses in the UV-visible-IR spectral region. The required tools include a regenerative amplifier (RA) and pump laser, a frequency doubled and quadrupled optical parametric amplifier (OPA), and a difference frequency mixer to allow us to cover the entire spectral region from 200 nm to 15 microns. The pre-existing mode-locked Ti:S laser is used to pump the RA which in turn drives the OPA. The proposal was to develop this capability for wavelength tunable, sub-picosecond optical characterization of the important optoelectronic wide bandgap III-N materials and heterostructures. With the grant, a Quantronix Titan regenerative amplifier, driven by a Quantronix 527 DPH Nd:YLF laser was purchased. A Light Conversion TOPAS optical parametric amplifier was also purchased. The TOPAS can be frequency doubled, quadrupled, and mixed to allow continuous coverage of the spectral region 200 nm to 15 microns, encompassing most AlGaIn and all InGaIn-based materials.				
14. SUBJECT TERMS			15. NUMBER OF PAGES 3	
			16. PRICE CODE	
17. SECURITY CLASSIFICATION OF REPORT UNCLASSIFIED	18. SECURITY CLASSIFICATION OF THIS PAGE UNCLASSIFIED	19. SECURITY CLASSIFICATION OF ABSTRACT UNCLASSIFIED	20. LIMITATION OF ABSTRACT	

20011126 118

Final Report
AFOSR DURIP Grant

**Wavelength-Tunable, Sub-picosecond Pulsed Laser Diagnostic Tool
for III-Nitride Semiconductors**

Prof. Stephen Teitworth, Dept. of Physics, Duke Univ.

Prof. H. Craig Casey, Dept. of Electrical Engineering, Duke Univ.

Adj. Prof. Henry Everitt, Dept. of Physics, Duke Univ.

Duke University has an expanding research effort in III-V nitrides with three faculty (Prof. H. Craig Casey, ECE; Prof. Stephen Teitworth, Physics; Dr. Henry Everitt, Physics & Army Research Office) currently involved in various projects. Two laboratories have been recently converted to III-N work, one focusing on electrical characterization and one focusing on optical characterization. Notable facilities include :

- a pulsed femtosecond frequency doubled and tripled Ti:Sapphire laser system,
- CW HeCd photoluminescence (PL) system,
- Xe Lamp photoluminescence excitation (PLE) system,
- high-power pulsed current-voltage (IV) system,
- radio frequency (RF) and microwave sources and detectors, and
- access to high-power tunable pulsed infrared (IR) and ultraviolet (UV) light at the Duke Free Electron Laser Laboratory (DFELL).

The requested equipment would fill a tremendous void in Duke's capabilities, namely the ability to generate wavelength tunable sub-picosecond pulses in the UV-visible-IR spectral region. The required tools include a regenerative amplifier (RA) and pump laser, a frequency doubled and quadrupled optical parametric amplifier (OPA), and a difference frequency mixer to allow us to cover the entire spectral region from 200 nm to 15 microns. The pre-existing mode-locked Ti:S laser is used to pump the RA which in turn drives the OPA as seen in Figure 1.




Figure 1. Diagram of requested equipment. Pre-existing equipment is shaded in gray.

We proposed to develop this capability for wavelength tunable, sub-picosecond optical characterization of the important optoelectronic wide bandgap III-N materials and heterostructures. With the AFOSR DURIP grant, we purchased a Quantronix Titan regenerative amplifier, driven by a Quantronix 527 DPH Nd:YLF laser (Figure 2). We also purchased a Light Conversion TOPAS optical parametric amplifier. The TOPAS can be frequency doubled, quadrupled, and mixed to allow continuous coverage of the spectral region 200 nm to 15 microns, encompassing most AlGaIn and all InGaIn-based materials.

Originally, a Spectra Physics system was contemplated. However, Quantronix offered a superior capability (2 mJ per pulse vs. 1 mJ per pulse for Spectra Physics) and an acknowledged superior optical parametric amplifier for a lower price. The Quantronix / Light Conversion system was purchased in the spring of 2001 and installed in the summer of 2001. We have spent the remainder of the summer reconfiguring our laboratory to accommodate the new system, including two repairs that have had to be made since installation. Once fully operational, this laser system will provide a unique capability to investigate the ultrafast, microscopic dynamics and fundamental parameters of one of the most important classes of optoelectronic materials.




Figure 2. Installed laser system, including Quantronix Nd:YLF and Titan (background) and Light Conversion TOPAS (middle left).