

Investigation and Optimization of Miniature Fan-Based Impactors

²Jennifer L. Stepnowski, ¹Eric Chang, ¹Chris
Kendziora, ¹R. Andrew McGill, ³Hylton McWhinney,
⁴Viet Nguyen, ¹Stanley V. Stepnowski, ²Evgueni
Sokolovski, ²Derrick Thiel

U.S. Naval Research Laboratory, Code 6365, Washington DC 20375¹

Nova Research, Inc., Alexandria, VA 22304²

Prairie View A&M University, Prairie View, TX 77446³

Geo-Centers, Inc., Fort Washington, MD 20744⁴

Tel : (202) 767-8533; Email: jennifer.stepnowski@nrl.navy.mil



Report Documentation Page				Form Approved OMB No. 0704-0188	
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.					
1. REPORT DATE 17 NOV 2004		2. REPORT TYPE N/A		3. DATES COVERED -	
4. TITLE AND SUBTITLE Investigation and Optimization of Miniature Fan-Based Impactors				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S)				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) U.S. Naval Research Laboratory, Code 6365, Washington DC 20375				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release, distribution unlimited					
13. SUPPLEMENTARY NOTES See also ADM001849, 2004 Scientific Conference on Chemical and Biological Defense Research. Held in Hunt Valley, Maryland on 15-17 November 2004., The original document contains color images.					
14. ABSTRACT					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT UU	18. NUMBER OF PAGES 19	19a. NAME OF RESPONSIBLE PERSON
a. REPORT unclassified	b. ABSTRACT unclassified	c. THIS PAGE unclassified			

Background

- Collectors are an important component of a biodetection system to concentrate biological particles for detection
- Need a small and low power option to interface with increasingly smaller biosensors
- Other desirable features for biocollector design:
 - High collection efficiency in the 1 – 10 micron range
 - High sample volume
 - Inexpensive, disposable
 - Simple operation
 - Simple recovery

Fan-Based Impactor


Example Fan Specification:

Weight	15 – 34g
Rated Voltage	6 – 12V
Imax	600 mA
Power	0.2-9W
FR l/min	100 L/min
MTTF	up to 4000 hrs
T/°C	-20 to + 65



A high flow vane-axial fan utilized as an impactor provides many desirable biocollector features.


Testbed Design



- Test particles – 1.0 micron polystyrene latex spheres
- Flow rate of system – 120 LPM
- Met One 3315 Airborne Particle Counter with seven bin sizes (range 0.5 – 25 microns) – monitor bin 1.0 – 2.0 μm
- Collection efficiency calculated by measuring particles not collected
- % Collection Efficiency = $[(C_o - C_{\text{fan}})/C_o] * 100$;
- C_o = baseline particle count and C_{fan} = particles that pass through the fan

Mechanism of Fan Collection

Direction of Particle Motion



Fan showing fluorescent PS


Impaction on the fan blades and rim
is the mechanism of collection.

Collection Efficiency Experiments

Polystyrene Latex Spheres

- Single fan with rated voltage of 12V
- Fan washed with isopropanol and allowed to dry in a flow bench to minimize dust particles
- $67 \pm 2\%$ collection efficiency at 12V
- $72 \pm 3\%$ maximum collection efficiency achieved by overdriving to 15V

Effect of Fan Speed on Particle Collection



An increase in applied voltage increases the fan speed which leads to improved collection efficiency.

Preliminary Collection Experiments with BG Spores




The RAPID PCR

- Collection efficiency of greater than 80% for 1 μ m BG spores (with Jay Eversole and Cathy Scotto, Code 5611 at NRL)
- 500 μ l water used to process the polystyrene beads and BG spores off the fans
- Recovery and detection of BG spores confirmed with RAPID PCR system


Surface Modification Effects on Collection Efficiency

- As received, fan surface is a dense smooth collection substrate
- Modify fan surface to change surface area and energy
- Modifications: surface roughening, Al_2O_3 , Ti




*SEM image of
standard fan blade surface*

Surface Modification: Sand Blasted Collectors



*SEM image of
standard fan blade surface*




*SEM image of
sand blasted PPO fan blade*

- Preliminary experiments show 0 – 20% improvement
- Need more rigorous control of surface roughening

Surface Modification: Al_2O_3 and Ti Coated Collectors

- CVD used to grow aluminum oxide film directly on to PPO impellers
- XPS used to verify the oxide layer
- Titanium films were also investigated

Al_2O_3
coated
fan




XPS of aluminum oxide fan

Comparison of PPO, Al₂O₃ and Ti Collectors

Collection Efficiency (%) at 10,000 RPM		
PPO	Al ₂ O ₃	Ti
37 ± 6	36 ± 1	32 ± 6

- All the surfaces have the same collection efficiency within the error
- Due to high flow rates, the particles have high linear velocity and all surfaces appear the same

Coupon Impactor Design




- Fan used as impeller instead of impactor
- Decrease impact force for improved collection and less re-bounce with softer surface
- Concentrate particles in smaller footprint area for ease of recovery or analysis
- In some cases may be able to culture directly on coupon

Coupon Collector Proof-of-Concept

Experiment


- One micron fluorescent PSLs collected on coupon aft of the fan exhaust
- Four areas of collection because of geometry of fan




One micron fluorescent PSL spheres collected on planar coupon illuminated using a 245nm lamp

Simple Fan Biotrigger



Concept



Experimental Design



Fan Biotrigger Proof of Concept



Conclusions

- Fan-based impactor can be used as a low power, small, disposable biocollector.
- Collection efficiencies of $>70\%$ have been demonstrated for PSLs and BG spores.
- Increasing fan speed increases collection efficiency while surface material has little effect.
- Coupon collection shows promise as a secondary route to low power, higher efficiency collector.
- Low cost biotrigger is possible with fan coupled to inexpensive light source and detector.

Future Directions

- Optimize collection efficiency by examining variables
 - blade pitch
 - surface roughening
 - surface materials: softer surfaces
- Further investigate coupon collection to characterize and optimize collection efficiency
- Investigate viability

Acknowledgements

- Jerry Bottiger and RDECOM
- Jay Eversole, Cathy Scotto, Alan Huston, Vasanthi Sivaprakasam, NRL
- Safety Tech International and Micronel US

