

Registration No. 21143



Evaluation of Portable Near Infrared Fuel Analysis
Spectrometer

Joel Schmitigal

DISTRIBUTION STATEMENT A. Approved for public release;
distribution is unlimited.

AUGUST 2010

U.S. Army Tank Automotive Research,
Development, and Engineering Center
Detroit Arsenal
Warren, Michigan 48397-5000

Report Documentation Page			Form Approved OMB No. 0704-0188		
Public reporting burden for the collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden, to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to a penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.					
1. REPORT DATE 10 AUG 2010	2. REPORT TYPE N/A	3. DATES COVERED 01 JUN 2010 - 10 AUG 2010			
4. TITLE AND SUBTITLE Evaluation of Portable Near Infrared Fuel Analysis Spectrometer		5a. CONTRACT NUMBER			
		5b. GRANT NUMBER			
		5c. PROGRAM ELEMENT NUMBER			
6. AUTHOR(S) Joel Schmitgal		5d. PROJECT NUMBER			
		5e. TASK NUMBER			
		5f. WORK UNIT NUMBER			
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) US Army RDECOM-TARDEC 6501 E 11 Mile Rd Warren, MI 48397-5000, USA		8. PERFORMING ORGANIZATION REPORT NUMBER 21143			
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) US Army RDECOM-TARDEC 6501 E 11 Mile Rd Warren, MI 48397-5000, USA		10. SPONSOR/MONITOR'S ACRONYM(S) TACOM/TARDEC			
		11. SPONSOR/MONITOR'S REPORT NUMBER(S) 21143			
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release, distribution unlimited					
13. SUPPLEMENTARY NOTES The views, ipinion, and/or findings contained in this report are those of the authors and should not be construed as an official Department of the Army position, policy, or decision, unless so designated by other documents., The original document contains color images.					
14. ABSTRACT The objective of this effort was to evaluate the FuelEx portable Near Infrared (NIR) Spectrometer fuel analyzer, by Bruker Optics Inc., developed under guidance from TARDEC under Small Busines Innovative Research contract DAAE07-01-C-L008. The portable spectrometer was evaluted to determine the chemometric effectiveness of the spectrometer's models, providing for an analysis of chemometric model's ability to accurately correspond to known fuel property values as determined by standard ASTM test procedures, evaluate the repeatability of an instrument when measuring the same fuel multiple times and calculate the reproducibility to evaluate the variation of the four instruments being tested.					
15. SUBJECT TERMS Fuel, Diesel, JP-8, Chemometrics, Near Infrared Spectrometry					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT SAR	18. NUMBER OF PAGES 17	19a. NAME OF RESPONSIBLE PERSON
a. REPORT unclassified	b. ABSTRACT unclassified	c. THIS PAGE unclassified			

NOTICES

Disclaimers

The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorized documents.

Citation of manufacturer's or trade names does not constitute an official endorsement or approval of the use thereof.

Destroy this report when it is no longer needed. Do not return it to the originator.

U.S. Army Tank Automotive Research Development and Engineering Center

Warren, Michigan 48397-5000

Evaluation of Portable Near Infrared Fuel Analysis Spectrometer

Joel Schmitigal
Force Projection Technology

Approved for public release; distribution unlimited

Standard Form 298 (Rev. 8/98)

Prescribed by ANSI Std. Z39

Contents

Contents	iv
List of Tables	v
1. Introduction	1
2. Approach	1
3. Analysis	3
3.1. NIR Accuracy	3
3.2. Fuel Specification Categorization	3
3.3. Accuracy to ASTM Values	3
3.4. Standard Error of Prediction	4
3.5. NIR Precision	5
4. Conclusion	7
References	8
Appendix A	9
List of Symbols, Abbreviations, and Acronyms	12

List of Tables

Table 1. Test Fuels.....	2
Table 2. Summary of ASTM Repeatability and Reproducibility	4
Table 3. Standard Errors of Prediction for evaluated fuels.....	5
Table 4. NIR Repeatability of Jet fuel models.....	5
Table 5. NIR Repeatability of Diesel No. 2 fuel models.	6
Table 6. NIR Reproducibility of Jet fuel models.	6
Table 7. NIR Reproducibility of Diesel fuel models.	6
Table A1. NIR jet fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2004).....	9
Table A2. NIR jet fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2005).....	9
Table A3. NIR jet fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2006).....	9
Table A4. NIR jet fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2010).....	10
Table A5. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2004). 10	
Table A6. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2005). 10	
Table A7. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2006). 11	
Table A8. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2010). 11	

1. Introduction

The objective of this effort was to evaluate the FuelEx portable Near Infrared (NIR) Spectrometer fuel analyzer, by Bruker Optics Inc., developed under guidance from TARDEC under Small Business Innovative Research contract DAAE07-01-C-L008. The 4 Spectrometers evaluated under this effort were received by The U.S. Army Research Development and Engineering Command's Tank Automotive Research Development and Engineering Center (RDECOM-TARDEC) under contract W56HZV-08-C-L534, an effort that ruggedized the instrument to meet the Army's environmental requirements. The NIR spectrometer is being developed for utilization in the Army's Petroleum Test Kit (1) which will provide fuel handlers with a last quality check of the fuels usability, including fuel type, immediately prior to vehicle or aircraft fueling.

The FuelEX NIR spectrometer utilizes Partial Least Squares (PLS) chemometrics to predict the properties of the fuel and determines the usability of the fuel based on these values as has been detailed in previous reports (2) (3) (4) (5) (6).

2. Approach

The NIR fuel analyzer verification was performed to provide an analysis into three instrument factors. Firstly to evaluate the chemometric effectiveness of the spectrometer's models, providing for an analysis of chemometric model's ability to accurately correspond to known fuel property values as determined by standard ASTM International test procedures. Secondly to evaluate the repeatability of an instrument when measuring the same fuel multiple times. Lastly the reproducibility was calculated to evaluate the variation of the four instruments being testing.

The four NIR fuel analyzers were subjected to 32 fuel samples, consisting of Diesel No. 2 (DL-2), JP-8, JP-5, B100 Biodiesel, and gasoline as detailed in Table 1. Each fuel sample was measured 8 times, 4 times each under the Diesel No. 2 and Jet fuel property models, for a total of 1032 sample spectra analyzed. To establish the repeatability of the instrument five jet fuels and five diesel fuels were subjected to 50 additional spectral measurements each, for a total of 500 spectra. After all measurements were taken the data was analyzed to calculate the repeatability of the instruments, and the reproducibility between the instruments for comparison purposes to standard ASTM test procedures.

Sample ID	Fuel Type
737	JP-8
740	JP-8
756	JP-8
758	gasoline
759	Diesel No. 2
761	Diesel No. 2
763	JP-8
766	JP-8
769	Diesel No. 1
880	Diesel No. 2
881	Diesel No. 2
884	Diesel No. 2
890	JP-8
1036	Diesel No. 2
1038	Diesel No. 2
1041	JP-8
1043	Diesel No. 2
1238	JP-5
1304	Diesel No. 2
12770	JP-8
12854	Diesel No. 2
12855	Diesel No. 2
12988	JP-8
12996	JP-8
13014	Diesel No. 2
13119	B100 Biodiesel
13120	Diesel No. 2
13141	JP-8
13201	Diesel No. 2
13301	JP-8
13313	Diesel No. 2
13344	JP-8

Table 1. Test Fuels

3. Analysis

3.1. NIR Accuracy

NIR accuracy was analyzed by three methods, 1) fuel specification characterization; 2) NIR property prediction within 1.5 times of the published ASTM reproducibility; 3) Standard Error of Prediction (SEP). Method 1, fuel specification characterization, specifically provides the data output that is required of the NIR spectrometer in the Petroleum Test Kit application, that of providing the Soldier with a last quality check of the fuel; this includes providing a fuel specification characterization and acceptability for use check against published fuel specifications. Comparison of the instruments predicted values against the known ASTM values determined in a laboratory and the accepted reproducibility of the ASTM test methods provides an indication of the NIR method's ability to accurately predict the fuel's properties relative to ASTM accepted values. Method 3 of calculating the Standard Error of Prediction is a standard method for evaluating chemometric models in that it is the standard deviation of all the NIR predicted values from the accepted ASTM values.

3.2. Fuel Specification Categorization

The NIR fuel analyzer was able to accurately identify all 32 of the fuels analyzed, as required by the Petroleum Test Kits requirements documents (1) for each of the 1032 spectra analyzed. The NIR fuel analyzer correctly identified gasoline and biodiesel (B100), as not being an acceptable fuel type, either jet fuel or a DL-2, but was unable to correctly characterize and identify the fuels as there are not chemometric models for these fuels built into the instrument. The only instance where a fuel was incorrectly categorized was when the JP-5 sample was categorized as an acceptable jet fuel but, the samples flashpoint was predicted in the range of 53-57°C, which is below the required 60°C for a JP-5 fuel, the ASTM measured value was 63°C. This result is typical as chemometric modeling of spectroscopic data has not shown a high affinity for accurately predicting flashpoint, the incorrect result could also be due to using a model based on JP-8 fuels which typically have a lower flashpoint than JP-5. The Diesel No. 1 fuel that was analyzed was properly characterized as a Jet Fuel as well as Diesel No. 1 fuel.

3.3. Accuracy to ASTM Values

The NIR fuel analyzer's prediction accuracy to standard ASTM measurements for density (6), cetane index (7), aromatic content (8), distillation point temperatures (9), flashpoint (10), cloud point (11), freeze point (12) and viscosity at -20°C and 40°C (13) was determined to be acceptable for most properties, when utilizing 1.5 times ASTM's published reproducibility (1.5R) as the threshold for acceptability, as detailed in Appendix A. Table 2 details the ASTM repeatability and reproducibility values.

Density and viscosity measurements were found to be outside the 1.5R threshold over 50% of the time, despite those properties having a low Standard Error of Prediction (SEP) due to their ASTM techniques having such low repeatability and reproducibility it is very difficult for chemometric models to provide this level of precision. An average Standard Error or Prediction of 0.005g/mL for density, 0.3cSt for viscosities at -20°C and 40°C while not being within the 1.5R threshold, have

enough precision for qualifying fuel and allowing for the discrimination between fuel types, and to ensure compatibility for proper engine operation.

Property	Method	fuel type	r, repeatability	R, Reproducibility	Units
Density	ASTM D 1298	all	0.0005	0.0012	g/mL
Cetane Index	ASTM D 4737	all	N/A	N/A	
Aromatics	ASTM D 1319	all	1.4	3.0	%
Distillation*	ASTM D 86				°C
Distil. 10%		Diesel No. 2	3.1-5.3	6.6-10.6	°C
Distil. 90%		Diesel No. 2	2.6-3.6	5.5-7.9	°C
Distil. 10%		JP-8	1.9-3.7	4.3-7.6	°C
Distil. 90%		JP-8	2.1-2.9	4.3-6.3	°C
Flash Point	ASTM D 93	all	1.2-2.3	2.9-5.7	°C
Cloud Point	ASTM D 2500	Diesel No. 2	2.0	4.0	°C
Freeze Point	ASTM D 5972	JP-8	0.54	0.8	°C
Viscosity -20°C	ASTM D 445	JP-8	0.03	0.07-0.09	cSt
Viscosity 40°C	ASTM D 445	Diesel No. 2	.01-.02	.03	cSt

Table 2. Summary of ASTM Repeatability and Reproducibility

*Some ASTM methods provide for a calculated r and R based on the ASTM derived data.

In these cases the test fuel's r and R range are provided.

The flashpoint predictions fell outside of the 1.5R range 32% of the time, this lack of accuracy is not considered important as flashpoint has no impact on the usability of the fuel in diesel and turbine engine applications and the measurement has therefore been determined not to be important for the operational requirements of the Petroleum Test Kit.

3.4. Standard Error of Prediction

Standard Error of Prediction (SEP) is a standard method for evaluating a chemometric models accuracy in that it is the standard deviation of all the NIR predicted values from the accepted ASTM values. The average SEP, for all four instruments utilized in this evaluation, for the samples measured is given in Table 3.

Property	Average JP-8 SEP	Average Diesel SEP
Density (g/mL)	0.004	0.004
Cetane Index	2.3	0.9
Aromatics (%)	1.2	3.5
Distil. 10% (°C)	8.0	12.5
Distil. 90% (°C)	6.2	6.7
Flash Point (°C)	6.9	8.8
Cloud Point (°C)		2.7
Freeze Point (°C)	3.4	
Viscosity -20°C (cSt)	0.3	
Viscosity 40°C (cSt)		0.3

Table 3. Standard Errors of Prediction for evaluated fuels

3.5. NIR Precision

The repeatability and reproducibility of the 4 NIR fuel analyzers and chemometric models compared well to that of the published values of the standard ASTM measurement practices, as shown in tables 4-7. NIR repeatability was calculated using 95% probability and the average standard deviation of measurements performed using the on the same instrument of 50 measurements performed using the same instrument on 5 samples of each fuel type over an 8 hour period of time. The reproducibility was calculated using a 95% probability from the average standard deviation of the measurements for the 32 samples measured 4 times each.

Property	Average NIR Standard Deviation	NIR Repeatability	ASTM Repeatability
Density (g/mL)	0.0007	0.0018	0.0005
Cetane Index	0.2	0.6	N/A
Aromatics (%)	0.3	0.7	1.4
Distil. 10% (°C)	2.1	5.4	1.9-3.7
Distil. 90% (°C)	0.5	1.4	2.1-2.9
Flash Point (°C)	1.3	3.2	1.2-2.3
Freeze Point (°C)	0.48	1.25	0.54
Viscosity -20°C (cSt)	0.05	0.12	0.03

Table 4. NIR Repeatability of Jet fuel models.

Calculated using 95% probability and average standard deviation of 50 measurements of 5 jet fuel samples measured the same instrument over an 8 hour time period. ASTM values taken from ASTM method documents (www.astm.org).

Property	NIR Standard Deviation	NIR Reproducibility	ASTM Reproducibility
Density (g/mL)	0.0008	0.0020	0.0005
Cetane Index	0.1	0.4	N/A
Aromatics (%)	0.4	1.0	1.4
Distil. 10% (°C)	0.8	2.2	3.1-5.3
Distil. 90% (°C)	1.7	3.0	2.6-3.6
Flash Point (°C)	0.7	1.8	1.2-2.3
Cloud Point (°C)	0.4	1.0	2
Viscosity 40°C (cSt)	0.05	0.14	.01-.02

Table 5. NIR Repeatability of Diesel No. 2 fuel models.

Calculated using 95% probability and average standard deviation of 50 measurements of 5 jet fuel samples measured the same instrument over an 8 hour time period. ASTM values taken from ASTM method documents (www.astm.org).

Property	NIR Standard Deviation	NIR Reproducibility	ASTM Reproducibility
Density (g/mL)	0.3652	0.9495	0.0012
Cetane Index	0.3	0.8	N/A
Aromatics (%)	0.3	0.8	3.0
Distil. 10% (°C)	2.4	6.2	4.3-7.6
Distil. 90% (°C)	0.6	1.5	4.3-6.3
Flash Point (°C)	2.1	5.3	2.9-5.7
Freeze Point (°C)	1.1	2.8	0.8
Viscosity -20°C (cSt)	0.60	1.5	0.07-0.09

Table 6. NIR Reproducibility of Jet fuel models.

Calculated using 95% probability and average standard deviation of 15 jet fuel samples measured on 4 different instruments. ASTM values taken from ASTM method documents (www.astm.org).

Property	NIR Standard Deviation	NIR Reproducibility	ASTM Reproducibility
Density (g/mL)	0.0005	0.0012	0.0012
Cetane Index	0.1	0.2	N/A
Aromatics (%)	0.3	0.8	3.0
Distil. 10% (°C)	0.6	1.6	6.6-10.6
Distil. 90% (°C)	0.8	2.1	5.5-7.9
Flash Point (°C)	0.8	1.2	2.9-5.7
Cloud Point (°C)	0.3	0.7	4
Viscosity 40°C (cSt)	0.05	0.13	0.03

Table 7. NIR Reproducibility of Diesel fuel models.

Calculated using 95% probability and average standard deviation of 15 diesel samples measured on 4 different instruments. ASTM values taken from ASTM method documents (www.astm.org).

4. Conclusion

The NIR fuel analyzer was able to correctly identify all 32 of the fuels analyzed, as required by the Petroleum Test Kits requirements documents (1). The NIR fuel analyzer correctly identified gasoline and biodiesel, as not being an acceptable fuel type, either jet fuel or a DL-2, but was unable to identify characterize and identify the fuels as there are not chemometric models for these fuels built into the instrument. The only instance where a fuel was incorrectly categorized was when the JP-5 sample was categorized as a jet fuel but not as JP-5 as its flashpoint was predicted to be between 53-57°C which is below the required 60°C, the ASTM measured value for this sample was 63°C.

The NIR fuel analyzer was able to provide agreement with standard ASTM measurements, with the exception of density and viscosity measurements which have such low repeatability and reproducibility that it is very difficult for chemometric models to provide this level of precision.

The NIR fuel analyzer chemometric models had repeatability on par with that of utilizing standard ASTM measurement practices. The reproducibility between the four NIR fuel analyzers tested was observed be less than when utilizing standard ASTM measurement practices for most properties.

References

1. Operational Requirements Document for Petroleum Test Kit (PTK) ACAT III. December 11, 2002.
2. **Schmitigal, Joel.** *Near-Infrared Spectroscopy and Chemometrics Instrumentation and Methodology for Field Evaluation of Diesel and Aviation Fuels by the U.S. Army.* Warren : TARDEC, 2010.
3. *Design Improvements for the Army Portable NIR Fuel Analyzer.* **Cooper, John and Schmitigal, Joel.** New Orleans : Coordinating Research Council Inc., 2009.
4. *New Capabilities and Future Directions for the FUELex Portable Fuel Analyzer.* **Cooper, John and Schmitigal, Joel.** Alexandria : Coordinating Research Council Inc, 2010.
5. *Portable Fuel Analyzer for the Army Petroleum Test Kit: Current Design and Future Goals.* **Cooper, John and Schmitigal, Joel.** Alexandria : Coordinating Research Council, 2008.
6. *Near-Infrared Fuel Analysis.* **Schmitigal, Joel.** Alexandria : Coordinating Research Council, 2006.
7. *Standard Test Method for Density, Relative Density (Specific.* West Conshohocken : American Society of Testing and Materials, 1999. ASTM D 1298-99.
8. *Standard Test Method for Calculated Cetane Index by Four Variable Equation.* West Conshohocken : American Society of Testing and Materials, 2009. ASTM D 4737-09.
9. *Standard Test Method for Hydrocarbon Types in Liquid Petroleum Products by Fluorescent Indicator Adsorption.* West Conshohocken : American Society of Testing and Materials, 2010. ASTM D1319-10.
10. *Standard Test Method for Distillation of Petroleum Products at Atmospheric Pressure.* West Conshohocken : American Society of Testing and Materials, 2009. ASTM D 86-09.
11. *Standard Test Methods for Flash Point by Pensky-Martens Closed Cup Tester.* West Conshohocken : American Society of Testing and Materials, 2010. ASTM D 93-10.
12. *Standard Test Method for Cloud Point of Petroleum Products.* West Conshohocken : American Society of Testing and Materials, 2009. ASTM D 2005-09.
13. *Standard Test Method for Freezing Point of Aviation Fuels (Automatic Phase Transition Method).* West Conshohocken : American Society of Testing and Materials, 2010. ASTM D 5972-05.
14. *Standard Test Method for Kinematic Viscosity of Transparent and Opaque Liquids (and Calculation of Dynamic Viscosity).* West Conshohocken : American Society of Testing and Materials, 2009. ASTM D 445-09.

Appendix A

	Aromatics	Density	Distil. 10%	Distil. 90%	Flash Point	Freeze Point	Viscosity -20°C
13344	Yes	Yes	Yes	Yes	Yes	Yes	No
13301	Yes	Yes	Yes	No	Yes	Yes	No
13141		No	Yes	Yes	Yes	No	
12996	Yes	No	Yes	Yes	No	No	No
12988	Yes	No	Yes	Yes	No	No	No
1238		No			No		No
1041	Yes	Yes	Yes	Yes	Yes		Yes
890	Yes	No	No	Yes	Yes	No	
769		No			yes		
740	Yes	No	Yes	Yes	Yes	Yes	

Table A1. NIR jet fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2004)

	Aromatics	Density	Distil. 10%	Distil. 90%	Flash Point	Freeze Point	Viscosity -20°C
13344	Yes	No	Yes	Yes	Yes	No	Yes
13301	Yes	No	Yes	No	Yes	Yes	No
13141		Yes	Yes	Yes	Yes	No	
12996	Yes	No	No	Yes	No	No	No
12988	Yes	No	No	Yes	No	No	No
1238		Yes			Yes		Yes
1041	Yes	Yes	Yes	Yes	Yes		No
890	Yes	Yes	Yes	Yes	Yes	Yes	
769		No			yes		
740	No	Yes	Yes	Yes	No	No	

Table A2. NIR jet fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2005)

	Aromatics	Density	Distil. 10%	Distil. 90%	Flash Point	Freeze Point	Viscosity -20°C
13344	Yes	No	Yes	No	Yes	No	Yes
13301	Yes	No	Yes	No	Yes	Yes	No
13141		Yes	Yes	Yes	Yes	No	
12996	Yes	No	No	Yes	No	No	No
12988	Yes	No	Yes	Yes	No	No	No
1238		Yes			Yes		Yes
1041	Yes	No	Yes	Yes	Yes		No
890	Yes	Yes	Yes	Yes	Yes	No	
769		No			Yes		
740	Yes	Yes	Yes	Yes	Yes	Yes	

Table A3. NIR jet fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2006)

	Aromatics	Density	Distil. 10%	Distil. 90%	Flash Point	Freeze Point	Viscosity -20°C
13344	Yes	Yes	Yes	No	Yes	No	No
13301	Yes	No	Yes	No	Yes	Yes	No
13141		Yes	Yes	Yes	Yes	No	
12996	Yes	No	No	Yes	No	No	No
12988	Yes	No	Yes	Yes	No	No	No
1238		Yes			Yes		No
1041	Yes	Yes	Yes	Yes	Yes		No
890	Yes	No	No	Yes	Yes	No	
769		No			Yes		
740	Yes	Yes	Yes	Yes	Yes	Yes	

Table A4. NIR jet fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2010)

	Aromatics	Cloud Point	Density	Distil. 10%	Distil. 90%	Flash Point	Viscosity +40°C
13313	Yes	Yes	No	Yes	Yes	No	
13201	Yes	Yes	No	Yes	No	No	No
13120	Yes	Yes	No	Yes	Yes	No	No
13014	Yes	Yes	No	No	Yes	No	No
12855	Yes	Yes	Yes	No	Yes	Yes	No
12854	Yes		No	No	Yes	Yes	No
1038	No		No	Yes	Yes	Yes	Yes
1036		Yes	No	Yes	Yes	Yes	No
881	Yes		No			Yes	No
761	Yes		No	Yes	Yes	Yes	Yes
759	Yes		Yes	Yes	Yes	Yes	No

Table A5. NIR diesel fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2004)

	Aromatics	Cloud Point	Density	Distil. 10%	Distil. 90%	Flash Point	Viscosity +40°C
13313	Yes	Yes	No	Yes	Yes	No	
13201	Yes	Yes	Yes	Yes	No	No	No
13120	Yes	Yes	Yes	Yes	Yes	No	No
13014	Yes	Yes	No	No	Yes	No	No
12855	Yes	Yes	Yes	No	Yes	Yes	No
12854	Yes		No	Yes	Yes	Yes	No
1038	No		No	Yes	Yes	Yes	No
1036		Yes	Yes	Yes	Yes	Yes	No
881	Yes		Yes			Yes	No
761	Yes		Yes	Yes	Yes	Yes	No
759	Yes		No	Yes	Yes	Yes	No

Table A6. NIR diesel fuel property prediction overlap ASTM-value $\pm$ 1.5R (instrument 2005)

	Aromatics	Cloud Point	Density	Distil. 10%	Distil. 90%	Flash Point	Viscosity +40°C
13313	No	Yes	Yes	Yes	Yes	No	
13201	No	Yes	No	Yes	No	No	No
13120	No	Yes	No	Yes	Yes	No	No
13014	No	Yes	No	No	Yes	No	No
12855	No	Yes	No	No	Yes	Yes	No
12854	Yes		Yes	Yes	Yes	Yes	No
1038	Yes		No	Yes	Yes	Yes	No
1036		Yes	No	Yes	Yes	Yes	No
881	Yes		No			Yes	No
761	Yes		No	Yes	Yes	Yes	No
759	Yes		No	Yes	No	Yes	No

Table A7. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2006)

	Aromatics	Cloud Point	Density	Distil. 10%	Distil. 90%	Flash Point	Viscosity +40°C
13313	Yes	Yes	No	Yes	Yes	Yes	
13201	Yes	Yes	Yes	Yes	No	No	No
13120	No	Yes	Yes	Yes	Yes	No	No
13014	Yes	Yes	No	No	Yes	No	No
12855	Yes	Yes	Yes	No	Yes	Yes	No
12854	Yes		No	Yes	Yes	Yes	No
1038	Yes		Yes	Yes	Yes	Yes	No
1036		Yes	Yes	Yes	Yes	Yes	Yes
881	Yes		Yes			Yes	No
761	Yes		No	Yes	Yes	Yes	No
759	Yes		No	No	Yes	No	No

Table A8. NIR diesel fuel property prediction overlap ASTM-value $\pm 1.5R$ (instrument 2010)

List of Symbols, Abbreviations, and Acronyms

°C	degree(s) Celsius
cSt	centistokes
Distil.	Temperature at distillation boiling point
DL-1	Diesel Fuel No.1
DL-2	Diesel Fuel No.2
g	Grams
JP-5	Jet Propellant 5
JP-8	Jet Propellant 8
mL	Milliliter
NIR	Near Infrared
PLS	Partial Least Squares
r	Repeatability
R	Reproducibility
RDECOM	Research Development and Engineering Command
SEP	Standard Error of Prediction
TARDEC	Tank Automotive Research Development and Engineering Center