

Targeting Biological Sensing with Commercial SERS Substrates

by Mikella E. Farrell, Srikanth Singamaneni, and Paul M. Pellegrino

ARL-RP-0398

September 2012

A reprint from the Proc. of SPIE, Vol. 8366, 836602-1.

NOTICES

Disclaimers

The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorized documents.

Citation of manufacturer's or trade names does not constitute an official endorsement or approval of the use thereof.

Destroy this report when it is no longer needed. Do not return it to the originator.

Army Research Laboratory

Adelphi, MD 20783-1197

ARL-RP-0398

September 2012

Targeting Biological Sensing with Commercial SERS Substrates

Mikella E. Farrell and Paul M. Pellegrino
Sensors and Electron Devices Directorate, ARL

Srikanth Singamaneni
Department of Mechanical Engineering and Materials Science
Washington University, St. Louis, MO 63130

A reprint from the Proc. of SPIE, Vol. 8366, 836602-1.

REPORT DOCUMENTATION PAGE			Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing the burden, to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.				
1. REPORT DATE (DD-MM-YYYY) September 2012		2. REPORT TYPE Reprint		3. DATES COVERED (From - To)
4. TITLE AND SUBTITLE Targeting Biological Sensing with Commercial SERS Substrates		5a. CONTRACT NUMBER		
		5b. GRANT NUMBER		
		5c. PROGRAM ELEMENT NUMBER		
6. AUTHOR(S) Mikella E. Farrell, Srikanth Singamaneni, and Paul M. Pellegrino		5d. PROJECT NUMBER		
		5e. TASK NUMBER		
		5f. WORK UNIT NUMBER		
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) U.S. Army Research Laboratory ATTN: RDRL-SEE-E 2800 Powder Mill Road Adelphi, MD 20783-1197		8. PERFORMING ORGANIZATION REPORT NUMBER ARL-RP-0398		
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES)		10. SPONSOR/MONITOR'S ACRONYM(S)		
		11. SPONSOR/MONITOR'S REPORT NUMBER(S)		
12. DISTRIBUTION/AVAILABILITY STATEMENT Approved for public release; distribution unlimited.				
13. SUPPLEMENTARY NOTES A reprint from the <i>Proc. of SPIE</i> , Vol. 8366, 836602-1.				
14. ABSTRACT There is an increasing need and challenge for early rapid and accurate detection, identification, and quantification of chemical, biological, and energetic hazards in many fields of interest (e.g., medical, environmental, industrial, and defense applications). Increasingly to meet these challenges, researchers are turning interdisciplinary approaches combining spectroscopy with nanoscale platforms to create technologies that offer viable and novel solutions for today's sensing needs. One technology that has gained increasing popularity to meet these needs is surface enhanced Raman scattering (SERS). SERS is particularly advantageous as it does not suffer from interferences from water, requires little to no sample preparation is robust and can be used in numerous environments, is relatively insensitive to the wavelength of excitation employed and produces a narrow-band spectral signature unique to the molecular vibrations of the analyte. SERS enhancements (chemical and electromagnetic) are typically observed on metalized nanoscale roughened surfaces. For ideal SERS sensing, commercially available uniform and reproducible nanoscale surface demonstrating high sensitivity are desirable. Additionally, if these surfaces can be modified for the selective sensing of hazard materials, an ideal sensor platform for dynamic in field measurements can be imagined. In this proceedings paper, preliminary efforts towards the characterization and application of commercially available next generation Klarite substrates will be demonstrated. Additionally, efforts toward chemical modification of these substrates, through peptide recognition elements can be used for the targeting sensing of hazardous materials will be explored.				
15. SUBJECT TERMS SERS, sensor, chemical, biological, Raman, detection, peptide, biomimetic				
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT UU	18. NUMBER OF PAGES 18
a. REPORT Unclassified	b. ABSTRACT Unclassified	c. THIS PAGE Unclassified		
			19a. NAME OF RESPONSIBLE PERSON Mikella Farrell	
			19b. TELEPHONE NUMBER (Include area code) (301) 394-0948	

Targeting Biological Sensing with commercial SERS substrates

Mikella E. Farrell*, Srikanth Singamaneni□, and Paul M. Pellegrino*

*U.S. Army Research Laboratories, RDRL-SEE-E, 2800 Powder Mill Road, Adelphi, Maryland 20783

□Department of Mechanical Engineering and Materials Science; Washington University in St. Louis 318 Jolley Hall, 1 Brookings Drive., St. Louis, MO 63130

SUMMARY

There is an increasing need for rapid and accurate detection, identification, and quantification of chemical, biological, and energetic hazards in many fields of interest. To meet these challenges, researchers are combining spectroscopy with nanoscale platforms to create technologies that offer viable and novel solutions for today's sensing needs. One technology that has gained increasing popularity to meet these needs is surface enhanced Raman scattering (SERS). For ideal SERS sensing, commercially available uniform and reproducible nanoscale surface demonstrating high sensitivity are desirable. If these surfaces can be modified for the selective sensing of hazard materials, an ideal sensor platform for dynamic in field measurements can be imagined. In this proceedings paper, preliminary efforts towards the characterization and application of commercially available next generation Klarite substrates will be demonstrated and efforts towards selective sensing will be discussed.

ABSTRACT

There is an increasing need and challenge for early rapid and accurate detection, identification, and quantification of chemical, biological, and energetic hazards in many fields of interest (e.g., medical, environmental, industrial, and defense applications). Increasingly to meet these challenges, researchers are turning interdisciplinary approaches combining spectroscopy with nanoscale platforms to create technologies that offer viable and novel solutions for today's sensing needs. One technology that has gained increasing popularity to meet these needs is surface enhanced Raman scattering (SERS). SERS is particularly advantageous as it does not suffer from interferences from water, requires little to no sample preparation is robust and can be used in numerous environments, is relatively insensitive to the wavelength of excitation employed and produces a narrow-band spectral signature unique to the molecular vibrations of the analyte.

SERS enhancements (chemical and electromagnetic) are typically observed on metalized nanoscale roughened surfaces. For ideal SERS sensing, commercially available uniform and reproducible nanoscale surface demonstrating high sensitivity are desirable. Additionally, if these surfaces can be modified for the selective sensing of hazard materials, an ideal sensor platform for dynamic in field measurements can be imagined. In this proceedings paper, preliminary efforts towards the characterization and application of commercially available next generation Klarite substrates will be demonstrated. Additionally, efforts toward chemical modification of these substrates, through peptide recognition elements can be used for the targeting sensing of hazardous materials will be explored.

KEYWORDS: SERS, sensor, chemical, biological, Raman, detection, peptide, biomimetic

I. INTRODUCTION

There are many unique and exquisitely organized solutions for sensitive and selective sensing in biological systems. This innate biological ability to guide and assemble nanoscale components into precisely controlled and sophisticated structures from very basic components has motivated intense defense, industrial and academic research efforts to develop and harness these directed artificial methods that mimic or exploit the selective recognition capabilities and sensitive interactions inherently found in biological systems. Many of these approaches have resulted in the successful

merging of more traditional sensing techniques (Fluorescence⁽¹⁾, Raman⁽²⁾, Raman-based spectroscopic techniques⁽³⁻⁹⁾) in novel nano-scale synthetic platforms combined with biomolecular recognition capabilities, thus demonstrating viable and transformative solutions for today's sensing needs. In this proceedings paper, we discuss the development of a novel selective and sensitive sensor platform.

One of the main challenges in the design and fabrication of a sensor platform is analyte selectivity and sensitivity. A high degree of sensor selectivity can be achieved by using biological recognition elements (e.g. antibody), and thus fabricating a ⁽¹⁰⁻¹⁴⁾ sensor. Biomimetic sensing entails the adaptation of biological principles, designs, selective sensitive materials and signal processing schemes merged with artificial sensors. One type of biomimetic sensing recognition element demonstrating increased utility employs peptides as the biological recognition element.⁽¹⁵⁻¹⁹⁾ A peptide is a short (less than 50) chain of amino acids. Generally, in sensing platforms peptides are advantageous to use as they 1) are stable and robust in a myriad of environments, 2) can be easily synthesized, 3) are cost and time efficient, 4) can be easily modified to recognize a target, 5) are well characterized and 6) have been shown to easily immobilized to various surface platform materials (metals, plastics, fabrics, tissue sample). The ability to immobilize peptides to various platforms and targets has resulted in an upsurge in the development of ⁽²⁰⁻²⁴⁾ peptides constructed from a fusion of functional peptides. There are many examples in the literature of fusion peptides being used in biomimetic sensors in many areas of research (defense, medical, hazard detection, and environmental studies).

Fusion peptides have been identified following many different experimental methods. Historically, methods for peptide selection include identification from natural resources and most commonly using combinatorial peptide-phage display libraries selected via a technique called biopanning. The power of the phage display library is in its ability to efficiently and rapidly identify ligands with a desired target property from a large population of colonies displaying highly diverse surface peptides. Biopanning can generally be described as an affinity selection technique for choosing a peptide that can bind to a given target (demonstrating desired properties). The biopanning process typically involves four main steps: preparation of the phage displayed peptide libraries, capture of specific phage that binds to the target, washing of low affinity or non specific phages from the cell surface, and then finally recovery through elution of the enriched target binders for the next round of selection.⁽²⁵⁻²⁸⁾ Using these types of selection techniques, it has been possible to identify and select several different material specific binding peptide sequences (examples include metal compounds, semiconductor materials, minerals, and polymers)⁽²³⁾ as well as particular targets (chemical⁽²⁹⁻³²⁾, biological^(16, 33, 34), or energetic hazards⁽³⁵⁻³⁹⁾). The screening and selection of such phage displayed target binding peptides has attracted particular interest in the research areas of nanotechnology and sensor design. Thus the power of these high affinity target peptide binders are increasingly being utilized as the recognition element in new sensor platform designs.

With a new strategy for sensor design, there is a renewed effort to address the outstanding need and challenge for rapid, sensitive, accurate detection, identification, and quantification of chemical, biological, and energetic hazards in many fields of interest (e.g., medical, environmental, industrial, and defense applications).⁽⁶⁾ One spectroscopic detection method that is gaining significant popularity for meeting these sensitivity needs is surface enhanced Raman scattering (SERS). SERS-based techniques and platforms combine traditional spectroscopy with nanotechnology. SERS is particularly advantageous and an appropriate detection technique to utilize in biomimetic sensors as it does not suffer from interferences from water, requires little to no sample preparation, is robust and can be used in numerous environments, is relatively insensitive to the wavelength of excitation employed and produces a narrow-band spectral signature unique to the molecular vibrations of the analyte.

Since the discovery of SERS in the 1970's, it has been experimentally shown that the SERS signal enhancements are typically observed on metalized (typically silver or gold) nanoscale roughened surfaces.⁽⁴⁰⁻⁴²⁾ Although several open debates regarding the exact magnitude of contributions are still ongoing, the basis of the SERS mechanism is fairly well understood. In the SERS literature, the two mechanisms that control the enhancement of the Raman scattering (SERS) are considered to be the electromagnetic fields generated at or near nanoparticle surfaces and the physical (chemical) adsorption of a target analyte onto a surface.⁽⁴³⁻⁴⁶⁾ The electromagnetic enhancement (EM) is typically deemed to be the stronger contribution, with an enhancement factor (as compared to spontaneous Raman) ranging from approximately 10^4 to 10^{14} ,⁽⁴⁷⁾ while the chemical enhancement (CE) has been suggested to contribute at most 10^2 .⁽⁴⁸⁾ Because of the many sensing advantages of SERS-based techniques, significant research efforts (defense, industrial and academic) have been directed toward fabricating "better" SERS substrates for SERS-based sensor platforms.⁽⁴⁸⁻⁵⁷⁾ Some of these SERS platforms are fabricated from colloids⁽⁵⁸⁾, film over nanospheres⁽⁵⁹⁻⁶²⁾, fiber optic bundles⁽³⁾, nanoparticles,^(55, 63-66) and

lithographically⁽⁵⁾ produced structures. The more sensitive substrate platforms generally have a 15% relative standard deviation (RSD; the measure of the reproducibility of an analysis) from substrate-to-substrate and SERS signal enhancements of 7 to 8 orders of magnitude.^(60, 61) Consequently, many researchers in academia, industry, and government have focused concerted efforts toward increasing the signal enhancement ability, reproducibility, and mass production manufacturing of substrates to increase the utility of this technique.^(5, 67) For the Army and first responders, such a substrate platform with increased sensitivity and reliability would be very advantageous for the detection and identification of unknowns.^(68, 69)

Significant research efforts have also been concentrated on better directing the optimization of the substrate surface from which the SERS enhancement occurs.^(5, 70-73) Based in part on experimental and as well as theoretical efforts, the directed fabrication of SERS platforms has focused on modifying the feature size⁽⁷⁴⁻⁷⁶⁾, spacing between objects, geometry and shape of structures⁽⁷⁷⁻⁷⁹⁾, identity and incorporation of metals on the surface⁽⁸⁰⁾, feature height, and the character of the foundation layer^(59, 81-83) on which the architecture is fabricated.^(69, 84, 85) There are numerous examples in the literature detailing how variation in some of these parameters in some cases can result in very dramatic changes to the overall SERS enhancing capabilities of the substrate surface. Rigorous efforts continue to focus on developing an understanding about how these parameters can synergistically work together to result in a highly reproducible and sensitive SERS substrate. As focus continues on improving the overall sensing capabilities of the SERS surface, congruently research continues to push towards the development of a uniform reproducible mass produced platform necessary to facilitate widespread incorporation of SERS in viable biomimetic sensing platforms.

Some success fabricating both spectrally and physically reproducible SERS substrates has been demonstrated with commercially available standard Klarite™ substrates (Renishaw Diagnostics).^(4, 71, 86-88) These substrates were developed using Si-based semiconductor fabrication techniques.⁽⁴⁾ Klarite substrates are fabricated using a well defined silicon fabrication technique in which a silicon diode mask is defined by optical lithography, and then potassium hydroxide (KOH) surface etched. The process results in an array of highly reproducible inverted pyramid features.⁽⁴⁾ These array pyramids are reported to have “hot spots” or “trapped plasmons” located inside the wells.⁽⁴⁾ These substrates have been previously characterized.⁽⁸⁹⁾ Atomic force microscopy (AFM) images have demonstrated that the inverted pyramids features are about 1.47 μm in width and 1 μm in depth. These substrates have plasmon absorbance bands located at 577 nm and 749 nm, thus demonstrating the usefulness of this substrate for a range of excitation sources. Additionally, due to the fabrication process used, under ideal conditions these substrates have demonstrated typical RSDs ranging from 10-15% under drop and dry conditions. While these standard Klarite substrates do demonstrate a high degree of substrate reproducibility and very low substrate background (SERS signal and surface morphology), for many applications to real-world situations increased analyte sensitivity is still necessary. Recently new prototype Klarite based substrates have been fabricated by Renishaw Diagnostics with the intent to expand substrate sensing capabilities. The morphologies of these substrates dramatically differ in overall shape, pitch and spacing as compared to the standard Klarite substrate resulting in very interesting sensing capabilities.^(90, 91)

In this proceedings paper, preliminary efforts functionalizing commercially available SERS substrates for directed biomimetic hazard sensing will be discussed and demonstrated. These chemical modification methods are achieved using fusion peptide recognition elements for the targeting sensing of hazardous materials such as the energetic sample TNT. By demonstrating these preliminary results, the potential of these SERS substrates for future incorporation in fielded Raman systems for the selective sensing of hazard materials can be imagined.

II. EXPERIMENTAL

Peptide Samples.

Potential peptide recognition elements were selected following a comprehensive literature search. Peptide materials were commercially purchased from RS Synthesis (Louisville, KY 40270), Genscript (Piscataway, NJ 08854) or Peptide 2.0 (Chantilly, VA 20153) and used without additional modification (> 85% purity). Fusion peptides were created with a glycine-glycine-glycine (GGG) spacer between the material binder and the target recognition element. The material binder, target recognition sequence, and naming systems used are listed in Table 1. Upon receipt, all samples were stored at -20°C or below and kept dry. Utilizing the GenScript Peptide Property Calculator, peptides were determined to have a neutral charge, and were suspended in a water only solution. Concentrations of peptide used are listed in the text.

After suspension in water solution, peptide solutions were stored refrigerated while being used for a maximum of one week, and then fresh solutions were made.

Table 1. Example potential peptide recognition elements with material target, sequences investigated, and naming systems listed.

Material of Interest	Peptide Sequence	Literature Reference	Name
Gold Binder	VSGSSPDS	Science, 2006, 312, 885-888	
Gold Binder	LKAHLPPSRLPS	Science, 2006, 312, 885-888	
Gold Binder	TGTSVLIATPYV	Acta Biomater., 2010, 6, 2681-2689	
Gold Binder	WAGAKRIVLRR	J. of Coll. and Interf. Sci., 2012, 365, 97-102	
Gold Binder	AYSSGAPMPFF	Nat. Mater. 2002, 22, 201-204	A3 Binding domain
Gold Binder	AYSSGAPMPFF	Small 2005, 1, No. 1, 1048-1052	A3 Binding domain
Gold Binder	AYSSGAPMPFF	Adv. Mater, 2006, 18, 1988-1992	A3 Binding domain
Gold Binder	VSGSSPDS	Nanolett., 2005, 15 (7), 1429-1434	Au bind
Gold Binder	CGPWALRRSIRRSY GPC	Langmuir, 2008, 24 (21), 12440-12445	cAuBind2
TNT Binder	WFVI	J. of Neurosci., 1999, 19 (17), 4768-7475	ASP1
TNT Binder	WHRTPSTLWGVI	Analytica Chim., 2002, 457, 13-19	
TNT Binder	WHRTPSTLWFVI	ARL Stock	
TNT Binder	WHWQRPLMPVSI	Langmuir, 2008, 24 (9), 4938-4943	TNT-BP
DNT Binder	HPNFSKYILHQR	Langmuir, 2008, 24 (9), 4938-4943	DNT-BP
PA Binder	STLKILPCTFCIWPS	ARL stock	YZ72
PA Binder	VYNGPRYVCTFCTWN	ARL stock	YZ91
PA Binder	VSHILCTFCLWSIP	ARL stock	YZ128

Energetic Materials

Certified energetics material TNT was purchased from Cerrilant. Samples were received in an acetonitrile solution, and were resuspended at a required concentration in a water solution for biological applications.




Figure 1. Example SEM images of commercial SERS substrate shown at different magnifications (A) and (B).

Commercially available SERS substrates.

Commercially available slide mounted standard Klarite™ 302 SERS substrates were purchased from Renishaw Diagnostics, see Figure 1 for an example SEM image of a typical substrate. Slides were individually wrapped and vacuum sealed. The SERS active area on these slides is a small 4 mm x 4 mm wafer with a gold surface. The standard Klarite slides were only used once and opened just prior to measurement to reduce any possible surface fouling. Additionally, the substrate was submerged in ethanol to remove any possible contamination that may have accumulated on the surface. Typically, in SERS data collection on the standard substrates, five measurements across the substrate surface were collected. Most data in this proceedings paper will be presented as an average of a collected data set and the standard deviation error shown, unless otherwise indicated.

Instrumentation and Data Analysis.

Scanning electron microscope (SEM) images were obtained using a FEI environmental SEM (Quanta 200 FEG).

A Renishaw inVia Reflex Raman microscope was used for SERS and Raman spectra collection. Spectra were collected using the NIR 785 nm laser. The laser light was focused onto the sample using a 20X objective, over a range of 100–3200 cm^{-1} , exposures were 10 seconds in length, and 3 accumulations were collected per spot. Approximately 7 mW of power irradiated the surface of the substrate. A total of 10 spectra were collected from each spot on the substrate. Samples were positioned using a motorized XYZ translational stage internal to the microscope. Spectra were collected, and the instrument was run using Wire 3.2 software operating on a dedicated computer.

Data analysis was achieved using IgorPro 6.0 software (Wavemetrics).⁽⁹⁰⁾

III. RESULTS AND DISCUSSION

Gold binding capabilities of peptide

To assess the utility of the fusion peptides, it is critical to first evaluate the gold binding capabilities of the peptide recognition elements. To do this, the dry peptide was resuspended in a water and a solution of 0.8 mg/mL of peptide was made. The solution was then vortexed for 2 minutes to assure complete mixing, and then a 4 μL aliquot of analyte solution was deposited both on the active and non active areas of the Klarite substrate. The sample was then allowed to incubate covered in the refrigerator for 20 minutes to assure ample time for binding. The SERS substrate surface was then repeatedly washed with a 3 mL aliquot of water 4 times to remove any incompletely bound peptide, and then suspended in a water solution to prevent the peptide from drying. Using the Renishaw InVia microscope system both the nonactive and active areas of the SERS substrate were interrogated.




Figure 2. Example spectra from Au binder (A) and Au binder + TNT binder (B). Spectra are shown on active and non active areas of SERS substrate and after multiple washes to remove any unbound sample.

An example of a gold binding peptide commonly encountered in the literature has the sequence AYSSGAPPMPPF. Since this peptide has been well characterized and demonstrated to bind effectively to gold, we wanted to complete a proof of principle study to assure binding and detection of a SERS. In Figure 2A, the SERS and Raman response are shown. In Figure 2A, SERS bands are located at 641 cm^{-1} , 721 cm^{-1} , 1001.8 cm^{-1} , 1031 cm^{-1} , 1204 cm^{-1} and 1443 cm^{-1} . For analysis, the 1001.8 cm^{-1} band was analyzed to determine a signal-to-noise (SNR) ratio. For Figure 2A, the SERS SNR is 30.54 and the Raman SNR is 0. Thus indicating that the peptide has successfully bound to the substrate surface

and does exhibit a distinguishable SERS spectrum. Repeating these studies, it was possible to detect the Au binder on the surface after multiple washings deposited at a concentration as low as millimolar levels.

Recognition elements- TNT binder

For the past decade, significant defense, industry and academic resources and research efforts have focused on the detection and recognition of hazardous materials. Efforts have been centered on such tasks as securing public locations and transportation, maintaining the integrity of checkpoint stations at military sites, and assuring public safety in food chain supplies and the environment. In many of these scenarios, the hazardous material of interest can be energetic in nature. In particular, the detection of high explosives such as 2,4,6-trinitrotoluene (TNT) and their byproducts like dinitrotoluene (DNT) have garnered significant attention because of its proliferations mainly due to availability and ease of use. Increasingly to detect and identify this threat, research agencies are focusing on new innovative bio-inspired technologies for possible solutions.

There are several examples of bio-inspired technologies being used for the detection of explosive materials. In the past decade there have been several example publications that have discussed using various peptides as a capture mechanism for TNT and TNT byproducts. Some examples in the literature that have been highlighted, include an article published in 1999 by Danty et.al.⁽⁹²⁾, in which a binding protein found in an olfactory mechanism of the honeybee was determined to be specific for TNT. The olfactory peptide sequence is WVFI. In an article published in 2002 by Goldman et. al.⁽⁹³⁾, phage displayed peptides were selected for the detection of TNT. From this paper, the sequence for TNT peptide detection was determined to be WHRTPSTLWGVI. In a 2008 article by Jaworski et al.⁽⁹⁴⁾, evolutionary screening of peptides for the selective detection of explosives was conducted. From this screening process, the following peptide sequences specific for TNT and DNT selectivity were determined, WHWQRLMPVSI and HPNFSKYILHQR.




Figure3. Example spectra from (A) A3 binding domain +GGG- ASP1 + TNT, (B) TNT, and (C) A3binding domain –GGG-ASP1 (no TNT). Note several common bands found in (A) are clearly in (B) and (C).

To determine the effectiveness of these peptides for capturing TNT on the commercial SERS substrate, peptide sequences consisting of a gold binding motif, a spacer sequence and the previously mentioned TNT binding peptides were prepared. In this proceedings paper, we will present results from one combination of this fusion peptide. These fusion peptides were immobilized on the Klarite substrate surface and washed several times to insure monolayer coverage. The peptide containing SERS substrate was then exposed to a high concentration (3.67×10^{-2} M) TNT in a water/alcohol/some residual acetonitrile solution. The TNT solution was in a mostly water solution to prevent any

potential damage to the peptide sequence from acetonitrile. A high concentration of TNT was originally selected for these proof-of-principle demonstrations. The TNT solution sat on the SERS substrate surface for 20 minutes to ensure complete binding between the peptide and TNT. The surface was next washed with 2 mL of deionized water to remove any excess and unbound TNT from the surface. See Figure 3 for example SERS spectra from the peptides bound to the surface, and with the addition of TNT. In Figure 3, (C) has the characteristic peptide band located at 1001.8cm^{-1} , while (B) TNT exhibits SERS bands located at 793 cm^{-1} , 823 cm^{-1} , and 1356 cm^{-1} . In Figure 3A, the A3 binding domain-GGG-ASP1 sequence with TNT added demonstrates SERS bands located at 793 cm^{-1} , 823 cm^{-1} , 1001.8cm^{-1} and 1356 cm^{-1} , thus demonstrating that the peptide remained bound and captured TNT onto the sensing surface.

After a demonstration of preliminary sensing with the targeted peptide sequence, it was necessary to determine if the gold binding motif was necessary for immobilizing the peptide on the surface. For these experiments, the Au binding capabilities of the A3 binding domain as compared to a non material specific binding domain were assessed. Peptides of the sequence A3 binding domain-GGG-ASP1 and Graphene binding -GGG-ASP1 were immobilized onto different Klarite substrates overnight. After immobilization, the surface was washed with 2 mL of water, and 5 μL of TNT was added to the surface for 1 hour. Following TNT exposure, the surface was washed with 2 mL of water and SERS measurements were collected. In Figure 4, examples of results collected are shown. In Figure 4A, bands common to TNT are clearly shown. In Figure 4B, there are some minor TNT bands shown, however the overall signal analyzing the TNT 1356 cm^{-1} band is determined to be almost 20X greater. From this Figure, it can be clearly concluded that the A3 binding domain is necessary for peptide immobilization onto the Klarite surface.




Figure 4. Example spectra TNT exposed to (A) A3 binding domain +GGG- ASP1 and the (B) Graphene binder-GGG-ASP1. From these results it can be concluded that A3 binding domain is necessary for immobilizing the peptide onto the gold Klarite surface.

After preliminary efforts to bind the A3-GGG-ASP1 binding motif to the Klarite surface and detect TNT were demonstrated, and it was conclusively shown that the A3 binding domain aids in peptide immobilization the specificity of the peptide for TNT detection was preliminarily assessed. A proof-of-principle experiment was conducted to determine if the peptide binder caused preferential binding of TNT to the surface as compared to non-specific binding. For these experiments, the peptides A3 binding domain-GGG-ASP1 (TNT binder) and A3 binding domain-GGG-Graphene binder (non specific to TNT) were immobilized on the Klarite substrate surface overnight, and washed with 2 mL of water to remove any unbound peptide. The SERS Klarite surface was then exposed to 5 μL of TNT for one hour. Following TNT exposure, the surface was washed with 2 mL of water and SERS measurements were collected. In Figure 5, the results for this binding study are shown. In Figure 5A the A3 binding domain-GGG-ASP1 peptide exposed to TNT is shown. In this figure the peaks observed are located at 793 cm^{-1} , 823 cm^{-1} , 1001.8cm^{-1} and 1356 cm^{-1} , thus demonstrating that the peptide remained bound and captured TNT onto the sensing surface. In Figure 5B, an example spectrum from the non specific peptide exposed to TNT is shown. In this figure, it can be clearly seen that the peptide

band early discussed at 1001.8cm^{-1} is clearly shown, however no SERS bands for TNT are observed. Thus it can be seen that the ASP1 binding domain is necessary for TNT binding to the peptide and being detected.




Figure 5. Example spectra TNT exposed to (A) A3 binding domain +GGG- ASP1 and the (B) A3binding domain –GGG-Graphene binder. From these results it can be concluded that TNT is preferentially bound to the ASP1 containing peptide.

IV. CONCLUSIONS

In conclusion, we have shown a preliminary proof-of-principle demonstration that using specially targeted fusion peptides it is possible to selectively bind hazard materials (like the energetic TNT) to a SERS Klarite sensing surface. Additional studies need to be conducted to determine the optimized conditions needed for higher sensitivity detection of energetic and other Army relevant hazard materials. We also plan to transition these biological sensing schematics onto next generation Klarite sensing.

ACKNOWLEDGEMENTS

Thank you to team members Dimitra Stratis-Cullum, Josh Kogot and Deborah Sarkes for initial assistance with peptide experiments.

REFERENCES

- [1] Hansen, E.H., Miro, M., "Interfacing microfluidic handling with spectroscopic detection for real-life applications via the lab-on-valve platform: A review," *Applied Spectroscopy Reviews*, 335-357 (2008).
- [2] Small, E.W., Peticolas, W.L., "Conformational dependence of the Raman scattering intensities from polynucleotides," *Biopolymers*, 69 (1971).
- [3] Hankus, M.E., Cullum, B.M., "SERS probes for the detection and imaging of biochemical species on the nanoscale," *Proceedings of the SPIE - The International Society for Optical Engineering*, 638004-638001-638012 (2006).
- [4] Netti, M.C., Zoorob, M.E., Charlton, M.C.B., Ayliffe, P., Mahnkopf, S., Stopford, P., Todd, K., Lincoln, J.R., Perney, N.M.B., Baumberg, J.J., "Probing molecules by Surface-Enhanced Raman Spectroscopy," *SPIE*, (2006).
- [5] Tripp, R.A., Dluhy, R.A., Zhao, Y.P., "Novel nanostructures for SERS biosensing," *Nano Today*, 31-37 (2008).
- [6] Bantz, K.C., Meyer, A.F., Wittenberg, N.J., Im, H., Kurtulus, O., Lee, S.H., Lindquist, N.C., Oh, S.H., Haynes, C.L., "Recent progress in SERS biosensing," *Physical Chemistry Chemical Physics*, 11551-11567 (2011).
- [7] Fan, M.K., Andrade, G.F.S., Brolo, A.G., "A review on the fabrication of substrates for surface enhanced Raman spectroscopy and their applications in analytical chemistry," *Analytica Chimica Acta*, 7-25 (2011).

- [8] Peng, H.I., Miller, B.L., "Recent advancements in optical DNA biosensors: Exploiting the plasmonic effects of metal nanoparticles," *Analyst*, 436-447 (2011).
- [9] Yan, B., Boriskina, S.V., Reinhard, B.M., "Design and Implementation of Noble Metal Nanoparticle Cluster Arrays for Plasmon Enhanced Biosensing," *Journal of Physical Chemistry C*, 24437-24453 (2011).
- [10] Goyal, A.K., Johal, E.S., Rath, G., "Nanotechnology for Water Treatment," *Current Nanoscience*, 640-654 (2011).
- [11] Kowalczyk, S.W., Blosser, T.R., Dekker, C., "Biomimetic nanopores: learning from and about nature," *Trends in Biotechnology*, 607-614 (2011).
- [12] Pinheiro, A.V., Han, D.R., Shih, W.M., Yan, H., "Challenges and opportunities for structural DNA nanotechnology," *Nature Nanotechnology*, 763-772 (2011).
- [13] Schoen, A.P., Schoen, D.T., Huggins, K.N.L., Arunagirinathan, M.A., Heilshorn, S.C., "Template Engineering Through Epitope Recognition: A Modular, Biomimetic Strategy for Inorganic Nanomaterial Synthesis," *Journal of the American Chemical Society*, 18202-18207 (2011).
- [14] Pongprayoon, P., Beckstein, O., Sansom, M.S.P., "Biomimetic Design of a Brush-Like Nanopore: Simulation Studies," *Journal of Physical Chemistry B*, 462-468 (2012).
- [15] Kumar, A., Ma, H.L., Zhang, X., Huang, K.Y., Jin, S.B., Liu, J., Wei, T., Cao, W.P., Zou, G.Z., Liang, X.J., "Gold nanoparticles functionalized with therapeutic and targeted peptides for cancer treatment," *Biomaterials*, 1180-1189 (2012).
- [16] Parveen, S., Misra, R., Sahoo, S.K., "Nanoparticles: a boon to drug delivery, therapeutics, diagnostics and imaging," *Nanomedicine-Nanotechnology Biology and Medicine*, 147-166 (2012).
- [17] Wang, C.L., Wei, L.Y., Yuan, C.J., Hwang, K.C., "Reusable Amperometric Biosensor for Measuring Protein Tyrosine Kinase Activity," *Analytical Chemistry*, 971-977 (2012).
- [18] Wang, Y.L., Cui, Y.Y., Zhao, Y.L., Liu, R., Sun, Z.P., Li, W., Gao, X.Y., "Bifunctional peptides that precisely biomineralize Au clusters and specifically stain cell nuclei," *Chemical Communications*, 871-873 (2012).
- [19] Zanolli, L.M., D'Agata, R., Spoto, G., "Functionalized gold nanoparticles for ultrasensitive DNA detection," *Analytical and Bioanalytical Chemistry*, 1759-1771 (2012).
- [20] Elbayoumi, T., Weissig, V., "Implications of Intracellular Distribution of Nanovesicles for Bioimaging Studies," *Journal of Biomedical Nanotechnology*, 620-633 (2009).
- [21] Song, L., Liu, Y.Y., Zhang, Z.F., Wang, X., Chen, J.C., "Bio-inorganic Synthesis of ZnO Powders Using Recombinant His-tagged ZnO Binding Peptide as a Promoter," *Protein Journal*, 516-523 (2010).
- [22] Romano, N.H., Sengupta, D., Chung, C., Heilshorn, S.C., "Protein-engineered biomaterials: Nanoscale mimics of the extracellular matrix," *Biochimica Et Biophysica Acta-General Subjects*, 339-349 (2011).
- [23] Seker, U.O.S., Demir, H.V., "Material Binding Peptides for Nanotechnology," *Molecules*, 1426-1451 (2011).
- [24] Tarasov, S.G., Gaponenko, V., Howard, O.M.Z., Chen, Y.H., Oppenheim, J.J., Dyba, M.A., Subramaniam, S., Lee, Y., Michejda, C., Tarasova, N.I., "Structural plasticity of a transmembrane peptide allows self-assembly into biologically active nanoparticles," *Proceedings of the National Academy of Sciences of the United States of America*, 9798-9803 (2011).
- [25] Gazarian, K., Gazarian, T., Betancourt, J.I.S., Morales, R.A.A., "Immunogenic peptides from phage display libraries with potential of protecting mice against the Pseudorabies virus," *Veterinary Microbiology*, 29-36 (2011).
- [26] Hashimoto, K., Yoshinari, M., Matsuzaka, K., Shiba, K., Inoue, T., "Identification of peptide motif that binds to the surface of zirconia," *Dental Materials Journal*, 935-940 (2011).
- [27] Li, Z.K., Fan, J.S., Zhao, W.X., Jin, L., Ma, L., "The specific binding of peptide ligands to cardiomyocytes derived from mouse embryonic stem cells," *Journal of Peptide Science*, 771-782 (2011).
- [28] Hamzeh-Mivehroud, M., Mahmoudpour, A., Dastmalchi, S., "Identification of New Peptide Ligands for Epidermal Growth Factor Receptor Using Phage Display and Computationally Modeling their Mode of Binding," *Chemical Biology & Drug Design*, 246-259 (2012).
- [29] Lakshmanan, A., Hauser, C.A.E., "Ultrasmall Peptides Self-Assemble into Diverse Nanostructures: Morphological Evaluation and Potential Implications," *International Journal of Molecular Sciences*, 5736-5746 (2011).
- [30] Sakakibara, K., Hill, J.P., Ariga, K., "Thin-Film-Based Nanoarchitectures for Soft Matter: Controlled Assemblies into Two-Dimensional Worlds," *Small*, 1288-1308 (2011).

- [31] Tarabout, C., Roux, S., Gobeaux, F., Fay, N., Pouget, E., Meriadec, C., Ligeti, M., Thomas, D., Ijsselstijn, M., Bessellievre, F., Buisson, D.A., Verbavatz, J.M., Petitjean, M., Valery, C., Perrin, L., Rousseau, B., Artzner, F., Paternostre, M., Cintrat, J.C., "Control of peptide nanotube diameter by chemical modifications of an aromatic residue involved in a single close contact," *Proceedings of the National Academy of Sciences of the United States of America*, 7679-7684 (2011).
- [32] Yang, L., Zhang, L.J., Webster, T.J., "Nanobiomaterials: State of the Art and Future Trends," *Advanced Engineering Materials*, B197-B217 (2011).
- [33] Milane, L., Duan, Z.F., Amiji, M., "Pharmacokinetics and biodistribution of lonidamine/paclitaxel loaded, EGFR-targeted nanoparticles in an orthotopic animal model of multi-drug resistant breast cancer," *Nanomedicine-Nanotechnology Biology and Medicine*, 435-444 (2011).
- [34] Taratula, O., Garbuzenko, O.B., Chen, A.M., Minko, T., "Innovative strategy for treatment of lung cancer: targeted nanotechnology-based inhalation co-delivery of anticancer drugs and siRNA," *Journal of Drug Targeting*, 900-914 (2011).
- [35] Jong, H.J., Na, J.H., Jin, B.S., Lee, W.K., Lee, W.H., Jung, H.J., Kim, S.C., Lim, S.H., Yu, Y.G., "Identification of Dinitrotoluene Selective Peptides by Phage Display Cloning," *Bulletin of the Korean Chemical Society*, 3703-3706 (2010).
- [36] Heller, D.A., Pratt, G.W., Zhang, J.Q., Nair, N., Hansborough, A.J., Boghossian, A.A., Reuel, N.F., Barone, P.W., Strano, M.S., "Peptide secondary structure modulates single-walled carbon nanotube fluorescence as a chaperone sensor for nitroaromatics," *Proceedings of the National Academy of Sciences of the United States of America*, 8544-8549 (2011).
- [37] Hwang, K.S., Lee, M.H., Lee, J., Yeo, W.S., Lee, J.N., Kim, K.M., Kang, J.Y., Kim, T.S., "Peptide receptor-based selective dinitrotoluene detection using a microcantilever sensor," *Biosensors & Bioelectronics*, 249-254 (2011).
- [38] Jaworski, J., Yokoyama, K., Zueger, C., Chung, W.J., Lee, S.W., Majumdar, A., "Polydiacetylene Incorporated with Peptide Receptors for the Detection of Trinitrotoluene Explosives," *Langmuir*, 3180-3187 (2011).
- [39] Yildirim, A., Acar, H., Erkal, T.S., Bayindir, M., Guler, M.O., "Template-Directed Synthesis of Silica Nanotubes for Explosive Detection," *Acs Applied Materials & Interfaces*, 4159-4164 (2011).
- [40] M. Fleischmann, Hendra, P.J., McQuillan, a.A.J., *Chem. Phys.Lett.*, 163-166 (1974).
- [41] Jeanmaire, D.L., Vanduyne, R.P., "Surface Raman Spectroelectrochemistry. 1. Heterocyclic, Aromatic, and Aliphatic-Amines Adsorbed on Anodized Silver Electrode," *J. Electroanal. Chem.*, 1 (1977).
- [42] Albrecht, M.G., Creighton, J.A., "Anomalous Intense Raman-Spectra of Pyridine at a Silver Electrode," *J. Am. Chem. Soc.*, 5215 (1977).
- [43] Dornhaus, R., Benner, R.E., Chang, R.K., Chabay, I., "SURFACE-PLASMON CONTRIBUTION TO SERS," *Surface Science*, 367-373 (1980).
- [44] Kerker, M., Siiman, O., Bumm, L.A., Wang, D.S., "SURFACE ENHANCED RAMAN-SCATTERING (SERS) OF CITRATE ION ADSORBED ON COLLOIDAL SILVER," *Applied Optics*, 3253-3255 (1980).
- [45] Kerker, M., Wang, D.S., Chew, H., "SURFACE ENHANCED RAMAN-SCATTERING (SERS) BY MOLECULES ADSORBED AT SPHERICAL-PARTICLES," *Applied Optics*, 3373-3388 (1980).
- [46] Wang, D.S., Chew, H., Kerker, M., "ENHANCED RAMAN-SCATTERING AT THE SURFACE (SERS) OF A SPHERICAL-PARTICLE," *Applied Optics*, 2256-2257 (1980).
- [47] Kneipp, K., Kneipp, H., Dasari, R.R., Feld, M.S., "Single molecule Raman spectroscopy using silver and gold nanoparticles," *Indian Journal of Physics and Proceedings of the Indian Association for the Cultivation of Science-Part B*, 39-47 (2003).
- [48] Campion, A., Kambhampati, P., "Surface-enhanced Raman scattering," *Chemical Society Reviews*, 241-250 (1998).
- [49] Cotton, T.M., Kim, J.H., Chumanov, G.D., "APPLICATION OF SURFACE-ENHANCED RAMAN-SPECTROSCOPY TO BIOLOGICAL-SYSTEMS," *Journal of Raman Spectroscopy*, 729-742 (1991).
- [50] Kneipp, K., Kneipp, H., Itzkan, I., Dasari, R.R., Feld, M.S., "Surface-enhanced Raman scattering and biophysics," *Journal of Physics-Condensed Matter*, R597-R624 (2002).
- [51] Baker, G.A., Moore, D.S., "Progress in plasmonic engineering of surface-enhanced Raman-scattering substrates toward ultra-trace analysis," *Analytical and Bioanalytical Chemistry*, 1751-1770 (2005).

- [52] Yonzon, C.R., Stuart, D.A., Zhang, X.Y., McFarland, A.D., Haynes, C.L., Van Duyne, R.P., "Towards advanced chemical and biological nanosensors - An overview," *Talanta*, 438-448 (2005).
- [53] Willets, K.A., Van Duyne, R.P., Localized surface plasmon resonance spectroscopy and sensing, in: *Annual Review of Physical Chemistry*, 2007, pp. 267-297.
- [54] Banholzer, M.J., Millstone, J.E., Qin, L.D., Mirkin, C.A., "Rationally designed nanostructures for surface-enhanced Raman spectroscopy," *Chemical Society Reviews*, 885-897 (2008).
- [55] Kneipp, J., Kneipp, H., Kneipp, K., "SERS - a single-molecule and nanoscale tool for bioanalytics," *Chemical Society Reviews*, 1052-1060 (2008).
- [56] Qian, X.M., Nie, S.M., "Single-molecule and single-nanoparticle SERS: from fundamental mechanisms to biomedical applications," *Chemical Society Reviews*, 912-920 (2008).
- [57] Stiles, P.L., Dieringer, J.A., Shah, N.C., Van Duyne, R.R., Surface-Enhanced Raman Spectroscopy, in: *Annual Review of Analytical Chemistry*, 2008, pp. 601-626.
- [58] Qian, X.M., Peng, X.H., Ansari, D.O., Yin-Goen, Q., Chen, G.Z., Shin, D.M., Yang, L., Young, A.N., Wang, M.D., Nie, S.M., "In vivo tumor targeting and spectroscopic detection with surface-enhanced Raman nanoparticle tags," *Nature Biotechnology*, 83-90 (2008).
- [59] Li, H., Baum, C.E., Cullum, B.M., "Characterization of novel Gold SERS Substrates with Multilayer Enhancements," *SPIE*, (2006).
- [60] Li, H., Baum, C.E., Sun, J., Cullum, B.M., "Multilayer enhanced SERS active materials: fabrication, characterization, and application to trace chemical detection," *SPIE*, 621804 (2006).
- [61] Li, H., Patel, P.H., Cullum, B.M., "Novel multilayered SERS substrates for trace chemical and biochemical analysis," *SPIE*, (2004).
- [62] Vo-Dinh, T., Cullum, B.M., Stokes, D.L., *Nanosensors and biochips: frontiers in biomolecular diagnostics*, in, Elsevier Science Sa, 2001, pp. 2-11.
- [63] Cao, Y.W.C., Jin, R.C., Mirkin, C.A., "Nanoparticles with Raman Spectroscopic Fingerprints for DNA and RNA Detection," *Science*, 1536 (2002).
- [64] Chowdhury, M.H., Gant, V.A., Trache, A., Baldwin, A., Meininger, G.A., Cote, G.L., "Use of surface-enhanced Raman spectroscopy for the detection of human integrins," *Journal of Biomedical Optics*, 8 (2006).
- [65] Culha, M., Adiguzel, A., Yazici, M.M., Kahraman, M., Sahin, F., Gulluce, M., "Characterization of Thermophilic Bacteria Using Surface-Enhanced Raman Scattering," *Applied Spectroscopy*, 1226-1232 (2008).
- [66] Jarvis, R.M., Law, N., Shadi, I.T., Oâ€™Brien, P., Lloyd, J.R., Goodacre, R., "Surface-Enhanced Raman Scattering from Intracellular and Extracellular Bacterial Locations," *Analytical Chemistry*, 6741-6746 (2008).
- [67] Driskell, J.D., Shanmukh, S., Liu, Y.J., Hennigan, S., Jones, L., Zhao, Y.P., Dluhy, R.A., Krause, D.C., Tripp, R.A., "Infectious agent detection with SERS-active silver nanorod arrays prepared by oblique angle deposition," *IEEE Sensors Journal*, 863-870 (2008).
- [68] Holthoff, E.L., Stratis-Cullum, D.N., Hankus, M.E., "A Nanosensor for TNT Detection Based on Molecularly Imprinted Polymers and Surface Enhanced Raman Scattering," *Sensors*, 2700-2714 (2011).
- [69] Holthoff, E.L., Stratis-Cullum, D.N., Hankus, M.E., "Xerogel-Based Molecularly Imprinted Polymers for Explosives Detection," *Chemical, Biological, Radiological, Nuclear, and Explosives (Cbrne) Sensing XI*, (2010).
- [70] Norrod, K.L., Sudnik, L.M., Rousell, D., Rowlen, K.L., "Quantitative Comparison of Five SERS Substrates: Sensitivity and Limit of Detection," *Appl. Spectrosc.*, 994 (1997).
- [71] Alexander, T.A., Le, D.M., "Characterization of a commercialized SERS-active substrate and its application to the identification of intact *Bacillus* endospores," *Appl. Opt.*, (2007).
- [72] Guicheteau, J., Christesen, S., Emge, D., Wilcox, P., Fountain, A.W., III, "Assessing Metal Nanofabricated Substrates for Surface-Enhanced Raman Scattering (SERS) Activity and Reproducibility," *Applied Spectroscopy*, 144-151 (2011).
- [73] Kneipp, K., Kneipp, H., Itzkan, I., Dasari, R.R., Feld, M.S., "Surface-enhanced Raman scattering: A new tool for biomedical spectroscopy," *Current Science*, 915-924 (1999).
- [74] Jones, J.P., Fell, N.F., Alexander, T., Fountain, A.W., "Photonic Nanostructures as SERS Substrates for Reproducible Characterization of Bacterial Spores," *SPIE*, 94-104 (2004).

- [75] Kneipp, H., Kneipp, K., "Surface-enhanced hyper Raman scattering in silver colloidal solutions," *Journal of Raman Spectroscopy*, 551-554 (2005).
- [76] Kao, P., Malvadkar, N.A., Cetinkaya, M., Wang, H., Allara, D.L., Demirel, M.C., "Surface-enhanced Raman detection on metalized nanostructured poly(p-xylylene) films," *Advanced Materials*, 3562-+ (2008).
- [77] Becker, M., Sivakov, V., Goesele, U., Stelzner, T., Andra, G., Reich, H.J., Hoffmann, S., Michler, J., Christiansen, S.H., "Nanowires enabling signal-enhanced nanoscale Raman spectroscopy," *Small*, 398-404 (2008).
- [78] Xie, J., Zhang, Q., Lee, J.Y., Wang, D.I.C., "The Synthesis of SERS-Active Gold Nanoflower Tags for In Vivo Applications," *ACS Nano*, 2473-2480 (2008).
- [79] Braun, G., Pavel, I., Morrill, A.R., Seferos, D.S., Bazan, G.C., Reich, N.O., Moskovits, M., "Chemically Patterned Microspheres for Controlled Nanoparticle Assembly in the Construction of SERS Hot spots," *J. Am. Chem. Soc.*, 7760-7761 (2007).
- [80] Kho, K.W., Shen, Z.X., Zeng, H.C., Soo, K.C., Olivo, M., "Deposition Method for Preparing SERS-Active Gold Nanoparticle Substrates," *Analytical Chemistry*, 7462-7471 (2005).
- [81] Jian, S., Hankus, M.E., Cullum, B.M., "SERS based immuno-microwell arrays for multiplexed detection of foodborne pathogenic bacteria," *Proceedings of the SPIE - The International Society for Optical Engineering*, 73130K (73110 pp.) (2009).
- [82] Hankus, M.E., Gibson, G., Chandrasekharan, N., Cullum, B.M., "Surface-enhanced Raman scattering (SERS) - nanoimaging probes for biological analysis," *Smart Medical and Biomedical Sensor Technology II*, 106-116 (2004).
- [83] Kuncicky, D.M., Prevo, B.G., Velev, O.D., "Controlled assembly of SERS substrates templated by colloidal crystal films," *J. Mater. Chem.*, 1207-1211 (2006).
- [84] Kneipp, K., Kneipp, H., Kneipp, J., "Surface-enhanced Raman scattering in local optical fields of silver and gold nanoaggregates - From single-molecule Raman spectroscopy to ultrasensitive probing in live cells," *Accounts of Chemical Research*, 443-450 (2006).
- [85] Lee, C.H., Hankus, M.E., Tian, L., Pellegrino, P.M., Singamaneni, S., "Highly Sensitive SERS Substrates Based on Filter Paper Loaded with Plasmonic Nanostructures," *Anal. Chem.*, (2011).
- [86] Alexander, T.A., "Applications of Surface-Enhanced Raman Spectroscopy (SERS) for Biosensing: An Analysis of Reproducible, Commercially Available Substrates," *SPIE*, 600703 (2005).
- [87] Alexander, T.A., "Development of Methodology Based on commercialized SERS-active Substrates for Rapid Discrimination of Poxviridae Virions," *Anal. Chem.*, 2817-2825 (2008).
- [88] Alexander, T.A., Pellegrino, P.M., Gillespie, J.B., "Near-Infrared Surface-Enhanced-Raman-Scattering-Mediated Detection of Single Optically Trapped Bacterial Spores," *Applied Spectroscopy*, 1340-1345 (2003).
- [89] Hankus, M., Stratis-Cullum, D., Pellegrino, P., "Enabling Technologies for Point and Remote Sensing of Chemical and Biological Agents Using Surface Enhanced Raman Scattering (SERS) Techniques," *ARL Technical Report*, 132 (2009).
- [90] Hankus, M.E., Stratis-Cullum, D.N., Pellegrino, P.M., "Characterization of next-generation commercial surface-enhanced Raman scattering (SERS) substrates " *Proc. SPIE*, 80180P (2011).
- [91] Hankus, M.E., Stratis-Cullum, D.N., Pellegrino, P.M., "Surface enhanced Raman scattering (SERS)-based next generation commercially available substrate: physical characterization and biological application," *Proceedings of the SPIE - The International Society for Optical Engineering*, (2011).
- [92] Danty, E., Briand, L., Michard-Vanhee, C., Perez, V., Arnold, G., Gaudemer, O., Huet, D., Huet, J.C., Ouali, C., Masson, C., Pernollet, J.C., "Cloning and expression of a queen pheromone-binding protein in the honeybee: an olfactory-specific, developmentally regulated protein," *Journal of Neuroscience*, 7468-7475 (1999).
- [93] Goldman, E.R., Pazirandeh, M.P., Charles, P.T., Balighian, E.D., Anderson, G.P., "Selection of phage displayed peptides for the detection of 2,4,6-trinitrotoluene in seawater," *Analytica Chimica Acta*, 13-19 (2002).
- [94] Jaworski, J.W., Raorane, D., Huh, J.H., Majumdar, A., Lee, S.W., "Evolutionary screening of biomimetic coatings for selective detection of explosives," *Langmuir*, 4938-4943 (2008).

NO. OF COPIES	ORGANIZATION
1	ADMNSTR DEFNS TECHL INFO CTR ATTN DTIC OCP 8725 JOHN J KINGMAN RD STE 0944 FT BELVOIR VA 22060-6218
8	US ARMY RSRCH LAB ATTN IMAL HRA MAIL & RECORDS MGMT ATTN RDRL CIO LL TECHL LIB ATTN RDRL CIO LT TECHL PUB ATTN RDRL SEE E M FARRELL (5 HCS) ADELPHI MD 20783-1197

INTENTIONALLY LEFT BLANK.