



EDGEWOOD CHEMICAL BIOLOGICAL CENTER

U.S. ARMY RESEARCH, DEVELOPMENT AND ENGINEERING COMMAND
Aberdeen Proving Ground, MD 21010-5424

ECBC-TR-1142

SCREENING SAMPLES FOR ARSENIC BY INDUCTIVELY COUPLED PLASMA-MASS SPECTROMETRY FOR TREATY SAMPLES

Kristin K. Clark
Lynn D. Hoffland

RESEARCH AND TECHNOLOGY DIRECTORATE

February 2014

Approved for public release; distribution is unlimited.



Disclaimer

The findings in this report are not to be construed as an official Department of the Army position unless so designated by other authorizing documents.

REPORT DOCUMENTATION PAGE*Form Approved*
OMB No. 0704-0188

Public reporting burden for this collection of information is estimated to average 1 h per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. **PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.**

1. REPORT DATE (DD-MM-YYYY)

XX-02-2014

2. REPORT TYPE

Final

3. DATES COVERED (From - To)

Aug 2012 - Oct 2012

4. TITLE AND SUBTITLE

Screening Samples for Arsenic by Inductively Coupled Plasma-Mass Spectrometry for Treaty Samples

5a. CONTRACT NUMBER**5b. GRANT NUMBER****5c. PROGRAM ELEMENT NUMBER****6. AUTHOR(S)**

Clark, Kristin K.; and Hoffland, Lynn D.

5d. PROJECT NUMBER**5e. TASK NUMBER****5f. WORK UNIT NUMBER****7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES)**

Director, ECBC, ATTN: RDCB-DRC-F, APG, MD 21010-5424

8. PERFORMING ORGANIZATION REPORT NUMBER

ECBC-TR-1142

9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES)Organization for the Prohibition of Chemical Weapons
Mr. Hugh Gregg, Head OPCW Laboratory
Heulweg 28-30, 2288 GN Rijswijk The Netherlands**10. SPONSOR/MONITOR'S ACRONYM(S)**

OPCW

11. SPONSOR/MONITOR'S REPORT NUMBER(S)**12. DISTRIBUTION / AVAILABILITY STATEMENT**

Approved for public release; distribution unlimited.

13. SUPPLEMENTARY NOTES**14. ABSTRACT:**

Arsenic is a metal found in chemical warfare agents (CWA), lewisite, and the hydrolysis products chlorovinyl arsonous acid (CVAA). One of the missions of the Forensic Analytical Center is to screen samples for compliance with the international treaty to not proliferate CWAs. In order to determine if a sample may contain a CWA screening techniques need to be established and validated. This report describes the screening method used for potential arsenic containing samples when an organic solvent may be present to mask the arsenic.

15. SUBJECT TERMS

Chlorovinylarsonous acid

Arsenic

OPCW

Dichloromethane interference

Lewisite

Hexane interference

Inductively coupled plasma- mass spectrometry

16. SECURITY CLASSIFICATION OF:**17. LIMITATION OF ABSTRACT****18. NUMBER OF PAGES****19a. NAME OF RESPONSIBLE PERSON**

Renu B. Rastogi

a. REPORT**b. ABSTRACT****c. THIS PAGE**

U

U

U

UU

32

19b. TELEPHONE NUMBER (include area code)

(410) 436-7545

Standard Form 298 (Rev. 8-98)
Prescribed by ANSI Std. Z39.18

Blank

PREFACE

The work described in this report was started in August 2012 and completed in October 2012.

The use of either trade or manufacturers' names in this report does not constitute an official endorsement of any commercial products. This report may not be cited for purposes of advertisement.

This report has been approved for public release.

Blank

CONTENTS

1.	INTRODUCTION	1
1.1	Testing Requirements for OPCW Proficiency Test.....	1
1.2	Lewisite Structure and Hydrolyzed Products	2
2.	INSTRUMENTATION	3
2.1	Inductively Coupled Plasma-Mass Spectrometry	3
2.2	Self Aspiration Mode	3
2.3	Interferents	3
3.	MATERIALS AND METHODS.....	4
4.	RESULTS	5
4.1	Calibration Curve Range.....	5
4.2	Linearity	7
4.3	Performance and Analysis Studies (P&A).....	8
4.4	Measurement of Uncertainty.....	15
4.5	Analysis of Unknown Samples	16
5.	IMPLICATIONS FOR FUTURE APPLICATIONS.....	18
	LITERATURE CITED	19
	BIBLIOGRAPHY	20
	ACRONYMS AND ABBREVIATIONS	21

FIGURES

1.	Structure of L-1, L-2, L-3, and precursor arsenic trichloride	2
2.	Day one calibration curves for arsenic from 0 to 25 ppb showing the variability in the three modes of detection at low levels in the presence of (A) dichloromethane and (B) hexane	6
3.	Calibration levels from 0 to 1000 ppb for the three modes of analysis with (A) dichloromethane and (B) hexane as interferents	7

TABLES

1.	Method of Evaluating Laboratory Performance ¹	2
2.	Arsenic Containing CWA Compounds and Precursors Used in This Study ²	3
3.	Chemical Names and Corresponding CAS Numbers for the Chemicals Used	4
4.	Nominal Concentrations of Arsenic Calibration Standards, ppb = $\mu\text{g/L}$	5
5.	Linearity Measurements for P&A Day 1, Day 2 and Method Detection Limits (MDL)	8
6.	ICP-MS Data for the Arsenic P&A Study. "No Gas" Mode with Dichloromethane	9
7.	ICP-MS Data for the Arsenic P&A Study. Helium Mode with Dichloromethane	10
8.	ICP-MS Data for the Arsenic P&A Study. Hydrogen Mode with Dichloromethane	11
9.	ICP-MS Data for the Arsenic P&A Study. "No Gas" Mode with Hexane	12
10.	ICP-MS Data for the Arsenic P&A Study. Helium Mode with Hexane	13
11.	ICP-MS Data for the Arsenic P&A Study. Hydrogen Mode with Hexane	14
12.	Measurement Uncertainty Three Modes with Dichloromethane	15
13.	Measurement Uncertainty Three Modes with Hexane	16
14.	Measurement of Unknown Arsenic Solution with and without Calibration with Interfering Solvent	17

SCREENING SAMPLES FOR ARSENIC BY INDUCTIVELY COUPLED PLASMA-MASS SPECTROMETRY FOR TREATY SAMPLES

1. INTRODUCTION

The U.S. Army Edgewood Chemical Biological Center-Forensic Analytical Center (ECBC-FAC) is a designated lab under the Organization for the Prohibition of Chemical Weapons (OPCW). This organization is designed to implement the Chemical Weapons Convention treaty¹. Independent experts comprise the Scientific Advisory Board, which is charged with the task of maintaining current technologies, methodologies and equipment to address relevant expertise and analysis for reports under the OPCW treaty. ECBC-FAC has participated in the generation of methods for the official Blue Book of *Recommended Operating Procedures for Analysis in the Verification of Chemical Disarmament*².

As part of the evaluation and actual screening process ECBC-FAC developed a new method to determine if chemical warfare agents (CWAs) containing arsenic are present. Arsenic is found in Arsenic trichloride, and the following chemical warfare agents and their degradation products: 1, 2-Chlorovinylchloroarsine (Lewisite, L); Bis(2-chlorovinyl)chloroarsine (Lewisite 2, L-2); and Tris(2-chlorovinyl)arsine (Lewisite 3, L-3). L-2 and L-3 are often found as impurities in L-1.

Past tests and samples included interferents to make the analysis and positive identification of the compound(s) more challenging and realistic. The main interferents utilized in screening for L were the solvents used to dissolve the solid agents. Hexane and dichloromethane readily accompanied L in commercial formulations and test samples received for proficiency testing. The 6th, 9th, 21st, 25th, and 31st Official OPCW Proficiency Tests all contained arsenic-containing L-spiking chemicals. The hydrolyzed products, for example, chlorovinylarsonous acid (CVAA), are also important compounds to screen for in these samples, as they could indicate that L had been present.

1.1 Testing Requirements for OPCW Proficiency Test

The criteria under C-I/DEC.65 stipulates that a designated laboratory maintains a quality system in accordance with International Organization for Standardization/International Electrotechnical Commission (ISO/IEC) 17025:2005 standards and participates in the annual proficiency-testing to maintain a rating of three As or two As and one B in consecutive tests. Evaluation of a laboratory's performance is based on the number of correctly identified chemicals minus misidentified chemicals to generate a performance rating (Table 1).

Table 1. Method of Evaluating Laboratory Performance¹

Performance Criteria Fulfilled	Identification of Chemicals	Performance Scoring	Performance Rating
Yes	Laboratory identifies all chemicals	Maximum score	A
Yes	Laboratory identifies all chemicals except one	Maximum score minus two	B
Yes	Laboratory identifies more than half of the chemicals	Score between zero and maximum minus two	C
Yes	Laboratory misses more chemicals than it identifies	Negative score	D
No		No score	Failure

1.2 Lewisite Structure and Hydrolyzed Products

L is a mixture of *cis* and *trans* isomers that are formed from the mixture of arsenic trichloride and acetylene. The structure of the different isomers and compounds that form L are shown in Figure 1. Physical chemical properties are listed in Table 2.




Figure 1. Structure of L-1, L-2, L-3, and precursor arsenic trichloride.

Table 2. Arsenic Containing CWA Compounds and Precursors Used in This Study²

Compounds	Chemical Name	Structure	Molecular Weight	Percent Weight Arsenic
Lewisite 1	2-chlorovinylarsonous dichloride	$\text{AsCl}_3\text{C}_2\text{H}_2$	207.27	36.15%
Lewisite 2	Bis(2-chlorovinyl) arsinous chloride	$\text{AsCl}_3\text{C}_4\text{H}_4$	233.27	32.12%
Lewisite 3	Tris(2-chlorovinyl) arsine	$\text{AsCl}_3\text{C}_6\text{H}_6$	259.27	28.90%
Arsenic Trichloride	Arsenic Trichloride	AsCl_3	181.27	41.33%

2. INSTRUMENTATION

2.1 Inductively Coupled Plasma-Mass Spectrometry

An Agilent Technologies (Santa Clara, CA) 7500cx series inductively coupled plasma-mass spectrometry (ICP-MS) equipped with a CETAC Technologies (Omaha, NE) ASX500 autosampler was used for the method development and sample analysis. ChemStation Version B.03.06 (Agilent Technologies, Santa Clara, CA) 2007 software was used to process the data.

2.2 Self Aspiration Mode

The data was collected with the nebulizer running in self aspiration mode. It was found that using a nebulizer in self aspiration mode reduced the signal noise.

2.3 Interferents

Based on previous proficiency tests, the unknown samples are commonly in an organic solvent of hexane or dichloromethane³. When the sample is prepared for ICP-MS analysis, there is a trace level of organic solvent present. When diluting the sample in 3% nitric acid solution for ICP-MS analysis, if 250 μL of an unknown organic sample in pure solvent is diluted with 9.75 mL of 3% nitric acid, then 2.5% of the organic solvent will be in the autosampler vial. It is critical that all solutions analyzed with this sample contain the same percentage of the identical organic solvent. The process of aspirating the sample into the nebulizer and spray chamber for aerosolization is affected by the sample matrix. If the calibration standards do not have the same matrix of solvent and acid the arsenic levels detected may be inaccurate. All calibration curves were prepared with the same level of organic solvent as found in the test samples.

3. MATERIALS AND METHODS

Liquid argon, helium, and hydrogen gases were used in the collision and reaction chamber of the ICP-MS. All water was ultrapure 18 M Ω or double distilled water free of metals. High purity, distilled concentrated nitric acid (HNO₃), with minimum trace metals suitable for the intended detection level were purchased from VWR International, LLC (Radnor, PA). Agilent tuning solutions with 1 μ g/L (ppb) and 10 μ g/L (ppb) cerium, cobalt, lithium, thallium, and yttrium were used to verify the performance of the instrument at the start of each day.

All chemicals used for this report and their corresponding Chemical Abstracts Service (CAS) numbers are presented in Table 3. Solutions of the arsenic compounds were prepared by dissolving the neat standard in the corresponding organic solvent of hexane or dichloromethane in a glass volumetric flask. These “stock” solutions were then used to create the calibration standards and the “unknowns” for ICP-MS analysis. Stock solutions were made from Inorganic Ventures (Christiansburg, VA) ICP-MS standards. Table 4 outlines the preparation for the diluted stock solution created from the purchased stock solutions. Once all arsenic solution concentrations were adjusted, organic solvent was added to provide equal concentration in all samples and standards.

Table 3. Chemical Names and Corresponding CAS Numbers for the Chemicals Used

Chemical Name	CAS Number
Argon	7440-37-1
Arsenic	7440-38-2
Arsenic Trichloride	7784-34-1
Cerium	7440-45-1
Cobalt	7440-48-4
Dichloromethane	75-09-2
Hexane	110-54-3
Lewisite 1	541-25-3
Lewisite 2	40334-69-8
Lewisite 3	40334-70-1
Lithium	7439-93-2
Nitric Acid	7697-37-2
Thallium	7440-28-0
Water	7732-18-5
Yttrium	7440-65-5

Table 4. Nominal Concentrations of Arsenic Calibration Standards, ppb = µg/L

Stock Concentration of Arsenic (ppb)	Stock Volume (mL)	Final Volume (mL)	Concentration of Arsenic (ppb)
1.00E+06	0.1000	100.0	1000.0
1000	0.2500	50.00	5.000
1000	0.5000	50.00	10.00
1000	1.250	50.00	25.00
1000	2.500	50.00	50.00
1000	5.000	50.00	100.0
1000	12.50	50.00	250.0
1000	25.00	50.00	500.0
1000	37.50	50.00	750.0
1000	50.00	50.00	1000.0
1000	25.00	50.00	500.0
1000	0.000	50.00	0.000

4. RESULTS

4.1 Calibration Curve Range

Arsenic solutions prepared in dichloromethane and hexane were prepared at 1.0×10^6 ppb level. The samples were serially diluted as described above with 3% nitric acid. A range of calibration solutions were prepared from 0 to 1000 ppb.

At low levels, the measured values were 5% different from the expected values.

This was true for both interferrants shown in Figure 2. Figure 3 shows the full calibration range from 0 to 1000 ppb for each solvent.

Quantitative analysis of arsenic in solution was performed using a minimum of a five-point calibration curve and forcing the intercept through zero. Equation 1 was used to convert the signal into a concentration using the slope and intercept of the calibration curve

$$Y = \frac{(A/S) - B}{M} \quad (1)$$

where Y is the measured concentration (ppb), A is the signal of the analyte, S is the signal of the standard, B is the y-intercept, and M is the slope of the calibration curve. The measured concentration was converted to a final concentration by multiplying by a dilution factor

$$C_f = C_M \times (V_f/V_s) \quad (2)$$

where C_f is the final concentration (ppb), C_M is the measured concentration (ppb), V_f is the final volume of the solution analyzed, and V_s is the volume of the sample.




Figure 2. Day one calibration curves for arsenic from 0 to 25 ppb showing the variability in the three modes of detection at low levels in the presence of (A) dichloromethane and (B) hexane.




Figure 3. Calibration levels from 0 to 1000 ppb for the three modes of analysis with (A) dichloromethane and (B) hexane as interferents.

4.2 Linearity

An eight point calibration curve was generated using dichloromethane as the trace organic and a nine point calibration curve for hexane as the trace organic. A minimum of six points should be used for daily calibrations. The calibration curves were linear for performance and analysis studies (P&As) and measurement of uncertainty studies with hexane and dichloromethane (Table 5).

Table 5. Linearity Measurements for P&A Day 1, Day 2 and Method Detection Limits (MDL)

Analyte	Curve Type	Coefficient of Determination (r ²)		
		External Standard Mode		
		MDL	P&A Day 1	P&A Day 2
Mode:Hydrogen Solvent: Dichloromethane	Linear	0.9986	0.9996	0.9986
Mode:Hydrogen Solvent: Hexane	Linear	0.9996	1.0000	0.9996
Mode:Helium Solvent: Dichloromethane	Linear	0.9990	0.9992	0.9990
Mode:Helium Solvent: Hexane	Linear	0.9976	1.0000	0.9976
Mode:No Gas Solvent: Dichloromethane	Linear	1.0000	1.0000	1.0000
Mode:No Gas Solvent: Hexane	Linear	0.9998	0.9998	0.9998

4.3 Performance and Analysis Studies (P&A)

The method reproducibility was conducted over two days using six solutions at each calibration level. A different analyst prepared the solutions on each day to capture the variability of the analyst. The same stock solution was used to prepare the solutions so the error was from the dilutions and pipette technique of the analyst. Tables 6 through 11 show the different solvents under the different gas conditions in the collision cell.

The precision of the samples was analyzed by calculating the relative percent difference (RPD) between the measured and expected concentrations

$$\left(\frac{C_M - C_E}{\text{Average } (C_M + C_E)} \right) \times 100\% \quad (3)$$

where C_M is the concentration measured by the instrument and C_E is the expected concentration. RPDs were less than 25% for the calibration standards at the levels presented in Tables 6 through 11.

Table 6. ICP-MS Data for the Arsenic P&A Study. “No Gas” Mode with Dichloromethane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	4.049	9.766	25.10	51.08	110.6	274.3	525.8	800.6	1054
Replicate 2	3.974	9.744	24.50	54.48	109.2	273.6	520.1	818.2	1067
Replicate 3	4.052	9.248	24.99	50.92	115.6	274.1	545.2	795.6	1086
Replicate 4	5.327	10.11	26.26	51.49	106.5	256.8	511.8	784.9	1046
Replicate 5	5.232	9.642	24.26	47.67	105.3	262.6	511.5	736.3	995.8
Replicate 6	5.557	10.51	25.64	49.40	108.0	251.2	504.1	756.1	1002
Mean Value:	4.699	9.837	25.13	50.84	109.2	265.4	519.8	781.9	1042
Mean % Recovery:	93.97%	98.37%	100.5%	101.7%	109.2%	106.2%	104.0%	104.3%	104.2%
Standard Deviation:	0.7458	0.4304	0.7364	2.274	3.657	10.06	14.56	30.36	35.93
%RSD:	15.87%	4.375%	2.931%	4.473%	3.349%	3.788%	2.802%	3.883%	3.449%

RSD, relative standard deviation

Table 7. ICP-MS Data for the Arsenic P&A Study. Helium Mode with Dichloromethane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	5.394	11.34	25.63	57.82	107.3	278.8	544.5	831.9	1074
Replicate 2	5.609	11.19	26.08	52.58	105.0	267.2	526.1	816.4	1081
Replicate 3	5.255	10.79	27.01	56.66	110.9	271.5	541.1	819.3	1050
Replicate 4	5.444	10.48	25.72	49.67	98.39	251.0	516.5	787.2	1061
Replicate 5	6.124	10.94	26.49	53.02	100.5	265.1	508.8	797.5	1066
Replicate 6	5.806	11.38	27.70	53.15	104.7	256.5	523.9	802.0	1054
Mean Value:	5.605	11.02	26.44	53.82	104.5	265.0	526.8	809.1	1064
Mean % Recovery:	112.1%	110.2%	105.8%	107.6%	104.5%	106.0%	105.4%	107.9%	106.4%
Standard Deviation:	0.3171	0.3498	0.8024	2.964	4.526	10.06	13.83	16.38	11.81
%RSD:	5.657%	3.174%	3.035%	5.507%	4.332%	3.795%	2.626%	2.025%	1.110%

Table 8. ICP-MS Data for the Arsenic P&A Study. Hydrogen Mode with Dichloromethane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	5.979	12.15	27.00	55.92	107.5	261.9	540.8	791.6	1063
Replicate 2	6.116	12.04	27.61	55.70	106.4	271.7	529.4	823.8	1128
Replicate 3	5.939	11.93	28.46	54.33	108.6	269.8	540.2	828.5	1090
Replicate 4	5.539	10.75	25.65	50.72	101.3	255.6	531.5	816.1	1053
Replicate 5	6.242	11.17	26.61	50.48	102.9	262.1	509.1	801.4	1087
Replicate 6	5.928	11.86	28.77	53.38	104.6	261.3	523.7	827.1	1045
Mean Value:	5.957	11.650	27.35	53.42	105.2	263.7	529.1	814.8	1078
Mean % Recovery:	119.1%	116.5%	109.4%	106.8%	105.2%	105.5%	105.8%	108.6%	107.8%
Standard Deviation:	0.2379	0.5595	1.173	2.375	2.798	5.973	11.79	15.10	30.54
%RSD:	3.990%	4.802%	4.288%	4.446%	2.659%	2.265%	2.228%	1.853%	2.834%

Table 9. ICP-MS Data for the Arsenic P&A Study. "No Gas" Mode with Hexane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	4.709	9.676	25.84	49.71	99.65	266.0	525.0	763.6	1001
Replicate 2	4.895	9.954	25.65	50.89	101.1	272.4	540.8	797.2	1043
Replicate 3	5.230	10.30	26.22	51.94	102.3	272.2	538.6	772.1	1047
Replicate 4	4.173	9.617	24.66	47.62	91.88	224.9	497.8	734.3	996.7
Replicate 5	4.385	8.817	24.03	46.39	97.14	221.0	495.4	743.4	972.1
Replicate 6	4.590	9.041	24.15	46.49	92.65	220.9	494.0	762.2	976.4
Mean Value:	4.664	9.568	25.09	48.84	97.45	246.2	515.3	762.1	1006
Mean % Recovery:	93.27%	95.68%	100.4%	97.68%	97.45%	98.49%	103.1%	101.6%	100.6%
Standard Deviation:	0.3742	0.5551	0.9322	2.349	4.378	26.39	22.11	22.17	32.21
%RSD:	8.024%	5.802%	3.715%	4.809%	4.493%	10.719%	4.290%	2.908%	3.202%

Table 10. ICP-MS Data for the Arsenic P&A Study. Helium Mode with Hexane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	4.869	10.02	27.06	51.88	104.8	261.6	519.5	768.7	1016
Replicate 2	5.091	10.30	26.90	53.43	106.2	265.8	541.9	800.3	1116
Replicate 3	5.478	10.77	27.50	54.83	107.7	268.4	534.7	776.6	1133
Replicate 4	4.398	9.907	25.37	49.10	96.71	215.3	486.3	729.0	1054
Replicate 5	4.597	9.232	25.24	48.45	99.77	212.9	488.0	738.6	1041
Replicate 6	4.814	9.378	25.02	49.04	97.20	213.9	484.1	741.9	1045
Mean Value:	4.875	9.935	26.18	51.12	102.1	239.7	509.1	759.2	1068
Mean % Recovery:	97.49%	99.35%	104.7%	102.2%	102.1%	95.86%	101.8%	101.2%	106.8%
Standard Deviation:	0.3792	0.5732	1.0882	2.654	4.774	28.16	26.19	27.27	46.23
%RSD:	7.779%	5.770%	4.156%	5.191%	4.677%	11.75%	5.144%	3.592%	4.331%

Table 11. ICP-MS Data for the Arsenic P&A Study. Hydrogen Mode with Hexane

	Recovered Concentrations (µg/L)								
Level	1	2	3	4	5	6	7	8	9
Spiked Conc. (µg/L)	5.000	10.00	25.00	50.00	100.0	250.0	500.0	750.0	1000.0
Replicate 1	4.762	9.940	26.73	51.42	104.3	259.6	518.0	766.5	1004
Replicate 2	5.051	10.21	26.67	52.85	104.7	264.5	535.9	793.8	1047
Replicate 3	5.673	10.61	27.18	54.48	106.5	266.3	530.3	774.4	1053
Replicate 4	4.838	10.64	26.78	52.00	105.2	228.1	528.7	790.7	1048
Replicate 5	4.997	10.06	27.30	52.63	105.4	228.7	528.9	802.1	1051
Replicate 6	5.337	10.00	26.61	53.29	105.4	232.5	527.1	789.4	1058
Mean Value:	5.110	10.24	26.88	52.78	105.3	246.6	528.2	786.2	1044
Mean % Recovery:	102.2%	102.4%	107.5%	105.6%	105.3%	98.65%	105.6%	104.8%	104.4%
Standard Deviation:	0.3405	0.3091	0.2884	1.062	0.7503	18.65	5.825	13.18	19.75
%RSD:	6.665%	3.018%	1.073%	2.012%	0.7129%	7.562%	1.103%	1.676%	1.892%

4.4

Measurement of Uncertainty

Tables 12 and 13 show the performance quantitation limits (PQL) and the calculated measurements of uncertainty in the two day reproducibility test of the performance and analysis of six samples at each calibration level. Some levels of uncertainty were higher than expected. This is important for the treaty samples where the identification of a sample must be as precise and accurate as possible. Performing this analysis with the two trace level solvents as interferrants shows the importance of including this in the matrix of the calibration standards.

Table 12. Measurement Uncertainty Three Modes with Dichloromethane

Analyte	Concentration (mg/L)	Uncertainty
Arsenic (No Gas)	5.000	452
	10.00	226
	25.00*	90
	50.00	47
	100.0	5.0
	250.0	14
	500.0	9.0
	1000.0	10.0
Arsenic (Helium)	5.000	429
	10.00	214
	25.00*	86
	50.00	46
	100.0	24
	250.0	14
	500.0	8.0
	1000.0	4.0
Arsenic (Hydrogen)	5.000	447
	10.00	224
	25.00*	90.0
	50.00	46
	100.0	23
	250.0	11
	500.0	8.0
	1000.0	8.0

* Denotes PQL identified through P&A, and measurement uncertainty (MU).

Table 13. Measurement Uncertainty Three Modes with Hexane

Analyte	Concentration (mg/L)	Uncertainty (%)
Arsenic (No Gas)	5.000	179
	10.00*	90.0
	25.00	37
	50.00	21
	100.0	14
	250.0	27
	500.0	11
	750.0	8.0
	1000.0	8.0
Arsenic (Helium)	5.000	546
	10.00	273
	25.00	109
	50.00*	6.0
	100.0	30.0
	250.0	31
	500.0	15
	750.0	10.0
	1000.0	12
Arsenic (Hydrogen)	5.000	227
	10.00	114
	25.00*	45
	50.00	23
	100.0	11
	250.0	20.0
	500.0	4.0
	750.0	5.0
	1000.0	5.0

* Denotes PQL identified through P&A, and MU.

4.5 Analysis of Unknown Samples

Table 14 shows the representative samples used to simulate unknown treaty samples. Samples of L were prepared in dichloromethane and hexane representative of previous samples received. These unknowns were analyzed against calibration curves with the solvent to determine the level of accuracy in the reporting of the compound for identification purposes.

Table 14. Measurement of Unknown Arsenic Solution with and without Calibration with Interfering Solvent

Analyte	Expected [AsCl ₃] ppm	Expected [As] ppm	Measured [As] ppb			Percent Recovery		
			H ₂ Mode	Helium Mode	No Gas Mode	H ₂ Mode	Helium Mode	No Gas Mode
AsCl ₃	5.000	20.67	19.92	19.77	18.59	96.39	95.67	89.96
AsCl ₃	50.00	206.6	173.3	166.7	152.1	83.86	80.67	73.60
2009 OPCW Sample	20.00	8.260	7.496	7.355	6.804	90.75	89.04	82.37

5. IMPLICATIONS FOR FUTURE APPLICATIONS

The ICP-MS in hydrogen mode has a high percent recovery when detecting arsenic in real world samples. Helium is the recommended mode to measure arsenic ^{75}As to distinguish m/z 75 from polyatomic interference by argon chloride $^{40}\text{Ar}^{35}\text{Cl}$ m/z 75 in the no gas mode⁵. Either hydrogen or helium appear to be sufficient at distinguishing arsenic in simulated treaty samples. Addition of 0.250 mL of solvent in the calibration standards significantly improved the detection of arsenic in samples with solvent. Arsenic levels measured 0 ppb without the solvent in the calibration standards and arsenic recoveries of 90% or better were detected with the solvent present. This simplified correction for detecting arsenic in samples avoids the tedious process of derivitization needed for gas chromatography-mass spectrometry (GC-MS), and allows for a quick screening technique in time sensitive samples. Increased sensitivity of detection and speciation can be achieved with an optional oxygen line and hyphenated techniques such as the addition of an IC or high-performance liquid chromatography (HPLC) to the front end of the ICP-MS for chromatographic separation of the different arsenic species⁶. For the purposes of rapidly screening samples for presence and absence this method was validated and found to have detection levels suitable for screening OPCW samples.

LITERATURE CITED

1. *Standard Operating Procedure for the Organisation of OPCW Proficiency Tests, Rev 2, Issue 2*, QDOC/LAB/SOP/PT01; Organisation for the Prohibition of Chemical Weapons: The Hague, The Netherlands, 2012.
2. *Recommended Operating Procedures for Analysis in the Verification of Chemical Disarmament*; Vanninenm P., Ed.; The Ministry for Foreign Affairs of Finland: University of Helsinki, 2011.
3. *Evaluation of Results Twenty-Fifth Official OPCW Proficiency Test*; Organisation for the Prohibition of Chemical Weapons: The Hague, The Netherlands, 2009.
4. Organisation for the Prohibition of Chemical Weapons - Basic Facts on Chemical Disarmament. <<http://www.opcw.org/news-publications/publications/history-of-the-chemical-weapons-convention/#c4135>> (accessed November 23, 2012).
5. Getting the Most from What You've Got: How to Improve Data Quality with Any ICP-MS. ICP-MS Training, provided November 4, 2011 at University of Maryland, Baltimore County. Thermo Fisher Scientific: Waltham, MA.
6. Raber, G.; Raml, R.; Goessler, W.; Francesconi, K.A. Quantitative Speciation of Arsenic Compounds when Using Organic Solvent Gradients in HPLC-MS. *J. Anal. At. Spectrom.* **2010**, 25, 570-576.

BIBLIOGRAPHY

Agilent 7500 Series ICP-MS Chem Station Operator Manual, Rev A. G3270-90115; Agilent Technologies: Santa Clara, CA, September 2007.

Agilent 7500 Series ICP-MS Customer Maintenance Parts List, Rev A. G3270-90124; Agilent Technologies: Santa Clara, CA, April 2007.

Agilent 7500 Series ICP-MS Hardware Manual, Rev C. G3270-90104; Agilent Technologies: Santa Clara, CA, October 2007.

Agilent 7500 Series ICP-MS Installation Guide, Rev B. G3270-90213; Agilent Technologies: Santa Clara, CA, June 2007.

Agilent 7500 Series ICP-MS Option Instruction Manual, Rev C. G3270-90140; Agilent Technologies: Santa Clara, CA, April 2007.

Agilent 7500 Series ICP-MS Maintenance Video (DVD), Rev A. G3270-65100; Agilent Technologies: Santa Clara, CA, September 2005.

Agilent ASX-500 Series Autosampler (CD Rev A. G3286-90200; Agilent Technologies, CA, March 2006.

Environmental Protection Agency, 2007. *Method 6020A, Revision 1, Inductively Coupled Plasma-Mass Spectrometry.* <www.epa.gov/osw/hazard/testmethods/sw846/pdfs/6020a.pdf>

ACRONYMS AND ABBREVIATIONS

CAS	Chemical Abstract Service
CVAA	chlorovinyl arsonous acid
CWA	chemical warfare agents
ECBC	U.S. Army Edgewood Chemical Biological Center
FAC	Forensic Analytical Center
GC-MS	gas chromatography-mass spectrometry
HPLC	high-performance liquid chromatography
ICP-MS	inductively coupled plasma-mass spectrometry
ISO/IEC	International Organization for Standardization/International Electrotechnical Commission
L	lewisite
MDL	method detection limit
MU	measurement uncertainty
OPCW	Organization for the Prohibition of Chemical Weapons
P&A	performance and analysis
PQL	performance quantitation limits
RPD	relative percent difference
RSD	relative standard deviation

