

REPORT DOCUMENTATION PAGE

Form Approved
OMB No. 0704-0188

The public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing the burden, to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number.

PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.

1. REPORT DATE (DD-MM-YYYY) 14/11/2013		2. REPORT TYPE Final Report		3. DATES COVERED (From - To) 3/2010-7/2013	
4. TITLE AND SUBTITLE SafePort Proposal - Henry Laboratory 2010				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S) Charles Henry				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) Dept. of Chemistry, 1872 Campus Delivery, Colorado State University, Fort Collins, CO 80523				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING/MONITORING AGENCY NAME(S) AND ADDRESS(ES) Office of Naval Research				10. SPONSOR/MONITOR'S ACRONYM(S) ONR	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION/AVAILABILITY STATEMENT 20131125202					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT Microfluidic devices for analyzing ionic pollutants in water were developed and characterized. Different device materials were tested with cyclic olefin copolymer (COC) being determined as the most suitable for long term depolyment. Subsequent testing involved challenging the devices with samples containing high ionic strength samples. A method was developed for removing the majority of the interfering high concentration species using solid phase extraction. Using the method, detection limits were approached for appropriate detection of perchlorate in ground water.					
15. SUBJECT TERMS					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT	18. NUMBER OF PAGES	19a. NAME OF RESPONSIBLE PERSON
a. REPORT	b. ABSTRACT	c. THIS PAGE			19b. TELEPHONE NUMBER (Include area code)

Final Report for W9132T-10-2-0012

Objectives:

- 1) Develop modular glass microchips for perchlorate analysis.

We attempted to make modular chips for this application but were unsuccessful due to the problems with the high bonding temperature of glass and the melting point of the metals we used. As a result, we focused more on studies done comparing PDMS, PMMA, PC, and COC microchips for contact conductivity detection for ion separations.

- 2) Compare modular glass microchips to PDMS and PMMA microchips.

Once learning we were not able to make glass microchips, we focused on comparing the performance of PDMS, PMMA, PC, and COC (cyclic olefin copolymer). These polymers were chosen because they are the most common polymers used in microfluidics and can be manufactured via a wide range of methods including molding and hot embossing. In our initial comparison steps, we compared separation efficiency, migration time reproducibility, and peak area reproducibility. PDMS and COC proved to be the most consistent with PMMA and PC showing substantial variability from batch to batch of the material. As a result, further evaluations were made between PDMS and COC. In comparing COC to PDMS, two major trends were noted. First, the migration times of perchlorate under identical buffer conditions were always longer, by 20% for COC microchips versus PDMS. PDMS microchips also gave more defined (sharper) peaks than the COC. Both chip materials, however, were able to resolve perchlorate from competing ions with ease and thus it is proposed that COC microchips be used going forward because they are easier to manufacturing using injection molding on an industrial scale. They are also easier to work with because they are rigid polymers.

- 3) Challenge perchlorate separation with multiple concentrations of anions and include the ability to generate a calibration curve.

One major limitation of microchip electrophoresis is the potential for high ionic strength samples containing excess sulfate and/or chloride to prevent the system from functioning properly. As a result, we tested multiple approaches to quantifying part-per-billion levels of perchlorate in the presence of millimolar concentrations of sulfate and chloride. In the absence of any purification methods, perchlorate detection limits were 100 ppb with 1 mM each chloride and sulfate. To address this problem, we have explored use of OnGuard cartridges produced by ThermoFisher Scientific. The cartridges contain a resin with Ba and Ag. Ba reacts strongly with sulfate forming insoluble BaSO_4 , while Ag reacts with chloride to produce insoluble AgCl . The figure below shows electropherograms for before the sample was passed through the column (blue, bottom trace) and then after the material has been purified (red and black trace). The sample consisted of 1 mM each of chloride and sulfate and 25 ppb perchlorate. Two injection times were used to verify that the peaks were real and not system peaks. These results clearly show our ability to analyze perchlorate in these complex samples using the OnGuard columns.

