

Award Number:
W81XWH-04-1-0620

TITLE:
**Novel Approaches to Glucose Sensing Based on Polymerized
Crystalline Colloidal Array Hydrogel Sensors**

PRINCIPAL INVESTIGATOR:
Sanford A. Asher, Ph.D.

CONTRACTING ORGANIZATION:
**University of Pittsburgh
Pittsburgh, PA 15260**

REPORT DATE:
December 2006

TYPE OF REPORT:
Final Report

PREPARED FOR: U.S. Army Medical Research and Materiel Command
Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT:

Approved for public release; distribution unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

REPORT DOCUMENTATION PAGE				Form Approved OMB No. 0704-0188	
Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Department of Defense, Washington Headquarters Services, Directorate for Information Operations and Reports (0704-0188), 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302. Respondents should be aware that notwithstanding any other provision of law, no person shall be subject to any penalty for failing to comply with a collection of information if it does not display a currently valid OMB control number. PLEASE DO NOT RETURN YOUR FORM TO THE ABOVE ADDRESS.					
1. REPORT DATE (DD-MM-YYYY) 01-12-2006		2. REPORT TYPE Final		3. DATES COVERED (From - To) 1 JUL 2004 - 30 NOV 2006	
4. TITLE AND SUBTITLE Novel Approaches to Glucose Sensing Based on Polymerized Crystalline Colloidal Array Hydrogel Sensors				5a. CONTRACT NUMBER	
				5b. GRANT NUMBER W81XWH-04-1-0620	
				5c. PROGRAM ELEMENT NUMBER	
6. AUTHOR(S) Sanford A. Asher, Ph.D. asher@pitt.edu				5d. PROJECT NUMBER	
				5e. TASK NUMBER	
				5f. WORK UNIT NUMBER	
7. PERFORMING ORGANIZATION NAME(S) AND ADDRESS(ES) University of Pittsburgh Pittsburgh, PA 15260				8. PERFORMING ORGANIZATION REPORT NUMBER	
9. SPONSORING / MONITORING AGENCY NAME(S) AND ADDRESS(ES) U.S. Army Medical Research and Materiel Command Fort Detrick, Maryland 21702-5012				10. SPONSOR/MONITOR'S ACRONYM(S)	
				11. SPONSOR/MONITOR'S REPORT NUMBER(S)	
12. DISTRIBUTION / AVAILABILITY STATEMENT Approved for public release; distribution unlimited					
13. SUPPLEMENTARY NOTES					
14. ABSTRACT Final Progress Report: 1. We synthesized 150 nm highly monodisperse superparamagnetic colloidal particles with high surface charge. 2. We incorporated these particles into the glucose sensing hydrogel and demonstrated that it changed diffraction with changes in glucose concentration. 3. We built a hydrophone sound detection instrument that could monitor acoustic responses from the the hydrogel photonic crystal with the superparamagnetic particles. 4. Unfortunately, in spite of extensive efforts, we did not find any acoustic response when varying the magnetic field strength in the hydrogel material. 5. In the above studies we dramatically varied the hydrogel composition and the superparamagnetic particle composition, and the methods used to apply the magnetic field. No publications, references or appendices.					
15. SUBJECT TERMS None provided.					
16. SECURITY CLASSIFICATION OF:			17. LIMITATION OF ABSTRACT UU	18. NUMBER OF PAGES 4	19a. NAME OF RESPONSIBLE PERSON USAMRMC
a. REPORT U	b. ABSTRACT U	c. THIS PAGE U			19b. TELEPHONE NUMBER (include area code)

Table of Contents

Introduction.....4

Body.....4

Key Research Accomplishments.....4

Reportable Outcomes.....4

Conclusions.....4

References.....None

Appendices.....None

Final Progress Report

Award Number W81XWH-04-1-0620

"Novel Approaches to Glucose Sensing Based on Polymerized Crystalline Colloidal Array Hydrogel Sensors"

1. We synthesized 150 nm highly monodisperse superparamagnetic colloidal particles with high surface charge.
2. We incorporated these particles into the glucose sensing hydrogel and demonstrated that it changed diffraction with changes in glucose concentration.
3. We built a hydrophone sound detection instrument that could monitor acoustic responses from the the hydrogel photonic crystal with the superparamagnetic particles.
4. Unfortunately, in spite of extensive efforts, we did not find any acoustic response when varying the magnetic field strength in the hydrogel material.
5. In the above studies we dramatically varied the hydrogel composition and the superparamagnetic particle composition, and the methods used to apply the magnetic field.

No publications, references or appendices.